



Respironics, Inc.
Murrysville, PA

Rx ONLY



IP22



This is a medical device
100-240 V ~ 50/60 Hz, 2.0 - 1.0A,

USB Output:
5 ---, 1.5A



* Complies with 47 CFR Part 15
Class B of the FCC Rules
Contains FCC ID: TH01116426



Complies with
RTCA/D0-160G
section 21, category III

ETL Classified



Intertek
3184661
Medical Electrical
Equipment

DreamStation Go Auto CPAP

REF DSG500S11

SN JXXXXXXXXXXXXX



(01)00606859044463
(21)JXXXXXXXXXXXXX



2017-03-03
REV 00