



Philips Healthcare

1010 Murry Ridge Lane

Murrysville, PA 15668

Home Healthcare Solutions

SIMILARITY DECLARATION LETTER

Federal Communications Commission
Authorization and Evaluation Division
7435 Oakland Mills Road
Columbia MD 21046

May 26, 2020

FCC ID: THO1116426

Regarding a Class II Permissive Change for FCC ID: THO1116426 , Original Grant Date: 6/17/2015.

Applicant: Respironics Inc

1010 Murry Ridge Lane

Murrysville, PA 15668

The changes filed under this request are to create a new model: 100610C accessory for BASE UNIT FCC ID: THO1116426

Model Numbers: 100601C, 200601C, 100602C and 200602C are similar to new models: 100610C DreamStation Modem and 100650C DreamStation Modem

We, hereby declare the following Class 2 Permissive Change to the 100601C, 200601C, 100602C and 200602C related to the FCC regulations to the models applied for 9/19/19 adding a new model number 100610C.

- The Philips external mechanical and internal electrical construction are unchanged except as noted below:
 - No changes to the 100601C, 200601C, 100602C and 200602C PCB and the antenna excepting for the new model:
- The 100610C and 100650C use an EHS6 module with FCC ID: QIPEHS6 instead of an EHS5-US module FCC ID QIPEHS5-US.
 - The integrated cellular module uses the same chip set family, but the number of bands is increased.
 - The transmitter output power and spurious emission have been retested and evaluated to ensure compliance.
 - 100610C and 100650C only differ by the labels and the SIM internal data.

If you have any questions, please feel free to contact me at the address shown below.

Sincerely,

Jerry Shore

Tel: 724.387.4000

Tel: 800.345.6443 (toll free, US only)

www.philips.com/healthcare