



Philips Healthcare

1010 Murry Ridge Lane

Murrysville, PA 15668

Home Healthcare Solutions

Date: April 6, 2018

Applicant:

Respironics Inc.
1001 Murry Ridge Lane
Murrysville, PA 15668

Subject: Request for Class II Permissive Change

This is to request a Class II permissive change for FCC ID: THO-1116426.

The original grant is for a Limited Single Modular Approval for a radio module. This radio module is incorporated in our host equipment DreamStation CPAP containing FCC ID: THO 1116426.

With the creation of the new radio accessory modules with the identified host equipment, no changes were made to the Bluetooth radio module output power, modulation circuit, frequency determining circuitry, or by external adaptation circuitry that impacts the validity of the results of the original module's test reports.

The major changes filled under this application are changing the modem used from a SARA-R404 to a SARA-R410 on the 2 models of modem accessories 200603C and 200604C. The manufacturer FCC ID for these modules and will used for the accessory is XPY2AGQN4NNN. A single modem either Wi-Fi or cell modem may be used with the base Bluetooth radio.

The necessary testing reports for the collocated new modem type have been added to the DreamStation submittal by Cetecom.

The new cell modem models are identical except that the 200603C or 200604C unit has a number of parts not placed in a serial interface circuit:

R1, R4, R5, R6, R7, R8, R13, R16

C1, C4, C5, C6, C7, C8, C9, C10, C11, C12, C13, C14, C15, C19

U1, U2, U3, U4, U5

J3

D2, D3

FL1

Nick Cormas

Tel: 724.387.4000

Tel: 800.345.6443 (toll free, US only)

www.philips.com/healthcare