

Philips Healthcare

Respironics, Inc.
1001 Murry Ridge Lane
Murrysville, PA 15668-8550
724-387-7500
www.respironics.com

FCC-ID: THO1116426

**Specific Limited Single Modular Transmitter Installation Instructions
Restrictions to be respected for installation of the limited single modular transmitter with
FCC-ID THO1116426
(Bluetooth module as integrated on the mainboard of the Philips Dream Station)**

1. The main circuit board carrying the above identified limited modular transmitter has been tested and approved when installed and for operation within Philips Dream Station and may not be installed and operated in other host devices without prior re-assessment acc. to FCC rules;
2. Only trained assemblers at the Philips Respironics factory or authorized service centers are permitted to install the module, according to qualified Philips Respironics factory assembly instructions;
3. The module must not be easily removable, installable, replaceable by normal end-users; host product (Dream Station) instructions may not contain instructions to end-users about how to remove, install or replace the board;
4. The identified modular transmitter has been tested and approved when installed and for operation within Philips Dream Station and also when the Dream Station is operated together with either the Cellular Accessory (containing CDMA module with FCC ID QIPPCS3) or with the WiFi Accessory (containing WiFi module with FCC ID TLZ-CU288);
Co-location with any other transmitters is not permitted without prior assessment acc. to FCC rules;
5. A minimum distance of 20cm must be kept between the transmitter sections of the WiFi or the Cellular Accessory when the Dream Station is operated with one of those accessories.