

ID Label Sample & Location Information

Lumimed®
Medical LCD monitor
Model NO : MM20
AC 100-240V~,0.8-0.3A,50/60Hz

CLASSIFIED
UL
US
120V
MEDICAL EQUIPMENT
WITH RESPECT TO ELECTRIC SHOCK,
FIRE AND MECHANICAL HAZARDS ONLY
IN ACCORDANCE WITH UL 60601-1, AND
CAN/CSA C22.2 NO. 601.1
E298477

This device complies with part 15 of the FCC rules.
Operation is subject to the following two conditions:
(1) This device may not cause harmful interference, and
(2) This device must accept any interference received, including interference that may cause undesired operation.

FCC ID : TJLMM20

CAUTION : RISK OF ELECTRIC SHOCK, DO NOT OPEN.

Serial No

MADE IN CHINA P/N: 1502203

CE



