

## **Annex K. Calibration of Test Equipment List**

Calibration of Test Equipment List are shown as below.

| Equipment for SAR Test       |              |                   |            |               |               |
|------------------------------|--------------|-------------------|------------|---------------|---------------|
| Equipment                    | Manufacturer | Model             | SN         | Cal. Date     | Cal. Interval |
| System Validation Dipole     | SPEAG        | D2450V2           | 835        | Jun. 22, 2021 | 1 Year        |
| Dosimetric E-Field Probe     | SPEAG        | EX3DV4            | 7537       | Apr. 26, 2021 | 1 Year        |
| Data Acquisition Electronics | SPEAG        | DAE4              | 1585       | Apr. 15, 2021 | 1 Year        |
| Spectrum Analyzer            | R&S          | FSL6              | 102006     | Apr. 06, 2021 | 1 Year        |
| Universal Wireless Test Set  | Anritsu      | MT8870A/MU887000A | 6201699387 | Sep. 28, 2020 | 1 Year        |
| Thermometer                  | YFE          | YF-160A           | 120702365  | Aug. 12, 2020 | 1 Year        |
| Dielectric Assessment Kit    | SPEAG        | DAKS-3.5          | 0004       | Mar. 24, 2021 | 1 Year        |
| Powersource1                 | SPEAG        | SE_UMS_160 BA     | 4230       | Jun. 08, 2021 | 1 Year        |