



**ISO 17025**  
ACCREDITED LABORATORY



Accreditation certificate No. № BY/112 5.0065 of 09.01.2015

Certificate number 17-21 Date when calibrated 11.02.2021 Page 1 of 2

**Customer** Bureau Veritas Consumer Products Services (Hong Kong) Limited,  
Taoyuan Branch

**Method of calibration** GOST 20271.1, MK KL 8.2-16

*All measurements are traceable to the SI units which are realized by national measurement standards of NMI and state standards of RF. Gain measurements above 178 GHz are to confirm operation functionality and traceable only to MWM Lab standards and OML. This certificate shall not be reproduced, except in full. Any publication extracts from the calibration certificate requires written permission of the issuing calibration laboratory of microwave measuring equipment.*

### Authorising signature

/ Technical manager **Date of issue 11.02.2021**

# Calibration Certificate

Certificate number **17-21**

Page 2 of 2

## Calibration is performed by using

| Model   | Model Description      | Equipment ID | Cal Due Date      | Certificate Number | Trace Value |
|---------|------------------------|--------------|-------------------|--------------------|-------------|
| M 546   | Reference power meter  | 163          | 24 March 2022     | 1/111-174-20       | RF Power    |
| M 568   | Reference power meter  | 164          | 24 March 2022     | 1/111-175-20       | RF Power    |
| G4-161  | Signal generator       | 3            | 12 October 2021   | 20-20              | RF Power    |
| G4-186  | Signal generator       | 5            | 12 October 2021   | 21-20              | RF Power    |
| V7-34   | Universal voltmeter    | 0067787      | 23 September 2021 | 2742-42            | DC Voltage  |
| RCH3-72 | Frequency meter        | 931200       | 18 September 2021 | 2822-43            | Frequency   |
| P6-134  | Measuring horn antenna | 14002        | 23 September 2021 | 2372-43            | Gain        |

## Calibration conditions

Temperature: 22.2 °C.

Humidity: 37.0 %.

Pressure: 100.2 kPa.

## Calibration results are given in the measurement report # 17-21

| # | Parameter       | Specifications required | Specifications tested and measured |
|---|-----------------|-------------------------|------------------------------------|
| 1 | Frequency range | 50 – 75 GHz             | Corresponds                        |
| 2 | Antenna Gain    | 21.0* dBi               | Corresponds (Table 1)              |
| 3 | Antenna Factor  | 45 dB/m                 | Corresponds (Table 1)              |

\* – Expanded uncertainty of measurements 2.0 dB.

*The uncertainty evaluation has been performed in accordance with ISO/IEC Guide 98-3:2008 (GUM). The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor  $k$  such that the coverage probability corresponds to approximately 95 %. This probability corresponds to a coverage factor of  $k=2$  for a normal distribution.*

Signature of the person who has performed calibration



/ Engineer

**Calibration Laboratory of  
Microwave Measuring Equipment**

Accreditation certificate

No. BY/112 5.0065

Address: 6, P. Brovki str., Minsk  
220013, Belarus

Phone/Fax: +375 17 2938496



**MEASUREMENT REPORT # 17-21**

February 11, 2021

|                              |                                                                               |
|------------------------------|-------------------------------------------------------------------------------|
| Customer:                    | Bureau Veritas Consumer Products Services (Hong Kong) Limited, Taoyuan Branch |
| Item calibrated:             | <b>Antenna RCH015 # 2</b>                                                     |
| Method of calibration:       | GOST 20271.1, MK KL 8.2-16                                                    |
| Number of samples:           | One                                                                           |
| Delivery date of the sample: | 14.01.2021                                                                    |
| Date of calibration:         | From 14.01.2021 to 11.02.2021                                                 |

## MEASUREMENT CONDITIONS

|                      |                |                     |
|----------------------|----------------|---------------------|
| Temperature: 22.2 °C | Humidity: 37 % | Pressure: 100.2 kPa |
|----------------------|----------------|---------------------|

## MEASUREMENT EQUIPMENT

| Model   | Model Description      | Equipment ID | Cal Due Date      | Certificate Number | Trace Value |
|---------|------------------------|--------------|-------------------|--------------------|-------------|
| M 546   | Reference power meter  | 163          | 24 March 2022     | 1/111-174-20       | RF Power    |
| M 568   | Reference power meter  | 164          | 24 March 2022     | 1/111-175-20       | RF Power    |
| G4-161  | Signal generator       | 3            | 12 October 2021   | 20-20              | RF Power    |
| G4-186  | Signal generator       | 5            | 12 October 2021   | 21-20              | RF Power    |
| V7-34   | Universal voltmeter    | 0067787      | 23 September 2021 | 2742-42            | DC Voltage  |
| RCH3-72 | Frequency meter        | 931200       | 18 September 2021 | 2822-43            | Frequency   |
| P6-134  | Measuring horn antenna | 14002        | 23 September 2021 | 2372-43            | Gain        |

## MEASUREMENT RESULTS

Distance between tested and generating antenna 1.5 m.

Table 1

| Frequency, GHz                                           | 50    | 55    | 65    | 75    |
|----------------------------------------------------------|-------|-------|-------|-------|
| Power density of electromagnetic field, W/m <sup>2</sup> | 0.068 | 0.081 | 0.102 | 0.130 |
| Maximum level of measured power, dBm                     | -16.8 | -16.4 | -16.4 | -16.3 |
| Gain, dBi                                                | 20.3  | 20.8  | 21.2  | 21.5  |
| Expanded uncertainty, dB                                 | 2.0   | 2.0   | 2.0   | 2.0   |
| Antenna Factor, dB/m                                     | 43.9  | 44.2  | 45.3  | 46.2  |

*The uncertainty evaluation has been performed in accordance with ISO/IEC Guide 98-3:2008 (GUM). The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor  $k$  such that the coverage probability corresponds to approximately 95 %. This probability corresponds to a coverage factor of  $k=2$  for a normal distribution.*

Engineer



This measurement report issued in duplicate and sent to:

1. Bureau Veritas Consumer Products Services (Hong Kong) Limited, Taoyuan Branch
2. Calibration Laboratory of Microwave Measuring Equipment

Duplication of Measurement report (complete or partial) must be authorized by the laboratory.