
From: oetech@fcc.gov
Sent: Thursday, July 25, 2019 9:41 AM
To: Rod Peloquin
Subject: Response to Inquiry to FCC (Tracking Number 345183) (TCB)

Importance: High

Inquiry on 07/25/2019 :

Inquiry:

An applicant is seeking approval for a radio module that will be integrated into medical equipment, like patient monitors and Pulse CO-Oximeters that contain displays.

These types of products are not specifically listed in KDB 784748 Section 2.2 so we would like to know if they qualify to use e-labeling for the module.

Thanks

FCC response on 07/25/2019

Yes, if they have a display they qualify.

Attachment Details:

Do not reply to this message. Please select the [Reply to an Inquiry Response](#) link from the OET Inquiry System to add any additional information pertaining to this inquiry.