



MASIMO AMERICAS
Forty Parker
Irvine, CA 92618

Modification to Device Letter

4-12-2013

To whom it may concern,

This letter is to indicate that the following modifications were necessary for
Compatible Electronics to certify to FCC requirements for certification.

The following modifications will be implemented in production.

Modifications:

1. Reduced an internal control output power code for 802.11 a from 90 to 50.
2. Reduced an internal control output power code for 802.11 g from 90 to 60.

Sincerely,

A handwritten signature in black ink that reads 'Michael Clark'. The signature is written in a cursive, flowing style.

Michael Clark
Engineer Senior Test R&D Instruments