

8

7

6

5

4

3

2

1

NOTES: UNLESS OTHERWISE SPECIFIED:

1. INTERPRET THIS DRAWING PER ASME Y14.5M AND ALCON SPECIFICATION 701-026.
2. MATERIAL: 2.4 MIL POLYESTER FILM, GREY / THERMAL PRINTING, WITH PERMANENT ACRYLIC ADHESIVE.
3. DIMENSIONS ARE IN INCHES.
TOLERANCES:
 .XX ±.03
 .XXX ±.010
 ANGLE ± 1°
4. ARTWORK TO BE SUPPLIED BY ALCON. P/N: 721-2160-401 REV. P3.
5. ALL NOMENCLATURE TO BE 100% BLACK UNLESS OTHERWISE SPECIFIED.
6. INSPECT PER GENERIC QIP LABEL. 
7. THIS DRAWING CONTAINS A TOTAL OF 2 CONTROL DIMENSION NUMBERS. 
8. COLOR TO BE, YELLOW, RCO, COLOR 109, SUPPLIER SPECIFICATION. APPARENT INSPECTION YELLOW TO BE WITHIN: PMS 100 C - 102 C, 106 C - 108 C, 113 C - 116 C, 1225 C, 1235 C. BACKGROUND. ALL OTHER NOMENCLATURE AND STROKE IN SYMBOL TO BE BLACK. SYMBOL TO HAVE A WHITE STROKE ON OUTER EDGE.
9. COLOR TO BE, BLUE, RE2, COLOR 5005, SUPPLIER SPECIFICATION. APPARENT INSPECTION BLUE TO BE WITHIN: PMS 280 C, 281 C, 286 C, 287 C, 293 C, 294C, 300 C, 301 C, 2935 C, 2945 C. BACKGROUND. FIGURE IN SYMBOL TO BE WHITE. SYMBOL TO HAVE A WHITE STROKE ON OUTER EDGE.
10. MATERIAL CERTIFICATION REQUIRED, UPON REQUEST.
11. THIS IS PART OF ELECTRICAL OR ELECTRONIC EQUIPMENT AND SHALL COMPLY WITH THE RoHS DIRECTIVE PER ALCON DOCUMENT 908-0000-031.
12. SAFETY CRITICAL COMPONENT.

REVISION HISTORY		
REV	CN	DATE
P3	NOT RELEASED	05/01/14

The diagram shows the Luxor Foot Controller with the following details:

- Dimensions:**
 - Overall width: 4.00
 - Overall height: 2.00
- Callouts:**
 - Callout 8 points to the top-left corner of the device.
 - Callout 9 points to the bottom-left corner of the device.
- Device Labels and Text:**
 - Luxor™** (with logo)
 - Ophthalmic Microscope
 - FOOT CONTROLLER**
 - GENERAL WARNING** (with exclamation mark icon)
 - Follow Instructions for Use (with person icon)
 - Non-ionizing electromagnetic radiation (with radiation icon)
 - IPX6**
 - REF** 8065752566
 - For applicable patents, please see the "ABOUT" screen on the monitor during operation.
 - Use appropriate take-back system (with crossed-out trash can icon)
 - CE** 0123
 - Alcon®** (with logo)
 - Manufacturer
 - ALCON LABORATORIES, INC.
 - 6201 SOUTH FREEWAY
 - FORT WORTH, TX 76134-2099 USA
 - a Novartis company
 - MADE IN USA
 - © 2014 Novartis
 - 721-2160-001 REV. P3
 - This device complies with part 15 of the FCC Rules. Operation is subject to the following two conditions: (1) This device may not cause harmful interference, and (2) this device must accept any interference received including interference that may cause undesired operation.
 - FCC ID: VMCGNPFWSW1
 - IC : 7345A- NGPFWSW1
 - CONTAINS
 - FCC ID: RI7XE61
 - IC : 5131A-XE61
 - CE** 0981 (with exclamation mark icon)
 - UL (with logo)
 - 4 X R .125 (pointing to the bottom-right corner)

QTY		PART NUMBER		DESCRIPTION		MATERIAL/MANUFACTURER		ITEM	
BILL OF MATERIAL									
BY TIEN LAM		DATE 04/17/14		THIS DOCUMENT CONTAINS PROPRIETARY INFORMATION WHICH SHALL NOT BE REPRODUCED OR TRANSFERRED TO OTHER DOCUMENTS OR DISCLOSED TO OTHERS OR USED FOR MANUFACTURING OR ANY OTHER PURPOSE WITHOUT PRIOR WRITTEN PERMISSION OF ALCON.  Irvine, California 92618		DESCRIPTION LABEL.FT5W			
CHECKED LORETTA MCCLAUGHLIN		DATE 05/01/14				SIZE D DRAWING NUMBER 721-2160-001 REV P3			
APPROVED		DATE							
FORMAT REVISION 2/18/2009						SCALE: 2:1 SHEET 1 OF 1			

DRAWING NUMBER	REV
721-2160-001	P3

D


C

B

A

721-1321-5XX.asm
721-1321-501.asm

Label affixes
here



Label, FCC & IC

