

Federal Communication Commission
Equipment Authorization Division,
Application Processing Branch
7435 Oakland Mills Road
Columbia, MD 21048

25/01/2016

To Whom It May Concern

ID Label/ Label Location – Unitron Stride M Products

Pursuant to paragraph §2.925 of 47 C.F.R. the hearing instruments Unitron Stride M, which is subject of this filing, shall bear a permanently affixed, readily visible label listing the information as specified in §2.926 and §15.19(a) of the 47 C.F.R. Due to the specific use of the hearing instrument (as shown in figure 1) its external surface is in contact with the user's skin behind the ear.



Figure 1: Position of the Hearing Instrument on the User

Therefore, it is not appropriate to permanently affix a label on the outer surface of the device to prevent skin irritations. The only alternative place for the FCC ID and serial number of the device would be in the battery compartment.

In this case the size of the FCC ID will be smaller than 2 points, which does not satisfy the requirements for "readily visible" label as defined in §2.925 (d)(2) and §2.925(g).

Due to the small size of the devices and pursuant to paragraph § 15.19(a)(5) of 47 C.F.R., the FCC ID and the statement specified in §15.19(a)(3) are placed in the user manual as shown below.

USA	FCC ID	VMY-UWNB4
Canada	IC ID:	2756A-UWNB4

B. Matcheski

Brian Matcheski
Manager-Quality Assurance
Unitron Hearing
20 Beasley Drive Kitchener, ON
519-895-0100 x2110