



8.2.3 Deleting a Basal Rate Profile

1




Tap the basal rate profile you want to delete (for example, Profile 3).

2



Tap the  symbol in the upper right corner of the screen.


3



If the active basal rate profile is to be deleted, the following display appears.

Tap **Yes** to delete the basal rate profile.

4



The selected basal rate profile was deleted from the list of available basal rate profiles.

8.3 Temporary Basal Rates

A Temporary Basal Rate (TBR) allows you to temporarily increase or decrease your active basal rate profile on a percentage basis for a specific duration. This helps you to better control your blood glucose level during illness, physical activity or in other situations. Temporary Basal Rates can be set in increments of 10 % over a period of 15 minutes to 24 hours.

TBR	Settings range
Decrease	0–90 %
Increase	110–250 %


Note

- ▶ A TBR can only be programmed when the micropump is running.
- ▶ Stopping the pump (Stop mode) stops TBR delivery as well as any bolus delivery.
- ▶ When the duration of the TBR has expired, you are informed that the programmed basal rate has finished.

8.4 Creating and Editing a TBR


A Temporary Basal Rate can be programmed, edited and deleted in different ways. Start by using one of the following two options:

1




On the Status screen, tap the basal rate profile.

or



In the Main menu, tap the **Basal rate** menu.


2



Tap **Temporary basal rate**.


8.4.1 Programming a TBR

1



Tap **Standard TBR**.

2



Tap **Percentage** to enter the percentage of the Temporary Basal Rate.


3



Use **−** and **+** to set the percentage for adjusting the Temporary Basal Rate.


Tap **Save**.

4



Tap **Duration** to enter the running time for the Temporary Basal Rate.


5



Use $-$ and $+$ to set the hours and minutes for the duration of the standard TBR.

Tap **Save**.

6



Tap **Activate**.


Note

You can only activate the TBR if the percentage is less than or greater than 100 % (for example, 90 % or 110 %).

Tapping $\leftarrow$ cancels the activation of the standard TBR and takes you to the previous display.

The Standard TBR that was previously active, remains active.

7





To confirm this step and start the standard TBR, press the insulin button lit up in green  below the diabetes manager screen.

8.4.2 Programming a Customised TBR


You can program and save customised Temporary Basal Rates for recurring situations that change your insulin needs. For a customised TBR, the percentage and the duration are saved. These settings are used as default values each time you select this TBR. You also have the option of entering a name for a customised TBR.

Example

You want to run for 1 hour. You know that your body needs 25 % less insulin during this activity and the subsequent recovery phase of 2 hours. You program a TBR of 75 % for the next 3 hours.



2




Tap the symbol.

Enter the desired name for the **Customised TBR** (for example, Run). The name may have up to 12 characters.

Confirm your entry with .

3




Tap **Percentage** to enter the **Percentage** for the **Customised TBR**.

Note


A customised TBR less than 100 % is indicated by this symbol . A TBR greater than 100 % is indicated by this symbol .

4




Use and to set the **Percentage** for adjusting the **Customised TBR**.

Tap **Save**.




Tap **Duration** to enter the running time for the **Customised TBR**.




Use **-** and **+** to set the hours and minutes for the duration of the **Customised TBR**.

Tap **Save**.




To save your settings for the **Customised TBR** without starting it, tap **Save**.



The newly programmed **Customised TBR** appears in the list of Temporary Basal Rates.


8.4.3 Activating a TBR

1




To save and start the TBR straight away, tap **Activate**.

2




To activate a saved TBR, tap the desired entry in the list of Temporary Basal Rates.

3



To confirm this step and start the TBR, press the insulin button lit up in green  below the diabetes manager screen.


4



The activated TBR is displayed on the Status screen.


8.4.4 Cancelling a TBR

1




Tap **Cancel TBR?**.

2



Tap **Yes**. The active TBR is cancelled.

3



The TBR is deleted from the Status screen.

9 Replacing System Components

In this chapter you will learn how and when the infusion assembly, the reservoir and the pump base should be replaced.

Replace the system components early in the morning so that you can check your blood glucose level 1 to 3 hours later.

The following table contains guidelines on the period of use of these system components:

System component	To be replaced
Pump base	every 4 months (maximum 120 days)
Reservoir	every 3-4 days (maximum 96 hours)
Infusion assembly	every 2-3 days (maximum 72 hours)


The system components must be disposed of according to local regulations after use.

Note

- ▶ When system components are replaced, the micropump is automatically stopped. Note that this cancels any ongoing boluses and Temporary Basal Rates. If you want to avoid this, delay system component replacement to a later point in time.
- ▶ The average life expectancy of the battery used to supply the micropump with power is 4 days.
- ▶ You will find animated videos on replacing system components and operating the micropump in the Main menu of the diabetes manager under the [Help](#) menu item.

9.1 Starting the Replacement

1



Always start as follows when replacing the infusion assembly, reservoir or pump base:

On the Status screen, tap the area showing the reservoir symbol.

or

In the Main menu, tap the [Replace](#) menu.

9.2 Replacing the Infusion Assembly


Have the following system components at hand for this process:

- ▶ Micropump holder
- ▶ Cannula
- ▶ Insertion device
- ▶ Diabetes manager
- ▶ Disinfectant or sterile alcohol wipe

Start the infusion assembly replacement process by tapping the reservoir symbol  on the Status screen or the [Replace](#)  menu in the Main menu.

Note that any ongoing boluses and Temporary Basal Rates will be cancelled and must be programmed again.


1



The display for selecting the system components appears.

Tap **Infusion assembly**.

2



Tap **Replace**.

The micropump will be stopped.


Removing the used infusion assembly

3




Press the flap to detach the micropump and remove the pump from the infusion assembly.

4



Remove the infusion assembly by loosening the edges of the adhesive pad and pulling it off to the centre.


5



Dispose of the used infusion assembly according to local regulations.

Tap **Done**.

6



If you want to see a video on how to replace the infusion assembly, tap **Help**.

Tap **Done** when you have finished the handling step shown on the screen.

- ▶ Check the pulled out cannula to ensure that it has been completely removed.
- ▶ If you dampen the infusion assembly with warm water or apply an oily ointment, it becomes easier to pull off the adhesive pad.
- ▶ Before you attach the infusion assembly to the body, the new infusion site must be completely dry.
- ▶ A new infusion site should be at least 5 cm away from the last infusion site.

Attaching the infusion assembly to the body


7

See
Chapter 4
in this User's Manual.

Follow the handling step instructions in the section *Attaching the infusion assembly to the body*.

Also observe the warnings and notes given in this chapter as well as the instructions on the screen of the diabetes manager.

8




Attach the micropump to the infusion assembly.

Tap **Next**.

The infusion assembly is filled.

9



To confirm this step and then restart the micropump and return to the Status screen, press the insulin button lit up in green  below the diabetes manager screen.

9.3 Replacing the Reservoir


Have the following at hand for this process:

- ▶ Reservoir assembly
- ▶ Insulin vial with U100 insulin
- ▶ Disinfectant or sterile alcohol wipe


Always fill the reservoir with at least 80 U. The reservoir has a maximum holding capacity of 200 U (2.0 ml).

Initiate reservoir replacement by tapping the reservoir symbol  on the Status screen or the **Replace**  menu in the Main menu.

Note that any ongoing boluses and Temporary Basal Rates will be cancelled and must be programmed again.




1



The display for selecting the system components appears.

Tap **Reservoir**.

2



Tap **Replace**.

The micropump will be stopped.


Removing the reservoir

3



Press the flap to detach the micropump and remove the pump from the infusion assembly.

4



Remove the used reservoir from the pump base.


5



Dispose of the used reservoir according to local regulations.

Tap **Done**.


6



If you want to see an animated video on how to replace the reservoir, tap **Help**.

Once you have completed the steps shown on the display, tap **Done**.

7



Follow the instructions in the sections:

- ▶ *Filling a new reservoir with insulin*
- ▶ *Connecting the reservoir to the pump base*
- ▶ *Setting the reservoir fill amount*
- ▶ *Filling the reservoir needle*
- ▶ *Attaching the Micropump*
- ▶ *Activating the basal rate profile*

9.4 Replacing the Pump Base

The pump base can be used for approximately 4 months. It must not be used for longer than that because otherwise the delivery accuracy may be impaired. You will be reminded regularly to replace the pump base before the deadline expires. For information on the remaining running time of the micropump, see the settings in the [System information](#) menu. If you replace the pump base, you must also replace the reservoir.

Before starting this process, have the following components ready:

- ▶ A new pump base
- ▶ A new reservoir assembly
- ▶ An insulin vial with U100 insulin
- ▶ Disinfectant or a sterile alcohol wipe

Always fill the reservoir with at least 80 U. The reservoir has a maximum holding capacity of 200 U (2.0 ml).

Initiate pump base replacement by tapping the reservoir symbol  on the Status screen or the [Replace](#)  menu in the Main menu.

Note

- ▶ Refer to the information in Chapter 4 *Putting the Micropump Into Operation*.
- ▶ When you replace the micropump, all pump settings saved in the diabetes manager are preserved.
- ▶ **Order a new pump base in time before the operating life expires so that you always have a spare pump base available.**

1



The display for selecting the system components appears.

Tap **Micropump base**. The reservoir is automatically selected as well.

2



The display showing the selected system components appears.

Tap **Replace**.

The micropump will be stopped.


Removing the pump

3



Press the flap of the infusion assembly. Remove the used micropump from the infusion assembly.


4



Dispose of the used reservoir and the used pump base according to local regulations.

Tap **Done**.


5



Follow the instructions on the screen.

Tap **Done**.

6



Follow the handling step instructions in the sections:

- ▶ *Filling the reservoir with insulin*
- ▶ *Connecting the reservoir to the pump base*
- ▶ *Pairing the diabetes manager and micropump*
- ▶ *Setting the reservoir fill amount*
- ▶ *Filling the reservoir needle*
- ▶ *Attaching the micropump*
- ▶ *Activating the basal rate profile*

10 My Data

10.1 Overview

Analysing your therapy data saved in the diabetes manager is an effective way for you and your healthcare professional to determine how your diabetes is developing. This analysis is a valuable tool for making improvements to the treatment of your diabetes.

The diabetes manager generates charts and reports to help you analyse the information saved in the device. Each event in the logbook can be viewed separately. The diabetes manager can also display compilations of therapy data in the form of charts and overviews.

10.2 Logbook

You can display each single logbook entry on the screen of the diabetes manager. There, you will find all the information about blood glucose results along with time of test, carbohydrates, health events and boluses. In addition, you can change or add logbook entries.

The logbook saved in the diabetes manager replaces the handwritten blood glucose diary you may have kept in the past. The diabetes manager automatically stores up to 5,000 logbook entries with the time and date. You can view the most recent 250 logbook entries in the diabetes manager. If you are using a PC with compatible software, you can view all logbook entries.


A logbook entry can contain the following:
date and time, blood glucose result, time of test, carbohydrate intake, health events, bolus amounts, bolus type and notes

Note the following:

- ▶ The **Logbook** display shows the entries in the order in which they occurred with the most recent entry shown on top.
 - ▶ Blood glucose results cannot be entered manually or be subsequently adjusted.
 - ▶ If you want to add data to a logbook entry, you can also tap the **Add data** button on the Status screen or in the Main menu.
 - ▶ Bolus data from the micropump is automatically saved on the diabetes manager. However, the bolus advice feature will initially treat quick boluses that you deliver spontaneously as correction insulin. Therefore, you should edit the quick boluses recorded in the logbook with regard to bolus distribution (meal/correction insulin) and carbohydrates consumed.
 - ▶ You should enter any boluses that were delivered independently of bolus advice using an insulin pen or syringe as new data in the logbook.
 - ▶ The information remains saved in the logbook when the rechargeable battery is replaced. You must, however, check whether the time and date of the diabetes manager are still correct after replacing the battery.
 - ▶ Once 5,000 entries have been saved in the logbook, adding a new entry causes the oldest logbook entry to be deleted. Save the entries on a PC if you want to keep all entries.
 - ▶ Although control results are saved in the diabetes manager, they can only be viewed on a computer with suitable software.
- ▶ Before reviewing logbook entries on a PC, you first have to transfer the saved logbook entries to a PC that has specific diabetes management software.
 - ▶ Logbook data that has been used for bolus advice cannot be subsequently adjusted (except notes).

10.2.1 Understanding the Logbook

Main menu > My data > Logbook




- 1 **Column for displaying blood glucose information**
Blood glucose result range and blood glucose result.
- 2 **Column for displaying bolus information**
Bolus type, bolus advice and bolus amount.
- 3 **Column for displaying carbohydrate information**
Symbol for carbohydrates and carbohydrate amount.
- 4 Symbol for blood glucose result range information
- 5 Symbol for bolus type
- 6 Symbol for bolus advice
- 7 Symbol for carbohydrates
- 8 Carbohydrate amount
- 9 Bolus amount
- 10 Blood glucose result
- 11 Symbol for Stop mode
- 12 Symbol for health event
- 13 Symbol for general time of test
- 14 Symbol for note
- 15 Time of entry
- 16 Date of entry

My Data

Symbol	Symbol name	Description
	Blood glucose result range	The symbol colours have the following meaning: Blue  : above the target range; green  : within the target range Yellow  : below the target range; red  : below the hypo warning limit
	Standard bolus	Bolus insulin from a standard bolus
	Extended bolus	Bolus insulin from an extended bolus
	Multiwave bolus	Bolus insulin from a multiwave bolus
	Basal insulin	Basal insulin from an injection
	Manual bolus with pen/syringe	Bolus was delivered using an insulin pen or syringe.
	Bolus advice accepted	Bolus advice from the diabetes manager was accepted.
	Bolus advice not accepted	Bolus advice from the diabetes manager was changed prior to delivery.
	Carbohydrates	Carbohydrate data exists for the logbook entry.
	Time of test	Data regarding the time of test exists for the logbook entry.
	Health event	Health event data exists for the logbook entry.
	Pump stopped	The micropump was stopped.
	Notes	You entered a note.

10.2.2 Viewing and Adjusting Logbook Data


1



Scroll the display upwards or downwards to view additional logbook entries.

Tap a logbook entry if you want to view or adjust the details.

2




After tapping a logbook entry, the **Logbook entries** display shown above appears.

Tap the entry you want to view or adjust.

Time of test

3




On the **Logbook entries** display, tap **Time of test**. Tap a time of test (for example, **Before meal**).

Tap **Save**.

Carbohydrates

4




On the **Logbook entries** display, tap **Carbohydrates**. Enter the amount of carbohydrates you consumed.

Tap **Save**.

Health events

5




On the **Logbook entries** display, tap **Health events**.

Tap the appropriate entries. You can select up to 4 health events.

Tap **Save**.

Bolus

6



On the **Logbook entries** display, tap **Bolus**.


The display informs you about the bolus delivered.

Tap **OK**.

Note

The bolus advice feature initially treats quick boluses as correction insulin. Mark the quick boluses in the logbook as a meal bolus or correction bolus according to their purpose. Enter consumed carbohydrates in the logbook.

7




On the **Logbook entries** display, tap **Note**.

Type a note to save with this entry.

Tap **Done**.


10.2.3 Adding New Data

1



On the **Logbook** display, tap **+** to add new data to the logbook.

2



Tap the entries you want to add. Note the instructions on the next pages.


Then tap **Save**.

Note

You can also access the **Add data** display by tapping the **Add data** function button on the Status screen or in the Main menu.

10.3 Trend Graph


The diabetes manager uses the trend graph to display your blood glucose trends, basal rates, boluses, carbohydrates and other information. The logbook entries for the time period selected by you are used. Using the ◀▶ buttons, you can move backwards or forwards in time according to the selected time scale.



1	Time period	Illustrated time period
2	Events	Displays health events or indicates that the micropump was stopped.
3	Basal rate	Blue line: The distance from the top edge of the graph shows the basal rate.
4	Bolus	Pink bars: Show the amounts of bolus insulin
5	Blood glucose result	Crosses (x): Individual blood glucose results that are connected by lines
6	Carbohydrates	Brown bars: Show the carbohydrate amount consumed
7	Hypo warning limit	Red line: Warning limit for hypoglycaemia
8	Value below target range	Yellow line: Limit value for values that are below the target range but that are still above the hypo warning limit.
9	Blood glucose target range	Green area: Range between the lower and upper blood glucose threshold
10	Hyper warning limit	Blue line: Warning limit for hyperglycaemia
11	Selected time of test	Displays the logbook entries made for this time of test.

Basal insulin amount

Bolus values



Blood glucose results

Time
scaleCarbohydrate
amount

On the **upper left side (U)** of the graph, the basal amount is displayed. By means of the scale, you can read the basal insulin amounts, which are represented by the blue basal insulin line. Scaling depends on the largest basal insulin amount delivered during the selected time period.

Example

The scale comprises a range of 1, 2, 5, 10, 20 and 40 U. For example, if the largest basal insulin amount in the selected time period is 3 U, the scale will show the range from 0 to 5 U.

On the **lower left side (mmol/L)** of the graph, the blood glucose result is displayed. By means of the scale, you can read the blood glucose results, which are represented by a black line. The black line connects the individual test results indicated by a cross. Scaling depends on the largest blood glucose value measured during the selected time period.

On the **upper right side (U)** of the graph, the bolus amount is displayed. By means of the scale, you can read the bolus values of the pink bars. Scaling depends on the largest bolus delivered during the selected time period.

Example

The scale comprises a range of 1, 5, 15, 30 or 60 U. For example, if the largest bolus delivered during the selected time period is 8 U, the scale will show the range from 0 to 15 U.


On the **lower right side (g)** of the graph, the carbohydrate amount is displayed. By means of the scale, you can read the carbohydrate values, which are represented by the brown bars. Scaling depends on the largest carbohydrate amount consumed during the selected time period.

Example

The scale comprises a range of 40, 80, 120, 160, 200 and 240 g, or the equivalent scale for BE, KE or CC. For example, if the largest carbohydrate amount in the selected time period is 86 g, the scale will show the range from 0 to 120 g.


10.3.1 Viewing the Trend Graph

1



Tap ☒ to change the trend graph representation.


2



Tap the setting you want to change.

Time scale

3




Tap the desired time period.

Tap **Save**.

Graph options

4




Tap one or more of the available graph options (for example, **Bolus**) to be displayed in the graph.

Tap **Save**.

Time of test

5



Tap the appropriate time of test (for example, **After meal**).


Tap **Save**.

On the **Time of test** display you can select which logbook entries are to be represented. Only the logbook entries are displayed for which you entered a certain time of test.

10.4 Standard Week

Standard week graph

The diabetes manager uses the standard week graph to display your blood glucose averages, the individual tests and the standard deviations for the days of a standard week. Using the ◀▶ buttons, you can move backwards or forwards in time according to the selected time scale.




1	Time period	Represented time period, for example, 7 days
2	Selected time of test	Displays the logbook entries made for this time of test.
3	Warning limit for hyperglycaemia	Red line: Hyper warning limit
4	⊗	Average
5	x	Single blood glucose results
6	Blood glucose unit of measurement	mmol/L
7	Day of week	Monday to Sunday

The standard deviation indicates how the blood glucose results are scattered around the blood glucose average. A high standard deviation means that the blood glucose results are scattered away from the blood glucose average.

Standard week table

Tap  to switch to the standard week table. The standard week table displays the data of the standard week graph in table format. To return to the standard week graph, tap . Using the   buttons, you can move backwards or forwards in time according to the selected time scale.




	mmol/L				#
	Min	Ø	Max	SD	Tests
Mon	2.4	6.8	10.7	1.8	28
Tue	3.1	7	19.1	3	26
Wed	2.4	6.9	12.7	2.5	23
Thu	2.4	6.6	19.1	2.8	31
Fri	2.4	6.1	11.7	2.5	28
Sat	2.6	6.8	19.1	3.4	22
Sun	2.4	7.7	15.3	2.8	26

- 1 Time period
- 2 Selected time of test
- 3 Day of week
- 4 Lowest blood glucose result of the day of the standard week
- 5 Blood glucose average of the day of the standard week
- 6 Highest blood glucose result of the day of the standard week
- 7 Standard deviation (SD) of the day of the standard week
- 8 Number of tests on the day of the standard week


10.4.1 Displaying and Configuring the Standard Week

1



Tap to change the settings that determine how your standard week data is displayed in the graph.


2



Tap one of the entries available for selection.
Make the desired settings.
Tap **Done**.

Time scale


3



Tap the desired time period.
Tap **Save**.

Time of test

4




Tap the appropriate time of test.
Tap **Save**.

10.5 Target Range

The diabetes manager displays a pie chart and a table illustrating your blood glucose results for the time period and time of test you selected. The chart is divided into the following blood glucose result ranges: *Above*, *Within*, *Below* and *Hypo*. Using the ◀▶ buttons, you can move backwards or forwards in time according to the selected time period.


Target range chart



1	Time period	Illustrated time period
2	Selected time of test	Displays the logbook entries made for this time of test.
3	Target range chart	Shows the target ranges of blood glucose results as a pie chart.
4	Percentage distribution	Shows the percentage of blood glucose results in each blood glucose result range.
5	Blood glucose result ranges	Coloured representation of the blood glucose result ranges

Tap  to switch to the target range table. The target range table displays the target range data in table format. To return to the target range chart, tap . Using the   buttons, you can move backwards or forwards in time according to the selected time period.


Target range table



1	Time period	Illustrated time period
2	Buttons for scrolling left or right	Moves forwards or backwards in time according to the selected time period.
3	Blood glucose result ranges	Coloured representation of the blood glucose result ranges
4	Target range table	
5	Number	
6	Times of test	Shows the logbook entries for this time of test.


10.5.1 Displaying and Configuring Target Range Data

1



Tap ☒ to change the settings that determine how your target range data is displayed in the chart.

2




Tap one of the entries available for selection.

Make the desired settings and then tap **Done**.

Time scale

3




Tap the desired time period.

Tap **Save**.

Time of test

4



Tap the appropriate time of test (for example, **Bedtime**).

Tap **Save**.

10.6 BG Averages Table

The BG averages table displays your blood glucose averages and standard deviations (SD) for the time period and time of test you selected.


Blood glucose averages table

	mmol/L	SD	Tests
	888.8	888.8	888.8
	888.8	888.8	888.8
	888.8	888.8	888.8
	888.8	888.8	888.8
	888.8	888.8	888.8
	888.8	888.8	888.8

1	Time period	Illustrated time period
2	Buttons for scrolling left or right	Moves forwards or backwards in time according to the selected time scale.
3	Number of tests	Used to calculate the averages and standard deviation of each time of test.
4	Blood glucose standard deviation	
5	Blood glucose average	
6	Times of test	Data for all test results that are marked with the respective symbol for a certain time of test.

10.6.1 Viewing the BG Averages Table

1




Blood glucose averages

14 days
29 Mar 2018 - 12 Apr 2018

	mmol/L	SD	Tests
	888.8	888.8	888.8
	888.8	888.8	888.8
	888.8	888.8	888.8
	888.8	888.8	888.8
	888.8	888.8	888.8
	888.8	888.8	888.8

Tap to change the settings for the time period or time of test.

2



Time scale


- ☐ 7 days
- ☐ 14 days
- ☒ 30 days
- ☐ 60 days
- ☐ 90 days

Save

Tap the desired time period.

Tap **Save**.

3



Time of test

- ☐ All
- ☐ Before meal
- ☐ After meal
- ☒ Bedtime
- ☐ Fasting
- ☐ Other

Save

Tap the appropriate time of test (**All**, **Before meal**, **After meal**, **Bedtime**, **Fasting** or **Other**).

Tap **Save**.

11

Changing Settings

In the *Settings* menu, you can make changes to the factory settings or change the settings you made. This allows you to adjust the micropump system to your individual treatment requirements and your personal preferences.

For information on changing time blocks, health events and on settings for insulin sensitivity, carbohydrate ratio, meal rise, snack size, acting time and offset time, see the chapter *Bolus Advice*.

For information on reminder settings, see the chapter *Reminders*.

Note


When you edit a setting, any unsaved changes will be discarded when the diabetes manager turns off or a test strip is inserted into the test strip slot.

11.1 Making Settings

Main menu > Settings

WARNING


The therapy settings must be specified by your healthcare team and you may only change them after prior consultation. Otherwise, there is a risk of experiencing hyper- or hypoglycaemia.



11.1.1 Warning Limits


Main menu > Settings > Warning limits

You can set blood glucose warning limits for hypoglycaemia or hyperglycaemia. When your blood glucose result is below the hypo warning limit or above the hyper warning limit, the diabetes manager displays an appropriate warning.



Tap **Upper warning limit** or **Lower warning limit**.


Upper limit value



Use **-** and **+** to set the **Upper warning limit (hyper)**.


Tap **Save**.

Lower limit value



Use **-** and **+** to set the **Lower warning limit** (hypo).


Tap **Save**.



Tap **←** to return to the list of settings. If you want to make additional settings, tap the respective entry.

Remaining running time

The warning limit for the remaining running time of the micropump indicates the number of **days** as of which you want to receive a warning regarding the remaining running time of the micropump.




Tap **Micropump remaining running time**.

Reservoir level

The warning limit for the reservoir level indicates the number of insulin units as of which you want to receive a warning regarding the remaining insulin amount. The remaining insulin amount is the insulin amount that is still available in the micropump reservoir.


2



Use **-** and **+** to set the number of **days** as of which you want to receive a warning.

Tap **Save**.

1



Tap **Reservoir level**.

2



Use **-** and **+** to set the number of insulin units as of which you want to receive a warning.

Tap **Save**.


Automatic off




The automatic off function is a safety feature for emergency situations. If you have not touched any button on your micropump and not operated the diabetes manager for the specified number of hours, the micropump stops insulin delivery. In this case, it can be assumed that you are no longer able to stop the pump yourself, for example, due to severe hypoglycaemia.

The automatic off function is turned off by default.

Note

It is advisable to set a time span for the automatic off function that is longer than your usual bedtime period.






11.1.2 Time and Date


Main menu > Settings > Time and date

WARNING

Having the time and date set precisely is essential in order for your micropump system to function properly. A wrong time setting can have serious consequences.




3



Use $-$ and $+$ to set the hours and minutes for the current time.

Tap **Save**.


4



Use $-$ and $+$ to set the day, month and year.

Tap **Save**.

5



Then tap **Save**.

11.1.3 Bolus Settings

Main menu > Settings > Bolus settings

This section provides information on the following settings:

- ▶ Quick bolus buttons
- ▶ Maximum quick bolus
- ▶ Quick bolus increment
- ▶ Maximum bolus amount
- ▶ Delivery lag time

Note


If you turn off bolus advice, it must be set up again after turning it on the next time.

Quick bolus buttons

A quick bolus is a bolus that can be programmed and delivered using the quick bolus buttons on the micropump.

The factory setting for the quick bolus buttons is *on*.

For more information on the quick bolus, see the section *Quick Bolus* in the chapter *Delivering a Bolus*.



Tap **Quick bolus buttons** to activate or deactivate the quick bolus buttons of the micropump.


If you activate the **Quick bolus buttons**, you can program and deliver a quick bolus directly from the micropump, for example.

Maximum quick bolus


You use this setting to define the maximum insulin amount that may be delivered with a quick bolus.

Note

You cannot define a maximum quick bolus that is greater than the maximum bolus.



Tap the **Maximum quick bolus** entry to set the maximum bolus amount that can be programmed.



Use **−** and **+** to set the maximum amount for the quick bolus.

Tap **Save**.

Quick bolus increment

The quick bolus increment indicates the amount by which your insulin dose is increased with each press of the quick bolus buttons while programming a quick bolus.


You can set the following quick bolus increments:

0.2 U
0.5 U
1.0 U
2.0 U

Note down the set quick bolus increment in the log *Settings for the Accu-Chek Solo micropump system*.

Example

You have set a quick bolus increment of 0.2 U. If you want to program a bolus of 2.0 U using this increment, you must press the quick bolus buttons on the micropump 10 times.



Tap **Quick bolus increment**.

Tap the desired **Quick bolus increment**.


Tap **Save**.

Maximum bolus amount

This setting specifies the maximum insulin amount that may be delivered in any bolus. A bolus that exceeds the maximum bolus amount requires additional confirmation or must be reduced.

Note

For the maximum bolus, you can set an amount between 0 U and a maximum of 25 U in increments between 0.05 U and 1.0 U (depending on the total amount).




Delivery lag time

In some situations (for example, if you have gastroparesis) it may be helpful to only start a meal bolus after you have started eating. For each meal bolus, you can use the delivery lag time setting to specify a delay of up to 60 minutes between programming a bolus and the actual start of bolus delivery.

Note

You can set the desired duration when programming a bolus. You can set a duration of 0, 15, 30, 45 or 60 minutes. Delivery of a bolus that contains correction insulin cannot be delayed.




11.1.4 Tone and Vibration

Main menu > Settings > Tone and vibration

You can define how the diabetes manager should attract your attention in case of an event (for example, a warning). You can choose whether the diabetes manager issues an acoustic signal, vibrates or both. The settings you make are called *signal mode* in the rest of this User's Manual.


You can also suspend the signals for warnings for a specific period of time (for example, over night).

You can make the tone and vibration settings in the **Settings** menu under the **Tone and vibration** entry. The **Signals**  menu in the Main menu allows you to easily access the appropriate **Signal mode** and the **Turn off signals** feature. You can select the **Quiet signal mode** in a quiet environment, for example.



Tap **Tone and vibration**.

Signal modes



Tap **Loud**, **Normal**, **Quiet** or **Vibration** to set the desired signal mode.

Then tap **Done**.

Volume

3



Set the volume for the default setting by moving the slider to this position: **Loud** to the right-hand side, **Normal** to the centre or **Quiet** to the left of the centre. Turn vibration on or off.


Tap **Save**.

Note

When the volume slider is set to 0 percent in the far left position, vibration is automatically turned on.

Vibration

4




Tap **Vibration**. The volume slider is on the far left.

Tap **Save**.

Blood glucose test

5




Tap **Blood glucose test** to turn the signals for a blood glucose test on or off.

When you have made the desired setting, tap **Done**.


Note**Signals during a blood glucose test:**

- ▶ Insert test strip
- ▶ Blood application detected
- ▶ Test completed

Touchscreen feedback


Tap **Touchscreen** to set how the diabetes manager reacts when you make a selection using the touchscreen.

When you have made the desired setting, tap **Done**.



Tap the desired touchscreen setting.

Tap **Save**.

Insulin delivery signal

Tap **Insulin delivery signal**.

If you turn on the signal, the diabetes manager issues a signal when you confirm delivery of a basal rate or bolus.

Tap **Done**.

Note

Signals for system messages:


- ▶ USB plug connected to USB socket
- ▶ USB plug removed from USB socket
- ▶ Diabetes manager restart
- ▶ *Bluetooth* communication completed

11.1.5 Turning Off Signals

You can turn off the signals for *warnings* for a certain period of time. However, since *error* and *maintenance* messages require your immediate attention, you cannot suspend the signals for these events. You can set up signal deactivation as a one-time event or as an event that is repeated at the same time every day.


Note

Warnings that occur while the signals are turned off are displayed once the diabetes manager is turned on or the signal deactivation period ends.




In the Main menu, tap **Signals**.

2




Tap **Turn off signals**.

3



Tap the **Status** switch to turn on the **Turn off signals** feature.

4




Tap **Start time** to set the start of signal deactivation.

5




Use **−** and **+** to set the hours and minutes for the start time.

Tap **Save**.




Tap **End time** to set the end of signal deactivation.



Use **-** and **+** to set the hours and minutes for the end time.

Tap **Save**.



Tap **Frequency** to set the repetition interval for turning off the signals.

Tap **Done**.

Note

With the **Once** setting, the signals for warnings are turned off only once for the time period you specified. With the **Repeat** setting, the signals for warnings are turned off daily in the time period you specified.

Once the signal deactivation time has expired, the signals for the warnings that occurred are issued again.


11.1.6 General Settings

Main menu > Settings > General settings


You can make the following settings in the General settings:

- ▶ **Language**
You can select the language for the texts displayed on the screen from a predefined list of languages.
- ▶ **Brightness**
You can adjust the brightness of the diabetes manager screen for different lighting conditions.
- ▶ **Check system functions**
For more information, see the chapter *Care and Maintenance*.

Language




Tap **Language** or **Brightness** to make the desired settings.



Tap the entry for the language in which you want the menus and texts to appear on the screen.

Brightness


1



Set the **Brightness** of the screen. Move the slider to the left to decrease the **Brightness** or to the right to increase the **Brightness**.

Tap **Save**.

2



Tap **Done** to return to the list of settings.

11.1.7 Screen Lock

Main menu > Settings > Screen lock

The diabetes manager is equipped with a screen lock, which can be used to protect the device against unauthorised access. You can define a personal identification number (PIN) to be used for access. The PIN is a secret number with four to eight digits that you can enter and change in the **Screen lock** menu. By factory default, the PIN is “0000”.



WARNING


To prevent unauthorised access, you should always leave the screen lock turned on.

Note

- ▶ The screen lock is turned on by factory default.
- ▶ Choose a PIN that you can easily memorise and enter.


Turning off the screen lock

1



Tap **Screen lock**.

2




Tap **Status (PIN)** to turn off the screen lock.

Note

If you want to change the PIN, you have to turn the screen lock off and on again.


3



When you turn off the screen lock, the above display appears for your information.


Tap **Yes** if you do **not** want to enter a PIN or want to change the PIN.

Turning on the screen lock



Tap **Status (PIN)** to turn on the screen lock.


Entering the PIN



Enter a PIN with 4 to 8 digits.

Tap **Save**.

Confirming the PIN



Enter the PIN again to confirm it.


Tap **Save**.

Note

If you have forgotten your PIN, you have to enter the PIN unlock code instead of the PIN to unlock the diabetes manager. You will find the PIN unlock code on the back of this User's Manual.

Changing the background image


1



Tap **Background image** to set the pattern for the background image of the active screen lock.

Then tap **Done**.

1



Tap the tile with the desired background image.

Tap **None** if you do not want to have a background image.

Tap **Done**.

11.1.8 System Information

Main menu > Settings > System information


The system information provides various details on the micropump system:

- ▶ Serial number of the diabetes manager
- ▶ Firmware of the diabetes manager
- ▶ Firmware of the blood glucose meter in the diabetes manager
- ▶ Radio frequency signal of the RFI firmware
- ▶ Selected language
- ▶ Status of rechargeable battery
- ▶ Date of manufacture
- ▶ Diabetes manager hardware update
- ▶ Number of resets/times switched on/off
- ▶ Number of blood glucose tests/control tests
- ▶ Serial number, micropump battery status, firmware version of micropump

In addition, you can read legal information and the License Terms and Conditions.


Some of this information is primarily used by technical customer support and may be requested by your Customer Support and Service Centre.

1



Tap **System information**.


2



The list of **System information** is displayed.


Scroll the display upwards to see additional **System information**.

3



Tap the **Legal information** entry to view the stored license agreements.

4



Scroll the display upwards to be able to read more text.

Tap ← to return to the previous display.

11.2 Travelling and Flight Mode

Time zone changes

Having the date and time set accurately is very important for the Accu-Chek Solo micropump system to work correctly. If you travel across different time zones, you may need to adjust the time setting and also possibly the date. Discuss the settings that must be changed for your journey to other time zones with your healthcare team in advance.

For information on how to change the date and time settings of the micropump system, see the section *Setting the time and date*. If you change the time of the micropump system, the basal rate will be delivered according to the time set. This also holds true for changing the clocks in summer and winter time.

Example

You change the time of the micropump system from 10:00 to 13:00. After the change, the micropump delivers the basal rate for that time at 13:00.

Flight mode


You can turn on flight mode for flights that do not allow *Bluetooth* wireless technology from electronic devices. Flight mode enables the diabetes manager to comply with the airline regulations and ensures that the wireless functions of the diabetes manager cannot be used.

If the quick bolus feature was activated, you can continue to deliver boluses using the quick bolus buttons on the micropump. As soon as flight mode is turned off, the diabetes manager and micropump synchronise and update the event data.

For information on using the micropump system with a pen or syringe, see the chapter *Injection Therapy Mode*.


11.2.1 Turning On Flight Mode

1



In the Main menu, tap **Flight mode**.


2



Tap **Status** so that the switch is in the **ON** position.


Tap **Done**.

3



Confirm the displayed information by tapping **OK**.

4



The  symbol in the status bar indicates that flight mode is turned on.


No current micropump data is displayed in flight mode.

11.2.2 Turning Off Flight Mode


Note

It is not possible to turn off flight mode on the micropump using the diabetes manager.

To turn off flight mode on the pump, you must use the quick bolus buttons.




In the Main menu, tap the **Flight mode** menu.




Tap **Status** so that the switch is in the **OFF** position.

Tap **Done**.



Confirm the displayed information by tapping **OK**.

Turning off flight mode on the micropump




Press both quick bolus buttons on the micropump simultaneously 3 consecutive times within 10 seconds.




When flight mode is turned off, the micropump issues a sequence of signals.

Tap **Next**.



Tap **Done**.



Flight mode is turned off.

The flight mode symbol is no longer displayed in the status bar.

12 Reminders

Main menu > Settings > Reminders

You can set reminders for specific appointments and events. This can be useful when you have to make specific preparations, for example, for replacing the infusion assembly. A message on the screen and a signal attract your attention to the respective reminder at the set time.

You can select any tone from the option list for each reminder. In the factory settings, all reminders are deactivated. You can turn the reminders on or off by using the ON and OFF switch.

You can make different settings for the various reminder types:

Setting	Explanation
	This switch position indicates that the reminder is turned off (deactivated).
	This switch position indicates that the reminder is turned on (activated).
Time	The time of day the reminder occurs.
Date	The date the reminder occurs.
Remind after	Period of time following an event (for example, blood glucose value being too high) after which a reminder is to occur.
Tone	The tone that is used for the reminder.
Frequency	One-time reminder or regular reminder that is to occur every day at the same time.

12.1 Types of Reminders

Reminder	Explanation
Replace infusion assembly	Reminds you to replace the infusion assembly after a specified number of days.
Alarm clock/Customised	The alarm clock sounds at the specified time. The alarm clock reminds you, for example, to test your blood glucose or about any other daily appointment.
Test blood glucose	Reminds you to test your blood glucose at a specified time.
After meal	Reminds you to test your blood glucose after eating if you have previously marked a blood glucose result as Before meal . The diabetes manager turns on after the specified amount of time and displays the reminder (as long as a test strip has not been inserted).
Test after low blood glucose result	Reminds you to test your blood glucose if your test result was below the hypo warning limit (●).
Test after high blood glucose result	Reminds you to test your blood glucose if your test result was above the target range (●).
Missed bolus	This reminder occurs if no bolus was delivered within 2 hours prior to the programmed time.
Inject basal insulin	Reminds you to deliver basal insulin (available in injection mode only).
Healthcare professional visit, Lab test	Reminds you of a healthcare professional visit or lab test.

12.2 Programming Reminders

Set the desired time and frequency for each reminder. If you choose **Repeat**, you will be reminded of the event at certain intervals, for example, daily.


Some reminders only appear when certain conditions are met, for example, depending on a previous test result.

You can add more reminders by tapping **+**. Once the maximum number of reminders has been reached, you will see the information that no more reminders can be added, instead of the **+** symbol.

By tapping , you can delete an existing reminder.

You can assign a tone from an option list to each programmed reminder. Tap  in the option list to listen to the tone.


Reminder tone



Tap the desired tone to use for the reminder.

Tap  to listen to the tone beforehand.

Tap **Save**.




Choose **Once** or **Repeat**.


If you choose **Repeat**, you will be reminded daily, every 2 days or every 3 days.

Tap **Save**.

12.2.2 Reminder: Replace Infusion Assembly




Tap **Replace infusion assembly**.




Tap the **Status** switch so that it is in the **ON** position. Tap **Recurrence**, **Time**, **Tone** to make the desired settings.

Once you have made all settings, tap **Done**.

12.2.1 Reminder: Alarm Clock




Tap **Alarm clock 1**.



Tap  to assign a name to the alarm clock reminder.

12.2.3 Reminder: Test Blood Glucose


3



Enter a name for the alarm clock reminder using the keyboard. The name can be up to 8 characters long.

Tap **Done**.

4




Tap the **Status** switch so that it is in the **ON** position. Tap **Time**, **Tone** or **Frequency** to make the desired settings.

Once you have made all settings, tap **Done**.

Note

When you test your blood glucose, the diabetes manager dismisses any blood glucose test reminders that are pending within the next 30 minutes. If necessary, a new reminder is scheduled based on the blood glucose result.

1



Tap **Test blood glucose**.