



June 2, 2017

Gregory Czumak
American Certification Body Inc.
6731 Whittier Ave
McLean, VA 22101

RE: Comments of April 21, 2017
APPLICATION: MEDTRONIC-ZEPHYR TECHNOLOGIES
FCC ID: VZ6-BH3

Gregory Czumak:

Below are the comments that you have provided regarding the application for certification referenced above. Our responses to those comments are in ***bold italic***. Many responses refer you to additional exhibit(s) which has been uploaded to the application folder at the ATCB website.

Thank you for your attention. Please feel free to contact us for any additional information that you may require.

Regards,

Kimberly Colbert
Washington Laboratories
Administrative Assistant

If the EUT contains a second transmitter, in addition to the one being authorized in this application, as is indicated in other exhibits submitted, and the two transmitters are capable of transmitting simultaneously, please confirm that the second transmitter was, indeed, transmitting while radiated measurements were performed on the EUT, as required by Section 15.31(h).

R. Transmitters are not capable of transmitting simultaneously.