


# Dipole Intermediate Check


2600 MHz

| Dipole D2600V2 SN: 1162 |                 |                      |                        |                |
|-------------------------|-----------------|----------------------|------------------------|----------------|
| Head Liquid             |                 |                      |                        |                |
| Test By: Allen Sun      |                 | Checked By: Bard Liu |                        |                |
| Date of Measurement     | Return Loss(dB) | $\Delta\%$           | Impedance ( $\Omega$ ) | $\Delta\Omega$ |
| 2022.08.22              | -23.2           | N/A                  | 47.7                   | N/A            |
| 2023.08.21              | -23.8           | 2.586                | 45.5                   | 2.2            |
| 2024.08.20              | -20.3           | 12.5                 | 46.5                   | 1.2            |

D2600V2 - SN: 1162 (Date of Measurement: 2023/08/21)



D2600V2 - SN: 1162 (Date of Measurement: 2024/08/20)




# Dipole Intermediate Check


5250 MHz

| Dipole D5GH2V2 SN: 1296 |                 |                      |                        |                |
|-------------------------|-----------------|----------------------|------------------------|----------------|
| Head Liquid             |                 |                      |                        |                |
| Test By: Allen Sun      |                 | Checked By: Brad Liu |                        |                |
| Date of Measurement     | Return Loss(dB) | $\Delta\%$           | Impedance ( $\Omega$ ) | $\Delta\Omega$ |
| 2022.08.17              | -30.1           | N/A                  | 48.5                   | N/A            |
| 2023.08.16              | -32.3           | 0.664                | 47.1                   | 1.4            |
| 2024.08.15              | -27.2           | 9.635                | 47.5                   | 1              |

D5GH2V2 - SN: 1296 (Date of Measurement: 2023/08/16)



D5GH2V2 - SN: 1296 (Date of Measurement: 2024/08/15)




# Dipole Intermediate Check

5750 MHz

| Dipole D5GH2V2 SN: 1296    |                 |                               |                        |                |
|----------------------------|-----------------|-------------------------------|------------------------|----------------|
| Head Liquid                |                 |                               |                        |                |
| Test By: <i>Kellen Sun</i> |                 | Checked By: <i>Brenda Lin</i> |                        |                |
| Date of Measurement        | Return Loss(dB) | $\Delta\%$                    | Impedance ( $\Omega$ ) | $\Delta\Omega$ |
| 2022.08.17                 | -30.1           | N/A                           | 49.2                   | N/A            |
| 2023.08.16                 | -30.7           | 1.993                         | 52.8                   | 3.6            |
| 2024.08.15                 | -29.3           | 2.658                         | 47.7                   | 1.5            |

D5GH2V2 - SN: 1296 (Date of Measurement: 2023/08/16)



D5GH2V2 - SN: 1296 (Date of Measurement: 2024/08/15)

