



1516 Centre Circle, Downers Grove IL 60515

RF EXPOSURE EXHIBIT		
Manufacturer Name	Roche Diabetes Care, Inc.	
Manufacturer Address	9115 Hague Road Indianapolis, IN 46256	
Model No.	Accu-Chek Instant	
Specifications	FCC "Code of Federal Regulations" Title 47 Part 1.1307(b) FCC "Code of Federal Regulations" Title 47 Part 2.1093	
Test Facility	Elite Electronic Engineering, Inc. 1516 Centre Circle, Downers Grove, IL 60515	FCC Reg. Number: 269750 IC Reg. Number: 2987A CAB Identifier: US0107
Test Procedure	FCC KDB 447498 D01 General RF Exposure Guidance v06	
Limits	The 1-g and 10-g SAR test exclusion thresholds for 100MHz to 6GHz at test separation distances ≤ 50 mm are determined by the following: [(max. power of chan., inc. tune-up tol., mW)/(min. test sep. distance, mm)]*[$\sqrt{f(\text{GHz})}$] f(GHz) = RF channel transmit frequency in GHz. power is rounded to the nearest mW distance is rounded to the nearest mm result is rounded to 1 decimal place for comparison	
This is a portable device. The maximum peak conducted output power measured -3dBm. The declared maximum output including acceptable tolerances is 0 dBm. Maximum antenna gain is 3.4 dBi Maximum peak EIRP = 0 dBm + 3.4 dBi = 3.4 dBm = 2.18mW Exclusion Threshold: (2.18mW / 5mm) * $\sqrt{2.402\text{GHz}}$ = 0.675 0.675 is ≤ 3.0 for 1-g SAR and ≤ 7.5 for 10-g extremity SAR This device meets the SAR exclusion requirement.		