

HAC_E_Dipole_2600**DUT: HAC Dipole 2600 MHz**

Communication System: UID 0, CW (0); Frequency: 2600 MHz; Duty Cycle: 1:1

Medium: Air Medium parameters used: $\sigma = 0$ S/m, $\epsilon_r = 1$; $\rho = 0$ kg/m³

Ambient Temperature : 23.3 °C;

DASY5 Configuration:

- Probe: EF3DV3 - SN4050; ConvF(1, 1, 1); Calibrated: 2023/1/24

- Sensor-Surface: (Fix Surface)

- Electronics: DAE4 Sn1649; Calibrated: 2023/4/24

- Phantom: HAC Test Arch with AMCC; Type: SD HAC P01 BA

- Measurement SW: DASY52, Version 52.10 (4); SEMCAD X Version 14.6.14 (7501)

E Scan - measurement distance from the probe sensor center to CD2600 =

15mm/Hearing Aid Compatibility Test at 15mm distance (41x181x1): Interpolated grid:

dx=0.5000 mm, dy=0.5000 mm

Device Reference Point: 0, 0, -6.3 mm

Reference Value = 67.58 V/m; Power Drift = -0.09 dB

PMR not calibrated. PMF = 1.000 is applied.

E-field emissions = 83.85 V/m

Average value of Total=(83.85+83.41)/2=83.63 V/m

PMF scaled E-field

Grid 1 M3 81.99 V/m	Grid 2 M3 83.41 V/m	Grid 3 M3 82.18 V/m
Grid 4 M3 77.52 V/m	Grid 5 M3 78.72 V/m	Grid 6 M3 78.22 V/m
Grid 7 M3 81.84 V/m	Grid 8 M3 83.85 V/m	Grid 9 M3 82.95 V/m

Total = 83.85 V/m

E Category: M3

Location: -0.5, 23, 9.7 mm

