



GN Hearing A/S

Ref: Justification for placing the FCC ID in the user manual and packaging for the following device:

- FCC ID: X26CABR80

There are two main reasons why device stated above does not have FCC ID:

- 1) Due to the specific use of a typical hearing instrument, its external surface is in contact with the user's skin in the ear. Therefore, it is not appropriate to permanently affix a label on the outer surface of the device to prevent skin irritations.
- 2) The only permanent location is the battery door, but it is too small to have the visible text (smaller than 4-6 points).

Therefore, the FCC ID information is included in the user manual and sales package together with the device.