



Dräger Polytron[®] 6100 EC WL, Dräger Polytron[®] Repeater ISA

Notes on Approval

Device marking

1 Device marking



Model	Product	Type of communica-tion
RC001	Polytron 6100 EC WL Polytron Repeater ISA	ISA100

Model	Product	Type of communica-tion
RC002	Polytron 6100 EC WL Polytron Repeater ISA	ISA100
RC003	Polytron 6100 EC WL Polytron Repeater Wire-lessHART	HART

Serial Number key

The third letter of the serial number specifies the manufacturing year: M = 2019, N = 2020, P = 2021, R = 2022, S = 2023, T = 2024, U = 2025, W = 2026, X = 2027, Y = 2028, Z = 2029, etc. (Letters G, I, O, Q are omitted)

Example: Serial Number ARMB-0001: the third letter is M, which means that the unit was manufactured in 2019.

Radio

Max. radiated power:

Bluetooth® LE: <20 dBm EIRP

ISA100: <20 dBm EIRP

Operating frequency:

Bluetooth® LE: 2402-2480 MHz

ISA100: 2405-2480 MHz



EU-Konformitätserklärung
EU-Declaration of Conformity



Dokument-Nr. / Document No. SE23498-00

Wir / we

Dräger Safety AG & Co. KGaA, Revalstraße 1, 23560 Lübeck, Germany

erklären in alleiniger Verantwortung, dass das Produkt
declares under our sole responsibility that the product

Gasmessgerät Typ ETR 06 (Polytron 6100 EC WL)**
*Gas Detection Instrument type ETR 06** (Polytron 6100 EC WL)*

mit der EU-Baumusterprüfbescheinigung / Expertise
is in conformity with the EU-Type Examination Certificate /
Expertise

KIMMA 19 ATEX 0036X

ausgestellt von der notifizierten
Stelle mit der Kennzahl:
issued by the notified body
with identification No.
0620

KIMA Nederland B.V.
Vismarkt 1
7300 AC Apeldoorn
Niederlande

und mit den folgenden Richtlinien unter Anwendung der aufgeführten Normen übereinstimmt
and is in compliance with the following directives by application of the listed standards

Bestimmungen der Richtlinie provisions of directive		Nummer sowie Ausgabedatum der Norm Number and date of issue of standard
2014/34/EU	ATEX-Richtlinie ATEX Directive	EN IEC 60079-0:2018, EN 60079-11:2012
2014/53/EU	RED-Richtlinie RE Directive	EN 301 489-1 V2.2.0, EN 301 489-17 V3.2.0 EN 301 328 V2.1.1, EN 301 329 V2.1.1, EN 62368-1:2014 EN 50270:2015-AC:2016 susceptibility Type 2 emission Type 1
2011/65/EU	RoHS-Richtlinie RoHS Directive	EN 50561:2012

Überprüfung der Qualitätssicherung
Modul D durch
Modul D durch
Production not. Module D by
0188

DEKRA Testing and
Certification GmbH
Handwerker-15
D-70865 Stuttgart

Lübeck, 2019-10-08

Ort und Datum (jjjj-mm-tt)
Place and date (YYYY-mm-dd)


Dr. Marcus Roubie
Head of Electronic Engineering
Safety Products
Research & Develop

2 EU-Declaration of Conformity



EU-Konformitätserklärung
EU-Declaration of Conformity



Dokument Nr. / Document No. SE2409-00

Wir / we

Dräger Safety AG & Co. KGaA, ReinstraÙe 1, 23560 Lübeck, Germany

erklären in alleiniger Verantwortung, dass das Produkt
declare under our sole responsibility that the product

Gasmessgerät Typ ICU 01 (Polytron Repeater ISA)**
*Gas Detection Instrument type ICU 01** (Polytron Repeater ISA)*

mit der EU-Baumusterprüfbescheinigung / Expertise
is in conformity with the EU-Type Examination Certificate / Expertise

KIWA 19 ATEX 0036X

ausgestellt von der notifizierten
Stelle mit der Kenn-Nr.
0001
mit der Bezeichnung
with Identification No.
0620

Kiwa Nederland B.V.
Willemsoord 50
7300 AC Apeldoorn
Niederlands

und mit den folgenden Richtlinien unter Anwendung der aufgeführten Normen übereinstimmt
and is in compliance with the following directives by application of the listed standards

Bestimmungen der Richtlinie provisions of directive		Nummer sowie Ausgabedatum der Norm Number and date of issue of standard
2014/34/EU	ATEX-Richtlinie ATEX Directive	EN IEC 60079-0:2018, EN 60079-11:2012
2014/53/EU	RED-Richtlinie RE Directive	EN 504 485 V2.2.0, EN 301 485-17 V3.2.0 EN 300 328 V2.1.1, EN 62311:2008 EN 62386-1:2014 EN 60270:2015+AC:2016 EN 60270:2015+AC:2016 Type 2 emission, Type 1
2011/65/EU	RoHS-Richtlinie RoHS Directive	EN 50581:2012

Überweisung der Qualitäts-
sicherung Produktion nach
DIN EN ISO 9001
Surveillance of Quality Assurance
Production acc. Module D by
D 170005 Stuttgart
0158

DEKRA Testing and
Certification GmbH
Hardenbergstr.18
D-70509 Stuttgart

Lübeck, 2019-10-08

Ort und Datum (jJJ-jmm-tt)
Place and date (yyy-mm-dd)


Dr. Marcus Romha
Technical Engineering
Head of Product Qualification
Safety Products
Research & Develop

3 USA

Changes or modifications not expressly approved by the party responsible for compliance could void the user's authority to operate the equipment.

This device complies with part 15 of the FCC Rules. Operation is subject to the following two conditions:

- (1) This device may not cause harmful interference, and
- (2) this device must accept any interference received, including interference that may cause undesired operation.

[i] This equipment has been tested and found to comply with the limits for a Class B digital device, pursuant to part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference in a residential installation. This equipment generates, uses and can radiate radio frequency energy and, if not installed and used in accordance with the instructions, may cause harmful interference to radio communications.

However, there is no guarantee that interference will not occur in a particular installation. If this equipment does cause harmful interference to radio or television reception, which can be determined by turning the equipment off and on, the user is encouraged to try to correct the interference by one or more of the following measures:

- Reorient or relocate the receiving antenna.
- Increase the separation between the equipment and receiver.
- Connect the equipment into an outlet on a circuit different from that to which the receiver is connected.
- Consult the dealer or an experienced radio/TV technician for help.

FCC responsible party:

Draeger Inc.

7256 S. Sam Houston W. Parkway

Suite 100

Houston, TX 77085 USA

phone: +1 346-802-6111

e-mail: DIHouston.Approvals@draeger.com

4 Canada

This device contains licence-exempt transmitter(s)/receiver(s) that comply with Innovation, Science and Economic Development Canada's licence-exempt RSS(s). Operation is subject to the following two conditions:

1. This device may not cause interference.
2. This device must accept any interference, including interference that may cause undesired operation of the device.

L'émetteur/récepteur exempt de licence contenu dans le présent appareil est conforme aux CNR d'Innovation, Sciences et Développement économique Canada applicables aux appareils radio exempts de licence. L'exploitation est autorisée aux deux conditions suivantes :

1. L'appareil ne doit pas produire de brouillage;
2. L'appareil doit accepter tout brouillage radioélectrique subi, même si le brouillage est susceptible d'en compromettre le fonctionnement.

■ Manufacturer
Dräger Safety AG & Co. KGaA
Revalstraße 1
D-23560 Lübeck
Germany
+49 451 8 82-0

9300580 – 4683.630 enUS
© **Dräger Safety AG & Co. KGaA**
Edition: 01 – 2021-10
Subject to alterations
www.draeger.com

