

Medtronic

Endoscopy Operating Unit

Given Imaging Inc.
15 Hampshire Street
Mansfield, MA 02048
USA

www.medtronic.com

Attestation Letter

January 02, 2024

Reference: FCC ID: XF6-B001P4V2P1

To whom it may concern:

This letter is confirmation of our request to apply for a C2PC for the Wi-Fi bgn, BT5.0 SIP Module, model RS9116-B0014, approved under FCC ID: XF6-B001P4V2P1.

This evaluation concerns the use of this radio module in Given Imaging Inc. host, PillCam Genius SB Capsule Endoscopy Kit, model Genius Link Device, FCC ID: O8PPATCH.

This letter is provided to state Given Imaging Inc. has not performed any modification and not changed any characteristics of radio module, including output power, frequency range and antenna.

This Class II Permissive Change addresses the exposure distance when the radio is housed in the Given Imaging Inc. host Genius Link Device. Originally, the minimum exposure distance was 20 cm. The exposure distance now is considered for body-worn accessory exposure conditions (5 mm) as the module is used in body-worn device.

The Test Report No: 74873RAN.002A1_FCC_Draft_SAR of DEKRA attached to this application.

Sincerely yours,



Avishag Metzger
Senior Manager Regulatory Affairs