

**CERTIFICATE OF COMPLIANCE**


Authorized under Declaration of Conformity according to 47 CFR, Part 2 and Part 15 of the FCC Rules. The product listed in follows was (were) tested in the BTL EMC Laboratory to comply with the criteria limits Class B of conducted and radiated emissions of the Technical Standards FCC Part 15, Subpart B, established by the FCC, USA.

Equipment MEDIVista 18.5 Bedside Terminal
Model Name PA79(D35)
Brand Name FLYTECH
Applicant FLYTECH Technology Co., Ltd.
Address 1F, No. 168, Sing-Ai Rd., NeiHu District 11494, Taipei, Taiwan

Standard(s) FCC Part 15, Subpart B: 2014
ANSI C63.4-2009
ICES-003 Issue 5: 2012
CISPR 22: 2008
CAN/CSA-CISPR 22-10

Report(s) BTL-FCCE-1-1504039

The test data, data evaluation, and equipment configuration contained in our test report(s) above was (were) obtained utilizing the test procedures, test instruments, test sites that has been accredited by the Authority of TAF according to the ISO-17025 quality assessment standard and technical standard(s). The test data contained in the referenced test report relate only to the EUT sample and item(s) tested.



Andy Chiu
Authorized Signatory**BTL INC.**

B1, No. 37, Lane 365, Yang-Guang St.,
Nei-Hu District, Taipei City 114, Taiwan.

TEL:+886-2-2657-3299

FAX:+886-2-2657-3331

