



Date: 2026.1.15

Subject: attestation letter per KDB 987594D01 Appendix B

FCC ID:Z7A-AMS-R2X

To Whom it May Concern

We, [Stryker Medical], as the manufacturer of FCC ID: Z7A-AMS-R2X, hereby, declare that,

1. Device Protocol Attestation Statement:

- a. Statement that the device will only associate and connect with a low-power indoor access point or subordinate device and never directly connect to other client devices.
- b. Statement that this device will always initiate transmission under the control of a low-power indoor AP or subordinate except for brief transmissions before joining a network. These short messages will only occur if the client has detected an indoor AP or subordinate operating on a channel. These brief messages will have a time-out mechanism such that they will not continually repeat the request if they do not receive a response from an AP.
- c. Statement that transmissions will be lower or equal to the power advertised by the indoor low-power access point or subordinate and never above the maximum output power allowed by the FCC grant for equipment class 6XD.

2. Statement acknowledging device restrictions:

- a. Control of or communications with unmanned aircraft systems is prohibited, including drones.

Should you have any questions or comments concerning the above, please contact the undersigned.

Sincerely,

A handwritten signature in cursive script that reads "Divya Murali".

Divya Murali
Director, Global Regulatory Affairs
Divya.Murali@stryker.com
Medical
3800 E. Centre Avenue, Portage, MI 49002 USA
P +1 269 567 8254