

Calibration Laboratory of  
Schmid & Partner  
Engineering AG  
Zeughausstrasse 43, 8004 Zurich, Switzerland



S Schweizerischer Kalibrierdienst  
C Service suisse d'étalonnage  
S Servizio svizzero di taratura  
S Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)  
The Swiss Accreditation Service is one of the signatories to the EA  
Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: SCS 0108

Client DT&C (Dymstec)

Certificate No: EX3-3916\_Apr18

## CALIBRATION CERTIFICATE

Object EX3DV4 - SN:3916

Calibration procedure(s) QA CAL-01.v9, QA CAL-14.v4, QA CAL-23.v5, QA CAL-25.v6  
Calibration procedure for dosimetric E-field probes

Calibration date: April 25, 2018

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI).  
The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature  $(22 \pm 3)^\circ\text{C}$  and humidity  $< 70\%$ .

Calibration Equipment used (M&TE critical for calibration)

| Primary Standards          | ID               | Cal Date (Certificate No.)        | Scheduled Calibration  |
|----------------------------|------------------|-----------------------------------|------------------------|
| Power meter NRP            | SN: 104778       | 04-Apr-18 (No. 217-02672/02673)   | Apr-19                 |
| Power sensor NRP-Z91       | SN: 103244       | 04-Apr-18 (No. 217-02672)         | Apr-19                 |
| Power sensor NRP-Z91       | SN: 103245       | 04-Apr-18 (No. 217-02673)         | Apr-19                 |
| Reference 20 dB Attenuator | SN: S5277 (20x)  | 04-Apr-18 (No. 217-02682)         | Apr-19                 |
| Reference Probe ES3DV2     | SN: 3013         | 30-Dec-17 (No. ES3-3013_Dec17)    | Dec-18                 |
| DAE4                       | SN: 660          | 21-Dec-17 (No. DAE4-660_Dec17)    | Dec-18                 |
| Secondary Standards        | ID               | Check Date (in house)             | Scheduled Check        |
| Power meter E4419B         | SN: GB41293874   | 06-Apr-16 (in house check Jun-16) | In house check: Jun-18 |
| Power sensor E4412A        | SN: MY41498087   | 06-Apr-16 (in house check Jun-16) | In house check: Jun-18 |
| Power sensor E4412A        | SN: 000110210    | 06-Apr-16 (in house check Jun-16) | In house check: Jun-18 |
| RF generator HP 8648C      | SN: US3642U01700 | 04-Aug-99 (in house check Jun-16) | In house check: Jun-18 |
| Network Analyzer HP 8753E  | SN: US37390585   | 18-Oct-01 (in house check Oct-17) | In house check: Oct-18 |

|                        |                         |                                   |               |
|------------------------|-------------------------|-----------------------------------|---------------|
| Calibrated by:         | Name<br>Claudio Leubler | Function<br>Laboratory Technician | Signature<br> |
| Approved by:           | Katja Pokovic           | Technical Manager                 |               |
| Issued: April 26, 2018 |                         |                                   |               |

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

**Calibration Laboratory of  
Schmid & Partner  
Engineering AG**  
Zeughausstrasse 43, 8004 Zurich, Switzerland



**S** Schweizerischer Kalibrierdienst  
**C** Service suisse d'étalonnage  
**S** Servizio svizzero di taratura  
**S** Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)

Accreditation No.: SCS 0108

The Swiss Accreditation Service is one of the signatories to the EA  
Multilateral Agreement for the recognition of calibration certificates

### Glossary:

|                       |                                                                                                                                                   |
|-----------------------|---------------------------------------------------------------------------------------------------------------------------------------------------|
| TSL                   | tissue simulating liquid                                                                                                                          |
| NORM <sub>x,y,z</sub> | sensitivity in free space                                                                                                                         |
| ConvF                 | sensitivity in TSL / NORM <sub>x,y,z</sub>                                                                                                        |
| DCP                   | diode compression point                                                                                                                           |
| CF                    | crest factor (1/duty_cycle) of the RF signal                                                                                                      |
| A, B, C, D            | modulation dependent linearization parameters                                                                                                     |
| Polarization $\phi$   | $\phi$ rotation around probe axis                                                                                                                 |
| Polarization $\theta$ | $\theta$ rotation around an axis that is in the plane normal to probe axis (at measurement center),<br>i.e., $\theta = 0$ is normal to probe axis |
| Connector Angle       | information used in DASY system to align probe sensor X to the robot coordinate system                                                            |

### Calibration is Performed According to the Following Standards:

- IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013
- IEC 62209-1, "Measurement procedure for the assessment of Specific Absorption Rate (SAR) from hand-held and body-mounted devices used next to the ear (frequency range of 300 MHz to 6 GHz)", July 2016
- IEC 62209-2, "Procedure to determine the Specific Absorption Rate (SAR) for wireless communication devices used in close proximity to the human body (frequency range of 30 MHz to 6 GHz)", March 2010
- KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

### Methods Applied and Interpretation of Parameters:

- NORM<sub>x,y,z</sub>**: Assessed for E-field polarization  $\theta = 0$  ( $f \leq 900$  MHz in TEM-cell;  $f > 1800$  MHz: R22 waveguide). NORM<sub>x,y,z</sub> are only intermediate values, i.e., the uncertainties of NORM<sub>x,y,z</sub> does not affect the  $E^2$ -field uncertainty inside TSL (see below ConvF).
- NORM(f)<sub>x,y,z</sub>** = NORM<sub>x,y,z</sub> \* frequency\_response (see Frequency Response Chart). This linearization is implemented in DASY4 software versions later than 4.2. The uncertainty of the frequency response is included in the stated uncertainty of ConvF.
- DCP<sub>x,y,z</sub>**: DCP are numerical linearization parameters assessed based on the data of power sweep with CW signal (no uncertainty required). DCP does not depend on frequency nor media.
- PAR**: PAR is the Peak to Average Ratio that is not calibrated but determined based on the signal characteristics
- A<sub>x,y,z</sub>; B<sub>x,y,z</sub>; C<sub>x,y,z</sub>; D<sub>x,y,z</sub>; VR<sub>x,y,z</sub>**: A, B, C, D are numerical linearization parameters assessed based on the data of power sweep for specific modulation signal. The parameters do not depend on frequency nor media. VR is the maximum calibration range expressed in RMS voltage across the diode.
- ConvF and Boundary Effect Parameters**: Assessed in flat phantom using E-field (or Temperature Transfer Standard for  $f \leq 800$  MHz) and inside waveguide using analytical field distributions based on power measurements for  $f > 800$  MHz. The same setups are used for assessment of the parameters applied for boundary compensation (alpha, depth) of which typical uncertainty values are given. These parameters are used in DASY4 software to improve probe accuracy close to the boundary. The sensitivity in TSL corresponds to NORM<sub>x,y,z</sub> \* ConvF whereby the uncertainty corresponds to that given for ConvF. A frequency dependent ConvF is used in DASY version 4.4 and higher which allows extending the validity from  $\pm 50$  MHz to  $\pm 100$  MHz.
- Spherical isotropy (3D deviation from isotropy)**: in a field of low gradients realized using a flat phantom exposed by a patch antenna.
- Sensor Offset**: The sensor offset corresponds to the offset of virtual measurement center from the probe tip (on probe axis). No tolerance required.
- Connector Angle**: The angle is assessed using the information gained by determining the NORM<sub>x</sub> (no uncertainty required).

EX3DV4 – SN:3916

April 25, 2018

# Probe EX3DV4

## SN:3916

Manufactured: December 18, 2012  
Calibrated: April 25, 2018

Calibrated for DASY/EASY Systems  
(Note: non-compatible with DASY2 system!)

EX3DV4- SN:3916

April 25, 2018

## DASY/EASY - Parameters of Probe: EX3DV4 - SN:3916

### Basic Calibration Parameters

|                                                           | Sensor X | Sensor Y | Sensor Z | Unc (k=2)     |
|-----------------------------------------------------------|----------|----------|----------|---------------|
| Norm ( $\mu\text{V}/(\text{V}/\text{m})^2$ ) <sup>A</sup> | 0.56     | 0.47     | 0.52     | $\pm 10.1 \%$ |
| DCP (mV) <sup>B</sup>                                     | 99.6     | 101.3    | 99.8     |               |

### Modulation Calibration Parameters

| UID | Communication System Name |   | A<br>dB | B<br>dB $\sqrt{\mu\text{V}}$ | C   | D<br>dB | VR<br>mV | Unc <sup>E</sup><br>(k=2) |
|-----|---------------------------|---|---------|------------------------------|-----|---------|----------|---------------------------|
| 0   | CW                        | X | 0.0     | 0.0                          | 1.0 | 0.00    | 148.6    | $\pm 3.5 \%$              |
|     |                           | Y | 0.0     | 0.0                          | 1.0 |         | 159.6    |                           |
|     |                           | Z | 0.0     | 0.0                          | 1.0 |         | 142.3    |                           |

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor  $k=2$ , which for a normal distribution corresponds to a coverage probability of approximately 95%.

<sup>A</sup> The uncertainties of Norm X,Y,Z do not affect the  $E^2$ -field uncertainty inside TSL (see Pages 5 and 6).

<sup>B</sup> Numerical linearization parameter: uncertainty not required.

<sup>E</sup> Uncertainty is determined using the max. deviation from linear response applying rectangular distribution and is expressed for the square of the field value.

EX3DV4– SN:3916

April 25, 2018

## DASY/EASY - Parameters of Probe: EX3DV4 - SN:3916

### Calibration Parameter Determined in Head Tissue Simulating Media

| f (MHz) <sup>c</sup> | Relative Permittivity <sup>f</sup> | Conductivity (S/m) <sup>f</sup> | ConvF X | ConvF Y | ConvF Z | Alpha <sup>g</sup> | Depth <sup>g</sup> (mm) | Unc (k=2) |
|----------------------|------------------------------------|---------------------------------|---------|---------|---------|--------------------|-------------------------|-----------|
| 2450                 | 39.2                               | 1.80                            | 7.72    | 7.72    | 7.72    | 0.36               | 0.85                    | ± 12.0 %  |
| 2600                 | 39.0                               | 1.96                            | 7.51    | 7.51    | 7.51    | 0.37               | 0.84                    | ± 12.0 %  |
| 5200                 | 36.0                               | 4.66                            | 5.38    | 5.38    | 5.38    | 0.35               | 1.80                    | ± 13.1 %  |
| 5300                 | 35.9                               | 4.76                            | 5.04    | 5.04    | 5.04    | 0.40               | 1.80                    | ± 13.1 %  |
| 5500                 | 35.6                               | 4.96                            | 5.01    | 5.01    | 5.01    | 0.40               | 1.80                    | ± 13.1 %  |
| 5600                 | 35.5                               | 5.07                            | 4.84    | 4.84    | 4.84    | 0.40               | 1.80                    | ± 13.1 %  |
| 5800                 | 35.3                               | 5.27                            | 4.94    | 4.94    | 4.94    | 0.40               | 1.80                    | ± 13.1 %  |

<sup>c</sup> Frequency validity above 300 MHz of ± 100 MHz only applies for DASY v4.4 and higher (see Page 2), else it is restricted to ± 50 MHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below 300 MHz is ± 10, 25, 40, 50 and 70 MHz for ConvF assessments at 30, 64, 128, 150 and 220 MHz respectively. Above 5 GHz frequency validity can be extended to ± 110 MHz.

<sup>f</sup> At frequencies below 3 GHz, the validity of tissue parameters ( $\epsilon$  and  $\sigma$ ) can be relaxed to ± 10% if liquid compensation formula is applied to measured SAR values. At frequencies above 3 GHz, the validity of tissue parameters ( $\epsilon$  and  $\sigma$ ) is restricted to ± 5%. The uncertainty is the RSS of the ConvF uncertainty for indicated target tissue parameters.

<sup>g</sup> Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than ± 1% for frequencies below 3 GHz and below ± 2% for frequencies between 3-6 GHz at any distance larger than half the probe tip diameter from the boundary.

EX3DV4– SN:3916

April 25, 2018

## DASY/EASY - Parameters of Probe: EX3DV4 - SN:3916

### Calibration Parameter Determined in Body Tissue Simulating Media

| f (MHz) <sup>C</sup> | Relative Permittivity <sup>F</sup> | Conductivity (S/m) <sup>F</sup> | ConvF X | ConvF Y | ConvF Z | Alpha <sup>G</sup> | Depth <sup>G</sup> (mm) | Unc (k=2) |
|----------------------|------------------------------------|---------------------------------|---------|---------|---------|--------------------|-------------------------|-----------|
| 2450                 | 52.7                               | 1.95                            | 7.69    | 7.69    | 7.69    | 0.36               | 0.90                    | ± 12.0 %  |
| 2600                 | 52.5                               | 2.16                            | 7.42    | 7.42    | 7.42    | 0.41               | 0.90                    | ± 12.0 %  |
| 5200                 | 49.0                               | 5.30                            | 4.66    | 4.66    | 4.66    | 0.50               | 1.90                    | ± 13.1 %  |
| 5300                 | 48.9                               | 5.42                            | 4.44    | 4.44    | 4.44    | 0.50               | 1.90                    | ± 13.1 %  |
| 5500                 | 48.6                               | 5.65                            | 4.23    | 4.23    | 4.23    | 0.50               | 1.90                    | ± 13.1 %  |
| 5600                 | 48.5                               | 5.77                            | 4.02    | 4.02    | 4.02    | 0.50               | 1.90                    | ± 13.1 %  |
| 5800                 | 48.2                               | 6.00                            | 4.31    | 4.31    | 4.31    | 0.50               | 1.90                    | ± 13.1 %  |

<sup>C</sup> Frequency validity above 300 MHz of ± 100 MHz only applies for DASY v4.4 and higher (see Page 2), else it is restricted to ± 50 MHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below 300 MHz is ± 10, 25, 40, 50 and 70 MHz for ConvF assessments at 30, 64, 128, 150 and 220 MHz respectively. Above 5 GHz frequency validity can be extended to ± 110 MHz.


<sup>F</sup> At frequencies below 3 GHz, the validity of tissue parameters ( $\epsilon$  and  $\sigma$ ) can be relaxed to ± 10% if liquid compensation formula is applied to measured SAR values. At frequencies above 3 GHz, the validity of tissue parameters ( $\epsilon$  and  $\sigma$ ) is restricted to ± 5%. The uncertainty is the RSS of the ConvF uncertainty for indicated target tissue parameters.

<sup>G</sup> Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than ± 1% for frequencies below 3 GHz and below ± 2% for frequencies between 3-6 GHz at any distance larger than half the probe tip diameter from the boundary.

EX3DV4– SN:3916

April 25, 2018

## Frequency Response of E-Field (TEM-Cell:ifi110 EXX, Waveguide: R22)





Uncertainty of Frequency Response of E-field:  $\pm 6.3\%$  ( $k=2$ )

EX3DV4- SN:3916

April 25, 2018

## Receiving Pattern ( $\phi$ ), $\theta = 0^\circ$





Uncertainty of Axial Isotropy Assessment:  $\pm 0.5\%$  ( $k=2$ )

EX3DV4- SN:3916

April 25, 2018

### Dynamic Range f(SAR<sub>head</sub>) (TEM cell , f<sub>eval</sub>= 1900 MHz)




Uncertainty of Linearity Assessment: ± 0.6% (k=2)

EX3DV4- SN:3916


April 25, 2018

## Conversion Factor Assessment



## Deviation from Isotropy in Liquid

Error ( $\phi$ ,  $\theta$ ),  $f = 900 \text{ MHz}$



EX3DV4- SN:3916

April 25, 2018

## DASY/EASY - Parameters of Probe: EX3DV4 - SN:3916

### Other Probe Parameters

|                                               |            |
|-----------------------------------------------|------------|
| Sensor Arrangement                            | Triangular |
| Connector Angle (°)                           | 88.3       |
| Mechanical Surface Detection Mode             | enabled    |
| Optical Surface Detection Mode                | disabled   |
| Probe Overall Length                          | 337 mm     |
| Probe Body Diameter                           | 10 mm      |
| Tip Length                                    | 9 mm       |
| Tip Diameter                                  | 2.5 mm     |
| Probe Tip to Sensor X Calibration Point       | 1 mm       |
| Probe Tip to Sensor Y Calibration Point       | 1 mm       |
| Probe Tip to Sensor Z Calibration Point       | 1 mm       |
| Recommended Measurement Distance from Surface | 1.4 mm     |

Calibration Laboratory of  
Schmid & Partner  
Engineering AG  
Zeughausstrasse 43, 8004 Zurich, Switzerland



S Schweizerischer Kalibrierdienst  
C Service suisse d'étalonnage  
S Servizio svizzero di taratura  
S Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)  
The Swiss Accreditation Service is one of the signatories to the EA  
Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: SCS 0108

Client DT&C (Dymstec)

Certificate No: EX3-3866\_May18

## CALIBRATION CERTIFICATE

Object EX3DV4 - SN:3866

Calibration procedure(s) QA CAL-01.v9, QA CAL-14.v4, QA CAL-23.v5, QA CAL-25.v6  
Calibration procedure for dosimetric E-field probes

Calibration date: May 31, 2018

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI).  
The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature ( $22 \pm 3$ )°C and humidity < 70%.

Calibration Equipment used (M&TE critical for calibration)

| Primary Standards          | ID               | Cal Date (Certificate No.)        | Scheduled Calibration  |
|----------------------------|------------------|-----------------------------------|------------------------|
| Power meter NRP            | SN: 104778       | 04-Apr-18 (No. 217-02672/02673)   | Apr-19                 |
| Power sensor NRP-Z91       | SN: 103244       | 04-Apr-18 (No. 217-02672)         | Apr-19                 |
| Power sensor NRP-Z91       | SN: 103245       | 04-Apr-18 (No. 217-02673)         | Apr-19                 |
| Reference 20 dB Attenuator | SN: S5277 (20x)  | 04-Apr-18 (No. 217-02682)         | Apr-19                 |
| Reference Probe ES3DV2     | SN: 3013         | 30-Dec-17 (No. ES3-3013_Dec17)    | Dec-18                 |
| DAE4                       | SN: 660          | 21-Dec-17 (No. DAE4-660_Dec17)    | Dec-18                 |
| Secondary Standards        | ID               | Check Date (in house)             | Scheduled Check        |
| Power meter E4419B         | SN: GB41293874   | 06-Apr-16 (in house check Jun-16) | In house check: Jun-18 |
| Power sensor E4412A        | SN: MY41498087   | 06-Apr-16 (in house check Jun-16) | In house check: Jun-18 |
| Power sensor E4412A        | SN: 000110210    | 06-Apr-16 (in house check Jun-16) | In house check: Jun-18 |
| RF generator HP 8648C      | SN: US3642U01700 | 04-Aug-99 (in house check Jun-16) | In house check: Jun-18 |
| Network Analyzer HP 8753E  | SN: US37390585   | 18-Oct-01 (in house check Oct-17) | In house check: Oct-18 |

|                                                                                                                 | Name           | Function              | Signature |
|-----------------------------------------------------------------------------------------------------------------|----------------|-----------------------|-----------|
| Calibrated by:                                                                                                  | Jeton Kastrati | Laboratory Technician |           |
| Approved by:                                                                                                    | Katja Pokovic  | Technical Manager     |           |
| Issued: May 31, 2018                                                                                            |                |                       |           |
| This calibration certificate shall not be reproduced except in full without written approval of the laboratory. |                |                       |           |

**Calibration Laboratory of  
Schmid & Partner  
Engineering AG**  
Zeughausstrasse 43, 8004 Zurich, Switzerland



**S** Schweizerischer Kalibrierdienst  
**C** Service suisse d'étalonnage  
**S** Servizio svizzero di taratura  
**S** Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)  
The Swiss Accreditation Service is one of the signatories to the EA  
Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 0108**

### Glossary:

|                          |                                                                                                                                                         |
|--------------------------|---------------------------------------------------------------------------------------------------------------------------------------------------------|
| TSL                      | tissue simulating liquid                                                                                                                                |
| NORM <sub>x,y,z</sub>    | sensitivity in free space                                                                                                                               |
| ConvF                    | sensitivity in TSL / NORM <sub>x,y,z</sub>                                                                                                              |
| DCP                      | diode compression point                                                                                                                                 |
| CF                       | crest factor (1/duty_cycle) of the RF signal                                                                                                            |
| A, B, C, D               | modulation dependent linearization parameters                                                                                                           |
| Polarization $\phi$      | $\phi$ rotation around probe axis                                                                                                                       |
| Polarization $\vartheta$ | $\vartheta$ rotation around an axis that is in the plane normal to probe axis (at measurement center),<br>i.e., $\vartheta = 0$ is normal to probe axis |
| Connector Angle          | information used in DASY system to align probe sensor X to the robot coordinate system                                                                  |

### Calibration is Performed According to the Following Standards:

- IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013
- IEC 62209-1, "Measurement procedure for the assessment of Specific Absorption Rate (SAR) from hand-held and body-mounted devices used next to the ear (frequency range of 300 MHz to 6 GHz)", July 2016
- IEC 62209-2, "Procedure to determine the Specific Absorption Rate (SAR) for wireless communication devices used in close proximity to the human body (frequency range of 30 MHz to 6 GHz)", March 2010
- KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

### Methods Applied and Interpretation of Parameters:

- NORM<sub>x,y,z</sub>**: Assessed for E-field polarization  $\vartheta = 0$  ( $f \leq 900$  MHz in TEM-cell;  $f > 1800$  MHz: R22 waveguide). NORM<sub>x,y,z</sub> are only intermediate values, i.e., the uncertainties of NORM<sub>x,y,z</sub> does not affect the  $E^2$ -field uncertainty inside TSL (see below **ConvF**).
- NORM(f)<sub>x,y,z</sub>** = NORM<sub>x,y,z</sub> \* frequency\_response (see Frequency Response Chart). This linearization is implemented in DASY4 software versions later than 4.2. The uncertainty of the frequency response is included in the stated uncertainty of **ConvF**.
- DCP<sub>x,y,z</sub>**: DCP are numerical linearization parameters assessed based on the data of power sweep with CW signal (no uncertainty required). DCP does not depend on frequency nor media.
- PAR**: PAR is the Peak to Average Ratio that is not calibrated but determined based on the signal characteristics
- A<sub>x,y,z</sub>; B<sub>x,y,z</sub>; C<sub>x,y,z</sub>; D<sub>x,y,z</sub>; VR<sub>x,y,z</sub>**: A, B, C, D are numerical linearization parameters assessed based on the data of power sweep for specific modulation signal. The parameters do not depend on frequency nor media. VR is the maximum calibration range expressed in RMS voltage across the diode.
- ConvF and Boundary Effect Parameters**: Assessed in flat phantom using E-field (or Temperature Transfer Standard for  $f \leq 800$  MHz) and inside waveguide using analytical field distributions based on power measurements for  $f > 800$  MHz. The same setups are used for assessment of the parameters applied for boundary compensation (alpha, depth) of which typical uncertainty values are given. These parameters are used in DASY4 software to improve probe accuracy close to the boundary. The sensitivity in TSL corresponds to NORM<sub>x,y,z</sub> \* ConvF whereby the uncertainty corresponds to that given for ConvF. A frequency dependent ConvF is used in DASY version 4.4 and higher which allows extending the validity from  $\pm 50$  MHz to  $\pm 100$  MHz.
- Spherical isotropy (3D deviation from isotropy)**: in a field of low gradients realized using a flat phantom exposed by a patch antenna.
- Sensor Offset**: The sensor offset corresponds to the offset of virtual measurement center from the probe tip (on probe axis). No tolerance required.
- Connector Angle**: The angle is assessed using the information gained by determining the NORM<sub>x</sub> (no uncertainty required).

EX3DV4 – SN:3866

May 31, 2018

# Probe EX3DV4

## SN:3866

Manufactured: February 2, 2012  
Calibrated: May 31, 2018

Calibrated for DASY/EASY Systems  
(Note: non-compatible with DASY2 system!)

EX3DV4– SN:3866

May 31, 2018

## DASY/EASY - Parameters of Probe: EX3DV4 - SN:3866

### Basic Calibration Parameters

|                                                           | Sensor X | Sensor Y | Sensor Z | Unc (k=2) |
|-----------------------------------------------------------|----------|----------|----------|-----------|
| Norm ( $\mu\text{V}/(\text{V}/\text{m})^2$ ) <sup>A</sup> | 0.43     | 0.32     | 0.35     | ± 10.1 %  |
| DCP (mV) <sup>B</sup>                                     | 98.7     | 101.4    | 105.4    |           |

### Modulation Calibration Parameters

| UID | Communication System Name |   | A<br>dB | B<br>dB $\sqrt{\mu\text{V}}$ | C   | D<br>dB | VR<br>mV | Unc <sup>E</sup><br>(k=2) |
|-----|---------------------------|---|---------|------------------------------|-----|---------|----------|---------------------------|
| 0   | CW                        | X | 0.0     | 0.0                          | 1.0 | 0.00    | 129.5    | ±3.3 %                    |
|     |                           | Y | 0.0     | 0.0                          | 1.0 |         | 142.9    |                           |
|     |                           | Z | 0.0     | 0.0                          | 1.0 |         | 132.3    |                           |

Note: For details on UID parameters see Appendix.

### Sensor Model Parameters

|   | C1<br>fF | C2<br>fF | $\alpha$<br>$\text{V}^{-1}$ | T1<br>$\text{ms}\cdot\text{V}^{-2}$ | T2<br>$\text{ms}\cdot\text{V}^{-1}$ | T3<br>ms | T4<br>$\text{V}^{-2}$ | T5<br>$\text{V}^{-1}$ | T6    |
|---|----------|----------|-----------------------------|-------------------------------------|-------------------------------------|----------|-----------------------|-----------------------|-------|
| X | 61.34    | 450.3    | 34.79                       | 20.71                               | 0.897                               | 5.071    | 0.953                 | 0.532                 | 1.007 |
| Y | 35.97    | 270.0    | 35.93                       | 7.616                               | 0.990                               | 4.996    | 0.120                 | 0.508                 | 1.005 |
| Z | 34.59    | 248.7    | 33.42                       | 8.463                               | 0.617                               | 4.987    | 2.000                 | 0.071                 | 1.005 |

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor  $k=2$ , which for a normal distribution corresponds to a coverage probability of approximately 95%.

<sup>A</sup> The uncertainties of Norm X,Y,Z do not affect the  $E^2$ -field uncertainty inside TSL (see Pages 5 and 6).

<sup>B</sup> Numerical linearization parameter: uncertainty not required.

<sup>E</sup> Uncertainty is determined using the max. deviation from linear response applying rectangular distribution and is expressed for the square of the field value.

EX3DV4- SN:3866

May 31, 2018

## DASY/EASY - Parameters of Probe: EX3DV4 - SN:3866

### Calibration Parameter Determined in Head Tissue Simulating Media

| f (MHz) <sup>C</sup> | Relative Permittivity <sup>F</sup> | Conductivity (S/m) <sup>F</sup> | ConvF X | ConvF Y | ConvF Z | Alpha <sup>G</sup> | Depth <sup>G</sup> (mm) | Unc (k=2) |
|----------------------|------------------------------------|---------------------------------|---------|---------|---------|--------------------|-------------------------|-----------|
| 750                  | 41.9                               | 0.89                            | 10.16   | 10.16   | 10.16   | 0.49               | 0.80                    | ± 12.0 %  |
| 835                  | 41.5                               | 0.90                            | 9.62    | 9.62    | 9.62    | 0.39               | 0.93                    | ± 12.0 %  |
| 900                  | 41.5                               | 0.97                            | 9.40    | 9.40    | 9.40    | 0.40               | 0.92                    | ± 12.0 %  |
| 1750                 | 40.1                               | 1.37                            | 8.38    | 8.38    | 8.38    | 0.34               | 0.84                    | ± 12.0 %  |
| 1900                 | 40.0                               | 1.40                            | 8.03    | 8.03    | 8.03    | 0.27               | 0.87                    | ± 12.0 %  |
| 2300                 | 39.5                               | 1.67                            | 7.86    | 7.86    | 7.86    | 0.30               | 0.85                    | ± 12.0 %  |
| 2450                 | 39.2                               | 1.80                            | 7.45    | 7.45    | 7.45    | 0.34               | 0.82                    | ± 12.0 %  |
| 2600                 | 39.0                               | 1.96                            | 7.22    | 7.22    | 7.22    | 0.38               | 0.85                    | ± 12.0 %  |
| 3500                 | 37.9                               | 2.91                            | 6.89    | 6.89    | 6.89    | 0.20               | 1.25                    | ± 13.1 %  |
| 5200                 | 36.0                               | 4.66                            | 5.14    | 5.14    | 5.14    | 0.40               | 1.80                    | ± 13.1 %  |
| 5300                 | 35.9                               | 4.76                            | 4.95    | 4.95    | 4.95    | 0.40               | 1.80                    | ± 13.1 %  |
| 5500                 | 35.6                               | 4.96                            | 4.61    | 4.61    | 4.61    | 0.40               | 1.80                    | ± 13.1 %  |
| 5600                 | 35.5                               | 5.07                            | 4.52    | 4.52    | 4.52    | 0.40               | 1.80                    | ± 13.1 %  |
| 5800                 | 35.3                               | 5.27                            | 4.69    | 4.69    | 4.69    | 0.40               | 1.80                    | ± 13.1 %  |

<sup>C</sup> Frequency validity above 300 MHz of ± 100 MHz only applies for DASY v4.4 and higher (see Page 2), else it is restricted to ± 50 MHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below 300 MHz is ± 10, 25, 40, 50 and 70 MHz for ConvF assessments at 30, 64, 128, 150 and 220 MHz respectively. Above 5 GHz frequency validity can be extended to ± 110 MHz.

<sup>F</sup> At frequencies below 3 GHz, the validity of tissue parameters ( $\epsilon$  and  $\sigma$ ) can be relaxed to ± 10% if liquid compensation formula is applied to measured SAR values. At frequencies above 3 GHz, the validity of tissue parameters ( $\epsilon$  and  $\sigma$ ) is restricted to ± 5%. The uncertainty is the RSS of the ConvF uncertainty for indicated target tissue parameters.

<sup>G</sup> Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than ± 1% for frequencies below 3 GHz and below ± 2% for frequencies between 3-6 GHz at any distance larger than half the probe tip diameter from the boundary.

EX3DV4- SN:3866

May 31, 2018

## DASY/EASY - Parameters of Probe: EX3DV4 - SN:3866

### Calibration Parameter Determined in Body Tissue Simulating Media

| f (MHz) <sup>C</sup> | Relative Permittivity <sup>F</sup> | Conductivity (S/m) <sup>F</sup> | ConvF X | ConvF Y | ConvF Z | Alpha <sup>G</sup> | Depth <sup>G</sup> (mm) | Unc (k=2) |
|----------------------|------------------------------------|---------------------------------|---------|---------|---------|--------------------|-------------------------|-----------|
| 750                  | 55.5                               | 0.96                            | 9.69    | 9.69    | 9.69    | 0.33               | 0.97                    | ± 12.0 %  |
| 835                  | 55.2                               | 0.97                            | 9.43    | 9.43    | 9.43    | 0.42               | 0.80                    | ± 12.0 %  |
| 900                  | 55.0                               | 1.05                            | 9.57    | 9.57    | 9.57    | 0.48               | 0.80                    | ± 12.0 %  |
| 1750                 | 53.4                               | 1.49                            | 7.95    | 7.95    | 7.95    | 0.39               | 0.80                    | ± 12.0 %  |
| 1900                 | 53.3                               | 1.52                            | 7.68    | 7.68    | 7.68    | 0.30               | 0.85                    | ± 12.0 %  |
| 2300                 | 52.9                               | 1.81                            | 7.50    | 7.50    | 7.50    | 0.39               | 0.85                    | ± 12.0 %  |
| 2450                 | 52.7                               | 1.95                            | 7.40    | 7.40    | 7.40    | 0.43               | 0.90                    | ± 12.0 %  |
| 2600                 | 52.5                               | 2.16                            | 7.28    | 7.28    | 7.28    | 0.25               | 1.05                    | ± 12.0 %  |
| 3500                 | 51.3                               | 3.31                            | 6.43    | 6.43    | 6.43    | 0.28               | 1.20                    | ± 13.1 %  |
| 5200                 | 49.0                               | 5.30                            | 4.69    | 4.69    | 4.69    | 0.50               | 1.90                    | ± 13.1 %  |
| 5300                 | 48.9                               | 5.42                            | 4.50    | 4.50    | 4.50    | 0.50               | 1.90                    | ± 13.1 %  |
| 5500                 | 48.6                               | 5.65                            | 3.95    | 3.95    | 3.95    | 0.50               | 1.90                    | ± 13.1 %  |
| 5600                 | 48.5                               | 5.77                            | 3.87    | 3.87    | 3.87    | 0.50               | 1.90                    | ± 13.1 %  |
| 5800                 | 48.2                               | 6.00                            | 4.16    | 4.16    | 4.16    | 0.50               | 1.90                    | ± 13.1 %  |

<sup>C</sup> Frequency validity above 300 MHz of ± 100 MHz only applies for DASY v4.4 and higher (see Page 2), else it is restricted to ± 50 MHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below 300 MHz is ± 10, 25, 40, 50 and 70 MHz for ConvF assessments at 30, 64, 128, 150 and 220 MHz respectively. Above 5 GHz frequency validity can be extended to ± 110 MHz.


<sup>F</sup> At frequencies below 3 GHz, the validity of tissue parameters ( $\epsilon$  and  $\sigma$ ) can be relaxed to ± 10% if liquid compensation formula is applied to measured SAR values. At frequencies above 3 GHz, the validity of tissue parameters ( $\epsilon$  and  $\sigma$ ) is restricted to ± 5%. The uncertainty is the RSS of the ConvF uncertainty for indicated target tissue parameters.

<sup>G</sup> Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than ± 1% for frequencies below 3 GHz and below ± 2% for frequencies between 3-6 GHz at any distance larger than half the probe tip diameter from the boundary.

EX3DV4– SN:3866

May 31, 2018

## Frequency Response of E-Field (TEM-Cell:ifi110 EXX, Waveguide: R22)





Uncertainty of Frequency Response of E-field:  $\pm 6.3\%$  ( $k=2$ )

EX3DV4- SN:3866

May 31, 2018

## Receiving Pattern ( $\phi$ ), $\theta = 0^\circ$





**Uncertainty of Axial Isotropy Assessment:  $\pm 0.5\%$  ( $k=2$ )**

EX3DV4- SN:3866

May 31, 2018


### Dynamic Range $f(\text{SAR}_{\text{head}})$ (TEM cell, $f_{\text{eval}} = 1900 \text{ MHz}$ )

Uncertainty of Linearity Assessment:  $\pm 0.6\%$  ( $k=2$ )

EX3DV4- SN:3866


May 31, 2018

## Conversion Factor Assessment



## Deviation from Isotropy in Liquid

Error ( $\phi, \theta$ ),  $f = 900 \text{ MHz}$



EX3DV4– SN:3866

May 31, 2018

## DASY/EASY - Parameters of Probe: EX3DV4 - SN:3866

### Other Probe Parameters

|                                               |            |
|-----------------------------------------------|------------|
| Sensor Arrangement                            | Triangular |
| Connector Angle (°)                           | 61.8       |
| Mechanical Surface Detection Mode             | enabled    |
| Optical Surface Detection Mode                | disabled   |
| Probe Overall Length                          | 337 mm     |
| Probe Body Diameter                           | 10 mm      |
| Tip Length                                    | 9 mm       |
| Tip Diameter                                  | 2.5 mm     |
| Probe Tip to Sensor X Calibration Point       | 1 mm       |
| Probe Tip to Sensor Y Calibration Point       | 1 mm       |
| Probe Tip to Sensor Z Calibration Point       | 1 mm       |
| Recommended Measurement Distance from Surface | 1.4 mm     |

EX3DV4– SN:3866

May 31, 2018

### Appendix: Modulation Calibration Parameters

| UID           | Communication System Name                     |   | A<br>dB | B<br>dB $\mu$ V | C     | D<br>dB | VR<br>mV | Max<br>Unc <sup>E</sup><br>(k=2) |
|---------------|-----------------------------------------------|---|---------|-----------------|-------|---------|----------|----------------------------------|
| 0             | CW                                            | X | 0.00    | 0.00            | 1.00  | 0.00    | 129.5    | $\pm 3.3 \%$                     |
|               |                                               | Y | 0.00    | 0.00            | 1.00  |         | 142.9    |                                  |
|               |                                               | Z | 0.00    | 0.00            | 1.00  |         | 132.3    |                                  |
| 10010-<br>CAA | SAR Validation (Square, 100ms, 10ms)          | X | 4.96    | 74.03           | 14.55 | 10.00   | 20.0     | $\pm 9.6 \%$                     |
|               |                                               | Y | 1.96    | 62.67           | 8.25  |         | 20.0     |                                  |
|               |                                               | Z | 1.98    | 63.61           | 8.75  |         | 20.0     |                                  |
| 10011-<br>CAB | UMTS-FDD (WCDMA)                              | X | 1.46    | 74.36           | 19.19 | 0.00    | 150.0    | $\pm 9.6 \%$                     |
|               |                                               | Y | 0.84    | 66.93           | 14.18 |         | 150.0    |                                  |
|               |                                               | Z | 1.06    | 69.91           | 16.41 |         | 150.0    |                                  |
| 10012-<br>CAB | IEEE 802.11b WiFi 2.4 GHz (DSSS, 1 Mbps)      | X | 1.27    | 66.19           | 17.07 | 0.41    | 150.0    | $\pm 9.6 \%$                     |
|               |                                               | Y | 1.01    | 63.39           | 14.61 |         | 150.0    |                                  |
|               |                                               | Z | 1.12    | 64.44           | 15.48 |         | 150.0    |                                  |
| 10013-<br>CAB | IEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 6 Mbps) | X | 5.02    | 66.95           | 17.37 | 1.46    | 150.0    | $\pm 9.6 \%$                     |
|               |                                               | Y | 4.56    | 66.54           | 16.75 |         | 150.0    |                                  |
|               |                                               | Z | 4.61    | 66.83           | 16.87 |         | 150.0    |                                  |
| 10021-<br>DAC | GSM-FDD (TDMA, GMSK)                          | X | 100.00  | 116.34          | 28.99 | 9.39    | 50.0     | $\pm 9.6 \%$                     |
|               |                                               | Y | 4.35    | 71.51           | 13.58 |         | 50.0     |                                  |
|               |                                               | Z | 10.49   | 82.17           | 17.30 |         | 50.0     |                                  |
| 10023-<br>DAC | GPRS-FDD (TDMA, GMSK, TN 0)                   | X | 100.00  | 116.24          | 28.99 | 9.57    | 50.0     | $\pm 9.6 \%$                     |
|               |                                               | Y | 4.08    | 70.51           | 13.19 |         | 50.0     |                                  |
|               |                                               | Z | 7.34    | 77.92           | 15.91 |         | 50.0     |                                  |
| 10024-<br>DAC | GPRS-FDD (TDMA, GMSK, TN 0-1)                 | X | 100.00  | 114.37          | 27.13 | 6.56    | 60.0     | $\pm 9.6 \%$                     |
|               |                                               | Y | 2.47    | 68.27           | 11.00 |         | 60.0     |                                  |
|               |                                               | Z | 99.64   | 104.22          | 21.52 |         | 60.0     |                                  |
| 10025-<br>DAC | EDGE-FDD (TDMA, 8PSK, TN 0)                   | X | 7.29    | 85.63           | 33.51 | 12.57   | 50.0     | $\pm 9.6 \%$                     |
|               |                                               | Y | 3.34    | 62.89           | 20.63 |         | 50.0     |                                  |
|               |                                               | Z | 4.59    | 72.89           | 26.66 |         | 50.0     |                                  |
| 10026-<br>DAC | EDGE-FDD (TDMA, 8PSK, TN 0-1)                 | X | 19.51   | 108.37          | 37.98 | 9.56    | 60.0     | $\pm 9.6 \%$                     |
|               |                                               | Y | 6.99    | 84.48           | 28.68 |         | 60.0     |                                  |
|               |                                               | Z | 7.40    | 87.18           | 30.26 |         | 60.0     |                                  |
| 10027-<br>DAC | GPRS-FDD (TDMA, GMSK, TN 0-1-2)               | X | 100.00  | 114.69          | 26.54 | 4.80    | 80.0     | $\pm 9.6 \%$                     |
|               |                                               | Y | 1.47    | 65.78           | 9.10  |         | 80.0     |                                  |
|               |                                               | Z | 100.00  | 103.55          | 20.47 |         | 80.0     |                                  |
| 10028-<br>DAC | GPRS-FDD (TDMA, GMSK, TN 0-1-2-3)             | X | 100.00  | 116.57          | 26.68 | 3.55    | 100.0    | $\pm 9.6 \%$                     |
|               |                                               | Y | 0.75    | 62.53           | 6.91  |         | 100.0    |                                  |
|               |                                               | Z | 100.00  | 103.86          | 19.98 |         | 100.0    |                                  |
| 10029-<br>DAC | EDGE-FDD (TDMA, 8PSK, TN 0-1-2)               | X | 10.84   | 94.12           | 31.96 | 7.80    | 80.0     | $\pm 9.6 \%$                     |
|               |                                               | Y | 4.68    | 76.74           | 24.63 |         | 80.0     |                                  |
|               |                                               | Z | 4.76    | 77.76           | 25.40 |         | 80.0     |                                  |
| 10030-<br>CAA | IEEE 802.15.1 Bluetooth (GFSK, DH1)           | X | 100.00  | 113.28          | 26.21 | 5.30    | 70.0     | $\pm 9.6 \%$                     |
|               |                                               | Y | 1.50    | 64.87           | 8.87  |         | 70.0     |                                  |
|               |                                               | Z | 14.61   | 85.51           | 16.17 |         | 70.0     |                                  |
| 10031-<br>CAA | IEEE 802.15.1 Bluetooth (GFSK, DH3)           | X | 100.00  | 120.46          | 26.88 | 1.88    | 100.0    | $\pm 9.6 \%$                     |
|               |                                               | Y | 0.28    | 60.00           | 3.77  |         | 100.0    |                                  |
|               |                                               | Z | 100.00  | 97.01           | 16.04 |         | 100.0    |                                  |

EX3DV4– SN:3866

May 31, 2018

|           |                                                     |   |        |        |       |       |       |         |
|-----------|-----------------------------------------------------|---|--------|--------|-------|-------|-------|---------|
| 10032-CAA | IEEE 802.15.1 Bluetooth (GFSK, DH5)                 | X | 100.00 | 134.94 | 31.61 | 1.17  | 100.0 | ± 9.6 % |
|           |                                                     | Y | 2.98   | 214.36 | 19.03 |       | 100.0 |         |
|           |                                                     | Z | 100.00 | 96.12  | 15.00 |       | 100.0 |         |
| 10033-CAA | IEEE 802.15.1 Bluetooth (PI/4-DQPSK, DH1)           | X | 100.00 | 129.71 | 35.52 | 5.30  | 70.0  | ± 9.6 % |
|           |                                                     | Y | 3.37   | 73.07  | 15.63 |       | 70.0  |         |
|           |                                                     | Z | 5.18   | 79.83  | 18.59 |       | 70.0  |         |
| 10034-CAA | IEEE 802.15.1 Bluetooth (PI/4-DQPSK, DH3)           | X | 38.25  | 116.38 | 31.11 | 1.88  | 100.0 | ± 9.6 % |
|           |                                                     | Y | 1.32   | 66.13  | 11.17 |       | 100.0 |         |
|           |                                                     | Z | 2.19   | 72.52  | 14.56 |       | 100.0 |         |
| 10035-CAA | IEEE 802.15.1 Bluetooth (PI/4-DQPSK, DH5)           | X | 10.07  | 97.58  | 26.00 | 1.17  | 100.0 | ± 9.6 % |
|           |                                                     | Y | 1.02   | 64.74  | 10.26 |       | 100.0 |         |
|           |                                                     | Z | 1.68   | 70.82  | 13.73 |       | 100.0 |         |
| 10036-CAA | IEEE 802.15.1 Bluetooth (8-DPSK, DH1)               | X | 100.00 | 130.10 | 35.71 | 5.30  | 70.0  | ± 9.6 % |
|           |                                                     | Y | 3.79   | 74.73  | 16.33 |       | 70.0  |         |
|           |                                                     | Z | 6.44   | 82.95  | 19.72 |       | 70.0  |         |
| 10037-CAA | IEEE 802.15.1 Bluetooth (8-DPSK, DH3)               | X | 33.36  | 114.28 | 30.54 | 1.88  | 100.0 | ± 9.6 % |
|           |                                                     | Y | 1.25   | 65.67  | 10.94 |       | 100.0 |         |
|           |                                                     | Z | 1.95   | 71.33  | 14.08 |       | 100.0 |         |
| 10038-CAA | IEEE 802.15.1 Bluetooth (8-DPSK, DH5)               | X | 11.00  | 99.37  | 26.64 | 1.17  | 100.0 | ± 9.6 % |
|           |                                                     | Y | 1.03   | 65.03  | 10.52 |       | 100.0 |         |
|           |                                                     | Z | 1.72   | 71.30  | 14.06 |       | 100.0 |         |
| 10039-CAB | CDMA2000 (1xRTT, RC1)                               | X | 4.41   | 85.41  | 21.99 | 0.00  | 150.0 | ± 9.6 % |
|           |                                                     | Y | 0.86   | 64.63  | 9.97  |       | 150.0 |         |
|           |                                                     | Z | 1.99   | 74.44  | 15.11 |       | 150.0 |         |
| 10042-CAB | IS-54 / IS-136 FDD (TDMA/FDM, PI/4-DQPSK, Halfrate) | X | 100.00 | 112.07 | 26.26 | 7.78  | 50.0  | ± 9.6 % |
|           |                                                     | Y | 2.24   | 65.83  | 9.99  |       | 50.0  |         |
|           |                                                     | Z | 4.60   | 73.72  | 13.31 |       | 50.0  |         |
| 10044-CAA | IS-91/EIA/TIA-553 FDD (FDMA, FM)                    | X | 0.01   | 122.05 | 4.07  | 0.00  | 150.0 | ± 9.6 % |
|           |                                                     | Y | 0.35   | 142.03 | 0.00  |       | 150.0 |         |
|           |                                                     | Z | 0.02   | 123.73 | 10.80 |       | 150.0 |         |
| 10048-CAA | DECT (TDD, TDMA/FDM, GFSK, Full Slot, 24)           | X | 100.00 | 117.95 | 31.07 | 13.80 | 25.0  | ± 9.6 % |
|           |                                                     | Y | 4.50   | 67.37  | 13.41 |       | 25.0  |         |
|           |                                                     | Z | 5.19   | 70.06  | 14.31 |       | 25.0  |         |
| 10049-CAA | DECT (TDD, TDMA/FDM, GFSK, Double Slot, 12)         | X | 100.00 | 116.36 | 29.33 | 10.79 | 40.0  | ± 9.6 % |
|           |                                                     | Y | 4.23   | 69.49  | 13.02 |       | 40.0  |         |
|           |                                                     | Z | 5.27   | 72.87  | 14.27 |       | 40.0  |         |
| 10056-CAA | UMTS-TDD (TD-SCDMA, 1.28 Mcps)                      | X | 77.81  | 121.32 | 33.78 | 9.03  | 50.0  | ± 9.6 % |
|           |                                                     | Y | 6.03   | 75.76  | 17.19 |       | 50.0  |         |
|           |                                                     | Z | 9.07   | 82.59  | 19.86 |       | 50.0  |         |
| 10058-DAC | EDGE-FDD (TDMA, 8PSK, TN 0-1-2-3)                   | X | 7.57   | 86.51  | 28.41 | 6.55  | 100.0 | ± 9.6 % |
|           |                                                     | Y | 3.72   | 73.02  | 22.40 |       | 100.0 |         |
|           |                                                     | Z | 3.78   | 73.63  | 22.92 |       | 100.0 |         |
| 10059-CAB | IEEE 802.11b WiFi 2.4 GHz (DSSS, 2 Mbps)            | X | 1.41   | 68.44  | 18.21 | 0.61  | 110.0 | ± 9.6 % |
|           |                                                     | Y | 1.03   | 64.26  | 15.02 |       | 110.0 |         |
|           |                                                     | Z | 1.14   | 65.37  | 15.93 |       | 110.0 |         |
| 10060-CAB | IEEE 802.11b WiFi 2.4 GHz (DSSS, 5.5 Mbps)          | X | 100.00 | 140.28 | 36.98 | 1.30  | 110.0 | ± 9.6 % |
|           |                                                     | Y | 5.52   | 92.10  | 22.15 |       | 110.0 |         |
|           |                                                     | Z | 23.32  | 116.45 | 30.29 |       | 110.0 |         |

EX3DV4- SN:3866

May 31, 2018

|           |                                                |   |       |        |       |      |       |         |
|-----------|------------------------------------------------|---|-------|--------|-------|------|-------|---------|
| 10061-CAB | IEEE 802.11b WiFi 2.4 GHz (DSSS, 11 Mbps)      | X | 32.15 | 121.96 | 35.02 | 2.04 | 110.0 | ± 9.6 % |
|           |                                                | Y | 2.04  | 75.39  | 19.12 |      | 110.0 |         |
|           |                                                | Z | 2.36  | 78.14  | 20.85 |      | 110.0 |         |
| 10062-CAC | IEEE 802.11a/h WiFi 5 GHz (OFDM, 6 Mbps)       | X | 4.83  | 66.99  | 16.83 | 0.49 | 100.0 | ± 9.6 % |
|           |                                                | Y | 4.37  | 66.55  | 16.24 |      | 100.0 |         |
|           |                                                | Z | 4.43  | 66.90  | 16.40 |      | 100.0 |         |
| 10063-CAC | IEEE 802.11a/h WiFi 5 GHz (OFDM, 9 Mbps)       | X | 4.85  | 67.11  | 16.95 | 0.72 | 100.0 | ± 9.6 % |
|           |                                                | Y | 4.38  | 66.62  | 16.31 |      | 100.0 |         |
|           |                                                | Z | 4.44  | 66.97  | 16.47 |      | 100.0 |         |
| 10064-CAC | IEEE 802.11a/h WiFi 5 GHz (OFDM, 12 Mbps)      | X | 5.19  | 67.41  | 17.17 | 0.86 | 100.0 | ± 9.6 % |
|           |                                                | Y | 4.62  | 66.81  | 16.50 |      | 100.0 |         |
|           |                                                | Z | 4.67  | 67.13  | 16.63 |      | 100.0 |         |
| 10065-CAC | IEEE 802.11a/h WiFi 5 GHz (OFDM, 18 Mbps)      | X | 5.05  | 67.34  | 17.29 | 1.21 | 100.0 | ± 9.6 % |
|           |                                                | Y | 4.49  | 66.66  | 16.55 |      | 100.0 |         |
|           |                                                | Z | 4.54  | 66.96  | 16.68 |      | 100.0 |         |
| 10066-CAC | IEEE 802.11a/h WiFi 5 GHz (OFDM, 24 Mbps)      | X | 5.08  | 67.39  | 17.47 | 1.46 | 100.0 | ± 9.6 % |
|           |                                                | Y | 4.50  | 66.65  | 16.68 |      | 100.0 |         |
|           |                                                | Z | 4.54  | 66.92  | 16.80 |      | 100.0 |         |
| 10067-CAC | IEEE 802.11a/h WiFi 5 GHz (OFDM, 36 Mbps)      | X | 5.35  | 67.39  | 17.83 | 2.04 | 100.0 | ± 9.6 % |
|           |                                                | Y | 4.79  | 66.90  | 17.13 |      | 100.0 |         |
|           |                                                | Z | 4.82  | 67.14  | 17.23 |      | 100.0 |         |
| 10068-CAC | IEEE 802.11a/h WiFi 5 GHz (OFDM, 48 Mbps)      | X | 5.44  | 67.63  | 18.14 | 2.55 | 100.0 | ± 9.6 % |
|           |                                                | Y | 4.82  | 66.81  | 17.26 |      | 100.0 |         |
|           |                                                | Z | 4.85  | 67.03  | 17.35 |      | 100.0 |         |
| 10069-CAC | IEEE 802.11a/h WiFi 5 GHz (OFDM, 54 Mbps)      | X | 5.51  | 67.49  | 18.27 | 2.67 | 100.0 | ± 9.6 % |
|           |                                                | Y | 4.89  | 66.85  | 17.46 |      | 100.0 |         |
|           |                                                | Z | 4.91  | 67.04  | 17.53 |      | 100.0 |         |
| 10071-CAB | IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 9 Mbps)  | X | 5.12  | 67.05  | 17.68 | 1.99 | 100.0 | ± 9.6 % |
|           |                                                | Y | 4.66  | 66.59  | 17.01 |      | 100.0 |         |
|           |                                                | Z | 4.70  | 66.85  | 17.11 |      | 100.0 |         |
| 10072-CAB | IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 12 Mbps) | X | 5.14  | 67.52  | 17.95 | 2.30 | 100.0 | ± 9.6 % |
|           |                                                | Y | 4.62  | 66.83  | 17.17 |      | 100.0 |         |
|           |                                                | Z | 4.65  | 67.08  | 17.27 |      | 100.0 |         |
| 10073-CAB | IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 18 Mbps) | X | 5.21  | 67.69  | 18.29 | 2.83 | 100.0 | ± 9.6 % |
|           |                                                | Y | 4.68  | 67.01  | 17.47 |      | 100.0 |         |
|           |                                                | Z | 4.71  | 67.23  | 17.56 |      | 100.0 |         |
| 10074-CAB | IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 24 Mbps) | X | 5.18  | 67.59  | 18.46 | 3.30 | 100.0 | ± 9.6 % |
|           |                                                | Y | 4.69  | 66.95  | 17.60 |      | 100.0 |         |
|           |                                                | Z | 4.71  | 67.17  | 17.70 |      | 100.0 |         |
| 10075-CAB | IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 36 Mbps) | X | 5.26  | 67.87  | 18.86 | 3.82 | 90.0  | ± 9.6 % |
|           |                                                | Y | 4.73  | 66.99  | 17.83 |      | 90.0  |         |
|           |                                                | Z | 4.74  | 67.18  | 17.92 |      | 90.0  |         |
| 10076-CAB | IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 48 Mbps) | X | 5.23  | 67.53  | 18.89 | 4.15 | 90.0  | ± 9.6 % |
|           |                                                | Y | 4.77  | 66.89  | 18.00 |      | 90.0  |         |
|           |                                                | Z | 4.78  | 67.06  | 18.08 |      | 90.0  |         |
| 10077-CAB | IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 54 Mbps) | X | 5.25  | 67.58  | 18.98 | 4.30 | 90.0  | ± 9.6 % |
|           |                                                | Y | 4.81  | 66.98  | 18.11 |      | 90.0  |         |
|           |                                                | Z | 4.81  | 67.15  | 18.19 |      | 90.0  |         |

EX3DV4- SN:3866

May 31, 2018

|           |                                                    |   |        |        |       |      |       |         |
|-----------|----------------------------------------------------|---|--------|--------|-------|------|-------|---------|
| 10081-CAB | CDMA2000 (1xRTT, RC3)                              | X | 1.61   | 75.86  | 18.26 | 0.00 | 150.0 | ± 9.6 % |
|           |                                                    | Y | 0.40   | 60.59  | 6.95  |      | 150.0 |         |
|           |                                                    | Z | 0.70   | 65.91  | 10.99 |      | 150.0 |         |
| 10082-CAB | IS-54 / IS-136 FDD (TDMA/FDM, PI/4-QPSK, Fullrate) | X | 0.97   | 60.00  | 5.20  | 4.77 | 80.0  | ± 9.6 % |
|           |                                                    | Y | 6.26   | 66.77  | 5.69  |      | 80.0  |         |
|           |                                                    | Z | 1.47   | 63.08  | 5.04  |      | 80.0  |         |
| 10090-DAC | GPRS-FDD (TDMA, GMSK, TN 0-4)                      | X | 100.00 | 114.42 | 27.18 | 6.56 | 60.0  | ± 9.6 % |
|           |                                                    | Y | 2.51   | 68.39  | 11.07 |      | 60.0  |         |
|           |                                                    | Z | 91.59  | 103.46 | 21.38 |      | 60.0  |         |
| 10097-CAB | UMTS-FDD (HSDPA)                                   | X | 2.10   | 70.41  | 17.58 | 0.00 | 150.0 | ± 9.6 % |
|           |                                                    | Y | 1.66   | 67.98  | 15.08 |      | 150.0 |         |
|           |                                                    | Z | 1.92   | 70.17  | 16.49 |      | 150.0 |         |
| 10098-CAB | UMTS-FDD (HSUPA, Subtest 2)                        | X | 2.06   | 70.42  | 17.57 | 0.00 | 150.0 | ± 9.6 % |
|           |                                                    | Y | 1.62   | 67.91  | 15.04 |      | 150.0 |         |
|           |                                                    | Z | 1.88   | 70.12  | 16.47 |      | 150.0 |         |
| 10099-DAC | EDGE-FDD (TDMA, 8PSK, TN 0-4)                      | X | 19.63  | 108.47 | 38.01 | 9.56 | 60.0  | ± 9.6 % |
|           |                                                    | Y | 7.02   | 84.55  | 28.70 |      | 60.0  |         |
|           |                                                    | Z | 7.45   | 87.29  | 30.29 |      | 60.0  |         |
| 10100-CAD | LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)           | X | 3.76   | 73.52  | 18.27 | 0.00 | 150.0 | ± 9.6 % |
|           |                                                    | Y | 2.83   | 69.91  | 16.35 |      | 150.0 |         |
|           |                                                    | Z | 3.08   | 71.35  | 17.24 |      | 150.0 |         |
| 10101-CAD | LTE-FDD (SC-FDMA, 100% RB, 20 MHz, 16-QAM)         | X | 3.49   | 68.90  | 16.77 | 0.00 | 150.0 | ± 9.6 % |
|           |                                                    | Y | 2.97   | 67.19  | 15.62 |      | 150.0 |         |
|           |                                                    | Z | 3.10   | 67.97  | 16.12 |      | 150.0 |         |
| 10102-CAD | LTE-FDD (SC-FDMA, 100% RB, 20 MHz, 64-QAM)         | X | 3.59   | 68.75  | 16.81 | 0.00 | 150.0 | ± 9.6 % |
|           |                                                    | Y | 3.08   | 67.24  | 15.76 |      | 150.0 |         |
|           |                                                    | Z | 3.21   | 67.97  | 16.22 |      | 150.0 |         |
| 10103-CAD | LTE-TDD (SC-FDMA, 100% RB, 20 MHz, QPSK)           | X | 8.41   | 79.82  | 22.06 | 3.98 | 65.0  | ± 9.6 % |
|           |                                                    | Y | 4.84   | 71.96  | 18.48 |      | 65.0  |         |
|           |                                                    | Z | 5.42   | 74.19  | 19.53 |      | 65.0  |         |
| 10104-CAD | LTE-TDD (SC-FDMA, 100% RB, 20 MHz, 16-QAM)         | X | 7.70   | 76.73  | 21.71 | 3.98 | 65.0  | ± 9.6 % |
|           |                                                    | Y | 5.32   | 71.47  | 19.01 |      | 65.0  |         |
|           |                                                    | Z | 5.48   | 72.25  | 19.42 |      | 65.0  |         |
| 10105-CAD | LTE-TDD (SC-FDMA, 100% RB, 20 MHz, 64-QAM)         | X | 7.04   | 74.96  | 21.27 | 3.98 | 65.0  | ± 9.6 % |
|           |                                                    | Y | 4.70   | 68.99  | 18.19 |      | 65.0  |         |
|           |                                                    | Z | 5.14   | 70.85  | 19.09 |      | 65.0  |         |
| 10108-CAE | LTE-FDD (SC-FDMA, 100% RB, 10 MHz, QPSK)           | X | 3.28   | 72.58  | 18.09 | 0.00 | 150.0 | ± 9.6 % |
|           |                                                    | Y | 2.43   | 69.27  | 16.18 |      | 150.0 |         |
|           |                                                    | Z | 2.65   | 70.70  | 17.10 |      | 150.0 |         |
| 10109-CAE | LTE-FDD (SC-FDMA, 100% RB, 10 MHz, 16-QAM)         | X | 3.17   | 68.85  | 16.81 | 0.00 | 150.0 | ± 9.6 % |
|           |                                                    | Y | 2.61   | 67.16  | 15.46 |      | 150.0 |         |
|           |                                                    | Z | 2.76   | 68.08  | 16.06 |      | 150.0 |         |
| 10110-CAE | LTE-FDD (SC-FDMA, 100% RB, 5 MHz, QPSK)            | X | 2.68   | 71.66  | 17.86 | 0.00 | 150.0 | ± 9.6 % |
|           |                                                    | Y | 1.91   | 68.41  | 15.56 |      | 150.0 |         |
|           |                                                    | Z | 2.13   | 70.09  | 16.68 |      | 150.0 |         |
| 10111-CAE | LTE-FDD (SC-FDMA, 100% RB, 5 MHz, 16-QAM)          | X | 2.94   | 70.01  | 17.44 | 0.00 | 150.0 | ± 9.6 % |
|           |                                                    | Y | 2.36   | 68.46  | 15.69 |      | 150.0 |         |
|           |                                                    | Z | 2.60   | 69.97  | 16.63 |      | 150.0 |         |

EX3DV4- SN:3866

May 31, 2018

|           |                                                |   |      |       |       |      |       |         |
|-----------|------------------------------------------------|---|------|-------|-------|------|-------|---------|
| 10112-CAE | LTE-FDD (SC-FDMA, 100% RB, 10 MHz, 64-QAM)     | X | 3.28 | 68.68 | 16.79 | 0.00 | 150.0 | ± 9.6 % |
|           |                                                | Y | 2.74 | 67.25 | 15.57 |      | 150.0 |         |
|           |                                                | Z | 2.89 | 68.13 | 16.13 |      | 150.0 |         |
| 10113-CAE | LTE-FDD (SC-FDMA, 100% RB, 5 MHz, 64-QAM)      | X | 3.09 | 69.94 | 17.46 | 0.00 | 150.0 | ± 9.6 % |
|           |                                                | Y | 2.52 | 68.70 | 15.88 |      | 150.0 |         |
|           |                                                | Z | 2.75 | 70.13 | 16.75 |      | 150.0 |         |
| 10114-CAC | IEEE 802.11n (HT Greenfield, 13.5 Mbps, BPSK)  | X | 5.23 | 67.48 | 16.65 | 0.00 | 150.0 | ± 9.6 % |
|           |                                                | Y | 4.84 | 67.00 | 16.28 |      | 150.0 |         |
|           |                                                | Z | 4.90 | 67.34 | 16.43 |      | 150.0 |         |
| 10115-CAC | IEEE 802.11n (HT Greenfield, 81 Mbps, 16-QAM)  | X | 5.59 | 67.72 | 16.77 | 0.00 | 150.0 | ± 9.6 % |
|           |                                                | Y | 5.08 | 67.08 | 16.32 |      | 150.0 |         |
|           |                                                | Z | 5.13 | 67.37 | 16.44 |      | 150.0 |         |
| 10116-CAC | IEEE 802.11n (HT Greenfield, 135 Mbps, 64-QAM) | X | 5.36 | 67.74 | 16.70 | 0.00 | 150.0 | ± 9.6 % |
|           |                                                | Y | 4.92 | 67.20 | 16.31 |      | 150.0 |         |
|           |                                                | Z | 4.98 | 67.52 | 16.45 |      | 150.0 |         |
| 10117-CAC | IEEE 802.11n (HT Mixed, 13.5 Mbps, BPSK)       | X | 5.24 | 67.49 | 16.68 | 0.00 | 150.0 | ± 9.6 % |
|           |                                                | Y | 4.83 | 66.95 | 16.28 |      | 150.0 |         |
|           |                                                | Z | 4.89 | 67.29 | 16.43 |      | 150.0 |         |
| 10118-CAC | IEEE 802.11n (HT Mixed, 81 Mbps, 16-QAM)       | X | 5.66 | 67.87 | 16.85 | 0.00 | 150.0 | ± 9.6 % |
|           |                                                | Y | 5.16 | 67.27 | 16.43 |      | 150.0 |         |
|           |                                                | Z | 5.20 | 67.52 | 16.53 |      | 150.0 |         |
| 10119-CAC | IEEE 802.11n (HT Mixed, 135 Mbps, 64-QAM)      | X | 5.33 | 67.68 | 16.69 | 0.00 | 150.0 | ± 9.6 % |
|           |                                                | Y | 4.92 | 67.19 | 16.31 |      | 150.0 |         |
|           |                                                | Z | 4.97 | 67.51 | 16.46 |      | 150.0 |         |
| 10140-CAD | LTE-FDD (SC-FDMA, 100% RB, 15 MHz, 16-QAM)     | X | 3.64 | 68.73 | 16.72 | 0.00 | 150.0 | ± 9.6 % |
|           |                                                | Y | 3.10 | 67.23 | 15.65 |      | 150.0 |         |
|           |                                                | Z | 3.23 | 67.98 | 16.12 |      | 150.0 |         |
| 10141-CAD | LTE-FDD (SC-FDMA, 100% RB, 15 MHz, 64-QAM)     | X | 3.75 | 68.73 | 16.84 | 0.00 | 150.0 | ± 9.6 % |
|           |                                                | Y | 3.23 | 67.43 | 15.88 |      | 150.0 |         |
|           |                                                | Z | 3.36 | 68.16 | 16.32 |      | 150.0 |         |
| 10142-CAD | LTE-FDD (SC-FDMA, 100% RB, 3 MHz, QPSK)        | X | 2.50 | 72.12 | 17.94 | 0.00 | 150.0 | ± 9.6 % |
|           |                                                | Y | 1.65 | 68.10 | 14.67 |      | 150.0 |         |
|           |                                                | Z | 1.94 | 70.53 | 16.23 |      | 150.0 |         |
| 10143-CAD | LTE-FDD (SC-FDMA, 100% RB, 3 MHz, 16-QAM)      | X | 2.95 | 71.49 | 17.71 | 0.00 | 150.0 | ± 9.6 % |
|           |                                                | Y | 2.12 | 68.46 | 14.53 |      | 150.0 |         |
|           |                                                | Z | 2.52 | 71.14 | 16.09 |      | 150.0 |         |
| 10144-CAD | LTE-FDD (SC-FDMA, 100% RB, 3 MHz, 64-QAM)      | X | 2.62 | 68.66 | 15.91 | 0.00 | 150.0 | ± 9.6 % |
|           |                                                | Y | 1.78 | 65.25 | 12.38 |      | 150.0 |         |
|           |                                                | Z | 2.00 | 66.87 | 13.49 |      | 150.0 |         |
| 10145-CAE | LTE-FDD (SC-FDMA, 100% RB, 1.4 MHz, QPSK)      | X | 2.08 | 72.58 | 16.71 | 0.00 | 150.0 | ± 9.6 % |
|           |                                                | Y | 0.62 | 60.00 | 6.54  |      | 150.0 |         |
|           |                                                | Z | 0.76 | 61.85 | 8.27  |      | 150.0 |         |
| 10146-CAE | LTE-FDD (SC-FDMA, 100% RB, 1.4 MHz, 16-QAM)    | X | 4.74 | 77.79 | 17.95 | 0.00 | 150.0 | ± 9.6 % |
|           |                                                | Y | 0.91 | 59.91 | 6.14  |      | 150.0 |         |
|           |                                                | Z | 1.03 | 60.93 | 6.75  |      | 150.0 |         |
| 10147-CAE | LTE-FDD (SC-FDMA, 100% RB, 1.4 MHz, 64-QAM)    | X | 8.02 | 85.30 | 20.79 | 0.00 | 150.0 | ± 9.6 % |
|           |                                                | Y | 0.95 | 60.19 | 6.39  |      | 150.0 |         |
|           |                                                | Z | 1.09 | 61.38 | 7.09  |      | 150.0 |         |

EX3DV4- SN:3866

May 31, 2018

|           |                                            |   |      |       |       |      |       |         |
|-----------|--------------------------------------------|---|------|-------|-------|------|-------|---------|
| 10149-CAD | LTE-FDD (SC-FDMA, 50% RB, 20 MHz, 16-QAM)  | X | 3.18 | 68.92 | 16.86 | 0.00 | 150.0 | ± 9.6 % |
|           |                                            | Y | 2.63 | 67.24 | 15.52 |      | 150.0 |         |
|           |                                            | Z | 2.78 | 68.16 | 16.12 |      | 150.0 |         |
| 10150-CAD | LTE-FDD (SC-FDMA, 50% RB, 20 MHz, 64-QAM)  | X | 3.29 | 68.74 | 16.83 | 0.00 | 150.0 | ± 9.6 % |
|           |                                            | Y | 2.75 | 67.32 | 15.62 |      | 150.0 |         |
|           |                                            | Z | 2.90 | 68.21 | 16.19 |      | 150.0 |         |
| 10151-CAD | LTE-TDD (SC-FDMA, 50% RB, 20 MHz, QPSK)    | X | 9.25 | 82.93 | 23.38 | 3.98 | 65.0  | ± 9.6 % |
|           |                                            | Y | 5.36 | 75.32 | 19.83 |      | 65.0  |         |
|           |                                            | Z | 5.77 | 76.93 | 20.64 |      | 65.0  |         |
| 10152-CAD | LTE-TDD (SC-FDMA, 50% RB, 20 MHz, 16-QAM)  | X | 7.33 | 77.04 | 21.65 | 3.98 | 65.0  | ± 9.6 % |
|           |                                            | Y | 4.81 | 71.16 | 18.40 |      | 65.0  |         |
|           |                                            | Z | 4.98 | 72.05 | 18.88 |      | 65.0  |         |
| 10153-CAD | LTE-TDD (SC-FDMA, 50% RB, 20 MHz, 64-QAM)  | X | 7.72 | 77.89 | 22.37 | 3.98 | 65.0  | ± 9.6 % |
|           |                                            | Y | 5.21 | 72.44 | 19.38 |      | 65.0  |         |
|           |                                            | Z | 5.39 | 73.30 | 19.82 |      | 65.0  |         |
| 10154-CAE | LTE-FDD (SC-FDMA, 50% RB, 10 MHz, QPSK)    | X | 2.79 | 72.40 | 18.27 | 0.00 | 150.0 | ± 9.6 % |
|           |                                            | Y | 1.96 | 68.90 | 15.86 |      | 150.0 |         |
|           |                                            | Z | 2.20 | 70.63 | 16.99 |      | 150.0 |         |
| 10155-CAE | LTE-FDD (SC-FDMA, 50% RB, 10 MHz, 16-QAM)  | X | 2.94 | 70.00 | 17.44 | 0.00 | 150.0 | ± 9.6 % |
|           |                                            | Y | 2.37 | 68.50 | 15.72 |      | 150.0 |         |
|           |                                            | Z | 2.60 | 70.02 | 16.66 |      | 150.0 |         |
| 10156-CAE | LTE-FDD (SC-FDMA, 50% RB, 5 MHz, QPSK)     | X | 2.45 | 73.08 | 18.25 | 0.00 | 150.0 | ± 9.6 % |
|           |                                            | Y | 1.44 | 67.52 | 13.85 |      | 150.0 |         |
|           |                                            | Z | 1.79 | 70.64 | 15.81 |      | 150.0 |         |
| 10157-CAE | LTE-FDD (SC-FDMA, 50% RB, 5 MHz, 16-QAM)   | X | 2.55 | 70.02 | 16.44 | 0.00 | 150.0 | ± 9.6 % |
|           |                                            | Y | 1.56 | 65.08 | 11.79 |      | 150.0 |         |
|           |                                            | Z | 1.83 | 67.35 | 13.31 |      | 150.0 |         |
| 10158-CAE | LTE-FDD (SC-FDMA, 50% RB, 10 MHz, 64-QAM)  | X | 3.10 | 70.01 | 17.51 | 0.00 | 150.0 | ± 9.6 % |
|           |                                            | Y | 2.53 | 68.81 | 15.95 |      | 150.0 |         |
|           |                                            | Z | 2.76 | 70.25 | 16.83 |      | 150.0 |         |
| 10159-CAE | LTE-FDD (SC-FDMA, 50% RB, 5 MHz, 64-QAM)   | X | 2.72 | 70.68 | 16.81 | 0.00 | 150.0 | ± 9.6 % |
|           |                                            | Y | 1.63 | 65.40 | 12.01 |      | 150.0 |         |
|           |                                            | Z | 1.94 | 67.86 | 13.59 |      | 150.0 |         |
| 10160-CAD | LTE-FDD (SC-FDMA, 50% RB, 15 MHz, QPSK)    | X | 3.08 | 70.55 | 17.46 | 0.00 | 150.0 | ± 9.6 % |
|           |                                            | Y | 2.48 | 68.64 | 16.02 |      | 150.0 |         |
|           |                                            | Z | 2.64 | 69.71 | 16.74 |      | 150.0 |         |
| 10161-CAD | LTE-FDD (SC-FDMA, 50% RB, 15 MHz, 16-QAM)  | X | 3.19 | 68.68 | 16.82 | 0.00 | 150.0 | ± 9.6 % |
|           |                                            | Y | 2.64 | 67.29 | 15.48 |      | 150.0 |         |
|           |                                            | Z | 2.80 | 68.24 | 16.09 |      | 150.0 |         |
| 10162-CAD | LTE-FDD (SC-FDMA, 50% RB, 15 MHz, 64-QAM)  | X | 3.29 | 68.68 | 16.85 | 0.00 | 150.0 | ± 9.6 % |
|           |                                            | Y | 2.76 | 67.53 | 15.64 |      | 150.0 |         |
|           |                                            | Z | 2.91 | 68.48 | 16.24 |      | 150.0 |         |
| 10166-CAE | LTE-FDD (SC-FDMA, 50% RB, 1.4 MHz, QPSK)   | X | 4.04 | 71.00 | 19.90 | 3.01 | 150.0 | ± 9.6 % |
|           |                                            | Y | 3.16 | 69.13 | 18.77 |      | 150.0 |         |
|           |                                            | Z | 3.39 | 70.78 | 19.66 |      | 150.0 |         |
| 10167-CAE | LTE-FDD (SC-FDMA, 50% RB, 1.4 MHz, 16-QAM) | X | 5.36 | 74.73 | 20.65 | 3.01 | 150.0 | ± 9.6 % |
|           |                                            | Y | 3.79 | 71.67 | 18.97 |      | 150.0 |         |
|           |                                            | Z | 4.56 | 75.64 | 20.76 |      | 150.0 |         |

EX3DV4- SN:3866

May 31, 2018

|           |                                            |   |       |        |       |      |       |         |
|-----------|--------------------------------------------|---|-------|--------|-------|------|-------|---------|
| 10168-CAE | LTE-FDD (SC-FDMA, 50% RB, 1.4 MHz, 64-QAM) | X | 6.03  | 77.24  | 22.03 | 3.01 | 150.0 | ± 9.6 % |
|           |                                            | Y | 4.43  | 75.05  | 20.89 |      | 150.0 |         |
|           |                                            | Z | 5.62  | 80.10  | 23.00 |      | 150.0 |         |
| 10169-CAD | LTE-FDD (SC-FDMA, 1 RB, 20 MHz, QPSK)      | X | 3.82  | 73.12  | 20.79 | 3.01 | 150.0 | ± 9.6 % |
|           |                                            | Y | 2.61  | 67.68  | 18.06 |      | 150.0 |         |
|           |                                            | Z | 2.91  | 70.50  | 19.54 |      | 150.0 |         |
| 10170-CAD | LTE-FDD (SC-FDMA, 1 RB, 20 MHz, 16-QAM)    | X | 6.54  | 82.37  | 24.08 | 3.01 | 150.0 | ± 9.6 % |
|           |                                            | Y | 3.53  | 73.51  | 20.46 |      | 150.0 |         |
|           |                                            | Z | 5.29  | 82.18  | 24.01 |      | 150.0 |         |
| 10171-AAD | LTE-FDD (SC-FDMA, 1 RB, 20 MHz, 64-QAM)    | X | 4.84  | 75.97  | 20.60 | 3.01 | 150.0 | ± 9.6 % |
|           |                                            | Y | 2.80  | 68.67  | 17.18 |      | 150.0 |         |
|           |                                            | Z | 3.59  | 74.12  | 19.63 |      | 150.0 |         |
| 10172-CAD | LTE-TDD (SC-FDMA, 1 RB, 20 MHz, QPSK)      | X | 30.51 | 113.43 | 34.95 | 6.02 | 65.0  | ± 9.6 % |
|           |                                            | Y | 3.52  | 75.21  | 21.66 |      | 65.0  |         |
|           |                                            | Z | 4.81  | 82.95  | 25.11 |      | 65.0  |         |
| 10173-CAD | LTE-TDD (SC-FDMA, 1 RB, 20 MHz, 16-QAM)    | X | 81.01 | 124.21 | 35.55 | 6.02 | 65.0  | ± 9.6 % |
|           |                                            | Y | 6.10  | 82.01  | 22.32 |      | 65.0  |         |
|           |                                            | Z | 15.21 | 99.20  | 28.02 |      | 65.0  |         |
| 10174-CAD | LTE-TDD (SC-FDMA, 1 RB, 20 MHz, 64-QAM)    | X | 42.22 | 110.92 | 31.54 | 6.02 | 65.0  | ± 9.6 % |
|           |                                            | Y | 3.34  | 72.20  | 18.19 |      | 65.0  |         |
|           |                                            | Z | 7.70  | 87.16  | 23.71 |      | 65.0  |         |
| 10175-CAE | LTE-FDD (SC-FDMA, 1 RB, 10 MHz, QPSK)      | X | 3.75  | 72.64  | 20.47 | 3.01 | 150.0 | ± 9.6 % |
|           |                                            | Y | 2.57  | 67.33  | 17.78 |      | 150.0 |         |
|           |                                            | Z | 2.86  | 70.07  | 19.23 |      | 150.0 |         |
| 10176-CAE | LTE-FDD (SC-FDMA, 1 RB, 10 MHz, 16-QAM)    | X | 6.55  | 82.40  | 24.09 | 3.01 | 150.0 | ± 9.6 % |
|           |                                            | Y | 3.54  | 73.53  | 20.47 |      | 150.0 |         |
|           |                                            | Z | 5.31  | 82.22  | 24.03 |      | 150.0 |         |
| 10177-CAG | LTE-FDD (SC-FDMA, 1 RB, 5 MHz, QPSK)       | X | 3.79  | 72.88  | 20.60 | 3.01 | 150.0 | ± 9.6 % |
|           |                                            | Y | 2.59  | 67.48  | 17.87 |      | 150.0 |         |
|           |                                            | Z | 2.89  | 70.25  | 19.33 |      | 150.0 |         |
| 10178-CAE | LTE-FDD (SC-FDMA, 1 RB, 5 MHz, 16-QAM)     | X | 6.40  | 81.92  | 23.88 | 3.01 | 150.0 | ± 9.6 % |
|           |                                            | Y | 3.51  | 73.32  | 20.36 |      | 150.0 |         |
|           |                                            | Z | 5.22  | 81.88  | 23.87 |      | 150.0 |         |
| 10179-CAE | LTE-FDD (SC-FDMA, 1 RB, 10 MHz, 64-QAM)    | X | 5.57  | 78.86  | 22.14 | 3.01 | 150.0 | ± 9.6 % |
|           |                                            | Y | 3.10  | 70.80  | 18.61 |      | 150.0 |         |
|           |                                            | Z | 4.30  | 77.75  | 21.59 |      | 150.0 |         |
| 10180-CAE | LTE-FDD (SC-FDMA, 1 RB, 5 MHz, 64-QAM)     | X | 4.81  | 75.81  | 20.51 | 3.01 | 150.0 | ± 9.6 % |
|           |                                            | Y | 2.79  | 68.61  | 17.14 |      | 150.0 |         |
|           |                                            | Z | 3.58  | 74.03  | 19.58 |      | 150.0 |         |
| 10181-CAD | LTE-FDD (SC-FDMA, 1 RB, 15 MHz, QPSK)      | X | 3.79  | 72.85  | 20.59 | 3.01 | 150.0 | ± 9.6 % |
|           |                                            | Y | 2.59  | 67.46  | 17.86 |      | 150.0 |         |
|           |                                            | Z | 2.88  | 70.23  | 19.32 |      | 150.0 |         |
| 10182-CAD | LTE-FDD (SC-FDMA, 1 RB, 15 MHz, 16-QAM)    | X | 6.39  | 81.89  | 23.87 | 3.01 | 150.0 | ± 9.6 % |
|           |                                            | Y | 3.50  | 73.29  | 20.34 |      | 150.0 |         |
|           |                                            | Z | 5.21  | 81.83  | 23.85 |      | 150.0 |         |
| 10183-AAC | LTE-FDD (SC-FDMA, 1 RB, 15 MHz, 64-QAM)    | X | 4.80  | 75.78  | 20.50 | 3.01 | 150.0 | ± 9.6 % |
|           |                                            | Y | 2.79  | 68.59  | 17.13 |      | 150.0 |         |
|           |                                            | Z | 3.57  | 74.00  | 19.56 |      | 150.0 |         |

EX3DV4- SN:3866

May 31, 2018

|           |                                               |   |      |       |       |      |       |         |
|-----------|-----------------------------------------------|---|------|-------|-------|------|-------|---------|
| 10184-CAD | LTE-FDD (SC-FDMA, 1 RB, 3 MHz, QPSK)          | X | 3.80 | 72.91 | 20.62 | 3.01 | 150.0 | ± 9.6 % |
|           |                                               | Y | 2.59 | 67.50 | 17.89 |      | 150.0 |         |
|           |                                               | Z | 2.90 | 70.28 | 19.35 |      | 150.0 |         |
| 10185-CAD | LTE-FDD (SC-FDMA, 1 RB, 3 MHz, 16-QAM)        | X | 6.43 | 81.99 | 23.91 | 3.01 | 150.0 | ± 9.6 % |
|           |                                               | Y | 3.52 | 73.38 | 20.39 |      | 150.0 |         |
|           |                                               | Z | 5.25 | 81.97 | 23.91 |      | 150.0 |         |
| 10186-AAD | LTE-FDD (SC-FDMA, 1 RB, 3 MHz, 64-QAM)        | X | 4.82 | 75.87 | 20.54 | 3.01 | 150.0 | ± 9.6 % |
|           |                                               | Y | 2.80 | 68.65 | 17.16 |      | 150.0 |         |
|           |                                               | Z | 3.60 | 74.09 | 19.61 |      | 150.0 |         |
| 10187-CAE | LTE-FDD (SC-FDMA, 1 RB, 1.4 MHz, QPSK)        | X | 3.81 | 72.95 | 20.67 | 3.01 | 150.0 | ± 9.6 % |
|           |                                               | Y | 2.61 | 67.59 | 17.98 |      | 150.0 |         |
|           |                                               | Z | 2.91 | 70.39 | 19.45 |      | 150.0 |         |
| 10188-CAE | LTE-FDD (SC-FDMA, 1 RB, 1.4 MHz, 16-QAM)      | X | 6.79 | 83.15 | 24.45 | 3.01 | 150.0 | ± 9.6 % |
|           |                                               | Y | 3.65 | 74.15 | 20.83 |      | 150.0 |         |
|           |                                               | Z | 5.58 | 83.26 | 24.51 |      | 150.0 |         |
| 10189-AAE | LTE-FDD (SC-FDMA, 1 RB, 1.4 MHz, 64-QAM)      | X | 4.99 | 76.54 | 20.90 | 3.01 | 150.0 | ± 9.6 % |
|           |                                               | Y | 2.86 | 69.07 | 17.45 |      | 150.0 |         |
|           |                                               | Z | 3.73 | 74.78 | 20.00 |      | 150.0 |         |
| 10193-CAC | IEEE 802.11n (HT Greenfield, 6.5 Mbps, BPSK)  | X | 4.67 | 66.95 | 16.48 | 0.00 | 150.0 | ± 9.6 % |
|           |                                               | Y | 4.24 | 66.64 | 15.97 |      | 150.0 |         |
|           |                                               | Z | 4.33 | 67.09 | 16.19 |      | 150.0 |         |
| 10194-CAC | IEEE 802.11n (HT Greenfield, 39 Mbps, 16-QAM) | X | 4.87 | 67.32 | 16.59 | 0.00 | 150.0 | ± 9.6 % |
|           |                                               | Y | 4.38 | 66.89 | 16.11 |      | 150.0 |         |
|           |                                               | Z | 4.47 | 67.32 | 16.32 |      | 150.0 |         |
| 10195-CAC | IEEE 802.11n (HT Greenfield, 65 Mbps, 64-QAM) | X | 4.91 | 67.33 | 16.59 | 0.00 | 150.0 | ± 9.6 % |
|           |                                               | Y | 4.42 | 66.91 | 16.13 |      | 150.0 |         |
|           |                                               | Z | 4.50 | 67.33 | 16.33 |      | 150.0 |         |
| 10196-CAC | IEEE 802.11n (HT Mixed, 6.5 Mbps, BPSK)       | X | 4.69 | 67.05 | 16.52 | 0.00 | 150.0 | ± 9.6 % |
|           |                                               | Y | 4.23 | 66.65 | 15.96 |      | 150.0 |         |
|           |                                               | Z | 4.31 | 67.09 | 16.18 |      | 150.0 |         |
| 10197-CAC | IEEE 802.11n (HT Mixed, 39 Mbps, 16-QAM)      | X | 4.89 | 67.34 | 16.60 | 0.00 | 150.0 | ± 9.6 % |
|           |                                               | Y | 4.39 | 66.89 | 16.12 |      | 150.0 |         |
|           |                                               | Z | 4.47 | 67.32 | 16.32 |      | 150.0 |         |
| 10198-CAC | IEEE 802.11n (HT Mixed, 65 Mbps, 64-QAM)      | X | 4.92 | 67.34 | 16.60 | 0.00 | 150.0 | ± 9.6 % |
|           |                                               | Y | 4.41 | 66.91 | 16.13 |      | 150.0 |         |
|           |                                               | Z | 4.49 | 67.32 | 16.33 |      | 150.0 |         |
| 10219-CAC | IEEE 802.11n (HT Mixed, 7.2 Mbps, BPSK)       | X | 4.64 | 67.08 | 16.49 | 0.00 | 150.0 | ± 9.6 % |
|           |                                               | Y | 4.18 | 66.68 | 15.93 |      | 150.0 |         |
|           |                                               | Z | 4.27 | 67.14 | 16.16 |      | 150.0 |         |
| 10220-CAC | IEEE 802.11n (HT Mixed, 43.3 Mbps, 16-QAM)    | X | 4.89 | 67.33 | 16.60 | 0.00 | 150.0 | ± 9.6 % |
|           |                                               | Y | 4.38 | 66.85 | 16.10 |      | 150.0 |         |
|           |                                               | Z | 4.46 | 67.28 | 16.31 |      | 150.0 |         |
| 10221-CAC | IEEE 802.11n (HT Mixed, 72.2 Mbps, 64-QAM)    | X | 4.92 | 67.27 | 16.59 | 0.00 | 150.0 | ± 9.6 % |
|           |                                               | Y | 4.43 | 66.85 | 16.12 |      | 150.0 |         |
|           |                                               | Z | 4.51 | 67.26 | 16.32 |      | 150.0 |         |
| 10222-CAC | IEEE 802.11n (HT Mixed, 15 Mbps, BPSK)        | X | 5.22 | 67.52 | 16.68 | 0.00 | 150.0 | ± 9.6 % |
|           |                                               | Y | 4.80 | 66.93 | 16.26 |      | 150.0 |         |
|           |                                               | Z | 4.86 | 67.28 | 16.41 |      | 150.0 |         |