



Accredited by the Swiss Accreditation Service (SAS)  
 The Swiss Accreditation Service is one of the signatories to the EA  
 Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 0108**

Client **PC Test**

Certificate No: **D835V2-4d119\_Apr15**

## CALIBRATION CERTIFICATE

Object **D835V2 - SN:4d119**

Calibration procedure(s) **QA CAL-05.v9**  
 Calibration procedure for dipole validation kits above 700 MHz

Calibration date: **April 13, 2015**

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI).  
 The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature ( $22 \pm 3$ )°C and humidity < 70%.

Calibration Equipment used (M&TE critical for calibration)

| Primary Standards           | ID #               | Cal Date (Certificate No.)        | Scheduled Calibration  |
|-----------------------------|--------------------|-----------------------------------|------------------------|
| Power meter EPM-442A        | GB37480704         | 07-Oct-14 (No. 217-02020)         | Oct-15                 |
| Power sensor HP 8481A       | US37292783         | 07-Oct-14 (No. 217-02020)         | Oct-15                 |
| Power sensor HP 8481A       | MY41092317         | 07-Oct-14 (No. 217-02021)         | Oct-15                 |
| Reference 20 dB Attenuator  | SN: 5058 (20k)     | 01-Apr-15 (No. 217-02131)         | Mar-16                 |
| Type-N mismatch combination | SN: 5047.2 / 06327 | 01-Apr-15 (No. 217-02134)         | Mar-16                 |
| Reference Probe ES3DV3      | SN: 3205           | 30-Dec-14 (No. ES3-3205_Dec14)    | Dec-15                 |
| DAE4                        | SN: 601            | 18-Aug-14 (No. DAE4-601_Aug14)    | Aug-15                 |
| Secondary Standards         | ID #               | Check Date (in house)             | Scheduled Check        |
| RF generator R&S SMT-06     | 100005             | 04-Aug-99 (in house check Oct-13) | In house check: Oct-16 |
| Network Analyzer HP 8753E   | US37390585 S4206   | 18-Oct-01 (in house check Oct-14) | In house check: Oct-15 |

Calibrated by: **Israe Elnaouq** **Laboratory Technician**

Approved by: **Katja Pokovic** **Technical Manager**

Signature

Issued: April 13, 2015

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.



Accredited by the Swiss Accreditation Service (SAS)  
The Swiss Accreditation Service is one of the signatories to the EA  
Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 0108**

**Glossary:**

|       |                                 |
|-------|---------------------------------|
| TSL   | tissue simulating liquid        |
| ConvF | sensitivity in TSL / NORM x,y,z |
| N/A   | not applicable or not measured  |

**Calibration is Performed According to the Following Standards:**

- IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013
- IEC 62209-1, "Procedure to measure the Specific Absorption Rate (SAR) for hand-held devices used in close proximity to the ear (frequency range of 300 MHz to 3 GHz)", February 2005
- KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

**Additional Documentation:**

- DASY4/5 System Handbook

**Methods Applied and Interpretation of Parameters:**

- Measurement Conditions:** Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL:** The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss:** These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay:** One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured:** SAR measured at the stated antenna input power.
- SAR normalized:** SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters:** The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor  $k=2$ , which for a normal distribution corresponds to a coverage probability of approximately 95%.

## Measurement Conditions

DASY system configuration, as far as not given on page 1.

|                              |                        |             |
|------------------------------|------------------------|-------------|
| DASY Version                 | DASY5                  | V52.8.8     |
| Extrapolation                | Advanced Extrapolation |             |
| Phantom                      | Modular Flat Phantom   |             |
| Distance Dipole Center - TSL | 15 mm                  | with Spacer |
| Zoom Scan Resolution         | dx, dy, dz = 5 mm      |             |
| Frequency                    | 835 MHz $\pm$ 1 MHz    |             |

## Head TSL parameters

The following parameters and calculations were applied.

|                                         | Temperature         | Permittivity   | Conductivity         |
|-----------------------------------------|---------------------|----------------|----------------------|
| Nominal Head TSL parameters             | 22.0 °C             | 41.5           | 0.90 mho/m           |
| Measured Head TSL parameters            | (22.0 $\pm$ 0.2) °C | 40.9 $\pm$ 6 % | 0.94 mho/m $\pm$ 6 % |
| Head TSL temperature change during test | < 0.5 °C            | ----           | ----                 |

## SAR result with Head TSL

| SAR averaged over 1 cm <sup>3</sup> (1 g) of Head TSL | Condition          |                              |
|-------------------------------------------------------|--------------------|------------------------------|
| SAR measured                                          | 250 mW input power | 2.43 W/kg                    |
| SAR for nominal Head TSL parameters                   | normalized to 1W   | 9.38 W/kg $\pm$ 17.0 % (k=2) |

| SAR averaged over 10 cm <sup>3</sup> (10 g) of Head TSL | condition          |                              |
|---------------------------------------------------------|--------------------|------------------------------|
| SAR measured                                            | 250 mW input power | 1.57 W/kg                    |
| SAR for nominal Head TSL parameters                     | normalized to 1W   | 6.11 W/kg $\pm$ 16.5 % (k=2) |

## Body TSL parameters

The following parameters and calculations were applied.

|                                         | Temperature         | Permittivity   | Conductivity         |
|-----------------------------------------|---------------------|----------------|----------------------|
| Nominal Body TSL parameters             | 22.0 °C             | 55.2           | 0.97 mho/m           |
| Measured Body TSL parameters            | (22.0 $\pm$ 0.2) °C | 55.4 $\pm$ 6 % | 1.01 mho/m $\pm$ 6 % |
| Body TSL temperature change during test | < 0.5 °C            | ----           | ----                 |

## SAR result with Body TSL

| SAR averaged over 1 cm <sup>3</sup> (1 g) of Body TSL | Condition          |                              |
|-------------------------------------------------------|--------------------|------------------------------|
| SAR measured                                          | 250 mW input power | 2.37 W/kg                    |
| SAR for nominal Body TSL parameters                   | normalized to 1W   | 9.20 W/kg $\pm$ 17.0 % (k=2) |

| SAR averaged over 10 cm <sup>3</sup> (10 g) of Body TSL | condition          |                              |
|---------------------------------------------------------|--------------------|------------------------------|
| SAR measured                                            | 250 mW input power | 1.55 W/kg                    |
| SAR for nominal Body TSL parameters                     | normalized to 1W   | 6.06 W/kg $\pm$ 16.5 % (k=2) |

## Appendix (Additional assessments outside the scope of SCS 0108)

### Antenna Parameters with Head TSL

|                                      |                                |
|--------------------------------------|--------------------------------|
| Impedance, transformed to feed point | 50.2 $\Omega$ - 2.2 j $\Omega$ |
| Return Loss                          | - 33.3 dB                      |

### Antenna Parameters with Body TSL

|                                      |                                |
|--------------------------------------|--------------------------------|
| Impedance, transformed to feed point | 47.7 $\Omega$ - 4.9 j $\Omega$ |
| Return Loss                          | - 25.1 dB                      |

### General Antenna Parameters and Design

|                                  |          |
|----------------------------------|----------|
| Electrical Delay (one direction) | 1.386 ns |
|----------------------------------|----------|

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard.

No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

### Additional EUT Data

|                 |               |
|-----------------|---------------|
| Manufactured by | SPEAG         |
| Manufactured on | June 29, 2010 |

## DASY5 Validation Report for Head TSL

Date: 13.04.2015

Test Laboratory: SPEAG, Zurich, Switzerland

**DUT: Dipole 835 MHz; Type: D835V2; Serial: D835V2 - SN:4d119**

Communication System: UID 0 - CW; Frequency: 835 MHz

Medium parameters used:  $f = 835 \text{ MHz}$ ;  $\sigma = 0.94 \text{ S/m}$ ;  $\epsilon_r = 40.9$ ;  $\rho = 1000 \text{ kg/m}^3$

Phantom section: Flat Section

Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2011)

DASY52 Configuration:

- Probe: ES3DV3 - SN3205; ConvF(6.2, 6.2, 6.2); Calibrated: 30.12.2014;
- Sensor-Surface: 3mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 18.08.2014
- Phantom: Flat Phantom 4.9L; Type: QD000P49AA; Serial: 1001
- DASY52 52.8.8(1222); SEMCAD X 14.6.10(7331)

**Dipole Calibration for Head Tissue/Pin=250 mW, d=15mm/Zoom Scan (7x7x7)/Cube 0:**


Measurement grid:  $dx=5\text{mm}$ ,  $dy=5\text{mm}$ ,  $dz=5\text{mm}$

Reference Value =  $56.77 \text{ V/m}$ ; Power Drift =  $0.01 \text{ dB}$

Peak SAR (extrapolated) =  $3.64 \text{ W/kg}$

**SAR(1 g) =  $2.43 \text{ W/kg}$ ; SAR(10 g) =  $1.57 \text{ W/kg}$**

Maximum value of SAR (measured) =  $2.85 \text{ W/kg}$



0 dB =  $2.85 \text{ W/kg}$  =  $4.55 \text{ dBW/kg}$



# Impedance Measurement Plot for Head TSL

CH1 S11 1 U FS 13 Apr 2015 13:42:59  
 2: 50.213  $\Omega$  -2.1602  $\Omega$  88.237 pF 835.000 000 MHz

\*  
 Del  
 CA

Avg  
 16

H1d



# DASY5 Validation Report for Body TSL

Date: 13.04.2015

Test Laboratory: SPEAG, Zurich, Switzerland

**DUT: Dipole 835 MHz; Type: D835V2; Serial: D835V2 - SN:4d119**

Communication System: UID 0 - CW; Frequency: 835 MHz

Medium parameters used:  $f = 835 \text{ MHz}$ ;  $\sigma = 1.01 \text{ S/m}$ ;  $\epsilon_r = 55.4$ ;  $\rho = 1000 \text{ kg/m}^3$

Phantom section: Flat Section

Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2011)

DASY52 Configuration:

- Probe: ES3DV3 - SN3205; ConvF(6.17, 6.17, 6.17); Calibrated: 30.12.2014;
- Sensor-Surface: 3mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 18.08.2014
- Phantom: Flat Phantom 4.9L; Type: QD000P49AA; Serial: 1001
- DASY52 52.8.8(1222); SEMCAD X 14.6.10(7331)

**Dipole Calibration for Body Tissue/Pin=250 mW,  $d=15\text{mm}$ /Zoom Scan (7x7x7)/Cube 0:**


Measurement grid:  $dx=5\text{mm}$ ,  $dy=5\text{mm}$ ,  $dz=5\text{mm}$

Reference Value = 54.44 V/m; Power Drift = -0.06 dB

Peak SAR (extrapolated) = 3.52 W/kg

**SAR(1 g) = 2.37 W/kg; SAR(10 g) = 1.55 W/kg**

Maximum value of SAR (measured) = 2.77 W/kg



0 dB = 2.77 W/kg = 4.42 dBW/kg



# Impedance Measurement Plot for Body TSL

CH1 S11 1 U FS 13 Apr 2015 10:53:33  
 2: 47.658  $\Omega$  -4.9043  $\Omega$  38.865 pF 835.000 000 MHz

\*  
 Del  
 CA

Avg  
 16

H1d





Accredited by the Swiss Accreditation Service (SAS)

Accreditation No.: **SCS 0108**

The Swiss Accreditation Service is one of the signatories to the EA  
 Multilateral Agreement for the recognition of calibration certificates

Client **PC Test**

Certificate No: **D1750V2-1051\_Apr15**

## CALIBRATION CERTIFICATE

Object **D1750V2 - SN:1051**

Calibration procedure(s) **QA CAL-05.v9**  
**Calibration procedure for dipole validation kits above 700 MHz**

PM ✓  
 4/29/15

Calibration date: **April 15, 2015**

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI).  
 The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature  $(22 \pm 3)^{\circ}\text{C}$  and humidity  $< 70\%$ .

Calibration Equipment used (M&TE critical for calibration)

| Primary Standards           | ID #               | Cal Date (Certificate No.)        | Scheduled Calibration  |
|-----------------------------|--------------------|-----------------------------------|------------------------|
| Power meter EPM-442A        | GB37480704         | 07-Oct-14 (No. 217-02020)         | Oct-15                 |
| Power sensor HP 8481A       | US37292783         | 07-Oct-14 (No. 217-02020)         | Oct-15                 |
| Power sensor HP 8481A       | MY41092317         | 07-Oct-14 (No. 217-02021)         | Oct-15                 |
| Reference 20 dB Attenuator  | SN: 5058 (20k)     | 01-Apr-15 (No. 217-02131)         | Mar-16                 |
| Type-N mismatch combination | SN: 5047.2 / 06327 | 01-Apr-15 (No. 217-02134)         | Mar-16                 |
| Reference Probe ES3DV3      | SN: 3205           | 30-Dec-14 (No. ES3-3205_Dec14)    | Dec-15                 |
| DAE4                        | SN: 601            | 18-Aug-14 (No. DAE4-601_Aug14)    | Aug-15                 |
| Secondary Standards         | ID #               | Check Date (in house)             | Scheduled Check        |
| RF generator R&S SMT-06     | 100005             | 04-Aug-99 (in house check Oct-13) | In house check: Oct-16 |
| Network Analyzer HP 8753E   | US37390585 S4206   | 18-Oct-01 (in house check Oct-14) | In house check: Oct-15 |

Calibrated by: **Jeton Kastrati** **Laboratory Technician**

Signature

Approved by: **Katja Pokovic** **Technical Manager**

Issued: April 15, 2015

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.



Accredited by the Swiss Accreditation Service (SAS)

The Swiss Accreditation Service is one of the signatories to the EA  
Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 0108**

**Glossary:**

|       |                                 |
|-------|---------------------------------|
| TSL   | tissue simulating liquid        |
| ConvF | sensitivity in TSL / NORM x,y,z |
| N/A   | not applicable or not measured  |

**Calibration is Performed According to the Following Standards:**

- IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013
- IEC 62209-1, "Procedure to measure the Specific Absorption Rate (SAR) for hand-held devices used in close proximity to the ear (frequency range of 300 MHz to 3 GHz)", February 2005
- KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

**Additional Documentation:**

- DASY4/5 System Handbook

**Methods Applied and Interpretation of Parameters:**

- Measurement Conditions:* Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL:* The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss:* These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay:* One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured:* SAR measured at the stated antenna input power.
- SAR normalized:* SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters:* The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor  $k=2$ , which for a normal distribution corresponds to a coverage probability of approximately 95%.

## Measurement Conditions

DASY system configuration, as far as not given on page 1.

|                              |                        |             |
|------------------------------|------------------------|-------------|
| DASY Version                 | DASY5                  | V52.8.8     |
| Extrapolation                | Advanced Extrapolation |             |
| Phantom                      | Modular Flat Phantom   |             |
| Distance Dipole Center - TSL | 10 mm                  | with Spacer |
| Zoom Scan Resolution         | dx, dy, dz = 5 mm      |             |
| Frequency                    | 1750 MHz $\pm$ 1 MHz   |             |

## Head TSL parameters

The following parameters and calculations were applied.

|                                         | Temperature         | Permittivity   | Conductivity         |
|-----------------------------------------|---------------------|----------------|----------------------|
| Nominal Head TSL parameters             | 22.0 °C             | 40.1           | 1.37 mho/m           |
| Measured Head TSL parameters            | (22.0 $\pm$ 0.2) °C | 38.9 $\pm$ 6 % | 1.35 mho/m $\pm$ 6 % |
| Head TSL temperature change during test | < 0.5 °C            | ----           | ----                 |

## SAR result with Head TSL

| SAR averaged over 1 cm <sup>3</sup> (1 g) of Head TSL | Condition          |                              |
|-------------------------------------------------------|--------------------|------------------------------|
| SAR measured                                          | 250 mW input power | 9.04 W/kg                    |
| SAR for nominal Head TSL parameters                   | normalized to 1W   | 36.2 W/kg $\pm$ 17.0 % (k=2) |

| SAR averaged over 10 cm <sup>3</sup> (10 g) of Head TSL | condition          |                              |
|---------------------------------------------------------|--------------------|------------------------------|
| SAR measured                                            | 250 mW input power | 4.80 W/kg                    |
| SAR for nominal Head TSL parameters                     | normalized to 1W   | 19.2 W/kg $\pm$ 16.5 % (k=2) |

## Body TSL parameters

The following parameters and calculations were applied.

|                                         | Temperature         | Permittivity   | Conductivity         |
|-----------------------------------------|---------------------|----------------|----------------------|
| Nominal Body TSL parameters             | 22.0 °C             | 53.4           | 1.49 mho/m           |
| Measured Body TSL parameters            | (22.0 $\pm$ 0.2) °C | 51.5 $\pm$ 6 % | 1.48 mho/m $\pm$ 6 % |
| Body TSL temperature change during test | < 0.5 °C            | ----           | ----                 |

## SAR result with Body TSL

| SAR averaged over 1 cm <sup>3</sup> (1 g) of Body TSL | Condition          |                              |
|-------------------------------------------------------|--------------------|------------------------------|
| SAR measured                                          | 250 mW input power | 9.32 W/kg                    |
| SAR for nominal Body TSL parameters                   | normalized to 1W   | 37.1 W/kg $\pm$ 17.0 % (k=2) |

| SAR averaged over 10 cm <sup>3</sup> (10 g) of Body TSL | condition          |                              |
|---------------------------------------------------------|--------------------|------------------------------|
| SAR measured                                            | 250 mW input power | 5.01 W/kg                    |
| SAR for nominal Body TSL parameters                     | normalized to 1W   | 20.0 W/kg $\pm$ 16.5 % (k=2) |

## Appendix (Additional assessments outside the scope of SCS 0108)

### Antenna Parameters with Head TSL

|                                      |                                |
|--------------------------------------|--------------------------------|
| Impedance, transformed to feed point | 51.3 $\Omega$ - 0.2 j $\Omega$ |
| Return Loss                          | - 37.5 dB                      |

### Antenna Parameters with Body TSL

|                                      |                                |
|--------------------------------------|--------------------------------|
| Impedance, transformed to feed point | 46.9 $\Omega$ + 0.3 j $\Omega$ |
| Return Loss                          | - 29.9 dB                      |

### General Antenna Parameters and Design

|                                  |          |
|----------------------------------|----------|
| Electrical Delay (one direction) | 1.221 ns |
|----------------------------------|----------|

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard.

No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

### Additional EUT Data

|                 |                   |
|-----------------|-------------------|
| Manufactured by | SPEAG             |
| Manufactured on | February 19, 2010 |

## DASY5 Validation Report for Head TSL

Date: 15.04.2015

Test Laboratory: SPEAG, Zurich, Switzerland

**DUT: Dipole 1750 MHz; Type: D1750V2; Serial: D1750V2 - SN:1051**

Communication System: UID 0 - CW; Frequency: 1750 MHz

Medium parameters used:  $f = 1750$  MHz;  $\sigma = 1.35$  S/m;  $\epsilon_r = 38.9$ ;  $\rho = 1000$  kg/m<sup>3</sup>

Phantom section: Flat Section

Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2011)

DASY52 Configuration:

- Probe: ES3DV3 - SN3205; ConvF(5.2, 5.2, 5.2); Calibrated: 30.12.2014;
- Sensor-Surface: 3mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 18.08.2014
- Phantom: Flat Phantom 5.0 (front); Type: QD000P50AA; Serial: 1001
- DASY52 52.8.8(1222); SEMCAD X 14.6.10(7331)

### **Dipole Calibration for Head Tissue/Pin=250 mW, d=10mm/Zoom Scan (7x7x7)/Cube 0:**



Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 94.99 V/m; Power Drift = 0.06 dB

Peak SAR (extrapolated) = 16.3 W/kg

**SAR(1 g) = 9.04 W/kg; SAR(10 g) = 4.8 W/kg**

Maximum value of SAR (measured) = 11.5 W/kg




0 dB = 11.5 W/kg = 10.61 dBW/kg

# Impedance Measurement Plot for Head TSL

CH1 S11 1 U FS 15 Apr 2015 12:25:31  
 3: 51.330  $\Omega$  -248.05 m $\Omega$  365.65 pF 1 750.000 000 MHz

\*  
 Del  
 C $\Delta$



Avg  
 15


H1d

CH2 S11 LOG 5 dB/REF -20 dB 3:-37.470 dB 1 750.000 000 MHz

C $\Delta$

Avg  
 15

H1d



## DASY5 Validation Report for Body TSL

Date: 15.04.2015

Test Laboratory: SPEAG, Zurich, Switzerland

**DUT: Dipole 1750 MHz; Type: D1750V2; Serial: D1750V2 - SN:1051**

Communication System: UID 0 - CW; Frequency: 1750 MHz

Medium parameters used:  $f = 1750$  MHz;  $\sigma = 1.48$  S/m;  $\epsilon_r = 51.5$ ;  $\rho = 1000$  kg/m<sup>3</sup>

Phantom section: Flat Section

Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2011)

DASY52 Configuration:

- Probe: ES3DV3 - SN3205; ConvF(4.88, 4.88, 4.88); Calibrated: 30.12.2014;
- Sensor-Surface: 3mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 18.08.2014
- Phantom: Flat Phantom 5.0 (back); Type: QD000P50AA; Serial: 1002
- DASY52 52.8.8(1222); SEMCAD X 14.6.10(7331)

### **Dipole Calibration for Body Tissue/Pin=250 mW, d=10mm/Zoom Scan (7x7x7)/Cube 0:**



Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 92.87 V/m; Power Drift = 0.01 dB

Peak SAR (extrapolated) = 16.0 W/kg

**SAR(1 g) = 9.32 W/kg; SAR(10 g) = 5.01 W/kg**

Maximum value of SAR (measured) = 11.7 W/kg




0 dB = 11.7 W/kg = 10.68 dBW/kg

# Impedance Measurement Plot for Body TSL


15 Apr 2015 12:23:57  
 [CH1] S11 1 U FS 3: 46.930  $\Omega$  0.3242  $\Omega$  29.486 pH 1 750.000 000 MHz

\*  
 Del  
 Ca  
 Avg  
 16  
 H1 d



CH2 S11 LOG 5 dB/REF -20 dB 3:-29.939 dB 1 750.000 000 MHz

Ca  
 Avg  
 16  
 H1 d



START 1 550.000 000 MHz

STOP 1 950.000 000 MHz



Accredited by the Swiss Accreditation Service (SAS)  
 The Swiss Accreditation Service is one of the signatories to the EA  
 Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 0108**

Client **PC Test**

Certificate No: **D1900V2-5d148\_Feb15**

## CALIBRATION CERTIFICATE

Object **D1900V2 - SN: 5d148**

Calibration procedure(s) **QA CAL-05.v9**  
**Calibration procedure for dipole validation kits above 700 MHz**

CC ✓  
 3/6/15

Calibration date: **February 18, 2015**

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI).  
 The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature  $(22 \pm 3)^{\circ}\text{C}$  and humidity  $< 70\%$ .

Calibration Equipment used (M&TE critical for calibration)

| Primary Standards           | ID #               | Cal Date (Certificate No.)        | Scheduled Calibration  |
|-----------------------------|--------------------|-----------------------------------|------------------------|
| Power meter EPM-442A        | GB37480704         | 07-Oct-14 (No. 217-02020)         | Oct-15                 |
| Power sensor HP 8481A       | US37292783         | 07-Oct-14 (No. 217-02020)         | Oct-15                 |
| Power sensor HP 8481A       | MY41092317         | 07-Oct-14 (No. 217-02021)         | Oct-15                 |
| Reference 20 dB Attenuator  | SN: 5058 (20k)     | 03-Apr-14 (No. 217-01918)         | Apr-15                 |
| Type-N mismatch combination | SN: 5047.2 / 06327 | 03-Apr-14 (No. 217-01921)         | Apr-15                 |
| Reference Probe ES3DV3      | SN: 3205           | 30-Dec-14 (No. ES3-3205_Dec14)    | Dec-15                 |
| DAE4                        | SN: 601            | 18-Aug-14 (No. DAE4-601_Aug14)    | Aug-15                 |
| Secondary Standards         | ID #               | Check Date (in house)             | Scheduled Check        |
| RF generator R&S SMT-06     | 100005             | 04-Aug-99 (in house check Oct-13) | In house check: Oct-16 |
| Network Analyzer HP 8753E   | US37390585 S4206   | 18-Oct-01 (in house check Oct-14) | In house check: Oct-15 |

Calibrated by: **Michael Weber**      Name: **Michael Weber**      Function: **Laboratory Technician**

Signature

Approved by: **Katja Pokovic**      Name: **Katja Pokovic**      Technical Manager

Issued: February 18, 2015

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.



Accredited by the Swiss Accreditation Service (SAS)

The Swiss Accreditation Service is one of the signatories to the EA  
Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 0108**

**Glossary:**

|       |                                 |
|-------|---------------------------------|
| TSL   | tissue simulating liquid        |
| ConvF | sensitivity in TSL / NORM x,y,z |
| N/A   | not applicable or not measured  |

**Calibration is Performed According to the Following Standards:**

- IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013
- IEC 62209-1, "Procedure to measure the Specific Absorption Rate (SAR) for hand-held devices used in close proximity to the ear (frequency range of 300 MHz to 3 GHz)", February 2005
- KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

**Additional Documentation:**

- DASY4/5 System Handbook

**Methods Applied and Interpretation of Parameters:**

- Measurement Conditions:** Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL:** The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss:** These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay:** One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured:** SAR measured at the stated antenna input power.
- SAR normalized:** SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters:** The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor  $k=2$ , which for a normal distribution corresponds to a coverage probability of approximately 95%.

## Measurement Conditions

DASY system configuration, as far as not given on page 1.

|                              |                        |             |
|------------------------------|------------------------|-------------|
| DASY Version                 | DASY5                  | V52.8.8     |
| Extrapolation                | Advanced Extrapolation |             |
| Phantom                      | Modular Flat Phantom   |             |
| Distance Dipole Center - TSL | 10 mm                  | with Spacer |
| Zoom Scan Resolution         | dx, dy, dz = 5 mm      |             |
| Frequency                    | 1900 MHz $\pm$ 1 MHz   |             |

## Head TSL parameters

The following parameters and calculations were applied.

|                                         | Temperature         | Permittivity   | Conductivity         |
|-----------------------------------------|---------------------|----------------|----------------------|
| Nominal Head TSL parameters             | 22.0 °C             | 40.0           | 1.40 mho/m           |
| Measured Head TSL parameters            | (22.0 $\pm$ 0.2) °C | 39.1 $\pm$ 6 % | 1.42 mho/m $\pm$ 6 % |
| Head TSL temperature change during test | < 0.5 °C            | ----           | ----                 |

## SAR result with Head TSL

|                                                       |                    |                              |
|-------------------------------------------------------|--------------------|------------------------------|
| SAR averaged over 1 cm <sup>3</sup> (1 g) of Head TSL | Condition          |                              |
| SAR measured                                          | 250 mW input power | 10.3 W/kg                    |
| SAR for nominal Head TSL parameters                   | normalized to 1W   | 40.6 W/kg $\pm$ 17.0 % (k=2) |

|                                                         |                    |                              |
|---------------------------------------------------------|--------------------|------------------------------|
| SAR averaged over 10 cm <sup>3</sup> (10 g) of Head TSL | condition          |                              |
| SAR measured                                            | 250 mW input power | 5.37 W/kg                    |
| SAR for nominal Head TSL parameters                     | normalized to 1W   | 21.3 W/kg $\pm$ 16.5 % (k=2) |

## Body TSL parameters

The following parameters and calculations were applied.

|                                         | Temperature         | Permittivity   | Conductivity         |
|-----------------------------------------|---------------------|----------------|----------------------|
| Nominal Body TSL parameters             | 22.0 °C             | 53.3           | 1.52 mho/m           |
| Measured Body TSL parameters            | (22.0 $\pm$ 0.2) °C | 53.1 $\pm$ 6 % | 1.53 mho/m $\pm$ 6 % |
| Body TSL temperature change during test | < 0.5 °C            | ----           | ----                 |

## SAR result with Body TSL

|                                                       |                    |                              |
|-------------------------------------------------------|--------------------|------------------------------|
| SAR averaged over 1 cm <sup>3</sup> (1 g) of Body TSL | Condition          |                              |
| SAR measured                                          | 250 mW input power | 10.1 W/kg                    |
| SAR for nominal Body TSL parameters                   | normalized to 1W   | 40.2 W/kg $\pm$ 17.0 % (k=2) |

|                                                         |                    |                              |
|---------------------------------------------------------|--------------------|------------------------------|
| SAR averaged over 10 cm <sup>3</sup> (10 g) of Body TSL | condition          |                              |
| SAR measured                                            | 250 mW input power | 5.40 W/kg                    |
| SAR for nominal Body TSL parameters                     | normalized to 1W   | 21.5 W/kg $\pm$ 16.5 % (k=2) |

## Appendix (Additional assessments outside the scope of SCS0108)

### Antenna Parameters with Head TSL

|                                      |                             |
|--------------------------------------|-----------------------------|
| Impedance, transformed to feed point | $52.4 \Omega + 6.2 j\Omega$ |
| Return Loss                          | - 23.8 dB                   |

### Antenna Parameters with Body TSL

|                                      |                             |
|--------------------------------------|-----------------------------|
| Impedance, transformed to feed point | $48.2 \Omega + 6.6 j\Omega$ |
| Return Loss                          | - 23.1 dB                   |

### General Antenna Parameters and Design

|                                  |          |
|----------------------------------|----------|
| Electrical Delay (one direction) | 1.198 ns |
|----------------------------------|----------|

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard.

No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

### Additional EUT Data

|                 |                |
|-----------------|----------------|
| Manufactured by | SPEAG          |
| Manufactured on | March 11, 2011 |

## DASY5 Validation Report for Head TSL

Date: 18.02.2015

Test Laboratory: SPEAG, Zurich, Switzerland

**DUT: Dipole 1900 MHz; Type: D1900V2; Serial: D1900V2 - SN: 5d148**

Communication System: UID 0 - CW; Frequency: 1900 MHz

Medium parameters used:  $f = 1900$  MHz;  $\sigma = 1.42$  S/m;  $\epsilon_r = 39.1$ ;  $\rho = 1000$  kg/m<sup>3</sup>

Phantom section: Flat Section

Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2011)

DASY52 Configuration:

- Probe: ES3DV3 - SN3205; ConvF(5, 5, 5); Calibrated: 30.12.2014;
- Sensor-Surface: 3mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 18.08.2014
- Phantom: Flat Phantom 5.0 (front); Type: QD000P50AA; Serial: 1001
- DASY52 52.8.8(1222); SEMCAD X 14.6.10(7331)

### **Dipole Calibration for Head Tissue/Pin=250 mW, d=10mm/Zoom Scan (7x7x7)/Cube 0:**


Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 98.30 V/m; Power Drift = 0.02 dB


Peak SAR (extrapolated) = 18.8 W/kg

**SAR(1 g) = 10.3 W/kg; SAR(10 g) = 5.37 W/kg**


Maximum value of SAR (measured) = 13.0 W/kg



# Impedance Measurement Plot for Head TSL



Impedance Measurement Plot for Body TSL



## DASY5 Validation Report for Body TSL

Date: 18.02.2015

Test Laboratory: SPEAG, Zurich, Switzerland

**DUT: Dipole 1900 MHz; Type: D1900V2; Serial: D1900V2 - SN: 5d148**

Communication System: UID 0 - CW; Frequency: 1900 MHz

Medium parameters used:  $f = 1900$  MHz;  $\sigma = 1.53$  S/m;  $\epsilon_r = 53.1$ ;  $\rho = 1000$  kg/m<sup>3</sup>

Phantom section: Flat Section

Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2011)

DASY52 Configuration:

- Probe: ES3DV3 - SN3205; ConvF(4.65, 4.65, 4.65); Calibrated: 30.12.2014;
- Sensor-Surface: 3mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 18.08.2014
- Phantom: Flat Phantom 5.0 (back); Type: QD000P50AA; Serial: 1002
- DASY52 52.8.8(1222); SEMCAD X 14.6.10(7331)

### **Dipole Calibration for Body Tissue/Pin=250 mW, d=10mm/Zoom Scan (7x7x7)/Cube 0:**



Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 95.79 V/m; Power Drift = -0.01 dB

Peak SAR (extrapolated) = 17.2 W/kg

**SAR(1 g) = 10.1 W/kg; SAR(10 g) = 5.4 W/kg**

Maximum value of SAR (measured) = 12.7 W/kg



0 dB = 12.7 W/kg = 11.04 dBW/kg



Accredited by the Swiss Accreditation Service (SAS)  
 The Swiss Accreditation Service is one of the signatories to the EA  
 Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 108**

Client **PC Test**

Certificate No: **ES3-3288\_Sep14/2**

# **CALIBRATION CERTIFICATE (Replacement of No:ES3-3288\_Sep14)**

Object **ES3DV3 - SN:3288**

Calibration procedure(s)

**QA CAL-01.v9, QA CAL-23.v5, QA CAL-25.v6**  
**Calibration procedure for dosimetric E-field probes**

**CC**  
**10/12/14**

Calibration date:

**September 24, 2014**

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI).  
 The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature  $(22 \pm 3)^{\circ}\text{C}$  and humidity  $< 70\%$ .

Calibration Equipment used (M&TE critical for calibration)

| Primary Standards          | ID              | Cal Date (Certificate No.)        | Scheduled Calibration  |
|----------------------------|-----------------|-----------------------------------|------------------------|
| Power meter E4419B         | GB41293874      | 03-Apr-14 (No. 217-01911)         | Apr-15                 |
| Power sensor E4412A        | MY41498087      | 03-Apr-14 (No. 217-01911)         | Apr-15                 |
| Reference 3 dB Attenuator  | SN: S5054 (3c)  | 03-Apr-14 (No. 217-01915)         | Apr-15                 |
| Reference 20 dB Attenuator | SN: S5277 (20x) | 03-Apr-14 (No. 217-01919)         | Apr-15                 |
| Reference 30 dB Attenuator | SN: S5129 (30b) | 03-Apr-14 (No. 217-01920)         | Apr-15                 |
| Reference Probe ES3DV2     | SN: 3013        | 30-Dec-13 (No. ES3-3013_Dec13)    | Dec-14                 |
| DAE4                       | SN: 660         | 13-Dec-13 (No. DAE4-660_Dec13)    | Dec-14                 |
| Secondary Standards        | ID              | Check Date (in house)             | Scheduled Check        |
| RF generator HP 8648C      | US3642U01700    | 4-Aug-99 (in house check Apr-13)  | In house check: Apr-16 |
| Network Analyzer HP 8753E  | US37390585      | 18-Oct-01 (in house check Oct-13) | In house check: Oct-14 |

|                                                                                                                 | Name          | Function              | Signature |
|-----------------------------------------------------------------------------------------------------------------|---------------|-----------------------|-----------|
| Calibrated by:                                                                                                  | Leif Klysner  | Laboratory Technician |           |
| Approved by:                                                                                                    | Katja Pokovic | Technical Manager     |           |
| Issued: November 3, 2014                                                                                        |               |                       |           |
| This calibration certificate shall not be reproduced except in full without written approval of the laboratory. |               |                       |           |



Accredited by the Swiss Accreditation Service (SAS)  
 The Swiss Accreditation Service is one of the signatories to the EA  
 Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 108**

### Glossary:

|                          |                                                                                                                                                         |
|--------------------------|---------------------------------------------------------------------------------------------------------------------------------------------------------|
| TSL                      | tissue simulating liquid                                                                                                                                |
| NORM <sub>x,y,z</sub>    | sensitivity in free space                                                                                                                               |
| ConvF                    | sensitivity in TSL / NORM <sub>x,y,z</sub>                                                                                                              |
| DCP                      | diode compression point                                                                                                                                 |
| CF                       | crest factor (1/duty_cycle) of the RF signal                                                                                                            |
| A, B, C, D               | modulation dependent linearization parameters                                                                                                           |
| Polarization $\phi$      | $\phi$ rotation around probe axis                                                                                                                       |
| Polarization $\vartheta$ | $\vartheta$ rotation around an axis that is in the plane normal to probe axis (at measurement center),<br>i.e., $\vartheta = 0$ is normal to probe axis |
| Connector Angle          | information used in DASY system to align probe sensor X to the robot coordinate system                                                                  |

### Calibration is Performed According to the Following Standards:

- IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013
- IEC 62209-1, "Procedure to measure the Specific Absorption Rate (SAR) for hand-held devices used in close proximity to the ear (frequency range of 300 MHz to 3 GHz)", February 2005

### Methods Applied and Interpretation of Parameters:

- NORM<sub>x,y,z</sub>**: Assessed for E-field polarization  $\vartheta = 0$  ( $f \leq 900$  MHz in TEM-cell;  $f > 1800$  MHz: R22 waveguide). NORM<sub>x,y,z</sub> are only intermediate values, i.e., the uncertainties of NORM<sub>x,y,z</sub> does not affect the E<sup>2</sup>-field uncertainty inside TSL (see below ConvF).
- NORM(f)<sub>x,y,z</sub>** = NORM<sub>x,y,z</sub> \* frequency\_response (see Frequency Response Chart). This linearization is implemented in DASY4 software versions later than 4.2. The uncertainty of the frequency response is included in the stated uncertainty of ConvF.
- DCP<sub>x,y,z</sub>**: DCP are numerical linearization parameters assessed based on the data of power sweep with CW signal (no uncertainty required). DCP does not depend on frequency nor media.
- PAR**: PAR is the Peak to Average Ratio that is not calibrated but determined based on the signal characteristics
- A<sub>x,y,z</sub>; B<sub>x,y,z</sub>; C<sub>x,y,z</sub>; D<sub>x,y,z</sub>; VR<sub>x,y,z</sub>**: A, B, C, D are numerical linearization parameters assessed based on the data of power sweep for specific modulation signal. The parameters do not depend on frequency nor media. VR is the maximum calibration range expressed in RMS voltage across the diode.
- ConvF and Boundary Effect Parameters**: Assessed in flat phantom using E-field (or Temperature Transfer Standard for  $f \leq 800$  MHz) and inside waveguide using analytical field distributions based on power measurements for  $f > 800$  MHz. The same setups are used for assessment of the parameters applied for boundary compensation (alpha, depth) of which typical uncertainty values are given. These parameters are used in DASY4 software to improve probe accuracy close to the boundary. The sensitivity in TSL corresponds to NORM<sub>x,y,z</sub> \* ConvF whereby the uncertainty corresponds to that given for ConvF. A frequency dependent ConvF is used in DASY version 4.4 and higher which allows extending the validity from  $\pm 50$  MHz to  $\pm 100$  MHz.
- Spherical isotropy (3D deviation from isotropy)**: in a field of low gradients realized using a flat phantom exposed by a patch antenna.
- Sensor Offset**: The sensor offset corresponds to the offset of virtual measurement center from the probe tip (on probe axis). No tolerance required.
- Connector Angle**: The angle is assessed using the information gained by determining the NORM<sub>x</sub> (no uncertainty required).

# Probe ES3DV3

## SN:3288

|               |                    |
|---------------|--------------------|
| Manufactured: | July 6, 2010       |
| Repaired:     | September 18, 2014 |
| Calibrated:   | September 24, 2014 |

Calibrated for DASY/EASY Systems  
(Note: non-compatible with DASY2 system!)

# DASY/EASY - Parameters of Probe: ES3DV3 - SN:3288

## Basic Calibration Parameters

|                                       | Sensor X | Sensor Y | Sensor Z | Unc (k=2)     |
|---------------------------------------|----------|----------|----------|---------------|
| Norm ( $\mu V/(V/m)^2$ ) <sup>A</sup> | 1.05     | 1.16     | 0.92     | $\pm 10.1 \%$ |
| DCP (mV) <sup>B</sup>                 | 105.1    | 104.6    | 106.7    |               |

## Modulation Calibration Parameters

| UID           | Communication System Name                     |   | A<br>dB | B<br>dB $\sqrt{\mu V}$ | C    | D<br>dB | VR<br>mV | Unc <sup>E</sup><br>(k=2) |
|---------------|-----------------------------------------------|---|---------|------------------------|------|---------|----------|---------------------------|
| 0             | CW                                            | X | 0.0     | 0.0                    | 1.0  | 0.00    | 195.8    | $\pm 3.5 \%$              |
|               |                                               | Y | 0.0     | 0.0                    | 1.0  |         | 175.9    |                           |
|               |                                               | Z | 0.0     | 0.0                    | 1.0  |         | 177.1    |                           |
| 10010-<br>CAA | SAR Validation (Square, 100ms, 10ms)          | X | 2.71    | 61.9                   | 11.4 | 10.00   | 40.3     | $\pm 2.2 \%$              |
|               |                                               | Y | 2.37    | 60.2                   | 11.2 |         | 42.6     |                           |
|               |                                               | Z | 1.54    | 56.6                   | 8.9  |         | 41.2     |                           |
| 10011-<br>CAB | UMTS-FDD (WCDMA)                              | X | 3.29    | 67.1                   | 18.4 | 2.91    | 133.8    | $\pm 0.5 \%$              |
|               |                                               | Y | 3.43    | 67.9                   | 18.9 |         | 139.5    |                           |
|               |                                               | Z | 3.45    | 68.1                   | 18.9 |         | 141.3    |                           |
| 10012-<br>CAA | IEEE 802.11b WiFi 2.4 GHz (DSSS, 1 Mbps)      | X | 2.99    | 68.9                   | 18.6 | 1.87    | 135.1    | $\pm 0.7 \%$              |
|               |                                               | Y | 3.59    | 72.4                   | 20.4 |         | 140.7    |                           |
|               |                                               | Z | 3.54    | 72.4                   | 20.3 |         | 143.0    |                           |
| 10013-<br>CAA | IEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 6 Mbps) | X | 11.15   | 70.8                   | 23.3 | 9.46    | 132.3    | $\pm 3.5 \%$              |
|               |                                               | Y | 11.29   | 70.8                   | 23.2 |         | 141.1    |                           |
|               |                                               | Z | 11.07   | 70.7                   | 23.2 |         | 139.2    |                           |
| 10021-<br>DAB | GSM-FDD (TDMA, GMSK)                          | X | 14.71   | 90.5                   | 24.5 | 9.39    | 149.0    | $\pm 1.9 \%$              |
|               |                                               | Y | 16.40   | 92.8                   | 26.0 |         | 131.3    |                           |
|               |                                               | Z | 11.34   | 87.2                   | 23.6 |         | 126.1    |                           |
| 10023-<br>DAB | GPRS-FDD (TDMA, GMSK, TN 0)                   | X | 15.91   | 92.2                   | 25.3 | 9.57    | 138.9    | $\pm 2.5 \%$              |
|               |                                               | Y | 21.25   | 96.9                   | 27.2 |         | 142.0    |                           |
|               |                                               | Z | 11.68   | 87.2                   | 23.5 |         | 145.9    |                           |
| 10024-<br>DAB | GPRS-FDD (TDMA, GMSK, TN 0-1)                 | X | 38.62   | 99.8                   | 24.7 | 6.56    | 123.8    | $\pm 2.2 \%$              |
|               |                                               | Y | 36.71   | 99.7                   | 25.2 |         | 128.1    |                           |
|               |                                               | Z | 36.56   | 99.4                   | 24.5 |         | 129.5    |                           |
| 10027-<br>DAB | GPRS-FDD (TDMA, GMSK, TN 0-1-2)               | X | 56.60   | 99.6                   | 22.6 | 4.80    | 138.8    | $\pm 1.9 \%$              |
|               |                                               | Y | 46.94   | 99.9                   | 23.7 |         | 149.9    |                           |
|               |                                               | Z | 51.17   | 99.8                   | 22.9 |         | 144.9    |                           |
| 10028-<br>DAB | GPRS-FDD (TDMA, GMSK, TN 0-1-2-3)             | X | 70.88   | 100.0                  | 21.6 | 3.55    | 147.5    | $\pm 1.9 \%$              |
|               |                                               | Y | 52.58   | 99.8                   | 22.6 |         | 129.4    |                           |
|               |                                               | Z | 76.98   | 99.8                   | 21.2 |         | 128.7    |                           |
| 10032-<br>CAA | IEEE 802.15.1 Bluetooth (GFSK, DH5)           | X | 98.89   | 99.5                   | 18.9 | 1.16    | 135.8    | $\pm 1.4 \%$              |
|               |                                               | Y | 78.39   | 99.6                   | 19.5 |         | 141.7    |                           |
|               |                                               | Z | 95.21   | 95.5                   | 17.1 |         | 143.4    |                           |
| 10039-<br>CAB | CDMA2000 (1xRTT, RC1)                         | X | 4.72    | 66.7                   | 18.9 | 4.57    | 133.7    | $\pm 0.9 \%$              |
|               |                                               | Y | 4.85    | 67.1                   | 19.1 |         | 137.7    |                           |
|               |                                               | Z | 4.81    | 67.4                   | 19.2 |         | 141.9    |                           |

|           |                                               |   |       |      |      |      |       |        |
|-----------|-----------------------------------------------|---|-------|------|------|------|-------|--------|
| 10081-CAB | CDMA2000 (1xRTT, RC3)                         | X | 3.91  | 66.3 | 18.6 | 3.97 | 129.5 | ±0.7 % |
|           |                                               | Y | 4.00  | 66.6 | 18.7 |      | 133.7 |        |
|           |                                               | Z | 3.99  | 66.8 | 18.8 |      | 137.5 |        |
| 10098-CAB | UMTS-FDD (HSUPA, Subtest 2)                   | X | 4.63  | 66.9 | 18.7 | 3.98 | 141.4 | ±0.7 % |
|           |                                               | Y | 4.78  | 67.5 | 19.0 |      | 147.7 |        |
|           |                                               | Z | 4.57  | 66.8 | 18.6 |      | 127.8 |        |
| 10100-CAB | LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)      | X | 6.59  | 68.2 | 20.1 | 5.67 | 149.2 | ±1.4 % |
|           |                                               | Y | 6.36  | 67.3 | 19.6 |      | 130.7 |        |
|           |                                               | Z | 6.36  | 67.5 | 19.6 |      | 133.6 |        |
| 10108-CAB | LTE-FDD (SC-FDMA, 100% RB, 10 MHz, QPSK)      | X | 6.44  | 67.8 | 20.0 | 5.80 | 146.6 | ±1.4 % |
|           |                                               | Y | 6.23  | 66.8 | 19.4 |      | 128.8 |        |
|           |                                               | Z | 6.24  | 67.1 | 19.6 |      | 131.4 |        |
| 10110-CAB | LTE-FDD (SC-FDMA, 100% RB, 5 MHz, QPSK)       | X | 6.08  | 67.1 | 19.6 | 5.75 | 143.2 | ±1.4 % |
|           |                                               | Y | 6.20  | 67.4 | 19.8 |      | 148.0 |        |
|           |                                               | Z | 5.92  | 66.6 | 19.3 |      | 128.5 |        |
| 10114-CAA | IEEE 802.11n (HT Greenfield, 13.5 Mbps, BPSK) | X | 10.32 | 69.3 | 21.5 | 8.10 | 137.0 | ±2.2 % |
|           |                                               | Y | 10.31 | 69.1 | 21.4 |      | 143.5 |        |
|           |                                               | Z | 10.37 | 69.5 | 21.6 |      | 146.1 |        |
| 10117-CAA | IEEE 802.11n (HT Mixed, 13.5 Mbps, BPSK)      | X | 10.35 | 69.4 | 21.6 | 8.07 | 138.3 | ±2.2 % |
|           |                                               | Y | 10.36 | 69.3 | 21.4 |      | 146.4 |        |
|           |                                               | Z | 10.42 | 69.6 | 21.6 |      | 149.0 |        |
| 10151-CAB | LTE-TDD (SC-FDMA, 50% RB, 20 MHz, QPSK)       | X | 9.95  | 75.7 | 26.2 | 9.28 | 134.9 | ±3.3 % |
|           |                                               | Y | 10.37 | 76.0 | 26.1 |      | 146.6 |        |
|           |                                               | Z | 9.77  | 75.4 | 26.0 |      | 142.5 |        |
| 10154-CAB | LTE-FDD (SC-FDMA, 50% RB, 10 MHz, QPSK)       | X | 6.12  | 67.2 | 19.7 | 5.75 | 144.9 | ±1.4 % |
|           |                                               | Y | 6.21  | 67.4 | 19.8 |      | 148.8 |        |
|           |                                               | Z | 5.91  | 66.5 | 19.3 |      | 128.7 |        |
| 10160-CAB | LTE-FDD (SC-FDMA, 50% RB, 15 MHz, QPSK)       | X | 6.28  | 66.7 | 19.4 | 5.82 | 125.5 | ±1.2 % |
|           |                                               | Y | 6.37  | 66.8 | 19.4 |      | 129.7 |        |
|           |                                               | Z | 6.36  | 67.1 | 19.6 |      | 132.9 |        |
| 10169-CAB | LTE-FDD (SC-FDMA, 1 RB, 20 MHz, QPSK)         | X | 5.08  | 67.6 | 20.2 | 5.73 | 147.0 | ±1.2 % |
|           |                                               | Y | 4.95  | 66.6 | 19.6 |      | 128.6 |        |
|           |                                               | Z | 4.91  | 66.9 | 19.8 |      | 131.2 |        |
| 10172-CAB | LTE-TDD (SC-FDMA, 1 RB, 20 MHz, QPSK)         | X | 8.18  | 77.2 | 27.2 | 9.21 | 123.4 | ±2.7 % |
|           |                                               | Y | 8.37  | 76.6 | 26.6 |      | 129.5 |        |
|           |                                               | Z | 7.97  | 76.7 | 26.9 |      | 128.7 |        |
| 10175-CAB | LTE-FDD (SC-FDMA, 1 RB, 10 MHz, QPSK)         | X | 5.05  | 67.4 | 20.1 | 5.72 | 146.2 | ±1.4 % |
|           |                                               | Y | 5.10  | 67.3 | 20.0 |      | 142.8 |        |
|           |                                               | Z | 4.87  | 66.7 | 19.6 |      | 129.6 |        |
| 10181-CAB | LTE-FDD (SC-FDMA, 1 RB, 15 MHz, QPSK)         | X | 5.04  | 67.4 | 20.0 | 5.72 | 145.5 | ±1.2 % |
|           |                                               | Y | 5.12  | 67.4 | 20.0 |      | 143.4 |        |
|           |                                               | Z | 4.87  | 66.7 | 19.6 |      | 129.9 |        |
| 10193-CAA | IEEE 802.11n (HT Greenfield, 6.5 Mbps, BPSK)  | X | 9.92  | 68.9 | 21.4 | 8.09 | 131.0 | ±2.2 % |
|           |                                               | Y | 9.84  | 68.5 | 21.1 |      | 130.0 |        |
|           |                                               | Z | 9.94  | 69.0 | 21.4 |      | 138.6 |        |

|           |                                                               |   |       |      |      |      |       |        |
|-----------|---------------------------------------------------------------|---|-------|------|------|------|-------|--------|
| 10196-CAA | IEEE 802.11n (HT Mixed, 6.5 Mbps, BPSK)                       | X | 9.90  | 68.9 | 21.4 | 8.10 | 130.8 | ±2.2 % |
|           |                                                               | Y | 9.81  | 68.4 | 21.0 |      | 131.4 |        |
|           |                                                               | Z | 9.95  | 69.1 | 21.5 |      | 140.5 |        |
| 10219-CAA | IEEE 802.11n (HT Mixed, 7.2 Mbps, BPSK)                       | X | 9.81  | 68.8 | 21.3 | 8.03 | 130.0 | ±2.2 % |
|           |                                                               | Y | 9.89  | 68.9 | 21.3 |      | 138.1 |        |
|           |                                                               | Z | 9.89  | 69.1 | 21.5 |      | 140.5 |        |
| 10222-CAA | IEEE 802.11n (HT Mixed, 15 Mbps, BPSK)                        | X | 10.25 | 69.2 | 21.4 | 8.06 | 137.1 | ±2.2 % |
|           |                                                               | Y | 10.30 | 69.2 | 21.4 |      | 144.4 |        |
|           |                                                               | Z | 10.38 | 69.6 | 21.6 |      | 148.4 |        |
| 10225-CAB | UMTS-FDD (HSPA+)                                              | X | 6.90  | 66.8 | 19.3 | 5.97 | 132.8 | ±1.4 % |
|           |                                                               | Y | 7.09  | 67.3 | 19.6 |      | 142.0 |        |
|           |                                                               | Z | 7.04  | 67.4 | 19.6 |      | 143.5 |        |
| 10237-CAB | LTE-TDD (SC-FDMA, 1 RB, 10 MHz, QPSK)                         | X | 9.61  | 81.9 | 29.6 | 9.21 | 149.3 | ±2.7 % |
|           |                                                               | Y | 8.66  | 77.6 | 27.1 |      | 133.7 |        |
|           |                                                               | Z | 8.20  | 77.5 | 27.3 |      | 132.2 |        |
| 10252-CAB | LTE-TDD (SC-FDMA, 50% RB, 10 MHz, QPSK)                       | X | 9.16  | 74.5 | 25.8 | 9.24 | 126.3 | ±3.0 % |
|           |                                                               | Y | 9.62  | 75.0 | 25.8 |      | 137.4 |        |
|           |                                                               | Z | 9.16  | 74.8 | 25.9 |      | 135.2 |        |
| 10267-CAB | LTE-TDD (SC-FDMA, 100% RB, 10 MHz, QPSK)                      | X | 9.97  | 75.7 | 26.3 | 9.30 | 133.7 | ±3.3 % |
|           |                                                               | Y | 10.38 | 75.9 | 26.1 |      | 146.1 |        |
|           |                                                               | Z | 9.91  | 75.7 | 26.3 |      | 143.8 |        |
| 10274-CAB | UMTS-FDD (HSUPA, Subtest 5, 3GPP Rel8.10)                     | X | 5.86  | 66.6 | 18.7 | 4.87 | 129.9 | ±0.9 % |
|           |                                                               | Y | 6.01  | 67.1 | 19.0 |      | 135.7 |        |
|           |                                                               | Z | 5.95  | 67.1 | 19.0 |      | 139.4 |        |
| 10275-CAB | UMTS-FDD (HSUPA, Subtest 5, 3GPP Rel8.4)                      | X | 4.40  | 66.7 | 18.6 | 3.96 | 136.4 | ±0.7 % |
|           |                                                               | Y | 4.55  | 67.3 | 19.0 |      | 138.3 |        |
|           |                                                               | Z | 4.56  | 67.6 | 19.1 |      | 144.3 |        |
| 10291-AAB | CDMA2000, RC3, SO55, Full Rate                                | X | 3.64  | 66.9 | 18.7 | 3.46 | 127.4 | ±0.5 % |
|           |                                                               | Y | 3.77  | 67.6 | 19.1 |      | 130.2 |        |
|           |                                                               | Z | 3.72  | 67.5 | 19.0 |      | 134.4 |        |
| 10292-AAB | CDMA2000, RC3, SO32, Full Rate                                | X | 3.58  | 67.0 | 18.7 | 3.39 | 128.4 | ±0.5 % |
|           |                                                               | Y | 3.73  | 67.7 | 19.1 |      | 132.7 |        |
|           |                                                               | Z | 3.69  | 67.8 | 19.1 |      | 136.1 |        |
| 10297-AAA | LTE-FDD (SC-FDMA, 50% RB, 20 MHz, QPSK)                       | X | 6.43  | 67.7 | 19.9 | 5.81 | 145.5 | ±1.4 % |
|           |                                                               | Y | 6.49  | 67.7 | 19.9 |      | 149.5 |        |
|           |                                                               | Z | 6.23  | 67.0 | 19.6 |      | 129.5 |        |
| 10311-AAA | LTE-FDD (SC-FDMA, 100% RB, 15 MHz, QPSK)                      | X | 6.74  | 67.3 | 19.8 | 6.06 | 126.7 | ±1.4 % |
|           |                                                               | Y | 6.83  | 67.5 | 19.8 |      | 132.9 |        |
|           |                                                               | Z | 6.81  | 67.6 | 19.9 |      | 135.8 |        |
| 10315-AAA | IEEE 802.11b WiFi 2.4 GHz (DSSS, 1 Mbps, 96pc duty cycle)     | X | 3.00  | 69.9 | 19.4 | 1.71 | 133.9 | ±0.5 % |
|           |                                                               | Y | 3.30  | 71.5 | 20.1 |      | 141.0 |        |
|           |                                                               | Z | 3.22  | 71.4 | 20.0 |      | 142.9 |        |
| 10316-AAA | IEEE 802.11g WiFi 2.4 GHz (ERP-OFDM, 6 Mbps, 96pc duty cycle) | X | 10.17 | 69.2 | 21.8 | 8.36 | 130.5 | ±2.5 % |
|           |                                                               | Y | 10.20 | 69.1 | 21.6 |      | 138.4 |        |
|           |                                                               | Z | 10.20 | 69.4 | 21.8 |      | 140.7 |        |

|           |                                                               |   |       |      |      |      |       |        |
|-----------|---------------------------------------------------------------|---|-------|------|------|------|-------|--------|
| 10403-AAB | CDMA2000 (1xEV-DO, Rev. 0)                                    | X | 4.75  | 68.3 | 18.8 | 3.76 | 138.5 | ±0.7 % |
|           |                                                               | Y | 5.00  | 69.1 | 19.2 |      | 146.7 |        |
|           |                                                               | Z | 4.92  | 69.2 | 19.1 |      | 148.5 |        |
| 10404-AAB | CDMA2000 (1xEV-DO, Rev. A)                                    | X | 4.73  | 68.6 | 18.9 | 3.77 | 136.3 | ±0.7 % |
|           |                                                               | Y | 4.97  | 69.4 | 19.4 |      | 143.7 |        |
|           |                                                               | Z | 4.91  | 69.6 | 19.3 |      | 146.0 |        |
| 10415-AAA | IEEE 802.11b WiFi 2.4 GHz (DSSS, 1 Mbps, 99pc duty cycle)     | X | 2.65  | 68.1 | 18.5 | 1.54 | 135.2 | ±0.5 % |
|           |                                                               | Y | 3.05  | 70.8 | 19.9 |      | 140.7 |        |
|           |                                                               | Z | 2.87  | 69.8 | 19.3 |      | 144.8 |        |
| 10416-AAA | IEEE 802.11g WiFi 2.4 GHz (ERP-OFDM, 6 Mbps, 99pc duty cycle) | X | 10.00 | 69.0 | 21.5 | 8.23 | 130.8 | ±2.2 % |
|           |                                                               | Y | 10.06 | 68.9 | 21.4 |      | 138.6 |        |
|           |                                                               | Z | 10.08 | 69.3 | 21.7 |      | 141.6 |        |

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor  $k=2$ , which for a normal distribution corresponds to a coverage probability of approximately 95%.

<sup>A</sup> The uncertainties of NormX,Y,Z do not affect the  $E^2$ -field uncertainty inside TSL (see Pages 8 and 9).

<sup>B</sup> Numerical linearization parameter: uncertainty not required.

<sup>E</sup> Uncertainty is determined using the max. deviation from linear response applying rectangular distribution and is expressed for the square of the field value.

## DASY/EASY - Parameters of Probe: ES3DV3 - SN:3288

### Calibration Parameter Determined in Head Tissue Simulating Media

| f (MHz) <sup>C</sup> | Relative Permittivity <sup>F</sup> | Conductivity (S/m) <sup>F</sup> | ConvF X | ConvF Y | ConvF Z | Alpha <sup>G</sup> | Depth <sup>G</sup> (mm) | Uct. (k=2) |
|----------------------|------------------------------------|---------------------------------|---------|---------|---------|--------------------|-------------------------|------------|
| 750                  | 41.9                               | 0.89                            | 6.81    | 6.81    | 6.81    | 0.37               | 1.70                    | ± 12.0 %   |
| 835                  | 41.5                               | 0.90                            | 6.51    | 6.51    | 6.51    | 0.45               | 1.52                    | ± 12.0 %   |
| 1750                 | 40.1                               | 1.37                            | 5.38    | 5.38    | 5.38    | 0.44               | 1.58                    | ± 12.0 %   |
| 1900                 | 40.0                               | 1.40                            | 5.17    | 5.17    | 5.17    | 0.80               | 1.18                    | ± 12.0 %   |
| 2450                 | 39.2                               | 1.80                            | 4.56    | 4.56    | 4.56    | 0.80               | 1.21                    | ± 12.0 %   |
| 2600                 | 39.0                               | 1.96                            | 4.44    | 4.44    | 4.44    | 0.80               | 1.22                    | ± 12.0 %   |

<sup>C</sup> Frequency validity above 300 MHz of ± 100 MHz only applies for DASY v4.4 and higher (see Page 2), else it is restricted to ± 50 MHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below 300 MHz is ± 10, 25, 40, 50 and 70 MHz for ConvF assessments at 30, 64, 128, 150 and 220 MHz respectively. Above 5 GHz frequency validity can be extended to ± 110 MHz.

<sup>F</sup> At frequencies below 3 GHz, the validity of tissue parameters ( $\epsilon$  and  $\sigma$ ) can be relaxed to ± 10% if liquid compensation formula is applied to measured SAR values. At frequencies above 3 GHz, the validity of tissue parameters ( $\epsilon$  and  $\sigma$ ) is restricted to ± 5%. The uncertainty is the RSS of the ConvF uncertainty for indicated target tissue parameters.

<sup>G</sup> Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than ± 1% for frequencies below 3 GHz and below ± 2% for frequencies between 3-6 GHz at any distance larger than half the probe tip diameter from the boundary.

## DASY/EASY - Parameters of Probe: ES3DV3 - SN:3288

### Calibration Parameter Determined in Body Tissue Simulating Media

| f (MHz) <sup>C</sup> | Relative Permittivity <sup>F</sup> | Conductivity (S/m) <sup>F</sup> | ConvF X | ConvF Y | ConvF Z | Alpha <sup>G</sup> | Depth <sup>G</sup> (mm) | Unct. (k=2) |
|----------------------|------------------------------------|---------------------------------|---------|---------|---------|--------------------|-------------------------|-------------|
| 750                  | 55.5                               | 0.96                            | 6.38    | 6.38    | 6.38    | 0.31               | 1.89                    | ± 12.0 %    |
| 835                  | 55.2                               | 0.97                            | 6.32    | 6.32    | 6.32    | 0.55               | 1.39                    | ± 12.0 %    |
| 1750                 | 53.4                               | 1.49                            | 5.03    | 5.03    | 5.03    | 0.57               | 1.44                    | ± 12.0 %    |
| 1900                 | 53.3                               | 1.52                            | 4.82    | 4.82    | 4.82    | 0.51               | 1.54                    | ± 12.0 %    |
| 2450                 | 52.7                               | 1.95                            | 4.36    | 4.36    | 4.36    | 0.71               | 1.07                    | ± 12.0 %    |
| 2600                 | 52.5                               | 2.16                            | 4.22    | 4.22    | 4.22    | 0.80               | 1.07                    | ± 12.0 %    |


<sup>C</sup> Frequency validity above 300 MHz of ± 100 MHz only applies for DASY v4.4 and higher (see Page 2), else it is restricted to ± 50 MHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below 300 MHz is ± 10, 25, 40, 50 and 70 MHz for ConvF assessments at 30, 64, 128, 150 and 220 MHz respectively. Above 5 GHz frequency validity can be extended to ± 110 MHz.

<sup>F</sup> At frequencies below 3 GHz, the validity of tissue parameters ( $\epsilon$  and  $\sigma$ ) can be relaxed to ± 10% if liquid compensation formula is applied to measured SAR values. At frequencies above 3 GHz, the validity of tissue parameters ( $\epsilon$  and  $\sigma$ ) is restricted to ± 5%. The uncertainty is the RSS of the ConvF uncertainty for indicated target tissue parameters.

<sup>G</sup> Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than ± 1% for frequencies below 3 GHz and below ± 2% for frequencies between 3-6 GHz at any distance larger than half the probe tip diameter from the boundary.

## Frequency Response of E-Field


(TEM-Cell:ifi110 EXX, Waveguide: R22)





Uncertainty of Frequency Response of E-field:  $\pm 6.3\%$  ( $k=2$ )

## Receiving Pattern ( $\phi$ ), $\theta = 0^\circ$

f=600 MHz,TEM





f=1800 MHz,R22




Uncertainty of Axial Isotropy Assessment:  $\pm 0.5\%$  ( $k=2$ )

## Dynamic Range $f(\text{SAR}_{\text{head}})$ (TEM cell , $f_{\text{eval}} = 1900 \text{ MHz}$ )





Uncertainty of Linearity Assessment:  $\pm 0.6\%$  ( $k=2$ )

## Conversion Factor Assessment



## Deviation from Isotropy in Liquid

Error ( $\phi, \theta$ ),  $f = 900 \text{ MHz}$



## DASY/EASY - Parameters of Probe: ES3DV3 - SN:3288

### Other Probe Parameters

|                                               |            |
|-----------------------------------------------|------------|
| Sensor Arrangement                            | Triangular |
| Connector Angle (°)                           | -110       |
| Mechanical Surface Detection Mode             | enabled    |
| Optical Surface Detection Mode                | disabled   |
| Probe Overall Length                          | 337 mm     |
| Probe Body Diameter                           | 10 mm      |
| Tip Length                                    | 10 mm      |
| Tip Diameter                                  | 4 mm       |
| Probe Tip to Sensor X Calibration Point       | 2 mm       |
| Probe Tip to Sensor Y Calibration Point       | 2 mm       |
| Probe Tip to Sensor Z Calibration Point       | 2 mm       |
| Recommended Measurement Distance from Surface | 3 mm       |



Accredited by the Swiss Accreditation Service (SAS)  
 The Swiss Accreditation Service is one of the signatories to the EA  
 Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 0108**

Client **PC Test**

Certificate No: **ES3-3318\_Jan15**

## CALIBRATION CERTIFICATE

Object **ES3DV3 - SN:3318**

Calibration procedure(s) **QA CAL-01.v9, QA CAL-23.v5, QA CAL-25.v6**  
**Calibration procedure for dosimetric E-field probes**

CC  
 1/30/15

Calibration date: **January 23, 2015**

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI).  
 The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature  $(22 \pm 3)^\circ\text{C}$  and humidity  $< 70\%$ .

Calibration Equipment used (M&TE critical for calibration)

| Primary Standards          | ID              | Cal Date (Certificate No.)        | Scheduled Calibration  |
|----------------------------|-----------------|-----------------------------------|------------------------|
| Power meter E4419B         | GB41293874      | 03-Apr-14 (No. 217-01911)         | Apr-15                 |
| Power sensor E4412A        | MY41498087      | 03-Apr-14 (No. 217-01911)         | Apr-15                 |
| Reference 3 dB Attenuator  | SN: S5054 (3c)  | 03-Apr-14 (No. 217-01915)         | Apr-15                 |
| Reference 20 dB Attenuator | SN: S5277 (20x) | 03-Apr-14 (No. 217-01919)         | Apr-15                 |
| Reference 30 dB Attenuator | SN: S5129 (30b) | 03-Apr-14 (No. 217-01920)         | Apr-15                 |
| Reference Probe ES3DV2     | SN: 3013        | 30-Dec-14 (No. ES3-3013_Dec14)    | Dec-15                 |
| DAE4                       | SN: 660         | 14-Jan-15 (No. DAE4-660_Jan15)    | Jan-16                 |
| Secondary Standards        | ID              | Check Date (in house)             | Scheduled Check        |
| RF generator HP 8648C      | US3642U01700    | 4-Aug-99 (in house check Apr-13)  | In house check: Apr-16 |
| Network Analyzer HP 8753E  | US37390585      | 18-Oct-01 (in house check Oct-14) | In house check: Oct-15 |

|                |                              |                                          |                          |
|----------------|------------------------------|------------------------------------------|--------------------------|
| Calibrated by: | Name<br><b>Israe Elnaouq</b> | Function<br><b>Laboratory Technician</b> | Signature<br>            |
| Approved by:   | <b>Katja Pokovic</b>         | <b>Technical Manager</b>                 |                          |
|                |                              |                                          | Issued: January 26, 2015 |

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.



Accredited by the Swiss Accreditation Service (SAS)

Accreditation No.: **SCS 0108**

The Swiss Accreditation Service is one of the signatories to the EA  
 Multilateral Agreement for the recognition of calibration certificates

### Glossary:

|                          |                                                                                                                                                         |
|--------------------------|---------------------------------------------------------------------------------------------------------------------------------------------------------|
| TSL                      | tissue simulating liquid                                                                                                                                |
| NORM <sub>x,y,z</sub>    | sensitivity in free space                                                                                                                               |
| ConvF                    | sensitivity in TSL / NORM <sub>x,y,z</sub>                                                                                                              |
| DCP                      | diode compression point                                                                                                                                 |
| CF                       | crest factor (1/duty_cycle) of the RF signal                                                                                                            |
| A, B, C, D               | modulation dependent linearization parameters                                                                                                           |
| Polarization $\phi$      | $\phi$ rotation around probe axis                                                                                                                       |
| Polarization $\vartheta$ | $\vartheta$ rotation around an axis that is in the plane normal to probe axis (at measurement center),<br>i.e., $\vartheta = 0$ is normal to probe axis |
| Connector Angle          | information used in DASY system to align probe sensor X to the robot coordinate system                                                                  |

### Calibration is Performed According to the Following Standards:

- IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013
- IEC 62209-1, "Procedure to measure the Specific Absorption Rate (SAR) for hand-held devices used in close proximity to the ear (frequency range of 300 MHz to 3 GHz)", February 2005

### Methods Applied and Interpretation of Parameters:

- NORM<sub>x,y,z</sub>**: Assessed for E-field polarization  $\vartheta = 0$  ( $f \leq 900$  MHz in TEM-cell;  $f > 1800$  MHz: R22 waveguide). NORM<sub>x,y,z</sub> are only intermediate values, i.e., the uncertainties of NORM<sub>x,y,z</sub> does not affect the E<sup>2</sup>-field uncertainty inside TSL (see below ConvF).
- NORM(f)<sub>x,y,z</sub>** = NORM<sub>x,y,z</sub> \* frequency\_response (see Frequency Response Chart). This linearization is implemented in DASY4 software versions later than 4.2. The uncertainty of the frequency response is included in the stated uncertainty of ConvF.
- DCP<sub>x,y,z</sub>**: DCP are numerical linearization parameters assessed based on the data of power sweep with CW signal (no uncertainty required). DCP does not depend on frequency nor media.
- PAR**: PAR is the Peak to Average Ratio that is not calibrated but determined based on the signal characteristics
- A<sub>x,y,z</sub>; B<sub>x,y,z</sub>; C<sub>x,y,z</sub>; D<sub>x,y,z</sub>; VR<sub>x,y,z</sub>**: A, B, C, D are numerical linearization parameters assessed based on the data of power sweep for specific modulation signal. The parameters do not depend on frequency nor media. VR is the maximum calibration range expressed in RMS voltage across the diode.
- ConvF and Boundary Effect Parameters**: Assessed in flat phantom using E-field (or Temperature Transfer Standard for  $f \leq 800$  MHz) and inside waveguide using analytical field distributions based on power measurements for  $f > 800$  MHz. The same setups are used for assessment of the parameters applied for boundary compensation (alpha, depth) of which typical uncertainty values are given. These parameters are used in DASY4 software to improve probe accuracy close to the boundary. The sensitivity in TSL corresponds to NORM<sub>x,y,z</sub> \* ConvF whereby the uncertainty corresponds to that given for ConvF. A frequency dependent ConvF is used in DASY version 4.4 and higher which allows extending the validity from  $\pm 50$  MHz to  $\pm 100$  MHz.
- Spherical isotropy (3D deviation from isotropy)**: in a field of low gradients realized using a flat phantom exposed by a patch antenna.
- Sensor Offset**: The sensor offset corresponds to the offset of virtual measurement center from the probe tip (on probe axis). No tolerance required.
- Connector Angle**: The angle is assessed using the information gained by determining the NORM<sub>x</sub> (no uncertainty required).

# Probe ES3DV3

## SN:3318

Manufactured: January 10, 2012  
Calibrated: January 23, 2015

Calibrated for DASY/EASY Systems  
(Note: non-compatible with DASY2 system!)

# DASY/EASY - Parameters of Probe: ES3DV3 - SN:3318

## Basic Calibration Parameters

|                                       | Sensor X | Sensor Y | Sensor Z | Unc (k=2)     |
|---------------------------------------|----------|----------|----------|---------------|
| Norm ( $\mu V/(V/m)^2$ ) <sup>A</sup> | 1.15     | 0.92     | 1.28     | $\pm 10.1 \%$ |
| DCP (mV) <sup>B</sup>                 | 106.4    | 109.2    | 103.4    |               |

## Modulation Calibration Parameters

| UID           | Communication System Name                     |   | A<br>dB | B<br>dB $\sqrt{\mu V}$ | C    | D<br>dB | VR<br>mV | Unc <sup>E</sup><br>(k=2) |
|---------------|-----------------------------------------------|---|---------|------------------------|------|---------|----------|---------------------------|
| 0             | CW                                            | X | 0.0     | 0.0                    | 1.0  | 0.00    | 200.6    | $\pm 3.5 \%$              |
|               |                                               | Y | 0.0     | 0.0                    | 1.0  |         | 185.3    |                           |
|               |                                               | Z | 0.0     | 0.0                    | 1.0  |         | 207.7    |                           |
| 10010-<br>CAA | SAR Validation (Square, 100ms, 10ms)          | X | 3.26    | 66.4                   | 14.0 | 10.00   | 41.4     | $\pm 1.2 \%$              |
|               |                                               | Y | 1.76    | 59.6                   | 9.8  |         | 36.1     |                           |
|               |                                               | Z | 1.82    | 57.7                   | 9.6  |         | 43.6     |                           |
| 10011-<br>CAB | UMTS-FDD (WCDMA)                              | X | 3.48    | 68.9                   | 19.9 | 2.91    | 120.2    | $\pm 0.5 \%$              |
|               |                                               | Y | 3.76    | 70.1                   | 19.9 |         | 146.0    |                           |
|               |                                               | Z | 3.11    | 66.0                   | 17.9 |         | 124.4    |                           |
| 10012-<br>CAB | IEEE 802.11b WiFi 2.4 GHz (DSSS, 1 Mbps)      | X | 3.71    | 74.2                   | 21.7 | 1.87    | 121.7    | $\pm 0.7 \%$              |
|               |                                               | Y | 3.65    | 73.3                   | 20.7 |         | 147.5    |                           |
|               |                                               | Z | 2.77    | 67.4                   | 17.8 |         | 126.6    |                           |
| 10013-<br>CAB | IEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 6 Mbps) | X | 10.68   | 69.5                   | 22.7 | 9.46    | 114.7    | $\pm 2.5 \%$              |
|               |                                               | Y | 10.82   | 70.4                   | 23.0 |         | 139.8    |                           |
|               |                                               | Z | 11.22   | 71.1                   | 23.7 |         | 122.2    |                           |
| 10021-<br>DAB | GSM-FDD (TDMA, GMSK)                          | X | 16.13   | 95.0                   | 26.6 | 9.39    | 122.7    | $\pm 2.2 \%$              |
|               |                                               | Y | 4.61    | 73.1                   | 17.2 |         | 130.8    |                           |
|               |                                               | Z | 15.10   | 92.0                   | 25.4 |         | 135.9    |                           |
| 10023-<br>DAB | GPRS-FDD (TDMA, GMSK, TN 0)                   | X | 17.03   | 96.8                   | 27.5 | 9.57    | 113.0    | $\pm 1.9 \%$              |
|               |                                               | Y | 4.15    | 71.7                   | 16.8 |         | 119.9    |                           |
|               |                                               | Z | 21.50   | 98.0                   | 27.5 |         | 130.9    |                           |
| 10024-<br>DAB | GPRS-FDD (TDMA, GMSK, TN 0-1)                 | X | 35.51   | 99.5                   | 24.5 | 6.56    | 147.6    | $\pm 2.7 \%$              |
|               |                                               | Y | 6.12    | 77.2                   | 17.1 |         | 118.1    |                           |
|               |                                               | Z | 38.50   | 99.7                   | 24.7 |         | 114.0    |                           |
| 10027-<br>DAB | GPRS-FDD (TDMA, GMSK, TN 0-1-2)               | X | 45.57   | 99.9                   | 23.2 | 4.80    | 113.3    | $\pm 1.7 \%$              |
|               |                                               | Y | 2.73    | 68.4                   | 12.6 |         | 133.3    |                           |
|               |                                               | Z | 54.59   | 99.9                   | 22.9 |         | 131.0    |                           |
| 10028-<br>DAB | GPRS-FDD (TDMA, GMSK, TN 0-1-2-3)             | X | 53.68   | 99.5                   | 21.9 | 3.55    | 123.0    | $\pm 3.0 \%$              |
|               |                                               | Y | 60.05   | 99.8                   | 21.1 |         | 144.9    |                           |
|               |                                               | Z | 66.60   | 99.6                   | 21.6 |         | 140.7    |                           |
| 10032-<br>CAA | IEEE 802.15.1 Bluetooth (GFSK, DH5)           | X | 20.92   | 99.4                   | 21.8 | 1.16    | 136.6    | $\pm 2.2 \%$              |
|               |                                               | Y | 95.40   | 88.3                   | 13.8 |         | 117.6    |                           |
|               |                                               | Z | 100.00  | 99.5                   | 18.7 |         | 110.1    |                           |
| 10100-<br>CAB | LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)      | X | 6.50    | 68.1                   | 20.2 | 5.67    | 130.5    | $\pm 1.2 \%$              |
|               |                                               | Y | 6.11    | 66.7                   | 19.2 |         | 107.2    |                           |
|               |                                               | Z | 6.55    | 68.2                   | 20.1 |         | 142.7    |                           |

|           |                                          |   |       |      |      |      |       |        |
|-----------|------------------------------------------|---|-------|------|------|------|-------|--------|
| 10103-CAB | LTE-TDD (SC-FDMA, 100% RB, 20 MHz, QPSK) | X | 9.76  | 74.8 | 25.9 | 9.29 | 116.0 | ±2.5 % |
|           |                                          | Y | 8.85  | 72.2 | 24.1 |      | 134.9 |        |
|           |                                          | Z | 10.83 | 77.4 | 27.2 |      | 131.5 |        |
| 10108-CAC | LTE-FDD (SC-FDMA, 100% RB, 10 MHz, QPSK) | X | 6.36  | 67.7 | 20.1 | 5.80 | 128.7 | ±1.2 % |
|           |                                          | Y | 5.92  | 66.1 | 19.0 |      | 106.6 |        |
|           |                                          | Z | 6.42  | 67.7 | 20.0 |      | 140.4 |        |
| 10117-CAB | IEEE 802.11n (HT Mixed, 13.5 Mbps, BPSK) | X | 10.20 | 69.1 | 21.6 | 8.07 | 118.1 | ±2.5 % |
|           |                                          | Y | 10.27 | 69.3 | 21.4 |      | 143.9 |        |
|           |                                          | Z | 10.43 | 69.7 | 21.8 |      | 131.0 |        |
| 10151-CAB | LTE-TDD (SC-FDMA, 50% RB, 20 MHz, QPSK)  | X | 9.09  | 73.7 | 25.5 | 9.28 | 112.0 | ±2.7 % |
|           |                                          | Y | 8.35  | 71.5 | 23.9 |      | 131.1 |        |
|           |                                          | Z | 9.58  | 74.4 | 25.6 |      | 126.8 |        |
| 10154-CAC | LTE-FDD (SC-FDMA, 50% RB, 10 MHz, QPSK)  | X | 6.01  | 67.0 | 19.8 | 5.75 | 126.4 | ±1.2 % |
|           |                                          | Y | 6.17  | 67.7 | 19.9 |      | 148.9 |        |
|           |                                          | Z | 6.07  | 67.1 | 19.7 |      | 137.2 |        |
| 10160-CAB | LTE-FDD (SC-FDMA, 50% RB, 15 MHz, QPSK)  | X | 6.41  | 67.4 | 19.9 | 5.82 | 130.9 | ±0.9 % |
|           |                                          | Y | 6.06  | 66.2 | 19.0 |      | 109.1 |        |
|           |                                          | Z | 6.54  | 67.7 | 20.0 |      | 142.6 |        |
| 10169-CAB | LTE-FDD (SC-FDMA, 1 RB, 20 MHz, QPSK)    | X | 4.79  | 66.5 | 19.8 | 5.73 | 109.4 | ±0.9 % |
|           |                                          | Y | 4.82  | 67.1 | 19.8 |      | 128.8 |        |
|           |                                          | Z | 4.85  | 66.4 | 19.5 |      | 119.0 |        |
| 10172-CAB | LTE-TDD (SC-FDMA, 1 RB, 20 MHz, QPSK)    | X | 8.44  | 79.3 | 28.7 | 9.21 | 125.1 | ±2.5 % |
|           |                                          | Y | 7.15  | 75.0 | 26.0 |      | 144.0 |        |
|           |                                          | Z | 10.13 | 83.8 | 30.8 |      | 141.9 |        |
| 10175-CAC | LTE-FDD (SC-FDMA, 1 RB, 10 MHz, QPSK)    | X | 5.13  | 68.2 | 20.8 | 5.72 | 146.5 | ±0.9 % |
|           |                                          | Y | 4.77  | 66.8 | 19.6 |      | 125.2 |        |
|           |                                          | Z | 4.81  | 66.2 | 19.4 |      | 118.5 |        |
| 10181-CAB | LTE-FDD (SC-FDMA, 1 RB, 15 MHz, QPSK)    | X | 5.11  | 68.1 | 20.7 | 5.72 | 146.4 | ±0.9 % |
|           |                                          | Y | 4.79  | 67.0 | 19.7 |      | 126.0 |        |
|           |                                          | Z | 4.88  | 66.6 | 19.7 |      | 118.9 |        |
| 10196-CAB | IEEE 802.11n (HT Mixed, 6.5 Mbps, BPSK)  | X | 9.63  | 68.3 | 21.2 | 8.10 | 108.2 | ±2.5 % |
|           |                                          | Y | 9.84  | 68.9 | 21.3 |      | 135.5 |        |
|           |                                          | Z | 9.99  | 69.2 | 21.7 |      | 124.0 |        |
| 10225-CAB | UMTS-FDD (HSPA+)                         | X | 6.99  | 67.3 | 19.7 | 5.97 | 134.8 | ±0.9 % |
|           |                                          | Y | 6.73  | 66.8 | 19.2 |      | 115.9 |        |
|           |                                          | Z | 6.71  | 66.2 | 19.0 |      | 106.3 |        |
| 10237-CAB | LTE-TDD (SC-FDMA, 1 RB, 10 MHz, QPSK)    | X | 7.79  | 76.4 | 27.0 | 9.21 | 126.4 | ±2.5 % |
|           |                                          | Y | 7.19  | 75.1 | 26.1 |      | 144.7 |        |
|           |                                          | Z | 10.12 | 83.9 | 30.9 |      | 142.0 |        |
| 10252-CAB | LTE-TDD (SC-FDMA, 50% RB, 10 MHz, QPSK)  | X | 8.19  | 71.9 | 24.7 | 9.24 | 103.3 | ±2.2 % |
|           |                                          | Y | 7.76  | 70.8 | 23.6 |      | 122.0 |        |
|           |                                          | Z | 9.31  | 75.2 | 26.4 |      | 119.1 |        |
| 10267-CAB | LTE-TDD (SC-FDMA, 100% RB, 10 MHz, QPSK) | X | 8.90  | 73.0 | 25.1 | 9.30 | 108.7 | ±2.2 % |
|           |                                          | Y | 8.38  | 71.6 | 24.0 |      | 129.7 |        |
|           |                                          | Z | 10.15 | 76.5 | 26.9 |      | 126.1 |        |

|           |                                                               |   |       |      |      |      |       |        |
|-----------|---------------------------------------------------------------|---|-------|------|------|------|-------|--------|
| 10275-CAB | UMTS-FDD (HSUPA, Subtest 5, 3GPP Rel8.4)                      | X | 4.42  | 67.2 | 19.2 | 3.96 | 119.1 | ±0.7 % |
|           |                                                               | Y | 4.71  | 68.5 | 19.5 |      | 143.8 |        |
|           |                                                               | Z | 4.39  | 66.7 | 18.6 |      | 131.7 |        |
| 10291-AAB | CDMA2000, RC3, SO55, Full Rate                                | X | 3.65  | 67.5 | 19.3 | 3.46 | 111.3 | ±0.5 % |
|           |                                                               | Y | 3.89  | 69.0 | 19.6 |      | 130.9 |        |
|           |                                                               | Z | 3.49  | 66.1 | 18.2 |      | 122.4 |        |
| 10292-AAB | CDMA2000, RC3, SO32, Full Rate                                | X | 3.60  | 67.6 | 19.3 | 3.39 | 114.4 | ±0.5 % |
|           |                                                               | Y | 3.85  | 69.1 | 19.7 |      | 133.4 |        |
|           |                                                               | Z | 3.45  | 66.2 | 18.2 |      | 123.7 |        |
| 10297-AAA | LTE-FDD (SC-FDMA, 50% RB, 20 MHz, QPSK)                       | X | 6.36  | 67.6 | 20.1 | 5.81 | 128.7 | ±1.2 % |
|           |                                                               | Y | 5.95  | 66.1 | 19.0 |      | 106.5 |        |
|           |                                                               | Z | 6.39  | 67.6 | 19.9 |      | 140.7 |        |
| 10311-AAA | LTE-FDD (SC-FDMA, 100% RB, 15 MHz, QPSK)                      | X | 6.98  | 68.4 | 20.6 | 6.06 | 134.9 | ±1.2 % |
|           |                                                               | Y | 6.52  | 66.7 | 19.3 |      | 111.3 |        |
|           |                                                               | Z | 7.06  | 68.6 | 20.5 |      | 146.2 |        |
| 10403-AAB | CDMA2000 (1xEV-DO, Rev. 0)                                    | X | 4.97  | 69.7 | 19.7 | 3.76 | 122.2 | ±0.5 % |
|           |                                                               | Y | 5.31  | 71.6 | 20.2 |      | 143.6 |        |
|           |                                                               | Z | 4.54  | 67.3 | 18.2 |      | 133.0 |        |
| 10404-AAB | CDMA2000 (1xEV-DO, Rev. A)                                    | X | 4.77  | 69.4 | 19.6 | 3.77 | 120.8 | ±0.5 % |
|           |                                                               | Y | 5.40  | 72.4 | 20.6 |      | 141.3 |        |
|           |                                                               | Z | 4.71  | 68.5 | 18.9 |      | 131.5 |        |
| 10415-AAA | IEEE 802.11b WiFi 2.4 GHz (DSSS, 1 Mbps, 99pc duty cycle)     | X | 3.07  | 71.7 | 20.7 | 1.54 | 120.5 | ±0.7 % |
|           |                                                               | Y | 3.52  | 73.8 | 21.0 |      | 142.0 |        |
|           |                                                               | Z | 2.38  | 66.1 | 17.4 |      | 129.6 |        |
| 10416-AAA | IEEE 802.11g WiFi 2.4 GHz (ERP-OFDM, 6 Mbps, 99pc duty cycle) | X | 9.73  | 68.3 | 21.2 | 8.23 | 114.7 | ±2.5 % |
|           |                                                               | Y | 9.99  | 69.2 | 21.5 |      | 138.0 |        |
|           |                                                               | Z | 10.10 | 69.4 | 21.9 |      | 125.3 |        |

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor  $k=2$ , which for a normal distribution corresponds to a coverage probability of approximately 95%.

<sup>A</sup> The uncertainties of NormX,Y,Z do not affect the  $E^2$ -field uncertainty inside TSL (see Pages 7 and 8).

<sup>B</sup> Numerical linearization parameter: uncertainty not required.

<sup>E</sup> Uncertainty is determined using the max. deviation from linear response applying rectangular distribution and is expressed for the square of the field value.

## DASY/EASY - Parameters of Probe: ES3DV3 - SN:3318

### Calibration Parameter Determined in Head Tissue Simulating Media

| f (MHz) <sup>C</sup> | Relative Permittivity <sup>F</sup> | Conductivity (S/m) <sup>F</sup> | ConvF X | ConvF Y | ConvF Z | Alpha <sup>G</sup> | Depth <sup>G</sup> (mm) | Unct. (k=2) |
|----------------------|------------------------------------|---------------------------------|---------|---------|---------|--------------------|-------------------------|-------------|
| 750                  | 41.9                               | 0.89                            | 6.58    | 6.58    | 6.58    | 0.36               | 1.73                    | ± 12.0 %    |
| 835                  | 41.5                               | 0.90                            | 6.39    | 6.39    | 6.39    | 0.80               | 1.14                    | ± 12.0 %    |
| 1750                 | 40.1                               | 1.37                            | 5.27    | 5.27    | 5.27    | 0.76               | 1.19                    | ± 12.0 %    |
| 1900                 | 40.0                               | 1.40                            | 5.05    | 5.05    | 5.05    | 0.44               | 1.55                    | ± 12.0 %    |
| 2300                 | 39.5                               | 1.67                            | 4.78    | 4.78    | 4.78    | 0.80               | 1.23                    | ± 12.0 %    |
| 2450                 | 39.2                               | 1.80                            | 4.50    | 4.50    | 4.50    | 0.55               | 1.49                    | ± 12.0 %    |
| 2600                 | 39.0                               | 1.96                            | 4.34    | 4.34    | 4.34    | 0.76               | 1.32                    | ± 12.0 %    |

<sup>C</sup> Frequency validity above 300 MHz of ± 100 MHz only applies for DASY v4.4 and higher (see Page 2), else it is restricted to ± 50 MHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below 300 MHz is ± 10, 25, 40, 50 and 70 MHz for ConvF assessments at 30, 64, 128, 150 and 220 MHz respectively. Above 5 GHz frequency validity can be extended to ± 110 MHz.

<sup>F</sup> At frequencies below 3 GHz, the validity of tissue parameters ( $\epsilon$  and  $\sigma$ ) can be relaxed to ± 10% if liquid compensation formula is applied to measured SAR values. At frequencies above 3 GHz, the validity of tissue parameters ( $\epsilon$  and  $\sigma$ ) is restricted to ± 5%. The uncertainty is the RSS of the ConvF uncertainty for indicated target tissue parameters.

<sup>G</sup> Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than ± 1% for frequencies below 3 GHz and below ± 2% for frequencies between 3-6 GHz at any distance larger than half the probe tip diameter from the boundary.

## DASY/EASY - Parameters of Probe: ES3DV3 - SN:3318

### Calibration Parameter Determined in Body Tissue Simulating Media

| f (MHz) <sup>C</sup> | Relative Permittivity <sup>F</sup> | Conductivity (S/m) <sup>F</sup> | ConvF X | ConvF Y | ConvF Z | Alpha <sup>G</sup> | Depth <sup>G</sup> (mm) | Unct. (k=2) |
|----------------------|------------------------------------|---------------------------------|---------|---------|---------|--------------------|-------------------------|-------------|
| 750                  | 55.5                               | 0.96                            | 6.22    | 6.22    | 6.22    | 0.67               | 1.28                    | ± 12.0 %    |
| 835                  | 55.2                               | 0.97                            | 6.23    | 6.23    | 6.23    | 0.80               | 1.19                    | ± 12.0 %    |
| 1750                 | 53.4                               | 1.49                            | 4.95    | 4.95    | 4.95    | 0.40               | 1.77                    | ± 12.0 %    |
| 1900                 | 53.3                               | 1.52                            | 4.76    | 4.76    | 4.76    | 0.60               | 1.48                    | ± 12.0 %    |
| 2300                 | 52.9                               | 1.81                            | 4.52    | 4.52    | 4.52    | 0.80               | 1.19                    | ± 12.0 %    |
| 2450                 | 52.7                               | 1.95                            | 4.37    | 4.37    | 4.37    | 0.72               | 1.23                    | ± 12.0 %    |
| 2600                 | 52.5                               | 2.16                            | 4.17    | 4.17    | 4.17    | 0.80               | 1.00                    | ± 12.0 %    |


<sup>C</sup> Frequency validity above 300 MHz of  $\pm 100$  MHz only applies for DASY v4.4 and higher (see Page 2), else it is restricted to  $\pm 50$  MHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below 300 MHz is  $\pm 10$ , 25, 40, 50 and 70 MHz for ConvF assessments at 30, 64, 128, 150 and 220 MHz respectively. Above 5 GHz frequency validity can be extended to  $\pm 110$  MHz.

<sup>F</sup> At frequencies below 3 GHz, the validity of tissue parameters ( $\epsilon$  and  $\sigma$ ) can be relaxed to  $\pm 10\%$  if liquid compensation formula is applied to measured SAR values. At frequencies above 3 GHz, the validity of tissue parameters ( $\epsilon$  and  $\sigma$ ) is restricted to  $\pm 5\%$ . The uncertainty is the RSS of the ConvF uncertainty for indicated target tissue parameters.

<sup>G</sup> Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than  $\pm 1\%$  for frequencies below 3 GHz and below  $\pm 2\%$  for frequencies between 3-6 GHz at any distance larger than half the probe tip diameter from the boundary.

## Frequency Response of E-Field


(TEM-Cell:ifi110 EXX, Waveguide: R22)





Uncertainty of Frequency Response of E-field:  $\pm 6.3\%$  ( $k=2$ )

## Receiving Pattern ( $\phi$ ), $\theta = 0^\circ$

f=600 MHz,TEM




f=1800 MHz,R22



Uncertainty of Axial Isotropy Assessment:  $\pm 0.5\%$  ( $k=2$ )


# Dynamic Range f(SAR<sub>head</sub>)

(TEM cell , f<sub>eval</sub>= 1900 MHz)





Uncertainty of Linearity Assessment:  $\pm 0.6\%$  (k=2)

## Conversion Factor Assessment



## Deviation from Isotropy in Liquid

Error ( $\phi, \theta$ ),  $f = 900 \text{ MHz}$



## DASY/EASY - Parameters of Probe: ES3DV3 - SN:3318

### Other Probe Parameters

|                                               |            |
|-----------------------------------------------|------------|
| Sensor Arrangement                            | Triangular |
| Connector Angle (°)                           | -104.4     |
| Mechanical Surface Detection Mode             | enabled    |
| Optical Surface Detection Mode                | disabled   |
| Probe Overall Length                          | 337 mm     |
| Probe Body Diameter                           | 10 mm      |
| Tip Length                                    | 10 mm      |
| Tip Diameter                                  | 4 mm       |
| Probe Tip to Sensor X Calibration Point       | 2 mm       |
| Probe Tip to Sensor Y Calibration Point       | 2 mm       |
| Probe Tip to Sensor Z Calibration Point       | 2 mm       |
| Recommended Measurement Distance from Surface | 3 mm       |



Accredited by the Swiss Accreditation Service (SAS)  
 The Swiss Accreditation Service is one of the signatories to the EA  
 Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 0108**

Client **PC Test**

Certificate No: **ES3-3319\_Mar15**

## CALIBRATION CERTIFICATE

Object **ES3DV3 - SN:3319**

Calibration procedure(s) **QA CAL-01.v9, QA CAL-23.v5, QA CAL-25.v6**  
**Calibration procedure for dosimetric E-field probes**

Calibration date: **March 19, 2015**

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI).  
 The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature ( $22 \pm 3$ )°C and humidity < 70%.

Calibration Equipment used (M&TE critical for calibration)

| Primary Standards          | ID              | Cal Date (Certificate No.)        | Scheduled Calibration  |
|----------------------------|-----------------|-----------------------------------|------------------------|
| Power meter E4419B         | GB41293874      | 03-Apr-14 (No. 217-01911)         | Apr-15                 |
| Power sensor E4412A        | MY41498087      | 03-Apr-14 (No. 217-01911)         | Apr-15                 |
| Reference 3 dB Attenuator  | SN: S5054 (3c)  | 03-Apr-14 (No. 217-01915)         | Apr-15                 |
| Reference 20 dB Attenuator | SN: S5277 (20x) | 03-Apr-14 (No. 217-01919)         | Apr-15                 |
| Reference 30 dB Attenuator | SN: S5129 (30b) | 03-Apr-14 (No. 217-01920)         | Apr-15                 |
| Reference Probe ES3DV2     | SN: 3013        | 30-Dec-14 (No. ES3-3013_Dec14)    | Dec-15                 |
| DAE4                       | SN: 660         | 14-Jan-15 (No. DAE4-660_Jan15)    | Jan-16                 |
| Secondary Standards        | ID              | Check Date (in house)             | Scheduled Check        |
| RF generator HP 8648C      | US3642U01700    | 4-Aug-99 (in house check Apr-13)  | In house check: Apr-16 |
| Network Analyzer HP 8753E  | US37390585      | 18-Oct-01 (in house check Oct-14) | In house check: Oct-15 |

|                                                                                                                 |                              |                                          |                        |
|-----------------------------------------------------------------------------------------------------------------|------------------------------|------------------------------------------|------------------------|
| Calibrated by:                                                                                                  | Name<br><b>Israe Elnaouq</b> | Function<br><b>Laboratory Technician</b> | Signature<br>          |
| Approved by:                                                                                                    | <b>Katja Pokovic</b>         | <b>Technical Manager</b>                 |                        |
|                                                                                                                 |                              |                                          | Issued: March 19, 2015 |
| This calibration certificate shall not be reproduced except in full without written approval of the laboratory. |                              |                                          |                        |



Accredited by the Swiss Accreditation Service (SAS)

Accreditation No.: **SCS 0108**

The Swiss Accreditation Service is one of the signatories to the EA  
 Multilateral Agreement for the recognition of calibration certificates

### Glossary:

|                          |                                                                                                                                                         |
|--------------------------|---------------------------------------------------------------------------------------------------------------------------------------------------------|
| TSL                      | tissue simulating liquid                                                                                                                                |
| NORM <sub>x,y,z</sub>    | sensitivity in free space                                                                                                                               |
| ConvF                    | sensitivity in TSL / NORM <sub>x,y,z</sub>                                                                                                              |
| DCP                      | diode compression point                                                                                                                                 |
| CF                       | crest factor (1/duty_cycle) of the RF signal                                                                                                            |
| A, B, C, D               | modulation dependent linearization parameters                                                                                                           |
| Polarization $\varphi$   | $\varphi$ rotation around probe axis                                                                                                                    |
| Polarization $\vartheta$ | $\vartheta$ rotation around an axis that is in the plane normal to probe axis (at measurement center),<br>i.e., $\vartheta = 0$ is normal to probe axis |
| Connector Angle          | information used in DASY system to align probe sensor X to the robot coordinate system                                                                  |

### Calibration is Performed According to the Following Standards:

- IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013
- IEC 62209-1, "Procedure to measure the Specific Absorption Rate (SAR) for hand-held devices used in close proximity to the ear (frequency range of 300 MHz to 3 GHz)", February 2005

### Methods Applied and Interpretation of Parameters:

- NORM<sub>x,y,z</sub>**: Assessed for E-field polarization  $\vartheta = 0$  ( $f \leq 900$  MHz in TEM-cell;  $f > 1800$  MHz: R22 waveguide). NORM<sub>x,y,z</sub> are only intermediate values, i.e., the uncertainties of NORM<sub>x,y,z</sub> does not affect the E<sup>2</sup>-field uncertainty inside TSL (see below ConvF).
- NORM(f)<sub>x,y,z</sub>** = NORM<sub>x,y,z</sub> \* frequency\_response (see Frequency Response Chart). This linearization is implemented in DASY4 software versions later than 4.2. The uncertainty of the frequency response is included in the stated uncertainty of ConvF.
- DCP<sub>x,y,z</sub>**: DCP are numerical linearization parameters assessed based on the data of power sweep with CW signal (no uncertainty required). DCP does not depend on frequency nor media.
- PAR**: PAR is the Peak to Average Ratio that is not calibrated but determined based on the signal characteristics
- A<sub>x,y,z</sub>; B<sub>x,y,z</sub>; C<sub>x,y,z</sub>; D<sub>x,y,z</sub>; VR<sub>x,y,z</sub>**: A, B, C, D are numerical linearization parameters assessed based on the data of power sweep for specific modulation signal. The parameters do not depend on frequency nor media. VR is the maximum calibration range expressed in RMS voltage across the diode.
- ConvF and Boundary Effect Parameters**: Assessed in flat phantom using E-field (or Temperature Transfer Standard for  $f \leq 800$  MHz) and inside waveguide using analytical field distributions based on power measurements for  $f > 800$  MHz. The same setups are used for assessment of the parameters applied for boundary compensation (alpha, depth) of which typical uncertainty values are given. These parameters are used in DASY4 software to improve probe accuracy close to the boundary. The sensitivity in TSL corresponds to NORM<sub>x,y,z</sub> \* ConvF whereby the uncertainty corresponds to that given for ConvF. A frequency dependent ConvF is used in DASY version 4.4 and higher which allows extending the validity from  $\pm 50$  MHz to  $\pm 100$  MHz.
- Spherical isotropy (3D deviation from isotropy)**: in a field of low gradients realized using a flat phantom exposed by a patch antenna.
- Sensor Offset**: The sensor offset corresponds to the offset of virtual measurement center from the probe tip (on probe axis). No tolerance required.
- Connector Angle**: The angle is assessed using the information gained by determining the NORM<sub>x</sub> (no uncertainty required).

# Probe ES3DV3

## SN:3319

Manufactured: January 10, 2012  
Calibrated: March 19, 2015

Calibrated for DASY/EASY Systems  
(Note: non-compatible with DASY2 system!)

# DASY/EASY - Parameters of Probe: ES3DV3 - SN:3319

## Basic Calibration Parameters

|                                       | Sensor X | Sensor Y | Sensor Z | Unc (k=2)     |
|---------------------------------------|----------|----------|----------|---------------|
| Norm ( $\mu V/(V/m)^2$ ) <sup>A</sup> | 1.12     | 1.08     | 1.15     | $\pm 10.1 \%$ |
| DCP (mV) <sup>B</sup>                 | 104.4    | 106.0    | 104.4    |               |

## Modulation Calibration Parameters

| UID       | Communication System Name                     |   | A<br>dB | B<br>dB $\sqrt{\mu V}$ | C    | D<br>dB | VR<br>mV | Unc <sup>E</sup><br>(k=2) |
|-----------|-----------------------------------------------|---|---------|------------------------|------|---------|----------|---------------------------|
| 0         | CW                                            | X | 0.0     | 0.0                    | 1.0  | 0.00    | 176.1    | $\pm 3.3 \%$              |
|           |                                               | Y | 0.0     | 0.0                    | 1.0  |         | 192.7    |                           |
|           |                                               | Z | 0.0     | 0.0                    | 1.0  |         | 174.6    |                           |
| 10010-CAA | SAR Validation (Square, 100ms, 10ms)          | X | 3.26    | 64.8                   | 13.4 | 10.00   | 41.7     | $\pm 1.9 \%$              |
|           |                                               | Y | 2.66    | 62.2                   | 11.7 |         | 39.5     |                           |
|           |                                               | Z | 3.51    | 64.8                   | 13.2 |         | 42.1     |                           |
| 10011-CAB | UMTS-FDD (WCDMA)                              | X | 3.47    | 68.1                   | 19.1 | 2.91    | 142.9    | $\pm 0.5 \%$              |
|           |                                               | Y | 3.37    | 67.9                   | 19.1 |         | 133.0    |                           |
|           |                                               | Z | 3.57    | 68.7                   | 19.4 |         | 138.6    |                           |
| 10012-CAB | IEEE 802.11b WiFi 2.4 GHz (DSSS, 1 Mbps)      | X | 3.48    | 71.8                   | 20.2 | 1.87    | 143.9    | $\pm 0.7 \%$              |
|           |                                               | Y | 3.23    | 70.9                   | 19.9 |         | 134.6    |                           |
|           |                                               | Z | 3.68    | 72.8                   | 20.6 |         | 140.5    |                           |
| 10013-CAB | IEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 6 Mbps) | X | 11.18   | 70.5                   | 23.1 | 9.46    | 143.4    | $\pm 3.3 \%$              |
|           |                                               | Y | 10.98   | 70.5                   | 23.2 |         | 129.9    |                           |
|           |                                               | Z | 11.19   | 70.6                   | 23.1 |         | 138.8    |                           |
| 10021-DAB | GSM-FDD (TDMA, GMSK)                          | X | 15.55   | 92.7                   | 26.1 | 9.39    | 126.5    | $\pm 1.7 \%$              |
|           |                                               | Y | 21.21   | 98.0                   | 27.2 |         | 142.0    |                           |
|           |                                               | Z | 19.50   | 96.1                   | 27.0 |         | 125.4    |                           |
| 10023-DAB | GPRS-FDD (TDMA, GMSK, TN 0)                   | X | 23.54   | 100.0                  | 28.4 | 9.57    | 142.6    | $\pm 2.2 \%$              |
|           |                                               | Y | 23.24   | 99.9                   | 28.0 |         | 137.4    |                           |
|           |                                               | Z | 23.57   | 99.6                   | 28.2 |         | 139.7    |                           |
| 10024-DAB | GPRS-FDD (TDMA, GMSK, TN 0-1)                 | X | 17.00   | 90.2                   | 22.7 | 6.56    | 128.9    | $\pm 2.2 \%$              |
|           |                                               | Y | 35.20   | 99.7                   | 24.9 |         | 148.2    |                           |
|           |                                               | Z | 33.12   | 99.6                   | 25.4 |         | 123.8    |                           |
| 10027-DAB | GPRS-FDD (TDMA, GMSK, TN 0-1-2)               | X | 44.20   | 99.6                   | 23.6 | 4.80    | 146.0    | $\pm 1.9 \%$              |
|           |                                               | Y | 49.99   | 99.9                   | 23.0 |         | 136.6    |                           |
|           |                                               | Z | 41.43   | 99.6                   | 23.9 |         | 141.4    |                           |
| 10028-DAB | GPRS-FDD (TDMA, GMSK, TN 0-1-2-3)             | X | 46.56   | 99.7                   | 22.7 | 3.55    | 127.7    | $\pm 2.2 \%$              |
|           |                                               | Y | 58.11   | 99.8                   | 21.9 |         | 145.3    |                           |
|           |                                               | Z | 55.65   | 99.6                   | 22.2 |         | 124.3    |                           |
| 10032-CAA | IEEE 802.15.1 Bluetooth (GFSK, DH5)           | X | 34.25   | 99.4                   | 21.1 | 1.16    | 140.3    | $\pm 1.7 \%$              |
|           |                                               | Y | 40.72   | 100.0                  | 20.6 |         | 135.7    |                           |
|           |                                               | Z | 45.39   | 100.0                  | 20.8 |         | 136.4    |                           |
| 10100-CAB | LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)      | X | 6.30    | 67.1                   | 19.5 | 5.67    | 127.4    | $\pm 1.4 \%$              |
|           |                                               | Y | 6.58    | 68.4                   | 20.3 |         | 149.0    |                           |
|           |                                               | Z | 6.55    | 68.0                   | 19.9 |         | 146.3    |                           |

|           |                                          |   |       |      |      |      |       |        |
|-----------|------------------------------------------|---|-------|------|------|------|-------|--------|
| 10103-CAB | LTE-TDD (SC-FDMA, 100% RB, 20 MHz, QPSK) | X | 10.47 | 75.6 | 25.8 | 9.29 | 146.6 | ±3.0 % |
|           |                                          | Y | 10.18 | 75.8 | 26.3 |      | 136.2 |        |
|           |                                          | Z | 10.38 | 75.3 | 25.6 |      | 140.8 |        |
| 10108-CAC | LTE-FDD (SC-FDMA, 100% RB, 10 MHz, QPSK) | X | 6.18  | 66.6 | 19.4 | 5.80 | 126.9 | ±1.4 % |
|           |                                          | Y | 6.40  | 67.8 | 20.1 |      | 147.0 |        |
|           |                                          | Z | 6.44  | 67.6 | 19.9 |      | 145.7 |        |
| 10117-CAB | IEEE 802.11n (HT Mixed, 13.5 Mbps, BPSK) | X | 10.24 | 69.0 | 21.3 | 8.07 | 142.7 | ±2.5 % |
|           |                                          | Y | 10.25 | 69.2 | 21.5 |      | 136.7 |        |
|           |                                          | Z | 10.16 | 68.8 | 21.2 |      | 136.6 |        |
| 10151-CAB | LTE-TDD (SC-FDMA, 50% RB, 20 MHz, QPSK)  | X | 9.85  | 74.8 | 25.6 | 9.28 | 140.8 | ±3.0 % |
|           |                                          | Y | 9.49  | 74.7 | 25.9 |      | 130.5 |        |
|           |                                          | Z | 9.90  | 74.8 | 25.6 |      | 136.8 |        |
| 10154-CAC | LTE-FDD (SC-FDMA, 50% RB, 10 MHz, QPSK)  | X | 6.13  | 67.1 | 19.7 | 5.75 | 146.6 | ±1.4 % |
|           |                                          | Y | 6.11  | 67.4 | 19.9 |      | 147.7 |        |
|           |                                          | Z | 6.12  | 67.1 | 19.7 |      | 142.3 |        |
| 10160-CAB | LTE-FDD (SC-FDMA, 50% RB, 15 MHz, QPSK)  | X | 6.33  | 66.7 | 19.4 | 5.82 | 128.9 | ±1.4 % |
|           |                                          | Y | 6.33  | 67.1 | 19.7 |      | 128.7 |        |
|           |                                          | Z | 6.57  | 67.6 | 19.9 |      | 147.4 |        |
| 10169-CAB | LTE-FDD (SC-FDMA, 1 RB, 20 MHz, QPSK)    | X | 4.89  | 66.4 | 19.5 | 5.73 | 127.5 | ±1.2 % |
|           |                                          | Y | 4.99  | 67.5 | 20.2 |      | 149.3 |        |
|           |                                          | Z | 5.09  | 67.3 | 20.0 |      | 145.1 |        |
| 10172-CAB | LTE-TDD (SC-FDMA, 1 RB, 20 MHz, QPSK)    | X | 7.99  | 75.8 | 26.3 | 9.21 | 127.6 | ±2.7 % |
|           |                                          | Y | 9.29  | 81.7 | 29.6 |      | 149.8 |        |
|           |                                          | Z | 8.04  | 75.8 | 26.3 |      | 123.6 |        |
| 10175-CAC | LTE-FDD (SC-FDMA, 1 RB, 10 MHz, QPSK)    | X | 5.08  | 67.3 | 20.0 | 5.72 | 149.3 | ±1.4 % |
|           |                                          | Y | 5.00  | 67.6 | 20.3 |      | 145.0 |        |
|           |                                          | Z | 5.09  | 67.3 | 20.0 |      | 145.0 |        |
| 10181-CAB | LTE-FDD (SC-FDMA, 1 RB, 15 MHz, QPSK)    | X | 5.08  | 67.3 | 20.0 | 5.72 | 148.5 | ±1.4 % |
|           |                                          | Y | 5.06  | 67.9 | 20.4 |      | 147.1 |        |
|           |                                          | Z | 5.11  | 67.4 | 20.0 |      | 144.8 |        |
| 10196-CAB | IEEE 802.11n (HT Mixed, 6.5 Mbps, BPSK)  | X | 9.89  | 68.7 | 21.2 | 8.10 | 134.6 | ±2.2 % |
|           |                                          | Y | 9.84  | 68.9 | 21.4 |      | 130.4 |        |
|           |                                          | Z | 9.82  | 68.5 | 21.1 |      | 130.4 |        |
| 10225-CAB | UMTS-FDD (HSPA+)                         | X | 7.02  | 67.1 | 19.5 | 5.97 | 138.0 | ±1.4 % |
|           |                                          | Y | 6.88  | 67.0 | 19.5 |      | 133.2 |        |
|           |                                          | Z | 7.01  | 67.1 | 19.5 |      | 134.6 |        |
| 10237-CAB | LTE-TDD (SC-FDMA, 1 RB, 10 MHz, QPSK)    | X | 8.01  | 75.9 | 26.4 | 9.21 | 128.0 | ±2.7 % |
|           |                                          | Y | 9.39  | 82.1 | 29.9 |      | 149.7 |        |
|           |                                          | Z | 8.34  | 76.9 | 26.9 |      | 129.1 |        |
| 10252-CAB | LTE-TDD (SC-FDMA, 50% RB, 10 MHz, QPSK)  | X | 9.05  | 73.6 | 25.1 | 9.24 | 130.6 | ±3.0 % |
|           |                                          | Y | 8.76  | 73.7 | 25.5 |      | 123.6 |        |
|           |                                          | Z | 9.10  | 73.6 | 25.1 |      | 127.8 |        |
| 10267-CAB | LTE-TDD (SC-FDMA, 100% RB, 10 MHz, QPSK) | X | 9.81  | 74.7 | 25.6 | 9.30 | 139.3 | ±3.0 % |
|           |                                          | Y | 9.50  | 74.8 | 25.9 |      | 130.7 |        |
|           |                                          | Z | 9.81  | 74.6 | 25.5 |      | 135.0 |        |

|           |                                                               |   |       |      |      |      |       |        |
|-----------|---------------------------------------------------------------|---|-------|------|------|------|-------|--------|
| 10275-CAB | UMTS-FDD (HSUPA, Subtest 5, 3GPP Rel8.4)                      | X | 4.49  | 67.1 | 18.9 | 3.96 | 140.1 | ±0.7 % |
|           |                                                               | Y | 4.46  | 67.2 | 19.0 |      | 137.6 |        |
|           |                                                               | Z | 4.52  | 67.1 | 18.9 |      | 137.1 |        |
| 10291-AAB | CDMA2000, RC3, SO55, Full Rate                                | X | 3.68  | 67.0 | 18.8 | 3.46 | 129.3 | ±0.7 % |
|           |                                                               | Y | 3.64  | 67.3 | 19.0 |      | 130.3 |        |
|           |                                                               | Z | 3.84  | 67.9 | 19.2 |      | 148.6 |        |
| 10292-AAB | CDMA2000, RC3, SO32, Full Rate                                | X | 3.64  | 67.2 | 18.8 | 3.39 | 131.8 | ±0.5 % |
|           |                                                               | Y | 3.60  | 67.4 | 19.1 |      | 128.2 |        |
|           |                                                               | Z | 3.71  | 67.5 | 19.0 |      | 128.0 |        |
| 10297-AAA | LTE-FDD (SC-FDMA, 50% RB, 20 MHz, QPSK)                       | X | 6.43  | 67.5 | 19.9 | 5.81 | 147.2 | ±1.7 % |
|           |                                                               | Y | 6.39  | 67.7 | 20.0 |      | 145.4 |        |
|           |                                                               | Z | 6.42  | 67.5 | 19.8 |      | 143.2 |        |
| 10311-AAA | LTE-FDD (SC-FDMA, 100% RB, 15 MHz, QPSK)                      | X | 6.73  | 67.1 | 19.7 | 6.06 | 129.7 | ±1.4 % |
|           |                                                               | Y | 6.75  | 67.5 | 19.9 |      | 130.8 |        |
|           |                                                               | Z | 6.75  | 67.3 | 19.7 |      | 126.2 |        |
| 10400-AAB | IEEE 802.11ac WiFi (20MHz, 64-QAM, 99pc duty cycle)           | X | 10.14 | 68.9 | 21.5 | 8.37 | 136.7 | ±2.5 % |
|           |                                                               | Y | 10.23 | 69.5 | 22.0 |      | 136.5 |        |
|           |                                                               | Z | 10.13 | 68.9 | 21.5 |      | 132.8 |        |
| 10403-AAB | CDMA2000 (1xEV-DO, Rev. 0)                                    | X | 4.97  | 69.2 | 19.3 | 3.76 | 143.5 | ±0.5 % |
|           |                                                               | Y | 4.87  | 69.3 | 19.4 |      | 141.0 |        |
|           |                                                               | Z | 5.02  | 69.2 | 19.3 |      | 139.6 |        |
| 10404-AAB | CDMA2000 (1xEV-DO, Rev. A)                                    | X | 4.91  | 69.3 | 19.4 | 3.77 | 139.8 | ±0.7 % |
|           |                                                               | Y | 4.67  | 68.9 | 19.1 |      | 138.9 |        |
|           |                                                               | Z | 4.89  | 69.1 | 19.3 |      | 137.1 |        |
| 10415-AAA | IEEE 802.11b WiFi 2.4 GHz (DSSS, 1 Mbps, 99pc duty cycle)     | X | 2.93  | 70.1 | 19.6 | 1.54 | 137.8 | ±0.7 % |
|           |                                                               | Y | 2.84  | 69.8 | 19.6 |      | 138.2 |        |
|           |                                                               | Z | 3.04  | 70.8 | 19.9 |      | 134.2 |        |
| 10416-AAA | IEEE 802.11g WiFi 2.4 GHz (ERP-OFDM, 6 Mbps, 99pc duty cycle) | X | 9.94  | 68.7 | 21.3 | 8.23 | 134.6 | ±2.2 % |
|           |                                                               | Y | 10.00 | 69.1 | 21.7 |      | 134.1 |        |
|           |                                                               | Z | 9.89  | 68.5 | 21.2 |      | 130.1 |        |

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor  $k=2$ , which for a normal distribution corresponds to a coverage probability of approximately 95%.

<sup>A</sup> The uncertainties of NormX,Y,Z do not affect the  $E^2$ -field uncertainty inside TSL (see Pages 7 and 8).

<sup>B</sup> Numerical linearization parameter: uncertainty not required.

<sup>E</sup> Uncertainty is determined using the max. deviation from linear response applying rectangular distribution and is expressed for the square of the field value.

## DASY/EASY - Parameters of Probe: ES3DV3 - SN:3319

### Calibration Parameter Determined in Head Tissue Simulating Media

| f (MHz) <sup>C</sup> | Relative Permittivity <sup>F</sup> | Conductivity (S/m) <sup>F</sup> | ConvF X | ConvF Y | ConvF Z | Alpha <sup>G</sup> | Depth <sup>G</sup> (mm) | Unct. (k=2) |
|----------------------|------------------------------------|---------------------------------|---------|---------|---------|--------------------|-------------------------|-------------|
| 750                  | 41.9                               | 0.89                            | 6.69    | 6.69    | 6.69    | 0.40               | 1.70                    | ± 12.0 %    |
| 835                  | 41.5                               | 0.90                            | 6.41    | 6.41    | 6.41    | 0.43               | 1.62                    | ± 12.0 %    |
| 1750                 | 40.1                               | 1.37                            | 5.29    | 5.29    | 5.29    | 0.80               | 1.16                    | ± 12.0 %    |
| 1900                 | 40.0                               | 1.40                            | 5.10    | 5.10    | 5.10    | 0.80               | 1.24                    | ± 12.0 %    |
| 2300                 | 39.5                               | 1.67                            | 4.77    | 4.77    | 4.77    | 0.64               | 1.38                    | ± 12.0 %    |
| 2450                 | 39.2                               | 1.80                            | 4.55    | 4.55    | 4.55    | 0.80               | 1.29                    | ± 12.0 %    |
| 2600                 | 39.0                               | 1.96                            | 4.39    | 4.39    | 4.39    | 0.80               | 1.31                    | ± 12.0 %    |

<sup>C</sup> Frequency validity above 300 MHz of ± 100 MHz only applies for DASY v4.4 and higher (see Page 2), else it is restricted to ± 50 MHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below 300 MHz is ± 10, 25, 40, 50 and 70 MHz for ConvF assessments at 30, 64, 128, 150 and 220 MHz respectively. Above 5 GHz frequency validity can be extended to ± 110 MHz.

<sup>F</sup> At frequencies below 3 GHz, the validity of tissue parameters ( $\epsilon$  and  $\sigma$ ) can be relaxed to ± 10% if liquid compensation formula is applied to measured SAR values. At frequencies above 3 GHz, the validity of tissue parameters ( $\epsilon$  and  $\sigma$ ) is restricted to ± 5%. The uncertainty is the RSS of the ConvF uncertainty for indicated target tissue parameters.

<sup>G</sup> Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than ± 1% for frequencies below 3 GHz and below ± 2% for frequencies between 3-6 GHz at any distance larger than half the probe tip diameter from the boundary.

## DASY/EASY - Parameters of Probe: ES3DV3 - SN:3319

### Calibration Parameter Determined in Body Tissue Simulating Media

| f (MHz) <sup>C</sup> | Relative Permittivity <sup>F</sup> | Conductivity (S/m) <sup>F</sup> | ConvF X | ConvF Y | ConvF Z | Alpha <sup>G</sup> | Depth (mm) <sup>G</sup> | Unct. (k=2) |
|----------------------|------------------------------------|---------------------------------|---------|---------|---------|--------------------|-------------------------|-------------|
| 750                  | 55.5                               | 0.96                            | 6.10    | 6.10    | 6.10    | 0.34               | 1.80                    | ± 12.0 %    |
| 835                  | 55.2                               | 0.97                            | 6.07    | 6.07    | 6.07    | 0.47               | 1.56                    | ± 12.0 %    |
| 1750                 | 53.4                               | 1.49                            | 4.83    | 4.83    | 4.83    | 0.70               | 1.36                    | ± 12.0 %    |
| 1900                 | 53.3                               | 1.52                            | 4.53    | 4.53    | 4.53    | 0.71               | 1.39                    | ± 12.0 %    |
| 2300                 | 52.9                               | 1.81                            | 4.24    | 4.24    | 4.24    | 0.80               | 1.26                    | ± 12.0 %    |
| 2450                 | 52.7                               | 1.95                            | 4.11    | 4.11    | 4.11    | 0.80               | 1.10                    | ± 12.0 %    |
| 2600                 | 52.5                               | 2.16                            | 3.90    | 3.90    | 3.90    | 0.80               | 1.11                    | ± 12.0 %    |


<sup>C</sup> Frequency validity above 300 MHz of ± 100 MHz only applies for DASY v4.4 and higher (see Page 2), else it is restricted to ± 50 MHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below 300 MHz is ± 10, 25, 40, 50 and 70 MHz for ConvF assessments at 30, 64, 128, 150 and 220 MHz respectively. Above 5 GHz frequency validity can be extended to ± 110 MHz.

<sup>F</sup> At frequencies below 3 GHz, the validity of tissue parameters ( $\epsilon$  and  $\sigma$ ) can be relaxed to ± 10% if liquid compensation formula is applied to measured SAR values. At frequencies above 3 GHz, the validity of tissue parameters ( $\epsilon$  and  $\sigma$ ) is restricted to ± 5%. The uncertainty is the RSS of the ConvF uncertainty for indicated target tissue parameters.

<sup>G</sup> Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than ± 1% for frequencies below 3 GHz and below ± 2% for frequencies between 3-6 GHz at any distance larger than half the probe tip diameter from the boundary.

## Frequency Response of E-Field

(TEM-Cell:ifi110 EXX, Waveguide: R22)





Uncertainty of Frequency Response of E-field:  $\pm 6.3\%$  ( $k=2$ )

Receiving Pattern ( $\phi$ ),  $\theta = 0^\circ$



f=600 MHz,TEM

f=1800 MHz,R22




Uncertainty of Axial Isotropy Assessment:  $\pm 0.5\%$  ( $k=2$ )

**Dynamic Range f(SAR<sub>head</sub>)**  
(TEM cell , f<sub>eval</sub>= 1900 MHz)





**Uncertainty of Linearity Assessment: ± 0.6% (k=2)**

## Conversion Factor Assessment



## Deviation from Isotropy in Liquid

Error ( $\phi, \theta$ ),  $f = 900 \text{ MHz}$



Uncertainty of Spherical Isotropy Assessment:  $\pm 2.6\%$  ( $k=2$ )

**DASY/EASY - Parameters of Probe: ES3DV3 - SN:3319****Other Probe Parameters**

|                                               |            |
|-----------------------------------------------|------------|
| Sensor Arrangement                            | Triangular |
| Connector Angle (°)                           | -120.2     |
| Mechanical Surface Detection Mode             | enabled    |
| Optical Surface Detection Mode                | disabled   |
| Probe Overall Length                          | 337 mm     |
| Probe Body Diameter                           | 10 mm      |
| Tip Length                                    | 10 mm      |
| Tip Diameter                                  | 4 mm       |
| Probe Tip to Sensor X Calibration Point       | 2 mm       |
| Probe Tip to Sensor Y Calibration Point       | 2 mm       |
| Probe Tip to Sensor Z Calibration Point       | 2 mm       |
| Recommended Measurement Distance from Surface | 3 mm       |

**Calibration Laboratory of**  
**Schmid & Partner**  
**Engineering AG**  
 Zeughausstrasse 43, 8004 Zurich, Switzerland



**S** Schweizerischer Kalibrierdienst  
**C** Service suisse d'étalonnage  
**S** Servizio svizzero di taratura  
 Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)  
 The Swiss Accreditation Service is one of the signatories to the EA  
 Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 108**

Client **PC Test**

Certificate No: **ES3-3333\_Oct14**

## CALIBRATION CERTIFICATE

Object **ES3DV3 - SN:3333**

Calibration procedure(s) **QA CAL-01.v9, QA CAL-23.v5, QA CAL-25.v6**  
**Calibration procedure for dosimetric E-field probes**

CC  
10/31/14

Calibration date: **October 24, 2014**

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI).  
 The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature ( $22 \pm 3$ )°C and humidity < 70%.

Calibration Equipment used (M&TE critical for calibration)

| Primary Standards          | ID              | Cal Date (Certificate No.)        | Scheduled Calibration  |
|----------------------------|-----------------|-----------------------------------|------------------------|
| Power meter E4419B         | GB41293874      | 03-Apr-14 (No. 217-01911)         | Apr-15                 |
| Power sensor E4412A        | MY41498087      | 03-Apr-14 (No. 217-01911)         | Apr-15                 |
| Reference 3 dB Attenuator  | SN: S5054 (3c)  | 03-Apr-14 (No. 217-01915)         | Apr-15                 |
| Reference 20 dB Attenuator | SN: S5277 (20x) | 03-Apr-14 (No. 217-01919)         | Apr-15                 |
| Reference 30 dB Attenuator | SN: S5129 (30b) | 03-Apr-14 (No. 217-01920)         | Apr-15                 |
| Reference Probe ES3DV2     | SN: 3013        | 30-Dec-13 (No. ES3-3013_Dec13)    | Dec-14                 |
| DAE4                       | SN: 660         | 13-Dec-13 (No. DAE4-660_Dec13)    | Dec-14                 |
| Secondary Standards        | ID              | Check Date (in house)             | Scheduled Check        |
| RF generator HP 8648C      | US3642U01700    | 4-Aug-99 (in house check Apr-13)  | In house check: Apr-16 |
| Network Analyzer HP 8753E  | US37390585      | 18-Oct-01 (in house check Oct-14) | In house check: Oct-15 |

|                |                              |                                   |                          |
|----------------|------------------------------|-----------------------------------|--------------------------|
| Calibrated by: | Name<br><b>Leif Klysner</b>  | Function<br>Laboratory Technician | Signature<br>            |
| Approved by:   | Name<br><b>Kalja Pokovic</b> | Technical Manager                 |                          |
|                |                              |                                   | Issued: October 24, 2014 |

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.



Accredited by the Swiss Accreditation Service (SAS)  
The Swiss Accreditation Service is one of the signatories to the EA  
Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 108**

### Glossary:

|                          |                                                                                                                                                         |
|--------------------------|---------------------------------------------------------------------------------------------------------------------------------------------------------|
| TSL                      | tissue simulating liquid                                                                                                                                |
| NORM <sub>x,y,z</sub>    | sensitivity in free space                                                                                                                               |
| ConvF                    | sensitivity in TSL / NORM <sub>x,y,z</sub>                                                                                                              |
| DCP                      | diode compression point                                                                                                                                 |
| CF                       | crest factor (1/duty_cycle) of the RF signal                                                                                                            |
| A, B, C, D               | modulation dependent linearization parameters                                                                                                           |
| Polarization $\phi$      | $\phi$ rotation around probe axis                                                                                                                       |
| Polarization $\vartheta$ | $\vartheta$ rotation around an axis that is in the plane normal to probe axis (at measurement center),<br>i.e., $\vartheta = 0$ is normal to probe axis |
| Connector Angle          | information used in DASY system to align probe sensor X to the robot coordinate system                                                                  |

### Calibration is Performed According to the Following Standards:

- IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013
- IEC 62209-1, "Procedure to measure the Specific Absorption Rate (SAR) for hand-held devices used in close proximity to the ear (frequency range of 300 MHz to 3 GHz)", February 2005

### Methods Applied and Interpretation of Parameters:

- NORM<sub>x,y,z</sub>**: Assessed for E-field polarization  $\vartheta = 0$  ( $f \leq 900$  MHz in TEM-cell;  $f > 1800$  MHz: R22 waveguide). NORM<sub>x,y,z</sub> are only intermediate values, i.e., the uncertainties of NORM<sub>x,y,z</sub> does not affect the E<sup>2</sup>-field uncertainty inside TSL (see below ConvF).
- NORM(f)<sub>x,y,z</sub>** = NORM<sub>x,y,z</sub> \* frequency\_response (see Frequency Response Chart). This linearization is implemented in DASY4 software versions later than 4.2. The uncertainty of the frequency response is included in the stated uncertainty of ConvF.
- DCP<sub>x,y,z</sub>**: DCP are numerical linearization parameters assessed based on the data of power sweep with CW signal (no uncertainty required). DCP does not depend on frequency nor media.
- PAR**: PAR is the Peak to Average Ratio that is not calibrated but determined based on the signal characteristics
- A<sub>x,y,z</sub>; B<sub>x,y,z</sub>; C<sub>x,y,z</sub>; D<sub>x,y,z</sub>; VR<sub>x,y,z</sub>**: A, B, C, D are numerical linearization parameters assessed based on the data of power sweep for specific modulation signal. The parameters do not depend on frequency nor media. VR is the maximum calibration range expressed in RMS voltage across the diode.
- ConvF and Boundary Effect Parameters**: Assessed in flat phantom using E-field (or Temperature Transfer Standard for  $f \leq 800$  MHz) and inside waveguide using analytical field distributions based on power measurements for  $f > 800$  MHz. The same setups are used for assessment of the parameters applied for boundary compensation (alpha, depth) of which typical uncertainty values are given. These parameters are used in DASY4 software to improve probe accuracy close to the boundary. The sensitivity in TSL corresponds to NORM<sub>x,y,z</sub> \* ConvF whereby the uncertainty corresponds to that given for ConvF. A frequency dependent ConvF is used in DASY version 4.4 and higher which allows extending the validity from  $\pm 50$  MHz to  $\pm 100$  MHz.
- Spherical Isotropy (3D deviation from isotropy)**: in a field of low gradients realized using a flat phantom exposed by a patch antenna.
- Sensor Offset**: The sensor offset corresponds to the offset of virtual measurement center from the probe tip (on probe axis). No tolerance required.
- Connector Angle**: The angle is assessed using the information gained by determining the NORM<sub>x</sub> (no uncertainty required).

# Probe ES3DV3

## SN:3333

Manufactured: January 24, 2012  
Calibrated: October 24, 2014

Calibrated for DASY/EASY Systems  
(Note: non-compatible with DASY2 system!)

## DASY/EASY - Parameters of Probe: ES3DV3 - SN:3333

### Basic Calibration Parameters

|                                       | Sensor X | Sensor Y | Sensor Z | Unc (k=2) |
|---------------------------------------|----------|----------|----------|-----------|
| Norm ( $\mu V/(V/m)^2$ ) <sup>A</sup> | 1.08     | 0.90     | 0.88     | ± 10.1 %  |
| DCP (mV) <sup>B</sup>                 | 102.7    | 107.7    | 106.3    |           |

### Modulation Calibration Parameters

| UID       | Communication System Name                     |   | A<br>dB | B<br>dB $\sqrt{\mu V}$ | C    | D<br>dB | VR<br>mV | Unc <sup>E</sup><br>(k=2) |
|-----------|-----------------------------------------------|---|---------|------------------------|------|---------|----------|---------------------------|
| 0         | CW                                            | X | 0.0     | 0.0                    | 1.0  | 0.00    | 190.7    | ±2.5 %                    |
|           |                                               | Y | 0.0     | 0.0                    | 1.0  |         | 183.3    |                           |
|           |                                               | Z | 0.0     | 0.0                    | 1.0  |         | 197.9    |                           |
| 10010-CAA | SAR Validation (Square, 100ms, 10ms)          | X | 3.17    | 61.7                   | 12.2 | 10.00   | 42.4     | ±1.9 %                    |
|           |                                               | Y | 3.16    | 63.7                   | 12.4 |         | 38.0     |                           |
|           |                                               | Z | 1.84    | 59.2                   | 10.5 |         | 39.9     |                           |
| 10011-CAB | UMTS-FDD (WCDMA)                              | X | 3.22    | 65.9                   | 17.6 | 2.91    | 128.5    | ±0.5 %                    |
|           |                                               | Y | 3.60    | 69.3                   | 19.8 |         | 146.7    |                           |
|           |                                               | Z | 3.51    | 68.1                   | 18.8 |         | 133.7    |                           |
| 10012-CAA | IEEE 802.11b WIFI 2.4 GHz (DSSS, 1 Mbps)      | X | 3.14    | 68.6                   | 18.2 | 1.87    | 132.6    | ±0.7 %                    |
|           |                                               | Y | 3.64    | 73.3                   | 21.1 |         | 127.5    |                           |
|           |                                               | Z | 3.50    | 71.4                   | 19.6 |         | 136.4    |                           |
| 10013-CAA | IEEE 802.11g WIFI 2.4 GHz (DSSS-OFDM, 6 Mbps) | X | 11.56   | 70.8                   | 23.0 | 9.46    | 135.8    | ±3.5 %                    |
|           |                                               | Y | 10.93   | 70.2                   | 23.0 |         | 122.3    |                           |
|           |                                               | Z | 10.93   | 70.0                   | 22.6 |         | 132.8    |                           |
| 10021-DAB | GSM-FDD (TDMA, GMSK)                          | X | 24.80   | 96.9                   | 27.6 | 9.39    | 147.6    | ±1.9 %                    |
|           |                                               | Y | 19.44   | 94.3                   | 26.1 |         | 148.6    |                           |
|           |                                               | Z | 9.58    | 82.7                   | 21.9 |         | 138.2    |                           |
| 10023-DAB | GPRS-FDD (TDMA, GMSK, TN 0)                   | X | 20.09   | 93.0                   | 26.4 | 9.57    | 141.7    | ±2.7 %                    |
|           |                                               | Y | 24.86   | 99.0                   | 27.9 |         | 143.5    |                           |
|           |                                               | Z | 11.74   | 86.4                   | 23.4 |         | 134.4    |                           |
| 10024-DAB | GPRS-FDD (TDMA, GMSK, TN 0-1)                 | X | 23.76   | 91.2                   | 23.1 | 6.56    | 147.8    | ±2.5 %                    |
|           |                                               | Y | 37.10   | 99.8                   | 25.3 |         | 149.9    |                           |
|           |                                               | Z | 16.01   | 88.1                   | 21.6 |         | 128.0    |                           |
| 10027-DAB | GPRS-FDD (TDMA, GMSK, TN 0-1-2)               | X | 36.24   | 94.5                   | 22.6 | 4.80    | 128.6    | ±2.5 %                    |
|           |                                               | Y | 47.57   | 99.9                   | 23.7 |         | 133.5    |                           |
|           |                                               | Z | 44.37   | 99.7                   | 23.6 |         | 140.1    |                           |
| 10028-DAB | GPRS-FDD (TDMA, GMSK, TN 0-1-2-3)             | X | 65.86   | 99.7                   | 22.7 | 3.55    | 133.1    | ±2.7 %                    |
|           |                                               | Y | 55.92   | 100.0                  | 22.6 |         | 142.0    |                           |
|           |                                               | Z | 59.41   | 100.0                  | 22.2 |         | 125.1    |                           |
| 10032-CAA | IEEE 802.15.1 Bluetooth (GFSK, DH5)           | X | 85.87   | 100.0                  | 20.1 | 1.16    | 138.3    | ±2.2 %                    |
|           |                                               | Y | 14.41   | 99.2                   | 23.3 |         | 130.5    |                           |
|           |                                               | Z | 85.82   | 99.8                   | 19.3 |         | 135.9    |                           |
| 10100-CAB | LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)      | X | 6.49    | 67.4                   | 19.4 | 5.67    | 144.6    | ±1.7 %                    |
|           |                                               | Y | 6.49    | 68.0                   | 20.1 |         | 139.9    |                           |
|           |                                               | Z | 6.54    | 67.9                   | 19.7 |         | 147.3    |                           |

|           |                                          |   |       |      |      |      |       |        |
|-----------|------------------------------------------|---|-------|------|------|------|-------|--------|
| 10103-CAB | LTE-TDD (SC-FDMA, 100% RB, 20 MHz, QPSK) | X | 10.81 | 74.7 | 24.9 | 9.29 | 122.0 | ±3.0 % |
|           |                                          | Y | 10.50 | 75.9 | 26.1 |      | 131.6 |        |
|           |                                          | Z | 9.76  | 73.5 | 24.5 |      | 138.6 |        |
| 10108-CAB | LTE-FDD (SC-FDMA, 100% RB, 10 MHz, QPSK) | X | 6.43  | 67.2 | 19.4 | 5.80 | 143.3 | ±1.7 % |
|           |                                          | Y | 6.37  | 67.7 | 20.0 |      | 138.0 |        |
|           |                                          | Z | 6.43  | 67.5 | 19.7 |      | 146.7 |        |
| 10117-CAA | IEEE 802.11n (HT Mixed, 13.5 Mbps, BPSK) | X | 10.19 | 68.6 | 20.9 | 8.07 | 136.2 | ±2.5 % |
|           |                                          | Y | 10.15 | 68.9 | 21.4 |      | 128.3 |        |
|           |                                          | Z | 10.12 | 68.7 | 21.0 |      | 137.9 |        |
| 10151-CAB | LTE-TDD (SC-FDMA, 50% RB, 20 MHz, QPSK)  | X | 11.48 | 77.0 | 26.1 | 9.28 | 147.5 | ±3.3 % |
|           |                                          | Y | 9.81  | 74.9 | 25.8 |      | 125.7 |        |
|           |                                          | Z | 9.22  | 72.8 | 24.3 |      | 133.2 |        |
| 10154-CAB | LTE-FDD (SC-FDMA, 50% RB, 10 MHz, QPSK)  | X | 6.10  | 66.5 | 19.1 | 5.75 | 140.0 | ±1.7 % |
|           |                                          | Y | 6.04  | 67.1 | 19.8 |      | 134.8 |        |
|           |                                          | Z | 6.12  | 67.1 | 19.5 |      | 143.2 |        |
| 10160-CAB | LTE-FDD (SC-FDMA, 50% RB, 15 MHz, QPSK)  | X | 6.57  | 67.2 | 19.4 | 5.82 | 146.3 | ±1.7 % |
|           |                                          | Y | 6.47  | 67.6 | 20.0 |      | 139.6 |        |
|           |                                          | Z | 6.56  | 67.6 | 19.7 |      | 148.5 |        |
| 10169-CAB | LTE-FDD (SC-FDMA, 1 RB, 20 MHz, QPSK)    | X | 5.16  | 66.7 | 19.4 | 5.73 | 145.8 | ±1.4 % |
|           |                                          | Y | 5.02  | 67.5 | 20.2 |      | 137.5 |        |
|           |                                          | Z | 5.07  | 67.2 | 19.7 |      | 147.1 |        |
| 10172-CAB | LTE-TDD (SC-FDMA, 1 RB, 20 MHz, QPSK)    | X | 10.07 | 79.2 | 27.3 | 9.21 | 136.5 | ±3.0 % |
|           |                                          | Y | 9.70  | 81.5 | 29.3 |      | 142.5 |        |
|           |                                          | Z | 7.63  | 74.3 | 25.3 |      | 125.0 |        |
| 10175-CAB | LTE-FDD (SC-FDMA, 1 RB, 10 MHz, QPSK)    | X | 5.13  | 66.6 | 19.3 | 5.72 | 145.9 | ±1.4 % |
|           |                                          | Y | 5.01  | 67.4 | 20.1 |      | 137.5 |        |
|           |                                          | Z | 5.04  | 67.1 | 19.7 |      | 146.3 |        |
| 10181-CAB | LTE-FDD (SC-FDMA, 1 RB, 15 MHz, QPSK)    | X | 5.14  | 66.6 | 19.3 | 5.72 | 145.7 | ±1.4 % |
|           |                                          | Y | 5.03  | 67.5 | 20.3 |      | 137.4 |        |
|           |                                          | Z | 5.06  | 67.2 | 19.7 |      | 146.6 |        |
| 10196-CAA | IEEE 802.11n (HT Mixed, 6.5 Mbps, BPSK)  | X | 9.88  | 68.3 | 20.8 | 8.10 | 130.9 | ±2.5 % |
|           |                                          | Y | 10.13 | 69.6 | 21.8 |      | 149.0 |        |
|           |                                          | Z | 9.77  | 68.4 | 20.9 |      | 131.6 |        |
| 10225-CAB | UMTS-FDD (HSPA+)                         | X | 6.98  | 66.5 | 19.0 | 5.97 | 132.9 | ±1.7 % |
|           |                                          | Y | 7.14  | 67.8 | 20.0 |      | 149.7 |        |
|           |                                          | Z | 7.02  | 67.2 | 19.4 |      | 134.3 |        |
| 10237-CAB | LTE-TDD (SC-FDMA, 1 RB, 10 MHz, QPSK)    | X | 10.13 | 79.4 | 27.4 | 9.21 | 137.5 | ±3.0 % |
|           |                                          | Y | 9.73  | 81.6 | 29.3 |      | 143.3 |        |
|           |                                          | Z | 7.59  | 74.1 | 25.1 |      | 125.6 |        |
| 10252-CAB | LTE-TDD (SC-FDMA, 50% RB, 10 MHz, QPSK)  | X | 10.80 | 76.4 | 25.9 | 9.24 | 140.0 | ±3.3 % |
|           |                                          | Y | 10.19 | 77.2 | 27.1 |      | 147.2 |        |
|           |                                          | Z | 8.55  | 71.8 | 23.9 |      | 124.9 |        |
| 10267-CAB | LTE-TDD (SC-FDMA, 100% RB, 10 MHz, QPSK) | X | 11.59 | 77.3 | 26.3 | 9.30 | 148.4 | ±3.5 % |
|           |                                          | Y | 9.87  | 75.1 | 25.9 |      | 126.0 |        |
|           |                                          | Z | 9.21  | 72.7 | 24.2 |      | 133.6 |        |

|           |                                                               |   |      |      |      |      |       |        |
|-----------|---------------------------------------------------------------|---|------|------|------|------|-------|--------|
| 10275-CAB | UMTS-FDD (HSUPA, Subtest 5, 3GPP Rel8.4)                      | X | 4.40 | 66.1 | 18.1 | 3.96 | 134.1 | ±0.7 % |
|           |                                                               | Y | 4.48 | 67.4 | 19.2 |      | 129.7 |        |
|           |                                                               | Z | 4.54 | 67.2 | 18.7 |      | 137.4 |        |
| 10291-AAB | CDMA2000, RC3, SO55, Full Rate                                | X | 3.59 | 65.7 | 17.7 | 3.46 | 127.5 | ±0.7 % |
|           |                                                               | Y | 3.85 | 68.4 | 19.7 |      | 143.4 |        |
|           |                                                               | Z | 3.78 | 67.6 | 18.8 |      | 129.7 |        |
| 10292-AAB | CDMA2000, RC3, SO32, Full Rate                                | X | 3.56 | 65.9 | 17.8 | 3.39 | 127.9 | ±0.7 % |
|           |                                                               | Y | 3.81 | 68.6 | 19.8 |      | 144.2 |        |
|           |                                                               | Z | 3.71 | 67.5 | 18.8 |      | 130.7 |        |
| 10297-AAA | LTE-FDD (SC-FDMA, 50% RB, 20 MHz, QPSK)                       | X | 6.44 | 67.1 | 19.4 | 5.81 | 143.0 | ±1.7 % |
|           |                                                               | Y | 6.37 | 67.6 | 20.0 |      | 137.9 |        |
|           |                                                               | Z | 6.43 | 67.5 | 19.7 |      | 146.5 |        |
| 10311-AAA | LTE-FDD (SC-FDMA, 100% RB, 15 MHz, QPSK)                      | X | 7.02 | 67.8 | 19.8 | 6.06 | 148.7 | ±1.9 % |
|           |                                                               | Y | 6.96 | 68.2 | 20.4 |      | 143.6 |        |
|           |                                                               | Z | 6.72 | 67.1 | 19.5 |      | 126.9 |        |
| 10403-AAB | CDMA2000 (1xEV-DO, Rev. 0)                                    | X | 4.73 | 67.0 | 17.9 | 3.76 | 140.2 | ±0.7 % |
|           |                                                               | Y | 4.96 | 69.4 | 19.5 |      | 130.7 |        |
|           |                                                               | Z | 5.05 | 69.3 | 19.1 |      | 140.9 |        |
| 10404-AAB | CDMA2000 (1xEV-DO, Rev. A)                                    | X | 4.70 | 67.2 | 18.1 | 3.77 | 138.1 | ±0.7 % |
|           |                                                               | Y | 4.85 | 69.5 | 19.6 |      | 129.6 |        |
|           |                                                               | Z | 5.14 | 70.1 | 19.5 |      | 139.3 |        |
| 10415-AAA | IEEE 802.11b WiFi 2.4 GHz (DSSS, 1 Mbps, 99pc duty cycle)     | X | 2.47 | 66.1 | 17.1 | 1.54 | 133.2 | ±0.7 % |
|           |                                                               | Y | 3.15 | 72.2 | 20.9 |      | 127.9 |        |
|           |                                                               | Z | 3.32 | 72.0 | 20.1 |      | 137.2 |        |
| 10416-AAA | IEEE 802.11g WiFi 2.4 GHz (ERP-OFDM, 6 Mbps, 99pc duty cycle) | X | 9.99 | 68.4 | 21.0 | 8.23 | 131.6 | ±2.5 % |
|           |                                                               | Y | 9.84 | 68.6 | 21.4 |      | 123.3 |        |
|           |                                                               | Z | 9.89 | 68.6 | 21.1 |      | 133.4 |        |

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor  $k=2$ , which for a normal distribution corresponds to a coverage probability of approximately 95%.

<sup>A</sup> The uncertainties of NormX,Y,Z do not affect the  $E^2$ -field uncertainty inside TSL (see Pages 7 and 8).

<sup>B</sup> Numerical linearization parameter: uncertainty not required.

<sup>E</sup> Uncertainty is determined using the max. deviation from linear response applying rectangular distribution and is expressed for the square of the field value.

## DASY/EASY - Parameters of Probe: ES3DV3 - SN:3333

### Calibration Parameter Determined in Head Tissue Simulating Media

| f (MHz) <sup>C</sup> | Relative Permittivity <sup>F</sup> | Conductivity (S/m) <sup>F</sup> | ConvF X | ConvF Y | ConvF Z | Alpha <sup>G</sup> | Depth <sup>G</sup> (mm) | Unct. (k=2) |
|----------------------|------------------------------------|---------------------------------|---------|---------|---------|--------------------|-------------------------|-------------|
| 750                  | 41.9                               | 0.89                            | 6.55    | 6.55    | 6.55    | 0.34               | 1.74                    | ± 12.0 %    |
| 835                  | 41.5                               | 0.90                            | 6.33    | 6.33    | 6.33    | 0.44               | 1.48                    | ± 12.0 %    |
| 1750                 | 40.1                               | 1.37                            | 5.26    | 5.26    | 5.26    | 0.73               | 1.21                    | ± 12.0 %    |
| 1900                 | 40.0                               | 1.40                            | 5.11    | 5.11    | 5.11    | 0.66               | 1.32                    | ± 12.0 %    |
| 2450                 | 39.2                               | 1.80                            | 4.53    | 4.53    | 4.53    | 0.62               | 1.40                    | ± 12.0 %    |
| 2600                 | 39.0                               | 1.96                            | 4.40    | 4.40    | 4.40    | 0.68               | 1.38                    | ± 12.0 %    |

<sup>C</sup> Frequency validity above 300 MHz of ± 100 MHz only applies for DASY v4.4 and higher (see Page 2), else it is restricted to ± 50 MHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below 300 MHz is ± 10, 25, 40, 50 and 70 MHz for ConvF assessments at 30, 64, 128, 150 and 220 MHz respectively. Above 5 GHz frequency validity can be extended to ± 110 MHz.

<sup>F</sup> At frequencies below 3 GHz, the validity of tissue parameters ( $\epsilon$  and  $\sigma$ ) can be relaxed to ± 10% if liquid compensation formula is applied to measured SAR values. At frequencies above 3 GHz, the validity of tissue parameters ( $\epsilon$  and  $\sigma$ ) is restricted to ± 5%. The uncertainty is the RSS of the ConvF uncertainty for indicated target tissue parameters.

<sup>G</sup> Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than ± 1% for frequencies below 3 GHz and below ± 2% for frequencies between 3-6 GHz at any distance larger than half the probe tip diameter from the boundary.

## DASY/EASY - Parameters of Probe: ES3DV3 - SN:3333

### Calibration Parameter Determined In Body Tissue Simulating Media


| f (MHz) <sup>C</sup> | Relative Permittivity <sup>F</sup> | Conductivity (S/m) <sup>F</sup> | ConvF X | ConvF Y | ConvF Z | Alpha <sup>G</sup> | Depth <sup>G</sup> (mm) | Uct. (k=2) |
|----------------------|------------------------------------|---------------------------------|---------|---------|---------|--------------------|-------------------------|------------|
| 750                  | 55.5                               | 0.96                            | 6.14    | 6.14    | 6.14    | 0.35               | 1.76                    | ± 12.0 %   |
| 835                  | 55.2                               | 0.97                            | 6.12    | 6.12    | 6.12    | 0.57               | 1.37                    | ± 12.0 %   |
| 1750                 | 53.4                               | 1.49                            | 4.89    | 4.89    | 4.89    | 0.80               | 1.24                    | ± 12.0 %   |
| 1900                 | 53.3                               | 1.52                            | 4.67    | 4.67    | 4.67    | 0.75               | 1.29                    | ± 12.0 %   |
| 2450                 | 52.7                               | 1.95                            | 4.26    | 4.26    | 4.26    | 0.80               | 1.01                    | ± 12.0 %   |
| 2600                 | 52.5                               | 2.16                            | 4.13    | 4.13    | 4.13    | 0.80               | 0.99                    | ± 12.0 %   |

<sup>C</sup> Frequency validity above 300 MHz of ± 100 MHz only applies for DASY v4.4 and higher (see Page 2), else it is restricted to ± 50 MHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below 300 MHz is ± 10, 25, 40, 50 and 70 MHz for ConvF assessments at 30, 64, 128, 150 and 220 MHz respectively. Above 5 GHz frequency validity can be extended to ± 110 MHz.

<sup>F</sup> At frequencies below 3 GHz, the validity of tissue parameters ( $\epsilon$  and  $\sigma$ ) can be relaxed to ± 10% if liquid compensation formula is applied to measured SAR values. At frequencies above 3 GHz, the validity of tissue parameters ( $\epsilon$  and  $\sigma$ ) is restricted to ± 5%. The uncertainty is the RSS of the ConvF uncertainty for indicated target tissue parameters.

<sup>G</sup> Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than ± 1% for frequencies below 3 GHz and below ± 2% for frequencies between 3-6 GHz at any distance larger than half the probe tip diameter from the boundary.


## Frequency Response of E-Field (TEM-Cell: Ifi110 EXX, Waveguide: R22)





Uncertainty of Frequency Response of E-field:  $\pm 6.3\%$  ( $k=2$ )

## Receiving Pattern ( $\phi$ ), $\theta = 0^\circ$


f=600 MHz,TEM



f=1800 MHz,R22


Uncertainty of Axial Isotropy Assessment:  $\pm 0.5\%$  (k=2)

# Dynamic Range $f(\text{SAR}_{\text{head}})$ (TEM cell, $f_{\text{eval}} = 1900 \text{ MHz}$ )





Uncertainty of Linearity Assessment:  $\pm 0.6\%$  ( $k=2$ )

## Conversion Factor Assessment



## Deviation from Isotropy in Liquid

Error ( $\phi, \theta$ ),  $f = 900 \text{ MHz}$



Uncertainty of Spherical Isotropy Assessment:  $\pm 2.6\%$  ( $k=2$ )

**DASY/EASY - Parameters of Probe: ES3DV3 - SN:3333****Other Probe Parameters**

|                                               |            |
|-----------------------------------------------|------------|
| Sensor Arrangement                            | Triangular |
| Connector Angle (°)                           | -34.9      |
| Mechanical Surface Detection Mode             | enabled    |
| Optical Surface Detection Mode                | disabled   |
| Probe Overall Length                          | 337 mm     |
| Probe Body Diameter                           | 10 mm      |
| Tip Length                                    | 10 mm      |
| Tip Diameter                                  | 4 mm       |
| Probe Tip to Sensor X Calibration Point       | 2 mm       |
| Probe Tip to Sensor Y Calibration Point       | 2 mm       |
| Probe Tip to Sensor Z Calibration Point       | 2 mm       |
| Recommended Measurement Distance from Surface | 3 mm       |

**Calibration Laboratory of**  
**Schmid & Partner**  
**Engineering AG**  
 Zeughausstrasse 43, 8004 Zurich, Switzerland



**S** Schweizerischer Kalibrierdienst  
**C** Service suisse d'étalonnage  
**S** Servizio svizzero di taratura  
 Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)  
 The Swiss Accreditation Service is one of the signatories to the EA  
 Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 108**

Client **PC Test**

Certificate No: **ES3-3334\_Dec14**

## CALIBRATION CERTIFICATE

Object **ES3DV3 - SN:3334**

Calibration procedure(s) **QA CAL-01.v9, QA CAL-23.v5, QA CAL-25.v6**  
 Calibration procedure for dosimetric E-field probes

CC  
 12/31/14

Calibration date: **December 16, 2014**

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI).  
 The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature  $(22 \pm 3)^{\circ}\text{C}$  and humidity  $< 70\%$ .

Calibration Equipment used (M&TE critical for calibration)

| Primary Standards          | ID              | Cal Date (Certificate No.)        | Scheduled Calibration  |
|----------------------------|-----------------|-----------------------------------|------------------------|
| Power meter E4419B         | GB41293874      | 03-Apr-14 (No. 217-01911)         | Apr-15                 |
| Power sensor E4412A        | MY41498087      | 03-Apr-14 (No. 217-01911)         | Apr-15                 |
| Reference 3 dB Attenuator  | SN: S5054 (3c)  | 03-Apr-14 (No. 217-01915)         | Apr-15                 |
| Reference 20 dB Attenuator | SN: S5277 (20x) | 03-Apr-14 (No. 217-01919)         | Apr-15                 |
| Reference 30 dB Attenuator | SN: S5129 (30b) | 03-Apr-14 (No. 217-01920)         | Apr-15                 |
| Reference Probe ES3DV2     | SN: 3013        | 30-Dec-13 (No. ES3-3013_Dec13)    | Dec-14                 |
| DAE4                       | SN: 789         | 30-Apr-14 (No. DAE4-789_Apr14)    | Apr-15                 |
| Secondary Standards        | ID              | Check Date (in house)             | Scheduled Check        |
| RF generator HP 8648C      | US3642U01700    | 4-Aug-99 (in house check Apr-13)  | In house check: Apr-16 |
| Network Analyzer HP 8753E  | US37390585      | 18-Oct-01 (in house check Oct-14) | In house check: Oct-15 |

|                |                      |                                   |                           |
|----------------|----------------------|-----------------------------------|---------------------------|
| Calibrated by: | Name<br>Leif Klysner | Function<br>Laboratory Technician | Signature<br>             |
| Approved by:   | Katja Pokovic        | Technical Manager                 |                           |
|                |                      |                                   | Issued: December 16, 2014 |

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.



Accredited by the Swiss Accreditation Service (SAS)  
The Swiss Accreditation Service is one of the signatories to the EA  
Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 108**

### Glossary:

|                          |                                                                                                                                                         |
|--------------------------|---------------------------------------------------------------------------------------------------------------------------------------------------------|
| TSL                      | tissue simulating liquid                                                                                                                                |
| NORM <sub>x,y,z</sub>    | sensitivity in free space                                                                                                                               |
| ConvF                    | sensitivity in TSL / NORM <sub>x,y,z</sub>                                                                                                              |
| DCP                      | diode compression point                                                                                                                                 |
| CF                       | crest factor (1/duty_cycle) of the RF signal                                                                                                            |
| A, B, C, D               | modulation dependent linearization parameters                                                                                                           |
| Polarization $\phi$      | $\phi$ rotation around probe axis                                                                                                                       |
| Polarization $\vartheta$ | $\vartheta$ rotation around an axis that is in the plane normal to probe axis (at measurement center),<br>i.e., $\vartheta = 0$ is normal to probe axis |
| Connector Angle          | Information used in DASY system to align probe sensor X to the robot coordinate system                                                                  |

### Calibration is Performed According to the Following Standards:

- IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013
- IEC 62209-1, "Procedure to measure the Specific Absorption Rate (SAR) for hand-held devices used in close proximity to the ear (frequency range of 300 MHz to 3 GHz)", February 2005

### Methods Applied and Interpretation of Parameters:

- NORM<sub>x,y,z</sub>:** Assessed for E-field polarization  $\vartheta = 0$  ( $f \leq 900$  MHz in TEM-cell;  $f > 1800$  MHz: R22 waveguide). NORM<sub>x,y,z</sub> are only intermediate values, i.e., the uncertainties of NORM<sub>x,y,z</sub> does not affect the E<sup>2</sup>-field uncertainty inside TSL (see below ConvF).
- NORM(f)<sub>x,y,z</sub> = NORM<sub>x,y,z</sub> \* frequency\_response** (see Frequency Response Chart). This linearization is implemented in DASY4 software versions later than 4.2. The uncertainty of the frequency response is included in the stated uncertainty of ConvF.
- DCP<sub>x,y,z</sub>:** DCP are numerical linearization parameters assessed based on the data of power sweep with CW signal (no uncertainty required). DCP does not depend on frequency nor media.
- PAR:** PAR is the Peak to Average Ratio that is not calibrated but determined based on the signal characteristics
- A<sub>x,y,z</sub>; B<sub>x,y,z</sub>; C<sub>x,y,z</sub>; D<sub>x,y,z</sub>; VR<sub>x,y,z</sub>:** A, B, C, D are numerical linearization parameters assessed based on the data of power sweep for specific modulation signal. The parameters do not depend on frequency nor media. VR is the maximum calibration range expressed in RMS voltage across the diode.
- ConvF and Boundary Effect Parameters:** Assessed in flat phantom using E-field (or Temperature Transfer Standard for  $f \leq 800$  MHz) and inside waveguide using analytical field distributions based on power measurements for  $f > 800$  MHz. The same setups are used for assessment of the parameters applied for boundary compensation (alpha, depth) of which typical uncertainty values are given. These parameters are used in DASY4 software to improve probe accuracy close to the boundary. The sensitivity in TSL corresponds to NORM<sub>x,y,z</sub> \* ConvF whereby the uncertainty corresponds to that given for ConvF. A frequency dependent ConvF is used in DASY version 4.4 and higher which allows extending the validity from  $\pm 50$  MHz to  $\pm 100$  MHz.
- Spherical Isotropy (3D deviation from isotropy):** in a field of low gradients realized using a flat phantom exposed by a patch antenna.
- Sensor Offset:** The sensor offset corresponds to the offset of virtual measurement center from the probe tip (on probe axis). No tolerance required.
- Connector Angle:** The angle is assessed using the information gained by determining the NORM<sub>x</sub> (no uncertainty required).

# Probe ES3DV3

## SN:3334

|               |                   |
|---------------|-------------------|
| Manufactured: | January 24, 2012  |
| Repaired:     | December 9, 2014  |
| Calibrated:   | December 16, 2014 |

Calibrated for DASY/EASY Systems  
(Note: non-compatible with DASY2 system!)

## DASY/EASY - Parameters of Probe: ES3DV3 - SN:3334

### Basic Calibration Parameters

|                                                           | Sensor X | Sensor Y | Sensor Z | Unc (k=2)     |
|-----------------------------------------------------------|----------|----------|----------|---------------|
| Norm ( $\mu\text{V}/(\text{V}/\text{m})^2$ ) <sup>A</sup> | 1.04     | 1.05     | 1.01     | $\pm 10.1 \%$ |
| DCP (mV) <sup>B</sup>                                     | 106.5    | 105.0    | 105.6    |               |

### Modulation Calibration Parameters

| UID       | Communication System Name                     |   | A<br>dB | B<br>dB $\sqrt{\mu\text{V}}$ | C    | D<br>dB | VR<br>mV | Unc <sup>E</sup><br>(k=2) |
|-----------|-----------------------------------------------|---|---------|------------------------------|------|---------|----------|---------------------------|
| 0         | CW                                            | X | 0.0     | 0.0                          | 1.0  | 0.00    | 188.0    | $\pm 3.0 \%$              |
|           |                                               | Y | 0.0     | 0.0                          | 1.0  |         | 183.2    |                           |
|           |                                               | Z | 0.0     | 0.0                          | 1.0  |         | 181.8    |                           |
| 10010-CAA | SAR Validation (Square, 100ms, 10ms)          | X | 4.61    | 67.2                         | 13.7 | 10.00   | 38.4     | $\pm 1.4 \%$              |
|           |                                               | Y | 20.36   | 82.7                         | 18.7 |         | 38.0     |                           |
|           |                                               | Z | 17.55   | 80.3                         | 17.6 |         | 37.0     |                           |
| 10011-CAB | UMTS-FDD (WCDMA)                              | X | 3.56    | 68.4                         | 19.1 | 2.91    | 148.4    | $\pm 0.7 \%$              |
|           |                                               | Y | 3.44    | 68.1                         | 19.2 |         | 146.9    |                           |
|           |                                               | Z | 3.52    | 68.3                         | 19.1 |         | 144.7    |                           |
| 10012-CAB | IEEE 802.11b WIFI 2.4 GHz (DSSS, 1 Mbps)      | X | 3.54    | 71.9                         | 20.0 | 1.87    | 148.0    | $\pm 0.7 \%$              |
|           |                                               | Y | 3.51    | 72.2                         | 20.5 |         | 148.9    |                           |
|           |                                               | Z | 3.80    | 73.3                         | 20.6 |         | 144.6    |                           |
| 10013-CAB | IEEE 802.11g WIFI 2.4 GHz (DSSS-OFDM, 6 Mbps) | X | 11.39   | 71.1                         | 23.3 | 9.46    | 149.8    | $\pm 3.8 \%$              |
|           |                                               | Y | 11.54   | 71.8                         | 24.0 |         | 149.5    |                           |
|           |                                               | Z | 11.11   | 70.5                         | 23.0 |         | 141.6    |                           |
| 10021-DAB | GSM-FDD (TDMA, GMSK)                          | X | 15.29   | 91.3                         | 25.0 | 9.39    | 131.9    | $\pm 1.7 \%$              |
|           |                                               | Y | 24.16   | 100.0                        | 28.4 |         | 142.8    |                           |
|           |                                               | Z | 13.05   | 89.2                         | 24.5 |         | 126.5    |                           |
| 10023-DAB | GPRS-FDD (TDMA, GMSK, TN 0)                   | X | 16.07   | 91.7                         | 25.1 | 9.57    | 144.0    | $\pm 2.2 \%$              |
|           |                                               | Y | 19.00   | 95.3                         | 26.8 |         | 136.4    |                           |
|           |                                               | Z | 13.93   | 89.8                         | 24.6 |         | 141.0    |                           |
| 10024-DAB | GPRS-FDD (TDMA, GMSK, TN 0-1)                 | X | 19.98   | 91.0                         | 22.4 | 6.56    | 134.2    | $\pm 1.9 \%$              |
|           |                                               | Y | 34.78   | 99.7                         | 25.5 |         | 145.0    |                           |
|           |                                               | Z | 29.89   | 96.8                         | 24.1 |         | 129.8    |                           |
| 10027-DAB | GPRS-FDD (TDMA, GMSK, TN 0-1-2)               | X | 56.30   | 99.7                         | 22.8 | 4.80    | 125.2    | $\pm 1.9 \%$              |
|           |                                               | Y | 41.16   | 99.6                         | 23.9 |         | 131.2    |                           |
|           |                                               | Z | 50.78   | 99.8                         | 23.1 |         | 147.6    |                           |
| 10028-DAB | GPRS-FDD (TDMA, GMSK, TN 0-1-2-3)             | X | 49.35   | 99.7                         | 22.5 | 3.55    | 133.2    | $\pm 2.2 \%$              |
|           |                                               | Y | 46.49   | 99.6                         | 22.9 |         | 139.2    |                           |
|           |                                               | Z | 58.21   | 99.7                         | 22.0 |         | 129.4    |                           |
| 10032-CAA | IEEE 802.15.1 Bluetooth (GFSK, DH5)           | X | 56.54   | 100.0                        | 20.2 | 1.16    | 128.0    | $\pm 1.7 \%$              |
|           |                                               | Y | 20.03   | 99.3                         | 22.4 |         | 130.3    |                           |
|           |                                               | Z | 84.01   | 100.0                        | 19.4 |         | 141.0    |                           |
| 10100-CAB | LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)      | X | 6.44    | 67.6                         | 19.6 | 5.67    | 138.5    | $\pm 1.4 \%$              |
|           |                                               | Y | 6.50    | 67.9                         | 20.0 |         | 142.1    |                           |
|           |                                               | Z | 6.31    | 67.2                         | 19.4 |         | 129.4    |                           |

|           |                                          |   |       |      |      |      |       |        |
|-----------|------------------------------------------|---|-------|------|------|------|-------|--------|
| 10103-CAB | LTE-TDD (SC-FDMA, 100% RB, 20 MHz, QPSK) | X | 9.77  | 73.6 | 24.6 | 9.29 | 129.6 | ±3.3 % |
|           |                                          | Y | 10.52 | 76.0 | 26.3 |      | 132.1 |        |
|           |                                          | Z | 10.21 | 75.0 | 25.4 |      | 147.7 |        |
| 10108-CAC | LTE-FDD (SC-FDMA, 100% RB, 10 MHz, QPSK) | X | 6.36  | 67.2 | 19.6 | 5.80 | 136.8 | ±1.4 % |
|           |                                          | Y | 6.31  | 67.3 | 19.8 |      | 137.2 |        |
|           |                                          | Z | 6.20  | 66.7 | 19.3 |      | 128.8 |        |
| 10117-CAB | IEEE 802.11n (HT Mixed, 13.5 Mbps, BPSK) | X | 9.96  | 68.3 | 20.8 | 8.07 | 126.5 | ±2.5 % |
|           |                                          | Y | 10.12 | 68.8 | 21.3 |      | 126.6 |        |
|           |                                          | Z | 10.22 | 69.0 | 21.2 |      | 143.7 |        |
| 10151-CAB | LTE-TDD (SC-FDMA, 50% RB, 20 MHz, QPSK)  | X | 9.29  | 73.0 | 24.4 | 9.28 | 125.3 | ±3.3 % |
|           |                                          | Y | 9.65  | 74.5 | 25.6 |      | 124.4 |        |
|           |                                          | Z | 9.65  | 74.3 | 25.2 |      | 141.1 |        |
| 10154-CAC | LTE-FDD (SC-FDMA, 50% RB, 10 MHz, QPSK)  | X | 6.03  | 66.7 | 19.3 | 5.75 | 132.7 | ±1.4 % |
|           |                                          | Y | 5.97  | 66.7 | 19.5 |      | 132.7 |        |
|           |                                          | Z | 6.17  | 67.3 | 19.7 |      | 148.3 |        |
| 10160-CAB | LTE-FDD (SC-FDMA, 50% RB, 15 MHz, QPSK)  | X | 6.47  | 67.2 | 19.5 | 5.82 | 138.1 | ±1.4 % |
|           |                                          | Y | 6.44  | 67.3 | 19.8 |      | 138.2 |        |
|           |                                          | Z | 6.27  | 66.6 | 19.2 |      | 126.8 |        |
| 10169-CAB | LTE-FDD (SC-FDMA, 1 RB, 20 MHz, QPSK)    | X | 5.03  | 66.9 | 19.6 | 5.73 | 137.2 | ±1.2 % |
|           |                                          | Y | 4.97  | 67.0 | 19.9 |      | 135.7 |        |
|           |                                          | Z | 4.91  | 66.5 | 19.5 |      | 127.1 |        |
| 10172-CAB | LTE-TDD (SC-FDMA, 1 RB, 20 MHz, QPSK)    | X | 8.53  | 77.4 | 26.9 | 9.21 | 142.4 | ±2.7 % |
|           |                                          | Y | 9.59  | 81.3 | 29.3 |      | 142.3 |        |
|           |                                          | Z | 7.78  | 75.0 | 25.7 |      | 126.7 |        |
| 10175-CAC | LTE-FDD (SC-FDMA, 1 RB, 10 MHz, QPSK)    | X | 5.02  | 67.0 | 19.7 | 5.72 | 131.8 | ±1.2 % |
|           |                                          | Y | 4.98  | 67.0 | 19.9 |      | 136.1 |        |
|           |                                          | Z | 4.95  | 66.8 | 19.6 |      | 128.1 |        |
| 10181-CAB | LTE-FDD (SC-FDMA, 1 RB, 15 MHz, QPSK)    | X | 4.99  | 66.8 | 19.6 | 5.72 | 131.2 | ±1.2 % |
|           |                                          | Y | 4.99  | 67.1 | 20.0 |      | 136.2 |        |
|           |                                          | Z | 4.92  | 66.6 | 19.5 |      | 127.9 |        |
| 10196-CAB | IEEE 802.11n (HT Mixed, 6.5 Mbps, BPSK)  | X | 9.98  | 68.8 | 21.2 | 8.10 | 141.7 | ±2.5 % |
|           |                                          | Y | 10.14 | 69.5 | 21.8 |      | 147.2 |        |
|           |                                          | Z | 9.85  | 68.6 | 21.1 |      | 137.5 |        |
| 10225-CAB | UMTS-FDD (HSPA+)                         | X | 7.17  | 67.5 | 19.6 | 5.97 | 146.0 | ±1.4 % |
|           |                                          | Y | 7.13  | 67.7 | 19.9 |      | 149.9 |        |
|           |                                          | Z | 7.12  | 67.5 | 19.6 |      | 142.9 |        |
| 10237-CAB | LTE-TDD (SC-FDMA, 1 RB, 10 MHz, QPSK)    | X | 8.29  | 76.6 | 26.5 | 9.21 | 136.1 | ±2.7 % |
|           |                                          | Y | 9.60  | 81.4 | 29.3 |      | 142.3 |        |
|           |                                          | Z | 7.98  | 75.8 | 26.1 |      | 132.9 |        |
| 10252-CAB | LTE-TDD (SC-FDMA, 50% RB, 10 MHz, QPSK)  | X | 9.27  | 74.1 | 25.1 | 9.24 | 139.1 | ±3.3 % |
|           |                                          | Y | 10.25 | 77.5 | 27.4 |      | 146.3 |        |
|           |                                          | Z | 9.07  | 73.7 | 25.0 |      | 135.8 |        |
| 10267-CAB | LTE-TDD (SC-FDMA, 100% RB, 10 MHz, QPSK) | X | 9.95  | 74.9 | 25.4 | 9.30 | 147.0 | ±3.3 % |
|           |                                          | Y | 9.80  | 75.0 | 25.9 |      | 125.9 |        |
|           |                                          | Z | 9.74  | 74.6 | 25.4 |      | 143.8 |        |

|           |                                                               |   |       |      |      |      |       |        |
|-----------|---------------------------------------------------------------|---|-------|------|------|------|-------|--------|
| 10275-CAB | UMTS-FDD (HSUPA, Sublest 5, 3GPP Rel8.4)                      | X | 4.63  | 67.6 | 19.0 | 3.96 | 147.5 | ±0.7 % |
|           |                                                               | Y | 4.41  | 66.9 | 18.9 |      | 129.5 |        |
|           |                                                               | Z | 4.61  | 67.6 | 19.1 |      | 148.1 |        |
| 10291-AAB | CDMA2000, RC3, SO55, Full Rate                                | X | 3.83  | 67.7 | 19.0 | 3.46 | 133.7 | ±0.7 % |
|           |                                                               | Y | 3.71  | 67.4 | 19.0 |      | 139.0 |        |
|           |                                                               | Z | 3.86  | 68.1 | 19.2 |      | 133.7 |        |
| 10292-AAB | CDMA2000, RC3, SO32, Full Rate                                | X | 3.85  | 68.2 | 19.2 | 3.39 | 136.7 | ±0.5 % |
|           |                                                               | Y | 3.67  | 67.5 | 19.1 |      | 141.3 |        |
|           |                                                               | Z | 3.75  | 67.8 | 19.0 |      | 136.2 |        |
| 10297-AAA | LTE-FDD (SC-FDMA, 50% RB, 20 MHz, QPSK)                       | X | 6.31  | 67.1 | 19.5 | 5.81 | 130.6 | ±1.4 % |
|           |                                                               | Y | 6.32  | 67.3 | 19.8 |      | 135.1 |        |
|           |                                                               | Z | 6.24  | 66.9 | 19.4 |      | 129.2 |        |
| 10311-AAA | LTE-FDD (SC-FDMA, 100% RB, 15 MHz, QPSK)                      | X | 6.85  | 67.5 | 19.8 | 6.06 | 135.1 | ±1.4 % |
|           |                                                               | Y | 6.90  | 67.9 | 20.2 |      | 141.5 |        |
|           |                                                               | Z | 6.82  | 67.5 | 19.8 |      | 135.1 |        |
| 10403-AAB | CDMA2000 (1xEV-DO, Rev. 0)                                    | X | 5.04  | 69.1 | 19.1 | 3.76 | 126.0 | ±0.5 % |
|           |                                                               | Y | 4.90  | 69.0 | 19.3 |      | 129.6 |        |
|           |                                                               | Z | 5.11  | 69.7 | 19.4 |      | 125.8 |        |
| 10404-AAB | CDMA2000 (1xEV-DO, Rev. A)                                    | X | 5.05  | 69.6 | 19.4 | 3.77 | 147.1 | ±0.7 % |
|           |                                                               | Y | 4.84  | 69.2 | 19.5 |      | 127.8 |        |
|           |                                                               | Z | 5.15  | 70.1 | 19.6 |      | 143.3 |        |
| 10415-AAA | IEEE 802.11b WIFI 2.4 GHz (DSSS, 1 Mbps, 99pc duty cycle)     | X | 3.13  | 71.2 | 19.9 | 1.54 | 144.5 | ±0.5 % |
|           |                                                               | Y | 2.93  | 70.4 | 19.9 |      | 149.8 |        |
|           |                                                               | Z | 3.18  | 71.6 | 20.1 |      | 141.4 |        |
| 10416-AAA | IEEE 802.11g WIFI 2.4 GHz (ERP-OFDM, 6 Mbps, 99pc duty cycle) | X | 10.11 | 69.0 | 21.4 | 8.23 | 144.3 | ±2.5 % |
|           |                                                               | Y | 10.21 | 69.6 | 21.9 |      | 148.3 |        |
|           |                                                               | Z | 9.99  | 68.9 | 21.3 |      | 141.1 |        |

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor  $k=2$ , which for a normal distribution corresponds to a coverage probability of approximately 95%.

<sup>A</sup> The uncertainties of NormX,Y,Z do not affect the  $E^2$ -field uncertainty inside TSL (see Pages 7 and 8).

<sup>B</sup> Numerical linearization parameter: uncertainty not required.

<sup>E</sup> Uncertainty is determined using the max. deviation from linear response applying rectangular distribution and is expressed for the square of the field value.

## DASY/EASY - Parameters of Probe: ES3DV3 - SN:3334

### Calibration Parameter Determined in Head Tissue Simulating Media

| f (MHz) <sup>c</sup> | Relative Permittivity <sup>F</sup> | Conductivity (S/m) <sup>F</sup> | ConvF X | ConvF Y | ConvF Z | Alpha <sup>G</sup> | Depth <sup>G</sup> (mm) | Unct. (k=2) |
|----------------------|------------------------------------|---------------------------------|---------|---------|---------|--------------------|-------------------------|-------------|
| 750                  | 41.9                               | 0.89                            | 6.51    | 6.51    | 6.51    | 0.80               | 1.17                    | ± 12.0 %    |
| 835                  | 41.5                               | 0.90                            | 6.25    | 6.25    | 6.25    | 0.38               | 1.58                    | ± 12.0 %    |
| 1750                 | 40.1                               | 1.37                            | 5.21    | 5.21    | 5.21    | 0.43               | 1.63                    | ± 12.0 %    |
| 1900                 | 40.0                               | 1.40                            | 5.03    | 5.03    | 5.03    | 0.53               | 1.45                    | ± 12.0 %    |
| 2450                 | 39.2                               | 1.80                            | 4.51    | 4.51    | 4.51    | 0.80               | 1.26                    | ± 12.0 %    |
| 2600                 | 39.0                               | 1.96                            | 4.31    | 4.31    | 4.31    | 0.79               | 1.27                    | ± 12.0 %    |

<sup>c</sup> Frequency validity above 300 MHz of ± 100 MHz only applies for DASY v4.4 and higher (see Page 2), else it is restricted to ± 50 MHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below 300 MHz is ± 10, 25, 40, 50 and 70 MHz for ConvF assessments at 30, 64, 128, 150 and 220 MHz respectively. Above 5 GHz frequency validity can be extended to ± 110 MHz.

<sup>F</sup> At frequencies below 3 GHz, the validity of tissue parameters ( $\epsilon$  and  $\sigma$ ) can be relaxed to ± 10% if liquid compensation formula is applied to measured SAR values. At frequencies above 3 GHz, the validity of tissue parameters ( $\epsilon$  and  $\sigma$ ) is restricted to ± 5%. The uncertainty is the RSS of the ConvF uncertainty for indicated target tissue parameters.

<sup>G</sup> Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than ± 1% for frequencies below 3 GHz and below ± 2% for frequencies between 3-6 GHz at any distance larger than half the probe tip diameter from the boundary.

## DASY/EASY - Parameters of Probe: ES3DV3 - SN:3334

### Calibration Parameter Determined in Body Tissue Simulating Media

| f (MHz) <sup>c</sup> | Relative Permittivity <sup>f</sup> | Conductivity (S/m) <sup>f</sup> | ConvF X | ConvF Y | ConvF Z | Alpha <sup>g</sup> | Depth <sup>d</sup> (mm) | Unct. (k=2) |
|----------------------|------------------------------------|---------------------------------|---------|---------|---------|--------------------|-------------------------|-------------|
| 750                  | 55.5                               | 0.96                            | 6.09    | 6.09    | 6.09    | 0.49               | 1.47                    | ± 12.0 %    |
| 835                  | 55.2                               | 0.97                            | 6.14    | 6.14    | 6.14    | 0.69               | 1.27                    | ± 12.0 %    |
| 1750                 | 53.4                               | 1.49                            | 4.94    | 4.94    | 4.94    | 0.80               | 1.24                    | ± 12.0 %    |
| 1900                 | 53.3                               | 1.52                            | 4.73    | 4.73    | 4.73    | 0.62               | 1.44                    | ± 12.0 %    |
| 2450                 | 52.7                               | 1.95                            | 4.28    | 4.28    | 4.28    | 0.80               | 1.13                    | ± 12.0 %    |
| 2600                 | 52.5                               | 2.16                            | 4.16    | 4.16    | 4.16    | 0.75               | 1.25                    | ± 12.0 %    |


<sup>c</sup> Frequency validity above 300 MHz of ± 100 MHz only applies for DASY v4.4 and higher (see Page 2), else it is restricted to ± 50 MHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below 300 MHz is ± 10, 25, 40, 50 and 70 MHz for ConvF assessments at 30, 64, 128, 150 and 220 MHz respectively. Above 5 GHz frequency validity can be extended to ± 110 MHz.

<sup>f</sup> At frequencies below 3 GHz, the validity of tissue parameters ( $\epsilon$  and  $\sigma$ ) can be relaxed to ± 10% if liquid compensation formula is applied to measured SAR values. At frequencies above 3 GHz, the validity of tissue parameters ( $\epsilon$  and  $\sigma$ ) is restricted to ± 5%. The uncertainty is the RSS of the ConvF uncertainty for indicated target tissue parameters.

<sup>g</sup> Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than ± 1% for frequencies below 3 GHz and below ± 2% for frequencies between 3-6 GHz at any distance larger than half the probe tip diameter from the boundary.

## Frequency Response of E-Field


(TEM-Cell: Ifi110 EXX, Waveguide: R22)





Uncertainty of Frequency Response of E-field:  $\pm 6.3\%$  ( $k=2$ )

Receiving Pattern ( $\phi$ ),  $\theta = 0^\circ$

f=600 MHz,TEM




f=1800 MHz,R22




Uncertainty of Axial Isotropy Assessment:  $\pm 0.5\%$  ( $k=2$ )

### Dynamic Range $f(\text{SAR}_{\text{head}})$ (TEM cell, $f_{\text{eval}} = 1900 \text{ MHz}$ )





Uncertainty of Linearity Assessment:  $\pm 0.6\%$  ( $k=2$ )

## Conversion Factor Assessment



## Deviation from Isotropy in Liquid

Error ( $\phi$ ,  $\theta$ ),  $f = 900 \text{ MHz}$



**DASY/EASY - Parameters of Probe: ES3DV3 - SN:3334****Other Probe Parameters**

|                                               |            |
|-----------------------------------------------|------------|
| Sensor Arrangement                            | Triangular |
| Connector Angle (°)                           | 18.5       |
| Mechanical Surface Detection Mode             | enabled    |
| Optical Surface Detection Mode                | disabled   |
| Probe Overall Length                          | 337 mm     |
| Probe Body Diameter                           | 10 mm      |
| Tip Length                                    | 10 mm      |
| Tip Diameter                                  | 4 mm       |
| Probe Tip to Sensor X Calibration Point       | 2 mm       |
| Probe Tip to Sensor Y Calibration Point       | 2 mm       |
| Probe Tip to Sensor Z Calibration Point       | 2 mm       |
| Recommended Measurement Distance from Surface | 3 mm       |

**Calibration Laboratory of  
Schmid & Partner  
Engineering AG**  
Zeughausstrasse 43, 8004 Zurich, Switzerland



**S** Schweizerischer Kalibrierdienst  
**C** Service suisse d'étalonnage  
**S** Servizio svizzero di taratura  
Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)  
The Swiss Accreditation Service is one of the signatories to the EA  
Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 108**

Client **PC Test**

Certificate No: **ES3-3332\_Sep14/2**

## CALIBRATION CERTIFICATE (Replacement of No: ES3-3332\_Sep14)

Object **ES3DV3 - SN:3332**

Calibration procedure(s) **QA CAL-01.v9, QA CAL-23.v5, QA CAL-25.v6**  
Calibration procedure for dosimetric E-field probes

CC  
12/14/14

Calibration date: **September 18, 2014**

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI).  
The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature  $(22 \pm 3)^{\circ}\text{C}$  and humidity  $< 70\%$ .

Calibration Equipment used (M&TE critical for calibration)

| Primary Standards          | ID              | Cal Date (Certificate No.)        | Scheduled Calibration  |
|----------------------------|-----------------|-----------------------------------|------------------------|
| Power meter E4419B         | GB41293874      | 03-Apr-14 (No. 217-01911)         | Apr-15                 |
| Power sensor E4412A        | MY41498087      | 03-Apr-14 (No. 217-01911)         | Apr-15                 |
| Reference 3 dB Attenuator  | SN: S5054 (3c)  | 03-Apr-14 (No. 217-01915)         | Apr-15                 |
| Reference 20 dB Attenuator | SN: S5277 (20x) | 03-Apr-14 (No. 217-01919)         | Apr-15                 |
| Reference 30 dB Attenuator | SN: S5129 (30b) | 03-Apr-14 (No. 217-01920)         | Apr-15                 |
| Reference Probe ES3DV2     | SN: 3013        | 30-Dec-13 (No. ES3-3013_Dec13)    | Dec-14                 |
| DAE4                       | SN: 660         | 13-Dec-13 (No. DAE4-660_Dec13)    | Dec-14                 |
| Secondary Standards        | ID              | Check Date (in house)             | Scheduled Check        |
| RF generator HP 8648C      | US3642U01700    | 4-Aug-99 (in house check Apr-13)  | In house check: Apr-16 |
| Network Analyzer HP 8753E  | US37390585      | 18-Oct-01 (in house check Oct-13) | In house check: Oct-14 |

|                                                                                                                 | Name           | Function              | Signature |
|-----------------------------------------------------------------------------------------------------------------|----------------|-----------------------|-----------|
| Calibrated by:                                                                                                  | Israe El-Naouq | Laboratory Technician |           |
| Approved by:                                                                                                    | Katja Pokovic  | Technical Manager     |           |
| Issued: November 3, 2014                                                                                        |                |                       |           |
| This calibration certificate shall not be reproduced except in full without written approval of the laboratory. |                |                       |           |



Accredited by the Swiss Accreditation Service (SAS)  
The Swiss Accreditation Service is one of the signatories to the EA  
Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 108**

### Glossary:

|                          |                                                                                                                                                         |
|--------------------------|---------------------------------------------------------------------------------------------------------------------------------------------------------|
| TSL                      | tissue simulating liquid                                                                                                                                |
| NORM <sub>x,y,z</sub>    | sensitivity in free space                                                                                                                               |
| ConvF                    | sensitivity in TSL / NORM <sub>x,y,z</sub>                                                                                                              |
| DCP                      | diode compression point                                                                                                                                 |
| CF                       | crest factor (1/duty_cycle) of the RF signal                                                                                                            |
| A, B, C, D               | modulation dependent linearization parameters                                                                                                           |
| Polarization $\varphi$   | $\varphi$ rotation around probe axis                                                                                                                    |
| Polarization $\vartheta$ | $\vartheta$ rotation around an axis that is in the plane normal to probe axis (at measurement center),<br>i.e., $\vartheta = 0$ is normal to probe axis |
| Connector Angle          | information used in DASY system to align probe sensor X to the robot coordinate system                                                                  |

### Calibration is Performed According to the Following Standards:

- IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013
- IEC 62209-1, "Procedure to measure the Specific Absorption Rate (SAR) for hand-held devices used in close proximity to the ear (frequency range of 300 MHz to 3 GHz)", February 2005

### Methods Applied and Interpretation of Parameters:

- NORM<sub>x,y,z</sub>**: Assessed for E-field polarization  $\vartheta = 0$  ( $f \leq 900$  MHz in TEM-cell;  $f > 1800$  MHz: R22 waveguide). NORM<sub>x,y,z</sub> are only intermediate values, i.e., the uncertainties of NORM<sub>x,y,z</sub> does not affect the E<sup>2</sup>-field uncertainty inside TSL (see below ConvF).
- NORM(f)<sub>x,y,z</sub>** = NORM<sub>x,y,z</sub> \* frequency\_response (see Frequency Response Chart). This linearization is implemented in DASY4 software versions later than 4.2. The uncertainty of the frequency response is included in the stated uncertainty of ConvF.
- DCP<sub>x,y,z</sub>**: DCP are numerical linearization parameters assessed based on the data of power sweep with CW signal (no uncertainty required). DCP does not depend on frequency nor media.
- PAR**: PAR is the Peak to Average Ratio that is not calibrated but determined based on the signal characteristics
- A<sub>x,y,z</sub>; B<sub>x,y,z</sub>; C<sub>x,y,z</sub>; D<sub>x,y,z</sub>; VR<sub>x,y,z</sub>**: A, B, C, D are numerical linearization parameters assessed based on the data of power sweep for specific modulation signal. The parameters do not depend on frequency nor media. VR is the maximum calibration range expressed in RMS voltage across the diode.
- ConvF and Boundary Effect Parameters**: Assessed in flat phantom using E-field (or Temperature Transfer Standard for  $f \leq 800$  MHz) and inside waveguide using analytical field distributions based on power measurements for  $f > 800$  MHz. The same setups are used for assessment of the parameters applied for boundary compensation (alpha, depth) of which typical uncertainty values are given. These parameters are used in DASY4 software to improve probe accuracy close to the boundary. The sensitivity in TSL corresponds to NORM<sub>x,y,z</sub> \* ConvF whereby the uncertainty corresponds to that given for ConvF. A frequency dependent ConvF is used in DASY version 4.4 and higher which allows extending the validity from  $\pm 50$  MHz to  $\pm 100$  MHz.
- Spherical Isotropy (3D deviation from Isotropy)**: in a field of low gradients realized using a flat phantom exposed by a patch antenna.
- Sensor Offset**: The sensor offset corresponds to the offset of virtual measurement center from the probe tip (on probe axis). No tolerance required.
- Connector Angle**: The angle is assessed using the information gained by determining the NORM<sub>x</sub> (no uncertainty required).

# Probe ES3DV3

## SN:3332

Manufactured: January 24, 2012  
Calibrated: September 18, 2014

Calibrated for DASY/EASY Systems  
(Note: non-compatible with DASY2 system!)

## DASY/EASY - Parameters of Probe: ES3DV3 - SN:3332

### Basic Calibration Parameters

|                                                           | Sensor X | Sensor Y | Sensor Z | Unc (k=2)     |
|-----------------------------------------------------------|----------|----------|----------|---------------|
| Norm ( $\mu\text{V}/(\text{V}/\text{m})^2$ ) <sup>A</sup> | 0.94     | 1.15     | 0.98     | $\pm 10.1 \%$ |
| DCP (mV) <sup>B</sup>                                     | 105.8    | 103.8    | 112.4    |               |

### Modulation Calibration Parameters

| UID       | Communication System Name                     |   | A<br>dB | B<br>dB $\sqrt{\mu\text{V}}$ | C    | D<br>dB | VR<br>mV | Unc <sup>C</sup><br>(k=2) |
|-----------|-----------------------------------------------|---|---------|------------------------------|------|---------|----------|---------------------------|
| 0         | CW                                            | X | 0.0     | 0.0                          | 1.0  | 0.00    | 178.7    | $\pm 3.0 \%$              |
|           |                                               | Y | 0.0     | 0.0                          | 1.0  |         | 199.5    |                           |
|           |                                               | Z | 0.0     | 0.0                          | 1.0  |         | 186.5    |                           |
| 10010-CAA | SAR Validation (Square, 100ms, 10ms)          | X | 55.60   | 92.4                         | 20.6 | 10.00   | 35.7     | $\pm 1.7 \%$              |
|           |                                               | Y | 2.80    | 61.2                         | 11.6 |         | 42.9     |                           |
|           |                                               | Z | 10.49   | 80.1                         | 18.0 |         | 36.1     |                           |
| 10011-CAB | UMTS-FDD (WCDMA)                              | X | 3.47    | 67.9                         | 18.8 | 2.91    | 141.3    | $\pm 0.7 \%$              |
|           |                                               | Y | 3.29    | 67.0                         | 18.4 |         | 138.2    |                           |
|           |                                               | Z | 3.78    | 70.4                         | 20.1 |         | 147.9    |                           |
| 10012-CAA | IEEE 802.11b WiFi 2.4 GHz (DSSS, 1 Mbps)      | X | 3.53    | 72.0                         | 20.1 | 1.87    | 141.7    | $\pm 0.7 \%$              |
|           |                                               | Y | 3.03    | 69.1                         | 18.8 |         | 141.1    |                           |
|           |                                               | Z | 4.06    | 75.5                         | 21.6 |         | 148.2    |                           |
| 10013-CAA | IEEE 802.11g WiFi 2.4 GHz (DSSS-OFDM, 6 Mbps) | X | 10.87   | 69.8                         | 22.6 | 9.46    | 137.3    | $\pm 3.5 \%$              |
|           |                                               | Y | 11.63   | 71.7                         | 23.9 |         | 141.9    |                           |
|           |                                               | Z | 10.51   | 69.6                         | 22.5 |         | 139.2    |                           |
| 10021-DAB | GSM-FDD (TDMA, GMSK)                          | X | 6.92    | 78.4                         | 20.1 | 9.39    | 137.0    | $\pm 2.5 \%$              |
|           |                                               | Y | 26.20   | 99.6                         | 27.8 |         | 141.5    |                           |
|           |                                               | Z | 5.13    | 78.3                         | 21.1 |         | 144.7    |                           |
| 10023-DAB | GPRS-FDD (TDMA, GMSK, TN 0)                   | X | 9.10    | 83.6                         | 22.5 | 9.57    | 144.0    | $\pm 2.5 \%$              |
|           |                                               | Y | 26.31   | 100.0                        | 28.1 |         | 136.7    |                           |
|           |                                               | Z | 6.15    | 81.6                         | 22.5 |         | 139.9    |                           |
| 10024-DAB | GPRS-FDD (TDMA, GMSK, TN 0-1)                 | X | 10.54   | 84.1                         | 20.4 | 6.56    | 141.8    | $\pm 2.5 \%$              |
|           |                                               | Y | 40.55   | 99.6                         | 24.9 |         | 142.2    |                           |
|           |                                               | Z | 6.45    | 81.5                         | 20.2 |         | 145.7    |                           |
| 10027-DAB | GPRS-FDD (TDMA, GMSK, TN 0-1-2)               | X | 28.34   | 94.6                         | 21.9 | 4.80    | 131.4    | $\pm 2.5 \%$              |
|           |                                               | Y | 52.22   | 99.6                         | 23.3 |         | 126.8    |                           |
|           |                                               | Z | 28.33   | 99.5                         | 23.9 |         | 140.7    |                           |
| 10028-DAB | GPRS-FDD (TDMA, GMSK, TN 0-1-2-3)             | X | 52.17   | 100.0                        | 22.2 | 3.55    | 147.0    | $\pm 1.7 \%$              |
|           |                                               | Y | 57.29   | 99.6                         | 22.4 |         | 133.0    |                           |
|           |                                               | Z | 25.84   | 99.5                         | 23.3 |         | 126.2    |                           |
| 10032-CAA | IEEE 802.15.1 Bluetooth (GFSK, DH5)           | X | 59.05   | 100.0                        | 19.9 | 1.16    | 135.5    | $\pm 1.9 \%$              |
|           |                                               | Y | 100.00  | 99.7                         | 19.2 |         | 143.5    |                           |
|           |                                               | Z | 34.97   | 100.0                        | 20.4 |         | 143.1    |                           |

|           |                                               |   |       |      |      |      |       |        |
|-----------|-----------------------------------------------|---|-------|------|------|------|-------|--------|
| 10039-CAB | CDMA2000 (1xRTT, RC1)                         | X | 4.78  | 66.9 | 18.9 | 4.57 | 134.6 | ±0.9 % |
|           |                                               | Y | 4.85  | 67.1 | 19.1 |      | 141.0 |        |
|           |                                               | Z | 4.76  | 67.8 | 19.4 |      | 140.7 |        |
| 10081-CAB | CDMA2000 (1xRTT, RC3)                         | X | 3.98  | 66.4 | 18.6 | 3.97 | 130.4 | ±0.7 % |
|           |                                               | Y | 3.98  | 66.5 | 18.7 |      | 136.2 |        |
|           |                                               | Z | 4.04  | 67.7 | 19.2 |      | 137.4 |        |
| 10098-CAB | UMTS-FDD (HSUPA, Subtest 2)                   | X | 4.75  | 67.3 | 18.8 | 3.98 | 144.4 | ±0.7 % |
|           |                                               | Y | 4.55  | 66.5 | 18.5 |      | 126.5 |        |
|           |                                               | Z | 4.72  | 67.9 | 19.0 |      | 128.1 |        |
| 10100-CAB | LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)      | X | 6.26  | 66.9 | 19.2 | 5.67 | 124.5 | ±1.2 % |
|           |                                               | Y | 6.38  | 67.4 | 19.7 |      | 131.7 |        |
|           |                                               | Z | 6.36  | 67.7 | 19.7 |      | 132.3 |        |
| 10108-CAB | LTE-FDD (SC-FDMA, 100% RB, 10 MHz, QPSK)      | X | 6.44  | 67.5 | 19.7 | 5.80 | 147.4 | ±1.4 % |
|           |                                               | Y | 6.31  | 67.2 | 19.7 |      | 130.2 |        |
|           |                                               | Z | 6.17  | 67.2 | 19.6 |      | 130.1 |        |
| 10110-CAB | LTE-FDD (SC-FDMA, 100% RB, 5 MHz, QPSK)       | X | 6.08  | 66.9 | 19.5 | 5.75 | 142.7 | ±1.4 % |
|           |                                               | Y | 5.97  | 66.6 | 19.4 |      | 127.3 |        |
|           |                                               | Z | 5.84  | 66.7 | 19.3 |      | 126.2 |        |
| 10114-CAA | IEEE 802.11n (HT Greenfield, 13.5 Mbps, BPSK) | X | 10.13 | 68.7 | 21.0 | 8.10 | 136.9 | ±2.5 % |
|           |                                               | Y | 10.57 | 69.9 | 21.9 |      | 146.3 |        |
|           |                                               | Z | 10.06 | 69.0 | 21.1 |      | 143.6 |        |
| 10117-CAA | IEEE 802.11n (HT Mixed, 13.5 Mbps, BPSK)      | X | 10.12 | 68.6 | 21.0 | 8.07 | 138.2 | ±2.5 % |
|           |                                               | Y | 10.60 | 69.9 | 21.9 |      | 148.0 |        |
|           |                                               | Z | 10.07 | 69.0 | 21.1 |      | 146.6 |        |
| 10151-CAB | LTE-TDD (SC-FDMA, 50% RB, 20 MHz, QPSK)       | X | 8.76  | 71.7 | 23.8 | 9.28 | 130.7 | ±3.0 % |
|           |                                               | Y | 10.03 | 75.2 | 25.9 |      | 121.5 |        |
|           |                                               | Z | 8.15  | 70.7 | 23.5 |      | 134.1 |        |
| 10164-CAB | LTE-FDD (SC-FDMA, 50% RB, 10 MHz, QPSK)       | X | 6.10  | 67.0 | 19.5 | 5.75 | 144.4 | ±1.4 % |
|           |                                               | Y | 5.98  | 66.6 | 19.4 |      | 127.8 |        |
|           |                                               | Z | 5.84  | 66.6 | 19.3 |      | 127.2 |        |
| 10160-CAB | LTE-FDD (SC-FDMA, 50% RB, 15 MHz, QPSK)       | X | 6.56  | 67.5 | 19.7 | 5.82 | 149.5 | ±1.7 % |
|           |                                               | Y | 6.41  | 67.1 | 19.6 |      | 132.5 |        |
|           |                                               | Z | 6.17  | 66.8 | 19.4 |      | 130.4 |        |
| 10169-CAB | LTE-FDD (SC-FDMA, 1 RB, 20 MHz, QPSK)         | X | 5.01  | 67.0 | 19.7 | 5.73 | 147.8 | ±1.2 % |
|           |                                               | Y | 5.01  | 66.9 | 19.8 |      | 132.1 |        |
|           |                                               | Z | 4.75  | 66.9 | 19.7 |      | 130.3 |        |
| 10172-CAB | LTE-TDD (SC-FDMA, 1 RB, 20 MHz, QPSK)         | X | 7.65  | 75.0 | 25.8 | 9.21 | 144.9 | ±2.7 % |
|           |                                               | Y | 10.17 | 82.4 | 29.7 |      | 136.4 |        |
|           |                                               | Z | 6.53  | 72.3 | 24.6 |      | 145.6 |        |
| 10175-CAB | LTE-FDD (SC-FDMA, 1 RB, 10 MHz, QPSK)         | X | 4.98  | 66.9 | 19.6 | 5.72 | 141.0 | ±1.2 % |
|           |                                               | Y | 4.98  | 66.7 | 19.7 |      | 130.5 |        |
|           |                                               | Z | 4.71  | 66.7 | 19.5 |      | 128.1 |        |

|           |                                              |   |       |      |      |      |       |        |
|-----------|----------------------------------------------|---|-------|------|------|------|-------|--------|
| 10181-CAB | LTE-FDD (SC-FDMA, 1 RB, 15 MHz, QPSK)        | X | 4.95  | 66.7 | 19.5 | 5.72 | 139.8 | ±1.2 % |
|           |                                              | Y | 4.97  | 66.7 | 19.7 |      | 129.5 |        |
|           |                                              | Z | 4.72  | 66.8 | 19.6 |      | 128.0 |        |
| 10193-CAA | IEEE 802.11n (HT Greenfield, 6.5 Mbps, BPSK) | X | 9.75  | 68.2 | 20.9 | 8.09 | 131.8 | ±2.5 % |
|           |                                              | Y | 10.16 | 69.4 | 21.7 |      | 139.2 |        |
|           |                                              | Z | 9.62  | 68.6 | 21.0 |      | 137.3 |        |
| 10196-CAA | IEEE 802.11n (HT Mixed, 6.5 Mbps, BPSK)      | X | 9.77  | 68.3 | 20.9 | 8.10 | 133.6 | ±2.5 % |
|           |                                              | Y | 10.17 | 69.4 | 21.8 |      | 140.1 |        |
|           |                                              | Z | 9.61  | 68.5 | 21.0 |      | 140.1 |        |
| 10219-CAA | IEEE 802.11n (HT Mixed, 7.2 Mbps, BPSK)      | X | 9.69  | 68.3 | 20.9 | 8.03 | 133.6 | ±2.5 % |
|           |                                              | Y | 10.05 | 69.3 | 21.7 |      | 139.2 |        |
|           |                                              | Z | 9.58  | 68.7 | 21.1 |      | 139.4 |        |
| 10222-CAA | IEEE 802.11n (HT Mixed, 15 Mbps, BPSK)       | X | 10.13 | 68.7 | 21.0 | 8.06 | 140.7 | ±2.5 % |
|           |                                              | Y | 10.61 | 69.8 | 21.8 |      | 145.1 |        |
|           |                                              | Z | 10.11 | 69.1 | 21.2 |      | 148.4 |        |
| 10225-CAB | UMTS-FDD (HSPA+)                             | X | 7.03  | 67.2 | 19.4 | 5.97 | 138.0 | ±1.4 % |
|           |                                              | Y | 7.07  | 67.2 | 19.6 |      | 140.2 |        |
|           |                                              | Z | 6.97  | 67.8 | 19.7 |      | 144.6 |        |
| 10237-CAB | LTE-TDD (SC-FDMA, 1 RB, 10 MHz, QPSK)        | X | 7.11  | 72.9 | 24.7 | 9.21 | 124.6 | ±2.7 % |
|           |                                              | Y | 10.04 | 82.0 | 29.5 |      | 135.7 |        |
|           |                                              | Z | 6.29  | 71.2 | 24.0 |      | 126.2 |        |
| 10252-CAB | LTE-TDD (SC-FDMA, 50% RB, 10 MHz, QPSK)      | X | 8.61  | 72.5 | 24.3 | 9.24 | 145.2 | ±3.3 % |
|           |                                              | Y | 10.53 | 77.8 | 27.4 |      | 136.7 |        |
|           |                                              | Z | 7.56  | 70.0 | 23.1 |      | 126.7 |        |
| 10267-CAB | LTE-TDD (SC-FDMA, 100% RB, 10 MHz, QPSK)     | X | 8.74  | 71.6 | 23.8 | 9.30 | 128.7 | ±3.3 % |
|           |                                              | Y | 11.51 | 79.1 | 28.0 |      | 147.2 |        |
|           |                                              | Z | 8.07  | 70.4 | 23.2 |      | 134.1 |        |
| 10274-CAB | UMTS-FDD (HSUPA, Subtest 5, 3GPP Rel8.10)    | X | 5.90  | 66.7 | 18.7 | 4.87 | 128.0 | ±0.9 % |
|           |                                              | Y | 5.93  | 66.8 | 18.9 |      | 134.5 |        |
|           |                                              | Z | 5.92  | 67.6 | 19.1 |      | 138.2 |        |
| 10275-CAB | UMTS-FDD (HSUPA, Subtest 5, 3GPP Rel8.4)     | X | 4.53  | 67.1 | 18.8 | 3.96 | 133.8 | ±0.7 % |
|           |                                              | Y | 4.48  | 67.0 | 18.8 |      | 139.6 |        |
|           |                                              | Z | 4.62  | 68.3 | 19.3 |      | 145.0 |        |
| 10291-AAB | CDMA2000, RC3, SO55, Full Rate               | X | 3.82  | 67.8 | 19.0 | 3.46 | 147.6 | ±0.7 % |
|           |                                              | Y | 3.66  | 67.0 | 18.8 |      | 131.7 |        |
|           |                                              | Z | 3.97  | 69.6 | 20.0 |      | 135.9 |        |
| 10292-AAB | CDMA2000, RC3, SO32, Full Rate               | X | 3.70  | 67.5 | 18.8 | 3.39 | 128.1 | ±0.7 % |
|           |                                              | Y | 3.60  | 66.9 | 18.7 |      | 132.5 |        |
|           |                                              | Z | 3.80  | 68.9 | 19.5 |      | 139.8 |        |
| 10297-AAA | LTE-FDD (SC-FDMA, 50% RB, 20 MHz, QPSK)      | X | 6.47  | 67.6 | 19.8 | 5.81 | 149.7 | ±1.7 % |
|           |                                              | Y | 6.24  | 66.9 | 19.5 |      | 126.3 |        |
|           |                                              | Z | 6.20  | 67.3 | 19.6 |      | 130.9 |        |

|           |                                                               |   |       |      |      |      |       |        |
|-----------|---------------------------------------------------------------|---|-------|------|------|------|-------|--------|
| 10311-AAA | LTE-FDD (SC-FDMA, 100% RB, 15 MHz, QPSK)                      | X | 6.72  | 67.1 | 19.5 | 6.06 | 128.8 | ±1.4 % |
|           |                                                               | Y | 6.85  | 67.7 | 20.0 |      | 132.4 |        |
|           |                                                               | Z | 6.75  | 67.7 | 19.8 |      | 136.6 |        |
| 10315-AAA | IEEE 802.11b WiFi 2.4 GHz (DSSS, 1 Mbps, 96pc duty cycle)     | X | 3.27  | 71.1 | 19.8 | 1.71 | 140.1 | ±0.7 % |
|           |                                                               | Y | 2.95  | 69.4 | 19.1 |      | 139.8 |        |
|           |                                                               | Z | 3.75  | 74.4 | 21.2 |      | 146.9 |        |
| 10316-AAA | IEEE 802.11g WiFi 2.4 GHz (ERP-OFDM, 6 Mbps, 96pc duty cycle) | X | 10.04 | 68.7 | 21.3 | 8.36 | 136.3 | ±2.5 % |
|           |                                                               | Y | 10.42 | 69.8 | 22.1 |      | 138.1 |        |
|           |                                                               | Z | 9.84  | 68.9 | 21.3 |      | 139.7 |        |
| 10403-AAB | CDMA2000 (1xEV-DO, Rev. 0)                                    | X | 5.01  | 69.3 | 19.2 | 3.76 | 144.3 | ±0.7 % |
|           |                                                               | Y | 4.79  | 68.1 | 18.7 |      | 146.3 |        |
|           |                                                               | Z | 5.40  | 72.5 | 20.8 |      | 146.7 |        |
| 10404-AAB | CDMA2000 (1xEV-DO, Rev. A)                                    | X | 4.97  | 69.5 | 19.3 | 3.77 | 141.3 | ±0.7 % |
|           |                                                               | Y | 4.72  | 68.2 | 18.8 |      | 143.1 |        |
|           |                                                               | Z | 5.12  | 71.8 | 20.5 |      | 144.4 |        |
| 10415-AAA | IEEE 802.11b WiFi 2.4 GHz (DSSS, 1 Mbps, 99pc duty cycle)     | X | 3.05  | 70.5 | 19.5 | 1.54 | 139.7 | ±0.7 % |
|           |                                                               | Y | 2.71  | 68.7 | 18.9 |      | 140.2 |        |
|           |                                                               | Z | 4.22  | 77.3 | 22.5 |      | 145.9 |        |
| 10416-AAA | IEEE 802.11g WiFi 2.4 GHz (ERP-OFDM, 6 Mbps, 99pc duty cycle) | X | 9.92  | 68.6 | 21.1 | 8.23 | 136.3 | ±2.5 % |
|           |                                                               | Y | 10.20 | 69.4 | 21.8 |      | 138.3 |        |
|           |                                                               | Z | 9.76  | 68.8 | 21.3 |      | 138.9 |        |

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor  $k=2$ , which for a normal distribution corresponds to a coverage probability of approximately 95%.

<sup>A</sup> The uncertainties of NormX,Y,Z do not affect the  $E^2$ -field uncertainty inside TSL (see Pages 8 and 9).

<sup>B</sup> Numerical linearization parameter: uncertainty not required.

<sup>E</sup> Uncertainty is determined using the max. deviation from linear response applying rectangular distribution and is expressed for the square of the field value.

## DASY/EASY - Parameters of Probe: ES3DV3 - SN:3332

### Calibration Parameter Determined in Head Tissue Simulating Media

| f (MHz) <sup>C</sup> | Relative Permittivity <sup>F</sup> | Conductivity (S/m) <sup>F</sup> | ConvF X | ConvF Y | ConvF Z | Alpha <sup>G</sup> | Depth <sup>G</sup> (mm) | Unct. (k=2) |
|----------------------|------------------------------------|---------------------------------|---------|---------|---------|--------------------|-------------------------|-------------|
| 750                  | 41.9                               | 0.89                            | 6.56    | 6.56    | 6.56    | 0.50               | 1.43                    | ± 12.0 %    |
| 835                  | 41.5                               | 0.90                            | 6.31    | 6.31    | 6.31    | 0.61               | 1.31                    | ± 12.0 %    |
| 1750                 | 40.1                               | 1.37                            | 5.17    | 5.17    | 5.17    | 0.62               | 1.33                    | ± 12.0 %    |
| 1900                 | 40.0                               | 1.40                            | 5.04    | 5.04    | 5.04    | 0.80               | 1.17                    | ± 12.0 %    |
| 2450                 | 39.2                               | 1.80                            | 4.49    | 4.49    | 4.49    | 0.77               | 1.24                    | ± 12.0 %    |
| 2600                 | 39.0                               | 1.96                            | 4.35    | 4.35    | 4.35    | 0.73               | 1.38                    | ± 12.0 %    |

<sup>C</sup> Frequency validity above 300 MHz of ± 100 MHz only applies for DASY v4.4 and higher (see Page 2), else it is restricted to ± 50 MHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below 300 MHz is ± 10, 25, 40, 50 and 70 MHz for ConvF assessments at 30, 64, 128, 150 and 220 MHz respectively. Above 5 GHz frequency validity can be extended to ± 110 MHz.

<sup>F</sup> At frequencies below 3 GHz, the validity of tissue parameters ( $\epsilon$  and  $\sigma$ ) can be relaxed to ± 10% if liquid compensation formula is applied to measured SAR values. At frequencies above 3 GHz, the validity of tissue parameters ( $\epsilon$  and  $\sigma$ ) is restricted to ± 5%. The uncertainty is the RSS of the ConvF uncertainty for indicated target tissue parameters.

<sup>G</sup> Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than ± 1% for frequencies below 3 GHz and below ± 2% for frequencies between 3-6 GHz at any distance larger than half the probe tip diameter from the boundary.

## DASY/EASY - Parameters of Probe: ES3DV3 - SN:3332

### Calibration Parameter Determined in Body Tissue Simulating Media

| f (MHz) <sup>C</sup> | Relative Permittivity <sup>F</sup> | Conductivity (S/m) <sup>F</sup> | ConvF X | ConvF Y | ConvF Z | Alpha <sup>G</sup> | Depth <sup>G</sup> (mm) | Unct. (k=2) |
|----------------------|------------------------------------|---------------------------------|---------|---------|---------|--------------------|-------------------------|-------------|
| 750                  | 55.5                               | 0.96                            | 6.24    | 6.24    | 6.24    | 0.50               | 1.50                    | ± 12.0 %    |
| 835                  | 55.2                               | 0.97                            | 6.21    | 6.21    | 6.21    | 0.45               | 1.59                    | ± 12.0 %    |
| 1750                 | 53.4                               | 1.49                            | 4.88    | 4.88    | 4.88    | 0.39               | 1.78                    | ± 12.0 %    |
| 1900                 | 53.3                               | 1.52                            | 4.64    | 4.64    | 4.64    | 0.61               | 1.47                    | ± 12.0 %    |
| 2450                 | 52.7                               | 1.95                            | 4.31    | 4.31    | 4.31    | 0.80               | 1.18                    | ± 12.0 %    |
| 2600                 | 52.5                               | 2.16                            | 4.11    | 4.11    | 4.11    | 0.68               | 0.99                    | ± 12.0 %    |


<sup>C</sup> Frequency validity above 300 MHz of ± 100 MHz only applies for DASY v4.4 and higher (see Page 2), else it is restricted to ± 50 MHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below 300 MHz is ± 10, 25, 40, 50 and 70 MHz for ConvF assessments at 30, 64, 128, 150 and 220 MHz respectively. Above 5 GHz frequency validity can be extended to ± 110 MHz.

<sup>F</sup> At frequencies below 3 GHz, the validity of tissue parameters ( $\epsilon$  and  $\sigma$ ) can be relaxed to ± 10% if liquid compensation formula is applied to measured SAR values. At frequencies above 3 GHz, the validity of tissue parameters ( $\epsilon$  and  $\sigma$ ) is restricted to ± 5%. The uncertainty is the RSS of the ConvF uncertainty for indicated target tissue parameters.

<sup>G</sup> Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than ± 1% for frequencies below 3 GHz and below ± 2% for frequencies between 3-6 GHz at any distance larger than half the probe tip diameter from the boundary.

## Frequency Response of E-Field

(TEM-Cell:ifi110 EXX, Waveguide: R22)





Uncertainty of Frequency Response of E-field:  $\pm 6.3\%$  ( $k=2$ )

Receiving Pattern ( $\phi$ ),  $\theta = 0^\circ$


f=600 MHz,TEM

f=1800 MHz,R22




Uncertainty of Axial Isotropy Assessment:  $\pm 0.5\%$  ( $k=2$ )

# Dynamic Range $f(\text{SAR}_{\text{head}})$ (TEM cell , $f_{\text{eval}} = 1900 \text{ MHz}$ )





Uncertainty of Linearity Assessment:  $\pm 0.6\%$  ( $k=2$ )

## Conversion Factor Assessment



## Deviation from Isotropy in Liquid

Error ( $\phi, \theta$ ),  $f = 900 \text{ MHz}$



**DASY/EASY - Parameters of Probe: ES3DV3 - SN:3332****Other Probe Parameters**

|                                               |            |
|-----------------------------------------------|------------|
| Sensor Arrangement                            | Triangular |
| Connector Angle (°)                           | -3.7       |
| Mechanical Surface Detection Mode             | enabled    |
| Optical Surface Detection Mode                | disabled   |
| Probe Overall Length                          | 337 mm     |
| Probe Body Diameter                           | 10 mm      |
| Tip Length                                    | 10 mm      |
| Tip Diameter                                  | 4 mm       |
| Probe Tip to Sensor X Calibration Point       | 2 mm       |
| Probe Tip to Sensor Y Calibration Point       | 2 mm       |
| Probe Tip to Sensor Z Calibration Point       | 2 mm       |
| Recommended Measurement Distance from Surface | 3 mm       |

## APPENDIX D: SAR TISSUE SPECIFICATIONS

Measurement Procedure for Tissue verification:

- 3) The network analyzer and probe system was configured and calibrated.
- 4) The probe was immersed in the tissue. The tissue was placed in a nonmetallic container. Trapped air bubbles beneath the flange were minimized by placing the probe at a slight angle.
- 5) The complex admittance with respect to the probe aperture was measured
- 6) The complex relative permittivity  $\epsilon$  can be calculated from the below equation (Pournaropoulos and Misra):

$$Y = \frac{j2\omega\epsilon_r\epsilon_0}{[\ln(b/a)]^2} \int_a^b \int_a^b \int_0^\pi \cos\phi' \frac{\exp[-j\omega r(\mu_0\epsilon_r'\epsilon_0)^{1/2}]}{r} d\phi' d\rho' d\rho$$

where  $Y$  is the admittance of the probe in contact with the sample, the primed and unprimed coordinates refer to source and observation points, respectively,  $r^2 = \rho^2 + \rho'^2 - 2\rho\rho'\cos\phi'$ ,  $\omega$  is the angular frequency, and  $j = \sqrt{-1}$ .

**Table D-I**  
**Composition of the Tissue Equivalent Matter**

| Frequency (MHz)           | 750  | 750  | 835   | 835   | 1750 | 1750 | 1900  | 1900  | 2450 | 2450 |
|---------------------------|------|------|-------|-------|------|------|-------|-------|------|------|
| Tissue                    | Head | Body | Head  | Body  | Head | Body | Head  | Body  | Head | Body |
| Ingredients (% by weight) |      |      |       |       |      |      |       |       |      |      |
| Bactericide               |      |      | 0.1   | 0.1   |      |      |       |       |      |      |
| DGBE                      |      |      |       |       | 47   | 31   | 44.92 | 29.44 |      | 26.7 |
| HEC                       |      |      | 1     | 1     |      |      |       |       |      |      |
| NaCl                      |      |      | 1.45  | 0.94  | 0.4  | 0.2  | 0.18  | 0.39  |      | 0.1  |
| Sucrose                   |      |      | 57    | 44.9  |      |      |       |       |      |      |
| Water                     |      |      | 40.45 | 53.06 | 52.6 | 68.8 | 54.9  | 70.17 |      | 73.2 |

|                                    |                                                                                     |                       |                                                                                       |                                 |
|------------------------------------|-------------------------------------------------------------------------------------|-----------------------|---------------------------------------------------------------------------------------|---------------------------------|
| FCC ID: ZNFH731                    |  | SAR EVALUATION REPORT |  | Reviewed by:<br>Quality Manager |
| Test Dates:<br>05/18/15 - 05/28/15 | DUT Type:<br>Portable Handset                                                       |                       |                                                                                       | APPENDIX D:<br>Page 1 of 4      |

## 2 Composition / Information on ingredients

The Item is composed of the following ingredients:

|                                                                  |                                                                                                                                                     |
|------------------------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------------------------------|
| H <sub>2</sub> O                                                 | Water, 35 – 58%                                                                                                                                     |
| Sucrose                                                          | Sugar, white, refined, 40 – 60%                                                                                                                     |
| NaCl                                                             | Sodium Chloride, 0 – 6%                                                                                                                             |
| Hydroxyethyl-cellulose                                           | Medium Viscosity (CAS# 9004-62-0), <0.3%                                                                                                            |
| Preventol-D7                                                     | Preservative: aqueous preparation, (CAS# 55965-84-9), containing 5-chloro-2-methyl-3(2H)-isothiazolone and 2-methyl-3(2H)-isothiazolone, 0.1 – 0.7% |
| Relevant for safety; Refer to the respective Safety Data Sheet*. |                                                                                                                                                     |

**Figure D-1**

### Composition of 750 MHz Head and Body Tissue Equivalent Matter

**Note:** 750MHz liquid recipes are proprietary SPEAG. Since the composition is approximate to the actual liquids utilized, the manufacturer tissue-equivalent liquid data sheets are provided below.

#### Measurement Certificate / Material Test

|              |                                          |
|--------------|------------------------------------------|
| Item Name    | Body Tissue Simulating Liquid (MSL750V2) |
| Product No.  | SL AAM 075 AA (Charge: 130828-1)         |
| Manufacturer | SPEAG                                    |

#### Measurement Method

TSL dielectric parameters measured using calibrated OCP probe.

#### Setup Validation

Validation results were within  $\pm 2.5\%$  towards the target values of Methanol.

#### Target Parameters

Target parameters as defined in the IEEE 1528 and IEC 62209 compliance standards.



#### Test Condition

|                 |                                                             |
|-----------------|-------------------------------------------------------------|
| Ambient         | Environment temperatur ( $22 \pm 3$ )°C and humidity < 70%. |
| TSL Temperature | 22°C                                                        |
| Test Date       | 28-Aug-13                                                   |
| Operator        | IEN                                                         |

#### Additional Information

|                   |                         |
|-------------------|-------------------------|
| TSL Density       | 1.212 g/cm <sup>3</sup> |
| TSL Heat-capacity | 3.006 kJ/(kg*K)         |

| f [MHz] | Measured |        |       | Target |       | Diff. to Target [%] |                 |
|---------|----------|--------|-------|--------|-------|---------------------|-----------------|
|         | HP-e'    | HP-e'' | sigma | eps    | sigma | $\Delta$ -eps       | $\Delta$ -sigma |
| 600     | 57.4     | 24.76  | 0.83  | 56.1   | 0.95  | 2.3                 | -13.2           |
| 625     | 57.1     | 24.42  | 0.85  | 56.0   | 0.95  | 2.0                 | -11.0           |
| 650     | 56.8     | 24.09  | 0.87  | 55.9   | 0.96  | 1.6                 | -8.9            |
| 675     | 56.6     | 23.80  | 0.89  | 55.8   | 0.96  | 1.3                 | -6.7            |
| 700     | 56.3     | 23.52  | 0.92  | 55.7   | 0.96  | 1.0                 | -4.5            |
| 725     | 56.1     | 23.27  | 0.94  | 55.6   | 0.96  | 0.8                 | -2.4            |
| 750     | 55.8     | 23.03  | 0.96  | 55.5   | 0.96  | 0.5                 | -0.3            |
| 775     | 55.6     | 22.87  | 0.99  | 55.4   | 0.97  | 0.2                 | 2.1             |
| 800     | 55.3     | 22.71  | 1.01  | 55.3   | 0.97  | -0.1                | 4.5             |
| 825     | 55.1     | 22.54  | 1.03  | 55.2   | 0.98  | -0.3                | 5.8             |
| 850     | 54.9     | 22.45  | 1.05  | 55.2   | 0.98  | -0.5                | 6.4             |
| 875     | 54.8     | 22.37  | 1.06  | 55.2   | 0.99  | -0.6                | 7.0             |
| 900     | 54.6     | 22.25  | 1.08  | 55.1   | 1.02  | -0.9                | 6.2             |
| 925     | 54.4     | 22.13  | 1.11  | 55.0   | 1.06  | -1.1                | 5.5             |
| 950     | 54.2     | 22.02  | 1.13  | 55.0   | 1.06  | -1.5                | 6.6             |
| 975     | 53.9     | 21.91  | 1.16  | 54.9   | 1.08  | -1.8                | 7.7             |
| 1000    | 53.7     | 21.84  | 1.18  | 54.9   | 1.09  | -2.2                | 9.0             |
|         | 53.5     | 21.77  | 1.21  | 54.8   | 1.10  | -2.5                | 10.3            |



**Figure D-2**  
750MHz Body Tissue Equivalent Matter

|                                    |                                                                                     |                       |                                                                                       |                                 |
|------------------------------------|-------------------------------------------------------------------------------------|-----------------------|---------------------------------------------------------------------------------------|---------------------------------|
| FCC ID: ZNFH731                    |  | SAR EVALUATION REPORT |  | Reviewed by:<br>Quality Manager |
| Test Dates:<br>05/18/15 - 05/28/15 | DUT Type:<br>Portable Handset                                                       |                       |                                                                                       | APPENDIX D:<br>Page 2 of 4      |

## Measurement Certificate / Material Test

Item Name **Head Tissue Simulating Liquid (HSL750V2)**  
 Product No. SL AAH 075 AA (Charge: 130826-2)  
 Manufacturer SPEAG

### Measurement Method

TSL dielectric parameters measured using calibrated OCP probe.

### Setup Validation

Validation results were within  $\pm 2.5\%$  towards the target values of Methanol.

### Target Parameters

Target parameters as defined in the IEEE 1528 and IEC 62209 compliance standards.

### Test Condition

Ambient Environment temperatur ( $22 \pm 3$ )°C and humidity < 70%.  
 TSL Temperature 22°C  
 Test Date 28-Aug-13  
 Operator IEN

### Additional Information

TSL Density 1.284 g/cm<sup>3</sup>  
 TSL Heat-capacity 2.701 kJ/(kg\*K)

| f [MHz] | Measured |        |       | Target |       | Diff. to Target [%] |         |
|---------|----------|--------|-------|--------|-------|---------------------|---------|
|         | HP-e'    | HP-e'' | sigma | eps    | sigma | Δ-eps               | Δ-sigma |
| 600     | 44.4     | 23.32  | 0.78  | 42.7   | 0.88  | 3.9                 | -11.7   |
| 625     | 44.0     | 23.05  | 0.80  | 42.6   | 0.88  | 3.4                 | -9.3    |
| 650     | 43.6     | 22.78  | 0.82  | 42.5   | 0.89  | 2.8                 | -7.0    |
| 675     | 43.3     | 22.54  | 0.85  | 42.3   | 0.89  | 2.2                 | -4.7    |
| 700     | 42.9     | 22.30  | 0.87  | 42.2   | 0.89  | 1.6                 | -2.4    |
| 725     | 42.6     | 22.10  | 0.89  | 42.1   | 0.89  | 1.2                 | 0.0     |
| 750     | 42.3     | 21.90  | 0.91  | 41.9   | 0.89  | 0.8                 | 2.3     |
| 775     | 41.9     | 21.74  | 0.94  | 41.8   | 0.90  | 0.2                 | 4.6     |
| 800     | 41.6     | 21.57  | 0.96  | 41.7   | 0.90  | -0.3                | 7.0     |
| 825     | 41.3     | 21.42  | 0.98  | 41.6   | 0.91  | -0.8                | 8.4     |
| 838     | 41.1     | 21.35  | 0.99  | 41.5   | 0.91  | -1.0                | 9.1     |
| 850     | 41.0     | 21.27  | 1.01  | 41.5   | 0.92  | -1.3                | 9.8     |
| 875     | 40.7     | 21.14  | 1.03  | 41.5   | 0.94  | -2.0                | 9.1     |
| 900     | 40.4     | 21.00  | 1.05  | 41.5   | 0.97  | -2.6                | 8.4     |
| 925     | 40.1     | 20.89  | 1.07  | 41.5   | 0.98  | -3.2                | 9.4     |
| 950     | 39.9     | 20.78  | 1.10  | 41.4   | 0.99  | -3.8                | 10.4    |
| 975     | 39.6     | 20.69  | 1.12  | 41.4   | 1.00  | -4.4                | 11.7    |
| 1000    | 39.3     | 20.60  | 1.15  | 41.3   | 1.01  | -4.9                | 12.9    |





Figure D-3  
750MHz Head Tissue Equivalent Matter

|                                    |                                                                                     |                       |                                                                                       |                                 |
|------------------------------------|-------------------------------------------------------------------------------------|-----------------------|---------------------------------------------------------------------------------------|---------------------------------|
| FCC ID: ZNFH731                    |  | SAR EVALUATION REPORT |  | Reviewed by:<br>Quality Manager |
| Test Dates:<br>05/18/15 - 05/28/15 | DUT Type:<br>Portable Handset                                                       |                       |                                                                                       | APPENDIX D:<br>Page 3 of 4      |

## 2 Composition / Information on ingredients

The Item is composed of the following ingredients:

|                                               |                                                                                                                                                                                           |
|-----------------------------------------------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| H <sub>2</sub> O                              | Water, 52 – 75%                                                                                                                                                                           |
| C <sub>8</sub> H <sub>18</sub> O <sub>3</sub> | Diethylene glycol monobutyl ether (DGBE), 25 – 48%<br>(CAS-No. 112-34-5, EC-No. 203-961-6, EC-index-No. 603-096-00-8)<br>Relevant for safety; Refer to the respective Safety Data Sheet*. |
| NaCl                                          | Sodium Chloride, <1.0%                                                                                                                                                                    |

Figure D-4

### Composition of 2.4 GHz Head Tissue Equivalent Matter

**Note:** 2.4 GHz head liquid recipes are proprietary SPEAG. Since the composition is approximate to the actual liquids utilized, the manufacturer tissue-equivalent liquid data sheets are provided below.

#### Measurement Certificate / Material Test

|              |                                           |
|--------------|-------------------------------------------|
| Item Name    | Head Tissue Simulating Liquid (HSL2450V2) |
| Product No.  | SL AAH 245 BA (Charge: 130926-1)          |
| Manufacturer | SPEAG                                     |

#### Measurement Method

TSL dielectric parameters measured using calibrated OCP probe.

#### Setup Validation

Validation results were within  $\pm 2.5\%$  towards the target values of Methanol.

#### Target Parameters

Target parameters as defined in the IEEE 1528 and IEC 62209 compliance standards.

#### Test Condition

|                 |                                                             |
|-----------------|-------------------------------------------------------------|
| Ambient         | Environment temperatur ( $22 \pm 3$ )°C and humidity < 70%. |
| TSL Temperature | 23°C                                                        |
| Test Date       | 2-Oct-13                                                    |
| Operator        | CL                                                          |

#### Additional Information

|                   |                         |
|-------------------|-------------------------|
| TSL Density       | 0.988 g/cm <sup>3</sup> |
| TSL Heat-capacity | 3.680 kJ/(kg*K)         |

| f [MHz] | Measured |        |       | Target |       | Diff.to Target [%] |         |
|---------|----------|--------|-------|--------|-------|--------------------|---------|
|         | HP-e'    | HP-e'' | sigma | eps    | sigma | Δ-eps              | Δ-sigma |
| 1900    | 40.3     | 11.88  | 1.26  | 40.0   | 1.40  | 0.9                | -10.3   |
| 1925    | 40.3     | 11.98  | 1.28  | 40.0   | 1.40  | 0.6                | -8.3    |
| 1950    | 40.2     | 12.08  | 1.31  | 40.0   | 1.40  | 0.4                | -6.4    |
| 1975    | 40.1     | 12.15  | 1.34  | 40.0   | 1.40  | 0.2                | -4.6    |
| 2000    | 40.0     | 12.23  | 1.36  | 40.0   | 1.40  | -0.1               | -2.8    |
| 2025    | 39.9     | 12.34  | 1.39  | 40.0   | 1.42  | -0.2               | -2.3    |
| 2050    | 39.8     | 12.45  | 1.42  | 39.9   | 1.44  | -0.3               | -1.7    |
| 2075    | 39.7     | 12.54  | 1.45  | 39.9   | 1.47  | -0.4               | -1.3    |
| 2100    | 39.6     | 12.64  | 1.48  | 39.8   | 1.49  | -0.5               | -0.8    |
| 2125    | 39.5     | 12.69  | 1.50  | 39.8   | 1.51  | -0.7               | -0.7    |
| 2150    | 39.4     | 12.75  | 1.52  | 39.7   | 1.53  | -0.8               | -0.6    |
| 2175    | 39.3     | 12.84  | 1.55  | 39.7   | 1.56  | -1.0               | -0.1    |
| 2200    | 39.2     | 12.94  | 1.58  | 39.6   | 1.58  | -1.2               | 0.4     |
| 2225    | 39.1     | 13.00  | 1.61  | 39.6   | 1.60  | -1.3               | 0.6     |
| 2250    | 39.0     | 13.07  | 1.64  | 39.6   | 1.62  | -1.4               | 0.8     |
| 2275    | 38.9     | 13.15  | 1.66  | 39.5   | 1.64  | -1.5               | 1.2     |
| 2300    | 38.8     | 13.22  | 1.69  | 39.5   | 1.67  | -1.7               | 1.5     |
| 2325    | 38.7     | 13.32  | 1.72  | 39.4   | 1.69  | -1.9               | 2.0     |
| 2350    | 38.6     | 13.42  | 1.75  | 39.4   | 1.71  | -2.0               | 2.5     |
| 2375    | 38.5     | 13.49  | 1.78  | 39.3   | 1.73  | -2.1               | 2.8     |
| 2400    | 38.4     | 13.56  | 1.81  | 39.3   | 1.76  | -2.3               | 3.1     |
| 2425    | 38.3     | 13.63  | 1.84  | 39.2   | 1.78  | -2.5               | 3.5     |
| 2450    | 38.2     | 13.71  | 1.87  | 39.2   | 1.80  | -2.6               | 3.8     |
| 2475    | 38.1     | 13.79  | 1.90  | 39.2   | 1.83  | -2.8               | 3.9     |
| 2500    | 38.0     | 13.87  | 1.93  | 39.1   | 1.85  | -3.0               | 4.0     |
| 2525    | 37.9     | 13.96  | 1.96  | 39.1   | 1.88  | -3.1               | 4.2     |
| 2550    | 37.8     | 14.05  | 1.99  | 39.1   | 1.91  | -3.3               | 4.4     |
| 2575    | 37.7     | 14.11  | 2.02  | 39.0   | 1.94  | -3.5               | 4.4     |
| 2600    | 37.5     | 14.17  | 2.05  | 39.0   | 1.96  | -3.8               | 4.4     |
| 2625    | 37.4     | 14.23  | 2.08  | 39.0   | 1.99  | -4.0               | 4.4     |
| 2650    | 37.3     | 14.30  | 2.11  | 38.9   | 2.02  | -4.2               | 4.5     |
| 2675    | 37.2     | 14.37  | 2.14  | 38.9   | 2.05  | -4.4               | 4.6     |
| 2700    | 37.1     | 14.45  | 2.17  | 38.9   | 2.07  | -4.6               | 4.7     |




Figure D-5

### 2.4 GHz Head Tissue Equivalent Matter

|                                    |                                                                                     |                       |                                                                                       |                                 |
|------------------------------------|-------------------------------------------------------------------------------------|-----------------------|---------------------------------------------------------------------------------------|---------------------------------|
| FCC ID: ZNFH731                    |  | SAR EVALUATION REPORT |  | Reviewed by:<br>Quality Manager |
| Test Dates:<br>05/18/15 - 05/28/15 | DUT Type:<br>Portable Handset                                                       |                       |                                                                                       | APPENDIX D:<br>Page 4 of 4      |

## APPENDIX E: SAR SYSTEM VALIDATION

Per FCC KDB 865664 D02v01, SAR system validation status should be documented to confirm measurement accuracy. The SAR systems (including SAR probes, system components and software versions) used for this device were validated against its performance specifications prior to the SAR measurements. Reference dipoles were used with the required tissue- equivalent media for system validation, according to the procedures outlined in FCC KDB 865664 D01 v01 and IEEE 1528-2013. Since SAR probe calibrations are frequency dependent, each probe calibration point was validated at a frequency within the valid frequency range of the probe calibration point, using the system that normally operates with the probe for routine SAR measurements and according to the required tissue-equivalent media.

A tabulated summary of the system validation status including the validation date(s), measurement frequencies, SAR probes and tissue dielectric parameters has been included.

**Table E-I**  
**SAR System Validation Summary**

| SAR SYSTEM # | FREQ. [MHz] | DATE       | PROBE SN | PROBE TYPE | PROBE CAL. POINT |      | COND. | PERM.             | CW VALIDATION |                 |                | MOD. VALIDATION |             |      |
|--------------|-------------|------------|----------|------------|------------------|------|-------|-------------------|---------------|-----------------|----------------|-----------------|-------------|------|
|              |             |            |          |            |                  |      | (σ)   | (ε <sub>r</sub> ) | SENSITIVITY   | PROBE LINEARITY | PROBE ISOTROPY | MOD. TYPE       | DUTY FACTOR | PAR  |
| K            | 750         | 10/10/2014 | 3288     | ES3DV3     | 750              | Head | 0.889 | 40.212            | PASS          | PASS            | PASS           | N/A             | N/A         | N/A  |
| G            | 835         | 5/21/2015  | 3318     | ES3DV3     | 835              | Head | 0.894 | 41.397            | PASS          | PASS            | PASS           | GMSK            | PASS        | N/A  |
| J            | 1750        | 5/20/2015  | 3319     | ES3DV3     | 1750             | Head | 1.371 | 39.404            | PASS          | PASS            | PASS           | N/A             | N/A         | N/A  |
| K            | 1900        | 10/16/2014 | 3288     | ES3DV3     | 1900             | Head | 1.454 | 39.506            | PASS          | PASS            | PASS           | GMSK            | PASS        | N/A  |
| G            | 2450        | 4/15/2015  | 3318     | ES3DV3     | 2450             | Head | 1.842 | 40.368            | PASS          | PASS            | PASS           | OFDM/TDD        | PASS        | PASS |
| J            | 750         | 5/27/2015  | 3319     | ES3DV3     | 750              | Body | 0.948 | 53.966            | PASS          | PASS            | PASS           | N/A             | N/A         | N/A  |
| C            | 835         | 11/11/2014 | 3333     | ES3DV3     | 835              | Body | 0.948 | 53.049            | PASS          | PASS            | PASS           | GMSK            | PASS        | N/A  |
| B            | 1750        | 1/4/2015   | 3334     | ES3DV3     | 1750             | Body | 1.527 | 51.873            | PASS          | PASS            | PASS           | N/A             | N/A         | N/A  |
| C            | 1900        | 11/19/2014 | 3333     | ES3DV3     | 1900             | Body | 1.547 | 50.934            | PASS          | PASS            | PASS           | GMSK            | PASS        | N/A  |
| E            | 2450        | 11/3/2014  | 3332     | ES3DV3     | 2450             | Body | 1.996 | 52.207            | PASS          | PASS            | PASS           | OFDM/TDD        | PASS        | PASS |

NOTE: While the probes have been calibrated for both CW and modulated signals, all measurements were performed using communication systems calibrated for CW signals only. Modulations in the table above represent test configurations for which the measurement system has been validated per FCC KDB Publication 865664 D01v01 for scenarios when CW probe calibrations are used with other signal types. SAR systems were validated for modulated signals with a periodic duty cycle, such as GMSK, or with a high peak to average ratio (>5 dB), such as OFDM according to KDB 865664.

|                                           |                                                                                                                                   |                                   |                                                                                       |                                        |
|-------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------------|-----------------------------------|---------------------------------------------------------------------------------------|----------------------------------------|
| FCC ID: ZNFH731                           |  <b>PCTEST</b><br>ENGINEERING LABORATORY, INC. | <b>SAR EVALUATION REPORT</b>      |  | <b>Reviewed by:</b><br>Quality Manager |
| <b>Test Dates:</b><br>05/18/15 - 05/28/15 | <b>DUT Type:</b><br>Portable Handset                                                                                              | <b>APPENDIX E:</b><br>Page 1 of 1 |                                                                                       |                                        |