試験機器リスト

Measurement equipment list

USE	Equipment	Company	Model No.	Serial No.	Calibrated by	Cal. Method	Cal. Due	Cal. Date
X	Spectrum Analyzer	ROHDE&SCHWARZ	FSV40	487/641405	JEPSI	C	2024.5.31	2023.5.5
X	RF test system	ROHDE&SCHWARZ	TS8997	487/391835	JEPSI	C	2024.11.30	2023.11.23

Note1: 測定機器の較正は、1年間有効です。
The calibration of measurement equipment is valid for one year period.

Note2: "X" は使用した測定機器です。
"X" used equipment.

Note3: 較正方法 ...
Cal.Method ...

イ): 独立行政法人情報通信研究機構（以下「機構」という。）又は第百二条の十八第一項の指定較正機関が行う較正
a): Calibration conducted by the National Institute of Information and Communications Technology(NICT)(hereinafter referred to as "NICT") or a designated calibration agency under Article 102-18 paragraph (1)

ロ): 計量法（平成四年法律第五十一号）第百三十五条 又は第百四十四条 の規定に基づく校正
b): Correction conducted pursuant to the provisions of Article 135 or Article 144 of the Measurement Law (Law No. 51 of 1992)

ハ): 外国において行う較正であつて、機構又は第百二条の十八第一項の指定較正機関が行う較正に相当するもの
c): Calibration conducted in foreign countries, which shall be equivalent to the calibration conducted by the NICT or a designated calibration agency under Article 102-18 paragraph (1)

ニ): 別表第三の下欄に掲げる測定器その他の設備であつて、イからハまでのいずれかに掲げる較正等を受けたものを用いて行う較正等
d): Calibration conducted by using measuring instruments and other equipment listed in the right column of Table No. 3 attached hereto, which shall have been given any of calibration, etc. listed above from a) to c)

複製禁止